



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM
qoshidagi Sinov laboratoriyalar majmuasi

Elektromagnit
moslashuvchanlik
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,
Piskent tumani,
Lola-ariq MFY,
O`zbekiston ko`chasi,
174-uy.

«Термогигрометр»



Model: TFA 30.5002

Zavod raqami: № H370

Ishlab chiqaruvchi: Germaniya.

O'lchov chegarasi, (10 - 99) %, -10 °C do + 60 °C

Xatolik darajasi: ± 5 %, ±1°C

Ish harorat, namlik oralig'I va atmosfera bosimi: (20 ±5) °C, (30-80) %, (84-106.7) kPa