



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM
qoshidagi Sinov laboratoriyalar majmuasi

Elektromagnit
moslashuvchanlik
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,
Piskent tumani,
Lola-ariq MFY,
O`zbekiston ko'chasi,
174-uy.

«Аттенюатор (Attenuateur)»

Model: 16-4459

Zavod raqami: № 47001-001

Ishlab chiqaruvchi: Франция.

O'lchov chegarasi : до 4 ГГц, 20 dB 40 W.

Xatolik darajasi: ± 0.5 dB

Ish harorat, namlik oralig'i va atmosfera bosimi: (20 ± 5) °C,
 $(30-80)$ %, $(84-106.7)$ kPa

