



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM  
qoshidagi Sinov laboratoriyalar majmuasi

Elektromagnit  
moslashuvchanlik  
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,  
Piskent tumani,  
Lola-ariq MFY,  
O`zbekiston ko'chasi,  
174-uy.

## «Источник переменного тока»

**Model:** ACS 500N3

**Zavod raqami:** № P2026241931

**Ishlab chiqaruvchi:** Shvetsariya.

**O'lchov chegarasi, Hz:** (15 Hz – 3 kHz)

**Xatolik darajasi:**  $\pm 0.4\%$

**Ta'minot kuchlanishi:** 220/230 V

**Ta'minot manbai chastotasi:** 50/60 Hz  $\pm 0.5$

**Ish harorat, namlik oralig'i va atmosfera bosimi:** (20  $\pm$  5) °C, (30-80) %, (84-106.7) kPa

