



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM
qoshidagi Sinov labaratoriylar majmuasi

Elektromagnit
moslashuvchanlik
laboratoriysi

Po`chta manzili:

111404, Toshkent viloyati,
Piskent tumani,
Lola-ariq MFY,
O`zbekiston ko‘chasi,
174-uy.

«Приемник»

Model: ESW8

Zavod raqami: № 101344

Ishlab chiqaruvchi: Germaniya.

O`lchov chegarasi, Hz: (2 Hz - 8 GHz)

Xatolik darajasi: $\pm 1 \times 10^{-7}$

Ta'minot kuchlanishi: 220/230 V $\pm 10\%$.

Ta'minot manbai chastotasi: 50 Hz ± 0.5

Ish harorat, namlik oralig'I va atmosfera bosimi: (20 ± 5) °C, (30-80) %, (84-106.7) kPa

