



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM
qoshidagi Sinov laboratoriyalar majmuasi

Elektromagnit
moslashuvchanlik
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,
Piskent tumani,
Lola-ariq MFY,
O`zbekiston ko'chasi,
174-uy.

«Анализатор гармоник и мерцания (фликера)»

Model: *DPA 500N*

Zavod raqami: № *P2007238464*

Ishlab chiqaruvchi: *Shvetsariya .*

O'lchov chegarasi, Hz: (15 Hz - 3 kHz)

Xatolik darajasi: $\pm 0.4 \%$

Ta'minot kuchlanishi: $220/230 V \pm 10 \%$.

Ta'minot manbai chastotasi: $50/60 Hz \pm 0.5$

Ish harorat, namlik oralig'i va atmosfera bosimi: $(20 \pm 5) ^\circ C$, $(30-80) \%$, $(84-106.7) kPa$

