



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM  
qoshidagi Sinov laboratoriyalar majmuasi

Elektromagnit  
moslashuvchanlik  
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,  
Piskent tumani,  
Lola-ariq MFY,  
O`zbekiston ko`chasi,  
174-uy.

## «Измеритель мощности»

**Model:** NRX

**Zavod raqami:** №102084

**Ishlab chiqaruvchi:** Chexiya.

**O'lchov chegarasi, dBm:**  $(-70 \text{ do } +45 \text{ dBm})$

**Xatolik darajasi:**  $\pm 5 \%$

**Ta'minot kuchlanishi:**  $220/230 \text{ V} \pm 10 \%$ .

**Ta'minot manbai chastotasi:**  $50/60 \text{ Hz} \pm 0.5$

**Ish harorat, namlik oralig'i va atmosfera bosimi:**  $(20 \pm 5) ^\circ\text{C}$ ,  $(30-80) \%$ ,  $(84-106.7) \text{ kPa}$

