



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM
qoshidagi Sinov laboratoriyalar majmuasi

Elektromagnit
moslashuvchanlik
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,
Piskent tumani,
Lola-ariq MFY,
O`zbekiston ko`chasi,
174-uy.

«Связующая развязывающая сеть»

Model: *CDN M1-10S*

Zavod raqami: № 57282

Ishlab chiqaruvchi: *Shvetsariya*

O'lchov chegarasi, Hz: (10 kHz - 80 MHz)

Xatolik darajasi: ± 1 dB

Ta'minot kuchlanishi: $20\text{ V} \pm 10\%$.

Ish harorat, namlik oralig'i va atmosfera bosimi: $(20 \pm 5)^\circ\text{C}$, (30-80) %, (84-106.7) kPa

