



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM
qoshidagi Sinov laboratoriyalar majmuasi

Elektromagnit
moslashuvchanlik
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,
Piskent tumani,
Lola-ariq MFY,
O`zbekiston ko`chasi,
174-uy.

«ЭКВИВАЛЕНТ СЕТИ»

Model: ENV432

Zavod raqami: № 101564

Ishlab chiqaruvchi: Germaniya.

O'lchov chegarasi, Hz: (9 kHz - 30 MHz)

Xatolik darajasi: ± 1 dB

Ta'minot kuchlanishi: 220/230 V ± 10 %.

Ta'minot manbai chastotasi: 50/60 Hz ± 0.5

Ish harorat, namlik oralig'i va atmosfera bosimi: (20 \pm 5) °C, (30-80) %, (84-106.7) kPa

