



Government of Nepal Ministry of Health and Population

Geta Hospital

Geta, Kailali BIDDING DOCUMENT

for

Supply, Installation and Commissioning of Infertility Treatment Equipments International Competitive Bidding (ICB) Single Stage Two Envelope

Issued on: 22 April 2025

Issued to: All Eligible Bidders

Invitation for Bids No.: GH/G/ICB-06/2081-82

ICB No.: 06

Contract Identification No.: GH/G/ICB-06/2081-82/Supply, Installation and

Commissioning of Infertility Treatment Equipments



BDS Bid Data Sheet
BD Bidding Document
DCS Delivery and Completion Schedule
DPDevelopment Partner
EQC Evaluation and Qualification Criteria
GCC General Conditions of Contract
GoN ¹ Government of Nepal
ICC International Chamber of Commerce
IFB Invitation for Bids
ITB Instructions to Bidders
LGRS List of Goods and Related Services
NCB National Competitive Bidding
PANPermanent Account Number
PPMOPublic Procurement Monitoring Office
SBD Standard Bidding Document
SBQ Schedule of Bidder Qualifications
SCC Special Conditions of Contract
SR Schedule of Requirements
TS Technical Specifications

VAT Value Added Tax

-

¹ "GoN" word indicates all public entities according to Public Procurement Act, 2063



Invitation for Bids

PART 1 – Bidding Procedures	
Section I. Instructions to Bidders	6
Section II. Bid Data Sheet	42
Section III. Evaluation and Qualification Criteria	48
Section IV. Bidding Forms	62
PART 2 – Supply Requirements Section V. Schedule of Requirements	87
PART 3 – Conditions of Contract and Contract Form	ıs
Section VI. General Conditions of Contract	97
Section VII. Special Conditions of Contract	117
Section VIII. Contract Forms	125



Government of Nepal Ministry of Health and Population

Geta Hospital Geta, Kailali, Nepal

Invitation for Electronic Bids for Supply, Installation and Commissioning of Infertility Treatment Equipments First Date of Publication: 2082/01/09 (22 April 2025)

- The Geta Hospital, Geta Kailali invites electronic bids from eligible bidders for Supply, Installation and Commissioning of Infertility Treatment Equipments listed below in the table to be delivered at the Geta Hospital, Kailali under International Competitive Bidding (ICB) – Single Stage Two Envelope.procedures specified in the Public Procurement Act (PPA) 2063 and Public Procurement Regulation (PPR) 2064.
- 2. Under the Single Stage, Two Envelope Procedure, Bidders are required to submit simultaneously two separate sealed envelopes, one containing (i) the Technical Bid and the other (ii) the Price Bid, both in turn enclosed in one sealed envelope as per the provision of ITB 23 of the Bidding Document.
- 3. Eligible Bidders may obtain further information and inspect the bidding documents at the office of **Geta Hospital**, **Geta Kailali** or may visit PPMO e-GP system https://bolpatra.gov.np/egp.

4. Interested bidders shall register in the e-GP system and deposit the cost of bidding document in the following bank:

Name of the Bank: Rastriya Banijya Bank, Dhangadhi, Kailali		
Name of the Office: Geta Hospital, Geta Kailai		
Office Code No.: 370617701 Revenue Head No.: 14229		
Account No.: 1000100200010000	Name of Account: Ko. Le. Ni. Ka., Kailali	

5. Bids must be valid for a period of **90 days** from the date of bid opening. The bids are single slice package and bidder must bid all items in a single package, only the total items quoted will be considered as the bid amount. The name and identification of the contracts are as follows:

			Cost of	Bid So	ecurity
S. N.	IFB No.	Description of Goods	Bidding Document (NRs)	Minimum Amount (NRs)	Minimum Validity Period
1	GH/G/ICB- 06/2081-82	Supply, Installation and Commissioning of Infertility Treatment Equipments	5,000	658,000	2082/06/20 (06 Oct. 2025)

6. The bidding schedule is as follows:

	Pre-Bid Meeting	Bid Submission Deadline	Bid Opening
Date and Time	2082/01/30 (13 May 2025)	2082/02/23 (06 June 2025)	2082/02/23 (06 June 2025)
	13.00 Hours	12.00 Hours	13.00 Hours
Venue	Geta Hospital, Geta Kailali		

- 7. If the last date of purchasing and /or submission falls on a government holiday, then the next working day shall be considered as the last date. In such case the validity period of the bid and bid security shall remain the same as specified for the original last date of bid submission.
- 8. For more detail and further information bidders may contact the office of Geta Hospital, Geta Kailali email address: procurement.getahospital@gmail.com or may visit PPMO e-GP system www.bolpatra.gov.np/egp.

Director



Section I. Instructions to Bidders

Table of Clauses

A.	General	8
1.	Scope of Bid	8
2.	Source of Funds	8
3.	Fraud and Corruption	8
4.	Eligible Bidders	11
5.	Eligible Goods and Related Services	13
6.	Site Visit	13
B.	Contents of Bidding Document	13
7.	Sections of the Bidding Document	13
8.	Clarification of Bidding Document/Pre-bid meeting	14
9.	Amendment of Bidding Document	14
C.	Preparation of Bids	15
10.	Cost of Bidding	15
11.	Language of Bid	15
12.	Documents Comprising the Bid	15
13.	Bid Submission Letter and Price Schedules	16
14.	Alternative Bids	16
15.	Bid Prices and Discounts	16
16.	Currencies of Bid	19
17.	Documents Establishing the Eligibility of the Bidder	20
18.	Documents Establishing the Conformity of the Goods and Related Services to the	
	Bidding Document	20
19.	Documents Establishing the Qualifications of the Bidder	20
20.	Period of Validity of Bids	22
21.	Bid Security	22
22.	Format and Signing of Bid	24
D.	Submission and Opening of Bids	24
23.	Sealing and Marking of Bids	25
24.	Deadline for Submission of Bids	28
25.	Late Bids	29
26.	Withdrawal, or Modification of Bids	29
27.	Bid Opening	30
E. E	valuation and Comparison of Bids	32
28.	Confidentiality	32
29.	Clarification of Bids	33
30.	Deviations, Reservations, and Omissions	33

I NEDO.	idi	
Ministry of Weelth and Popula		
Geta Hospital	Director Examination of Technical Bid	
Geta, K3\1\.	Examination of Technical Bid	33
32.	Determination of Responsiveness of Technical Bid	34
33.	Non-material Non-conformi-ties	35
34.	Qualification of the Bidder	35
35.	Examination of Price Bid	36
36.	Correction of Arithmetical Errors	36
37.	Conversion to Single Currency	36
38.	Goods manufactured in Nepal to be produced	36
39.	Evaluation and Comparison of Price Bids	37
40.	Purchaser's Right to Accept Any Bid, and to Reject Any or All Bids	38
F. A	ward of Contract	38
41.	Award Criteria	38
42.	Purchaser's Right to Vary Quantities at Time of Award	38
43.	Notification of Intention to Award	38
44.	Performance Security	39
45.	Signing of Contract	39
43.	Complaint and Review	40



Section I. Instructions to Bidders

A. General

1.	Scope of Bid	 In connection with the Invitation for Bids indicated in the Bid Data Sheet (BDS), the Purchaser as indicated in the BDS issues this Bidding Document for the supply of Goods and Related Services incidental thereto as specified in Section V, Schedule of Requirements. The name, identification, and number of contracts (packages/lots) are indicated in BDS. 1.2 Throughout this Bidding Document: 		
		(a) the term "in writing" means communicated in written form with proof of receipt;		
		(b) if the context so requires, singular means plural and vice versa; and		
		(c) "day" means calendar day.		
2.	Source of Funds	 GoN Funded: In accordance with its annual program and budget, approved by the GoN, the Purchaser intends to apply a portion of the allocated budget to eligible payments under the contract(s) indicated in the BDS for which this Bidding Document is issued. Or DP Funded: The GoN has applied for or received financing (hereinafter called "funds") from the Development Partner (hereinafter called "the DP") indicated in the BDS toward the cost of the project named in the BDS. The GoN intends to apply a portion of the funds to eligible payments under the contract(s) for which this Bidding Document is issued. 		
		2.2 DP Funded: Payment by the DP will be made only at the request of the GoN and upon approval by the DP in accordance with the terms and conditions of the financing agreement between the GoN and the DP (hereinafter called the "Loan Agreement"), and will be subject in all respects to the terms and conditions of that Loan Agreement. No party other than the GoN shall derive any rights from the Loan Agreement or have any claim to the funds.		
3.	Fraud and Corruption	3.1 Procuring Entities as well as Bidders, suppliers and contractors and their sub-contractors shall adhere to the highest standard of ethics during the procurement and execution of such contracts. In pursuance of this:;		

		Nepal	on Z	y m	
Gov Ministry G	vernment of Health eta Ho Geta,	Stigs of	Di	rect	9r

- (a) the Purchaser adopts, for the purposes of this provision, the terms as defined below:
 - (i) "corrupt practice" means the offering, giving, receiving, or soliciting, directly or indirectly, anything of value to influence improperly the actions of another party;
 - (ii) "fraudulent practice" means any act or omission, including a misrepresentation, that knowingly or recklessly misleads, or attempts to mislead, a party to obtain a financial or other benefit or to avoid an obligation;
 - (iii) "coercive practice" means impairing or harming, or threatening to impair or harm, directly or indirectly, any party or the property of the party to influence improperly the actions of a party;
 - (iv) "collusive practice" means an arrangement between two or more parties designed to achieve an improper purpose, including influencing improperly the actions of another party.
 - (v) "obstructive practice" means (a) deliberately destroying, falsifying, altering, or concealing of evidence material to an investigation; (b) making false statements to investigators in order to materially impede an investigation; (c) failing to comply with requests to provide information, documents, or records in connection with an investigation; (d) threatening, harassing, or intimidating any party to prevent it from disclosing its knowledge of matters relevant to the investigation or from pursuing the investigation; or (e) materially impeding GoN/DP's contractual rights of audit or access to information; and
 - (vi) "integrity violation" is any act which violates Anticorruption Policy, including (i) to (v) above and the following: abuse, conflict of interest, violations of GoN/DP sanctions, retaliation against whistleblowers or witnesses, and other violations of Anticorruption Policy, including failure to adhere to the highest ethical standard.
- (b) the Purchaser will reject a proposal for award if it determines that the Bidder recommended for award has, directly or through an agent, engaged in corrupt, fraudulent, collusive, coercive, or obstructive practices or other integrity violations in competing for the contract;
- (c) DP will cancel the portion of the financing allocated to a contract if it determines at any time that representative(s) of the GoN or of a beneficiary of DP-financing engaged in corrupt, fraudulent, collusive, or coercive practices or other integrity violations during the procurement or the execution of that contract, without the GoN having taken timely and appropriate action satisfactory to DP to remedy the situation.
- (d) DP will impose remedial actions on a firm or an individual, at any

Mapphion Honly	
evernment of Popular Profital Direct	
Geta Hospital Directo Geta, Kailali	time, in accordance with DP's Anticorruption Policy and related Guidelines (as amended from time to time), including declaring ineligible, either indefinitely or for a stated period of time, to participate in DP-financed, -administered, or -supported activities or to benefit from an DP-financed, -administered, or -supported contract, financially or otherwise, if it at any time determines that the firm or individual has, directly or through an agent, engaged in corrupt, fraudulent, collusive, coercive, or obstructive practices or other integrity violations; and
	(e) The Supplier shall permit the GoN/DP to inspect the Supplier's accounts and records relating to the performance of the Supplier and to have them audited by auditors appointed by the GoN/DP, if so required by the GoN/DP.
	 3.2 The Bidder shall not carry out or cause to carry out the following acts with an intention to influence the implementation of the procurement process or the procurement agreement: (a) give or propose improper inducement directly or indirectly, (b) distortion or misrepresentation of facts, (c) engaging in corrupt or fraudulent practice or involving in such act, (d) interference in participation of other competing bidders, (e) coercion or threatening directly or indirectly to cause harm to the person or the property of any person to be involved in the procurement proceedings,
	(f) collusive practice among bidders before or after submission of bids for distribution of works among bidders or fixing artificial/uncompetitive bid price with an intention to deprive the Purchaser the benefit of open competitive bid price,
	(g) Contacting the Purchaser with an intention to influence the Purchaser with regards to the bids or interference of any kind in examination and evaluation of the bids during the period from the time of opening of the bids until the notification of award of contract.
	 3.3 PPMO, on the recommendation of the Procuring Entity may blacklist a Bidder for a period of one (1) to three (3) years for its conduct including on the following grounds and seriousness of the act committed by the bidder: (a) if convicted by a court of law in a criminal offence which disqualifies the Bidder from participating in the contract, (b) if it is established that the contract agreement signed by the Bidder was based on false or misrepresentation of Bidder's qualification information, (c)if it at any time determines that the firm has, directly or through an agent, engaged in corrupt, fraudulent, collusive, coercive, or obstructive practices in competing for, or in executing, a GoN/DP-financed contract.

ry of Hea	Alth and Poppel)r		
Get	a, Kailali	(d) if the Successful Bidder fails to sign the Contract.		
		3.4 A bidder declared blacklisted and ineligible by the GoN, Public Procurement Monitoring Office (PPMO) and/or the DP in case of DP funded project, may be ineligible to bid for a contract during the period of time determined by the GoN, PPMO and/or the DP including credit information bureau of Nepal.		
		3.5 In case of a natural person or firm/institution/company which is already declared blacklisted and ineligible by the GoN, any other new or existing firm/institution/company owned partially or fully by such Natural person or Owner or Board of director of blacklisted firm/institution/company; shall not be eligible bidder.		
		3.6 Furthermore, Bidders shall be aware of the provisions of GCC 34.1(c).		
4.	Eligible Bidders	4.1 This Invitation for Bids is open to eligible Bidders from all countries, except for any <i>specified in the BDS</i> .		
		4.2 A Bidder may be a natural person, private entity, government-owned entity (subject to ITB 4.5) or any combination of them with a formal intent to enter into an agreement or under an existing agreement in the form of a Joint Venture (JV). Maximum number of parners in JV shall be as specified in BDS . In the case of a JV:		
		(a) all parties to the JV shall be jointly and severally liable; and		
		(b) a JV shall nominate a representative who shall have the authority to conduct all businesses for and on behalf of any and all the parties of the JV during the bidding process and, in the event the JV is awarded the Contract, during contract execution.		
		4.3 A Bidder shall not have a conflict of interest. Any Bidders found to have a conflict of interest shall be disqualified. A Bidder may be considered to be in a conflict of interest with one or more parties in this bidding process if, including but not limited to:		
		(a) have controlling shareholders in common;		
		(b) receive or have received any direct or indirect subsidy from any of them;		
		(c) have the same legal representative for purposes of this Bid;		
		(d) have a relationship with each other, directly or through common third parties, that puts them in a position to have access to information about or influence on the Bid of another Bidder, or influence the decisions of the Purchaser regarding this bidding process;		

Ministry of Health and Population Goto Hospital Direct	
Geta, Kailali	(e) a Bidder participates in more than one bid in this bidding process. Participation by a Bidder in more than one Bid will result in the disqualification of all Bids in which it is involved. However, this does not limit the inclusion of the same subcontractor, not otherwise participating as a Bidder, in more than one bid; or
	(f) a Bidder or any of its affiliates participated as a consultant in the preparation of the design or technical specifications of the goods and services that are the subject of the bid.
	 (g) a Bidder that has a close business or family relationship with a professional staff of the Procuring Entity. 4.4 A Bidder that is under a declaration of ineligibility by the GoN/DP in accordance with ITB 3.4, at the date of the deadline for bid submission or thereafter, shall be disqualified. The list of debarred firms is available at the electronic address specified in the BDS.
	4.5 A GoN-owned enterprise may also participate in the bid if it is legally and financially autonomous, it operates under commercial law, and it is not dependent agency of the Purchaser.
	4.6 Bidders shall provide such evidence of their continued eligibility satisfactory to the Purchaser, as the Purchaser shall reasonably request.
	4.7 Firms shall be excluded in any of the cases, if
	(a) by an act of compliance with a decision of the United Nations Security Council taken under Chapter VII of the Charter of the United Nations prohibits any import of goods or Contracting of works or services from that country or any payments to persons or entities in that country.
	(b) DP Funded: as a matter of law or official regulation, GoN prohibits commercial relations with that country, provided that the DP is satisfied that such exclusion does not preclude effective competition for the supply of goods or related services required;
	(c) DP Funded: a firm has been determined to be ineligible by the DP in

4.8 A bidder and all parties constituting the Bidder shall have the nationality of

relation to their guidelines or appropriate provisions on preventing and combating fraud and corruption in projects financed by them.

4.9 The domestic Bidder who has obtained Permanent Account Number (PAN) and Value Added Tax (VAT) registration certificate(s) and Tax clearance certificate or proof of submission of tax return from the Inland

Vernme v of Hea	th and Popula	
Geta (Hospital Directo	Revenue Office shall only be eligible. The foreign bidder submitting the documents <i>indicated in the BDS</i> at the time of bid submission and a declaration to submit the document(s) <i>indicated in the BDS</i> at the time of contract agreement shall only be eligible. 4.10 In case a prequalification process has been conducted prior to the bidding process, this bidding is open only to prequalified Bidders.
5.	Eligible Goods and Related Services	 5.1 All goods and related services to be supplied under the contract are eligible, unless their origin is from a country <i>specified in the BDS</i>. 5.2 For purposes of this clause, "origin" means the place where the goods are mined, grown, or produced, or the place from which the related services are supplied
		5.3 The nationality of the firm/Bidder that produces, assembles, distributes, or sells the goods shall not determine their origin.
6.	Site Visit	 6.1 For goods contracts requiring installation/ commissioning/ networking or similar services at site, the Bidder, at the Bidder's own responsibility and risk, is encouraged to visit and examine the Site and obtain all information that may be necessary for preparing the Bid and entering into a contract for the supply of goods and related services. 6.2 The Bidder should ensure that the Purchaser is informed of the visit in
		adequate time to allow it to make appropriate arrangements. 6.3 The costs of visiting the Site shall be at the Bidder's own expense.

B. Contents of Bidding Document

7.	Sections of the Bidding Document	al	The Bidding Document consist of Parts 1, 2, and 3, which include all the Sections indicated below, and should be read and construed in conjunction with any Addenda issued in accordance with ITB 9.	
		PA	PART 1 Bidding Procedures	
		,	Section I.	Instructions to Bidders (ITB)
		,	Section II.	Bid Data Sheet (BDS)
		,	Section III.	Evaluation and Qualification Criteria
			Section IV. Bidding Forms	
		PA	PART 2 Supply Requirements	
			Section V. Schedule of Requirements	
		PA	ART 3 Condi	tions of Contract and Contract Forms

Nepalion - Han	ly .		
by of Health and Popular Direct			
Geta, Kailali	Section VI. General Conditions of Contract (GCC)		
	Section VII. Special Conditions of Contract (SCC)		
	Section VIII. Contract Forms		
	7.2 The Purchaser will reject any Bid submission (in case of hard copy submission) if the Bidding Document was not purchased directly from the Purchaser, or through its assigned office as stated in the invitation for bids or has not deposited (in case of electronically submission) the cost of Bidding Document as stated in the invitation for bids.		
	7.3 The Bidder is expected to examine all instructions, forms, terms, and specifications in the Bidding Document as well as in Amendments, if any. Failure to furnish all information or documentation required by the Bidding Document may result in the rejection of the Bid.		
	7.4 The Invitation for Bids issued by the Purchaser is not part of the Bidding Document		
8. Clarification of Bidding Document/ Pre-bid meeting	8.1 A prospective Bidder requiring any clarification of the Bidding Document shall contact the Purchaser in writing at the Purchaser's address <i>indicated in the BDS</i> or raise any question or curiosity during the pre-bid meeting if provided for in accordance with ITB 8.2. The Purchaser will respond in writing to any request for clarification, provided that such request is received within the time limit <i>specified in the BDS</i> prior to the deadline for submission of Bids. The Purchaser shall forward copies of its response to all Bidders who have acquired the Bidding Document directly from it, including a description of the inquiry but without identifying its source. Should the Purchaser deem it necessary to amend the Bidding Document as a result of a clarification, it shall do so following the procedure under ITB 9 and 24.2.		
	8.2 The purchaser may organize a pre-bid meeting of Bidders at least fifteen (15) days before the deadline for submission of Bids at the place, date, and time as <i>specified in the BDS</i> to provide information relating to Bidding Documents, Technical specifications and the like matters. Should the purchaser deem it necessary to amend the Bidding Document as a result of a clarification, it shall do so following the procedure under ITB 9 and ITB 24.2.		
9. Amendment of Bidding Document	9.1 At any time prior to the deadline for submission of the Bids, the Purchaser may amend the Bidding Document by issuing addenda.		
Document	9.2 Any addendum issued shall be part of the Bidding Document and shall be communicated in writing to all who have obtained the Bidding Document directly from the Purchaser.		

Nepalion -	<u>eq.</u>
Government of bobiles	
Inistry of Health and Ital Dire	or
Geta, Kailall	9.3 To give prospective Bidders reasonable time in which to take an
	addendum into account in preparing their Bids, the Purchaser may, at
	its discretion, extend the deadline for the submission of the Bids,
	pursuant to ITB 24.2.

C. Preparation of Bids

10.	Cost of Bidding	10.1	The Bidder shall bear all costs associated with the preparation and submission of its Bid, and the Purchaser shall not be responsible or liable for those costs, regardless of the conduct or outcome of the bidding process.		
11.	Language of Bid	11.1	The Bid, as well as all correspondence and documents relating to the Bid exchanged by the Bidder and the Purchaser, shall be written in the language <i>specified in the BDS</i> . Supporting documents and printed literature that are part of the Bid may be in another language provided they are accompanied by an accurate translation of the relevant passages in the language <i>specified in the BDS</i> , in which case, for purposes of interpretation of the Bid, such translation shall govern.		
12.	Documents Comprising the Bid	12.1	The Bid shall comprise two envelopes submitted simultaneously, one called the Technical Bid containing the documents listed in ITB 12.2 and the other the Price Bid containing the documents listed in ITB 12.3, both envelopes enclosed together in an outer single envelope.		
		12.2	The 7	Technical Bid shall comprise the following:	
			(a) Letter of Technical Bid;		
			(b) Bid Security in accordance with ITB 21;		
			(c) alternative technical bids, at Bidder's option and if permissible, in accordance with ITB 14;		
			(d)	written confirmation authorizing the signatory of the Bid to commit the Bidder, in accordance with ITB 22;	
			(e)	(e) documentary evidence in accordance with ITB 17 establishing the Bidder's eligibility to bid;	
			(f) documentary evidence in accordance with ITB Clauses 18 and 31, that the Goods and Related Services conform to the Bidding Document;		
			(g)	documentary evidence in accordance with ITB 19 establishing the Bidder's qualifications to perform the contract if its Bid is accepted; and	
			(h)	any other required documents, which is not against the provision of Procurement Act/Regulation/Directives and Standard Bidding	

overnment of the	Philagio,				
Ty of Health and S	tal Directo)r			
Geta, Kailal	,	Document issued by PPMO, required in the BDS.			
		12.3 The Price Bid submitted by the Bidder shall comprise the following:			
		(a) Price Bid Submission Letter and the applicable Price Schedules, in accordance with ITB 13, ITB 15, and ITB 16;			
		(b) alternative Price Bid corresponding to the alternative Technical Bid, if permissible, in accordance with ITB 14; and			
		(c) any other document required in the BDS.			
		12.4 The Bidder is solely responsible for the authenticity of the submitted documents.			
		12.5 The Technical Bid shall not include any financial information related to the Price Bid. A Technical Bid containing such material financial information shall be declared non-responsive.			
Let Pri	omission ter and	13.1 The Bidder shall submit the Technical Bid Submission Letter and the Price Bid Submission Letter using the form furnished in Section IV, Bidding Forms. These forms must be completed without any alterations to their format, and no substitutes shall be accepted. All blank spaces shall be filled in with the information requested.			
		13.2 The Bidder shall submit, as part of the Price Bid, the Price Schedules for Goods and Related Services, according to their origin as appropriate, using the forms furnished in Section IV, Bidding Forms.			
14. Alto Bid	ernative s	14.1 Unless otherwise <i>indicated in the BDS</i> , alternative bids shall not be considered.			
15. Bid and	Prices Discounts	15.1 The prices and discounts quoted by the Bidder in the Letter of Price bid and in the Price Schedules shall conform to the requirements specified below.			
		15.2 The Bidder shall complete the appropriate Price Schedule and the sources of Goods schedules included herein, stating the unit prices, total cost per item, the total Bid amount, and the expected countries of origin of the Goods to be supplied under the contract.			
		15.3 All items in the Schedule of Supply must be listed and priced separately in the Price Schedules. If a Price Schedule shows items listed but not priced, their prices shall be assumed to be included in the prices of other items. Items not listed in the Price Schedule shall be assumed not to be included in the Bid, and provided that the Bid is substantially responsive, the corresponding adjustment shall be applied in accordance with ITB 39.3. Unit rates and prices for all items in the Schedule of Supply shall be expressed in positive values. If unit rates and prices are expressed in negative values, the bid will			

Government of Neralion

Ministry of Health and popular

Director

be rejected.

- 15.4 The price to be quoted in the Letter of Price Bid shall be the total price of the Bid excluding any discounts offered. Absence of the total bid price in the Letter of Price Bid may result in the rejection of the Bid.
- 15.5 If the Bidder intends to offer any unconditional discount, it shall always be expressed in fixed percentage and that shall not vary as the quantity varies and be applicable to each unit rate. The discount and methodology for its application shall be quoted in Letter of Price Bid.
- 15.6 The terms EXW, CIF, CIP, and other similar terms shall be governed by the rules prescribed in the current edition of Incoterms, published by The International Chamber of Commerce, at the date of the Invitation for Bids or as specified in the BDS.
- 15.7 Prices shall be quoted as specified in each Price Schedule included in **Section IV**, **Bidding Forms**. The disaggregation of price components is required solely for the purpose of facilitating the comparison of Bids by the Purchaser. This shall not in any way limit the Purchaser's right to contract on any of the terms offered. Prices shall be entered in the following manner:
 - (a) For Goods manufactured in Nepal:
 - (i) the price of the goods quoted EXW (ex works, ex factory, ex warehouse, ex showroom, or off-the-shelf, as applicable), including all customs duties, Value Added Tax and other taxes already paid or payable on the components and raw material used in the manufacture or assembly of goods quoted ex works or ex factory, or on the previously imported goods of foreign origin quoted ex warehouse, ex showroom, or off-the-shelf;
 - (ii) Value Added Tax and all other taxes applicable in Nepal and payable on the Goods if the Contract is awarded to the Bidder;
 - (iii) the price for inland transportation, insurance, and other local services required to convey the Goods to their final destination (Project Site) specified in the BDS.
 - (iv) the total price for the item.
 - (b) For Goods manufactured outside Nepal, to be imported:
 - i. the price of the goods quoted CIF (named port of destination), or CIP (border point), or CIP (named place of destination), named place of destination as **specified in**

inepal ion	- Honly
Ministry of Health and Popular	Director

Geta, Kailali

the BDS;

- ii. the price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination (Project Site) **specified in the BDS**;
- iii. in addition to the CIP prices specified in (b)(i) above, the price of the Goods to be imported may be quoted FOB port of shipment (or FCA, as the case may be) or CPT (named place of destination), if so **specified in the BDS**; and

iv.the total price for the item.

(c) For Goods manufactured outside Nepal, already imported:

[For previously imported Goods, the quoted price shall be distinguishable from the original import value of these Goods declared to customs and shall include any rebate or mark-up of the local agent or representative and all local costs except import duties and taxes, which have been and/or have to be paid by the Purchaser. For clarity the bidders are asked to quote the price including import duties, and additionally to provide the import duties and the price net of import duties which is the difference of those values.]

- (i) the price of the Goods, including the original import value of the Goods; plus any mark-up (or rebate); plus any other related local cost, and custom duties and other import taxes already paid or to be paid on the Goods already imported.
- (ii) the custom duties and other import taxes already paid (need to be supported with documentary evidence) or to be paid on the Goods already imported;
- (iii) the price of the Goods, obtained as the difference between (i) and (ii) above;
- (iv) any sales and other taxes which will be payable on the Goods if the contract is awarded to the Bidder;
- (v) the price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination (Project Site) specified in the BDS; and

Vernment of the population of	<u> </u>
Geta Hospital Directo	(vi) the total price for the item.
Geta, rui.	(d) For Related Services, other than inland transportation and other services required to convey the Goods to their final destination, whenever such Related Services are specified in the Schedule of Requirements:
	(i) the local currency cost component of each item comprising the Related Services; and
	(ii) the foreign currency cost component of each item comprising the Related Services, inclusive of all custom ruties, Value Added Tax and other taxes aplicable in the Purchaser's country, payable on the related services, if the contract is awarded to the Bidder.
	15.8 Prices quoted by the Bidder shall be fixed during the Bidder's performance of the Contract and not subject to variation on any account, unless otherwise <i>specified in the BDS</i> . A Bid submitted with an adjustable price quotation shall be treated as non responsive and shall be rejected, pursuant to ITB 32. However, if in <i>accordance with the BDS</i> , prices quoted by the Bidder shall be subject to adjustment during the performance of the Contract, a Bid submitted with a fixed price quotation shall not be rejected, but the price adjustment shall be treated as zero
	15.9 If so indicated in ITB 1.1, Bids are being invited for individual contracts (lots) or for any combination of contracts (packages). Bidders wishing to offer any price discount for the award of more than one Contract shall specify in their Price Bids the price discount applicable to each package, or alternatively, to individual Contracts within the package. Price discounts shall be submitted in accordance with ITB 15.7, provided the Price Bids for all lots are submitted and opened at the same time.
16. Currencies of Bid	 16.1 Bid prices shall be quoted in the following currencies: (a) Bidders may express their bid price in any fully convertible currency. If a Bidder wishes to be paid in a combination of amounts in different currencies, it may quote its price accordingly but shall use no more than three currencies in addition to the Nepalese currency.

All expenditures that are to be incurred in Nepal for i) inland transportation and related costs, ii) all taxes, and iii) local currency cost component of related services other than inland transportation and other services should be expressed in the Bid in Nepalese currency and will be payable in Nepalese

(b)

vernmer y of Healt	hand popular Directo		
	, Kailali	currency.	
17.	Documents Establishing	17.1 To establish their eligibility in accordance with ITB 4, Bidders shall:	
	the Eligibility of the Bidder	(a) complete the eligibility declarations in the Bid Submission Letter, included in Section IV, Bidding Forms; and	
		(b) if the Bidder is an existing or intended JV in accordance with ITB 4.2, submit a copy of the JV Agreement, or a letter of intent to enter into such an Agreement. The respective document shall be signed by all legally authorized signatories of all the parties to the existing or intended JV, as appropriate.	
		(c) submit the copy of the documents as specified in Section III, Evaluation and Qualification Criteria.	
18.	Documents Establishing the Conformity of the Goods	18.1 To establish the conformity of the Goods and Related Services to the Bidding Document, the Bidder shall furnish as part of its Technical Bid the documentary evidence that the Goods and Related Services conform to the requirements specified in Section V, Supply Requirements.	
the Bidding Document or data, and shall consist of a detailed item-by-item desc essential technical and performance characteristics of the Related Services, demonstrating substantial responsive Goods and Related Services to those requirements, and if		18.2 The documentary evidence may be in the form of literature, drawings or data, and shall consist of a detailed item-by-item description of the essential technical and performance characteristics of the Goods and Related Services, demonstrating substantial responsiveness of the Goods and Related Services to those requirements, and if applicable, a statement of deviations and exceptions to the provisions of Section V, Schedule of Requirements.	
		18.3 The Bidder shall also furnish a list giving full particulars, including available sources and current prices of spare parts, special tools, etc., necessary for the proper and continuing functioning of the Goods during the period specified in the BDS following commencement of the use of the goods by the Purchaser.	
		18.4 Standards for workmanship, process, material, and equipment, as well as references to brand names or catalogue numbers specified by the Purchaser in the Section V, Schedule of Requirements, are intended to be descriptive only and not restrictive. The Bidder may offer other standards of quality, brand names, and/or catalogue numbers, provided that it demonstrates, to the Purchaser's satisfaction, that the substitutions ensure substantial equivalence or are superior to those specified in Section V, Schedule of Requirements.	
	Documents Establishing the	19.1 To establish its qualifications to perform the Contract, the Bidder shall submit as part of its Technical Proposal the evidence indicated for each qualification criteria specified in Section III (Evaluation and	



Geta, Qualifications of the Bidder Qualification Criteria). The documentary evidence of the Bidder's qualifications to perform the contract, if its bid is accepted, shall establish to the Purchaser's satisfaction that the Bidder meets each of the qualification criterion specified.

- 19.2 If so *required in the BDS*, a Bidder that does not manufacture or produce the Goods it offers to supply shall submit the Manufacturer's Authorization using the form included in Section IV, Bidding Forms to demonstrate that it has been duly authorized by the manufacturer or producer of the Goods to supply these Goods in Nepal and take care of the warranty provided.
- 19.3 If so *required in the BDS*, a Bidder that does not conduct business within Nepal shall submit evidence that it will be represented by an Agent in Nepal equipped and able to carry out the Supplier's maintenance, repair and spare parts-stocking obligations prescribed in the Conditions of Contract and/or Technical Specifications.
- 19.4 A foreign Bidder wishing to have or already having a local agent shall state the following:
 - a. Name and address of the Agent/Representative,
 - b. The Agent/Representative providing type of services,
 - c. Amount of commission if the Agent/Representative is entitled to get such payment and if it participates in the procedure of payment,
 - d. Other agreement with Agent/Representative, if any,
 - e. Bidder shall certify in the Letter of Authorization as follows:

"We certify that the statement and disclosure made by us on the above are complete and true to the best of our knowledge and belief",

If the agent has not been appointed:

- a. Source of information about tender invitation,
- b. The remuneration given to the individual or firm/company or organization to work on its behalf for submitting tender, representation in the bid opening and other required action in connection with the tender,
- c. Transfer or handover an evidence of foreign currency exchanged which required to be submitted with the tender,
- d. If the bank account of any Nepali citizen has been used for the exchange of foreign currency specify the name of the individual and his address. If the foreign currency has been exchanged by self then the certificate of currency exchange.

Medal ion - and	7.	
ry of Health and Popular		
Sata HOSP: DIFECT	or	
Geta, Kailali	19.5 If a foreign Bidder in its Bid, has not provided the information mentioned in ITB 19.4 or has submitted its bid stating that the Bidder does not have a local agent and later it is proved that the bidder has a local agent or it is proved that the commission mentioned in the Bid is less than the commission received by the local agent then the Purchaser shall initiate proceedings to blacklist such bidder in accordance with ITB 3.3.	
20. Period of Validity of Bids	1 Bid shall remain valid for a period <i>specified in the BDS</i> after the bid submission deadline date prescribed by the purchaser. If the prescribed bid submission deadline date falls on a government holiday, then the next working day shall be considered as the bid submission deadline date. In such case the validity period of the bids shall be considered from the original bid submission deadline date. A bid valid for a shorter period shall be rejected by the purchaser as nonresponsive.	
	20.2 In exceptional circumstances, prior to the expiration of the bid validity period, the Purchaser may request Bidders to extend the period of validity of their Bids. The request and the responses shall be made in writing. If a Bid Security is requested in accordance with ITB 21, it shall also be extended for a corresponding period. A Bidder may refuse the request without forfeiting its Bid Security. A Bidder granting the request shall not be required or permitted to modify its Bid and to include any additional conditions against the provisions specified in Bid Documents.	
21. Bid Security	21.1 The Bidder shall furnish as part of its bid, in original form a Bid Security as <i>specified in the BDS</i> . In case of e-submission of bid, the Bidder shall upload scanned copy of Bid security letter at the time of electronic submission of the bid. The Bidder accepts that the scanned copy of the Bid security shall, for all purposes, be equal to the original. The details of original Bid Security and the scanned copy submitted with e-bid should be the same otherwise the bid shall be non-responsive.	
	1.2 If a bid security is specified pursuant to ITB 21.1, the bid security shall be a demand guarantee in any of the following forms at the Bidder's option:	
	(a) original copy of an unconditional bank guarantee from Commercial Bank or Financial Institution eligible to issue Bank Guarantee as per prevailing Law or;	
	(b) original copy of an unconditional bank guarantee from commercial foreign bank or;	



(c) original copy of cash deposit voucher in the Purchaser's Account as *specified in BDS*.

In case of a bank guarantee, the Bid Security shall be submitted either using the Bid Security Form included in Section IV, Bidding Forms or in another form acceptable to the purchaser. The form must include the complete name of the Bidder. The Bid Security shall be valid for minimum thirty (30) days beyond the end of the validity period of the bid, or beyond any period of extension if requested under ITB 20.2.

The bid security issued by any foreign Bank outside Nepal must be counter guaranteed by an Commercial Bank or Financial Institution eligible to issue Bank Guarantee as per prevailing Law in Nepal.

- 21.3 If a bid Security is required in accordance with ITB 21.1, any Bid not accompanied by an enforceable and substantially compliant Bid Security in accordance with ITB 21.2, shall be rejected by the Purchaser as nonresponsive. In case of e- Submission, if the scanned copy of an acceptable bid security letter is not uploaded with the electronic bid then bid shall be rejected.
- 21.4 If a Bid Security is specified pursuant to ITB 21.1, the Bid Security of unsuccessful Bidders shall be returned within three (3) days upon the successful Bidder's furnishing of the required performance security and signing of the Contract Agreement pursuant to ITB 44.1 and 45.1.
- 21.5 If a Bid Security is specified pursuant to ITB 21.1, the Bid Security of the successful Bidder shall be returned as promptly as possible once the successful Bidder has furnished the required Performance Security and signed the Contract Agreement.
- 21.6 The Bid Security may be forfeited:
- (a) a Bidder requests for withdrawal or modification of its bid, except as provided in ITB 20.2
 - (i) during the period of bid validity specified by the Bidder on the Letter of Bid, in case of electronic submission;
 - (ii) from the period twenty-four hours prior to bid submission deadline up to the period of bid validity specified by the Bidder on the Letter of Bid, in case of hard copy submission.
- (b) a Bidder changes the prices or substance of the bid while providing information pursuant to clause 29.1;
- (c) a Bidder involves in fraud and corruption pursuant to clause 3.1;

- (d) the successful Bidder fails to:
 - (i) furnish a performance security in accordance with ITB 44.1;
 - (ii) sign the Contract in accordance with ITB 45.1; or
 - (iii) accept the correction of arithmetical errors pursuant to clause 36.
- 21.7 The Bid Security of a JV must be in the name of the JV that submits the bid. If the JV has not been legally constituted at the time of bidding, the Bid Security shall be in the names of all future partners as named in the letter of intent mentioned in ITB 17.1 (b).

22. Format and Signing of Bid

22.1 The Bidder shall prepare one original set of the Technical Bid and one original set of the Price Bid as described in ITB 12 and clearly mark each "ORIGINAL - TECHNICAL BID" and "ORIGINAL - PRICE BID". Alternative bids, if permitted in accordance with ITB 14, shall be clearly marked "ALTERNATIVE". In addition, the Bidder shall submit copies of the Technical Bid and the Price Bid, in the number specified in the BDS and clearly mark them "COPY NO... - TECHNICAL BID" and "COPY NO... - PRICE BID". In the event of any discrepancy between the original and the copies, the original shall prevail.

In case of e-submission of bid, the Bidder shall submit his bid electronically in PDF or online forms files as specified in ITB Clause 23.1(b). If a Bidder submits both the electronic bid and a bid in hard copy within the bid submission deadline, then the submitted Bids shall be accepted for evaluation provided that the facts and figures in hard copy confirm to those in electronic bid. If there is any major discrepancy in fact and figures in the electronic bid and bid in hard copy, it shall be treated as two separate bids from one Bidder and both the Bids shall be disqualified, as per ITB Clause 4.3 (e).

- 22.2 The original and all copies of the bid shall be typed or written in indelible ink and shall be signed by a person duly authorized to sign on behalf of the Bidder. This authorization shall consist of a written confirmation as *specified in the BDS* and shall be attached to the Bid. The name and position held by each person signing the authorization must be typed or printed below the signature. All pages of the bid, except for un amended printed literature, shall be signed or initialed by the person signing the bid.
- 22.3 Any amendments such as interlineations, erasures, or overwriting shall be valid only if they are signed or initialed by the person signing the Bid.

D. Submission and Opening of Bids

23. Sealing and Marking of Bids

- 23.1 Unless otherwise **specified in BDS**, Bidders shall submit their bids by electronic or by mail/ by hand/ by courier. Procedures for submission, sealing and marking are as follows:
 - (a) Bidders submitting bids by mail, by hand or by Courier shall enclose the original of the Technical Bid, and the original of the Price Bid and each copy of the Technical Bid and Price Bid, including alternative bids, if permitted in accordance with ITB 14, in separate sealed envelopes, duly marking the envelopes as "ORIGINAL TECHNICAL BID", "ORIGINAL PRICE BID", "ALTERNATIVE" and "COPY No..... TECHNICAL BID" and "COPY NO..... PRICE BID" as appropriate. These envelopes containing the original and the copies shall then be enclosed in one single envelope.
 - (b) Bidders submitting Bids electronically shall follow the electronic bid submission procedure specified in this clause.
 - i. The bidder is required to register in the e-GP system https://www.bolpatra.gov.np/egp following the procedure specified in e-GP guideline.
 - ii. Interested bidders may either purchase the bidding document from the Employer's office as specified in the Invitation for Bid (IFB) or bidders may download the IFB and bidding document from e-GP system.
 - iii. The registered bidders need to maintain their profile data required during preparation of bids.
 - iv. In order to submit their bids the cost of the bidding document can be deposited as specified in IFB. In addition, electronic scanned copy (.pdf format) of the bank deposit voucher/cash receipt should also be submitted along with the technical bid.
 - v. The bidder can prepare their technical and price bids using data and documents maintained in bidder's profile and forms/format provided in bidding document by Employer. The bidder may submit bids as a single entity or as a joint venture. The bidder submitting bid in joint venture shall have to upload joint venture agreement along with partner(s) Bolpatra ID provided during bidder's registration.

Government of Nepplation

Ministry of Health and population

Director

- vi. Bidders (all partners in case of JV) should update their profile data and documents required during preparation and submission of their technical bids.
- vii. In case of bid submission in JV, the consent of the partners shall be obtained through the confirmation link sent to the registered email address and the partners shall have to acknowledge their confirmation.

The required forms and documents shall be part of technical bids.

No.	Document	Requirement	Remarks
		-	
1.	Letter of Technical	Mandatory	PDF
	Bid		
2.	Bid Security/Bank	Mandatory	PDF
	Guarantee		
3.	Company registration	Mandatory	PDF
	Certificate		
4.	VAT registration	Mandatory	PDF
	Certificate	(for domestic	
5.	Duain aga Dagiatuatian	bidders only) Mandatory	PDF
3.	Business Registration Certificate	Wiandatory	rDr
6.	Tax Clearance	Mandatory	PDF
0.	Certificate/Tax return	(for domestic	ԻՄՐ
	submission	bidders only)	
	evidence/evidence of		
7	time extension	M 1 - 4	DDE
7.	Power of Attorney of	Mandatory	PDF
0	Bid signatory	N 1 4	DDE
8.	Bank Voucher for	Mandatory	PDF
	cost of bid document	3.5 1	PDE
9.	Joint venture	Mandatory in case of	PDF
	agreement	JV Only	
10.	Qualification	Mandatory	PDF
	Documents		
11.	Technical	Mandatory	PDF or Online
	Specification		Forms
12.	Delivery and	Mandatory	PDF or Online
	Completion Schedule		Forms
13.	Additional	Mandatory (If	PDF
L	documents]	any)	

Government of Nepalation — An expension — Ministry of Health and Population — Director

specified in ITB 12.2	
(h)	

The required forms and documents shall be part of price bids.

No.	Document	Requirement	Remarks
1.	Letter of Price Bid	Mandatory	PDF
2.	Completed Price Schedule	Mandatory	Online Forms
3.	Additional Documents specified in ITB 12.3 (c)	Mandatory (If any)	PDF

Note:

- a) The documents specified as "Mandatory" should be included in e-submission and non-submission of the documents shall be considered as non-responsive bid.
- b) Bidders (all partners in case of JV) should verify/update their profile documents as appropriate for the specific bid before submitting their bid electronically.
- viii. After providing all the details and documents, two separate bid response documents i.e technical bids and price bids will be generated from the system. Bidders are advised to download and verify the response documents prior to bid submission.
- ix. For verifying the authentic user, the system will send one time password (OTP) in the registered e-mail address of the bidder. System will validate the OTP and allow bidder to submit their bid.
- x. Electronically submitted bids can be modified and/or withdrawn through system. The bidder may modify their bids multiple times online within bid submission date and time specified in e-GP system. Once a Bid is withdrawn, bidder won't be able to submit another bid response for the same bid.
- xi. The Bidder / Bid shall meet the following requirements and conditions for e-submission of bids;
 - aa) The e-submitted bids must be readable through PDF

Menalin - Hang	
overnment of Popular Director Health and Popular Director Hospital	
Geta, Kailali	reader. bb) The facility for submission of bid electronically through e-submission is to promote transparency, non-discrimination, equality of access, and open competition in the bidding process. The Bidders are fully responsible to use the e- submission facility properly in e-GP system as per specified procedures and in no case the Employer shall be held liable for Bidder's inability to use this facility. cc) When a bidder submits electronic bid through the PPMO e-GP portal, it is assumed that the bidder has
	prepared the bid by studying and examining the complete set of the Bidding documents including specifications, drawings and conditions of contract. 23.2 The inner and outer envelopes shall:
	(a) bear the name and address of the Bidder;
	(b) be addressed to the Purchaser in accordance with ITB 24.1; and
	(c) bear the specific identification of this bidding process indicated in BDS 1.1.
	23.3 The outer envelope and the inner envelope containing Technical Proposal shall bear a warning not to open before the time and date for the opening of Technical Bid in accordance with ITB 27.1.
	23.4 The inner envelope containing the Price Bid shall bear a warning not to open until advised by the Employer in accordance with ITB 27.7
	23.5 If all envelopes are not sealed and marked as required, the Purchaser will assume no responsibility for the misplacement or premature opening of the bid.
24. Deadline for Submission of Bids	24.1 Bids must be received by the Purchaser at the address and no later than the date and time <i>indicated in the BDS</i> . In case of e-submission, the standard time for e-submission is Nepal Standard Time as set out in the server. The e-procurement system will accept the e-submission of

- the server. The e-procurement system will accept the e-submission of bid from the date of publishing of notice and will automatically not allow the e-submission of bid after the deadline for submission of bid.
- 24.2 The Purchaser may, at its discretion, extend the deadline for the submission of Bids by amending the Bidding Document in accordance with ITB 9, in which case all rights and obligations of the Purchaser and Bidders previously subject to the deadline shall thereafter be subject to the deadline as extended. However, the time available to submit bids shall not be less than five (5) days since amendment in

Vernme y of Heal	th and Popplia. Hospital Direct	or
Get	a, Kailali	bidding document.
25.	Late Bids	25.1 The Purchaser shall not consider any Bid - Technical or Price - that arrives after the deadline for submission of Bids, in accordance with ITB 24. Any Bid received by the Purchaser after the deadline for submission of Bids shall be declared late, rejected, and returned unopened to the Bidder.
26.	Withdrawal, or Modification of Bids	26.1 A bidder may withdraw, or modify its bid after it has been submitted either in hard copy or by e-Submission. Once a Bid is withdrawn, bidder shall not be able to submit another bid for this bidding process. Procedures for withdrawal or modification of submitted bids are as follows:
		GoN Funded:
		(i) Bids submitted in hard Copy
		a) Bidders may withdraw or modify its bids by sending a written notice in a sealed envelope, duly signed by an authorized representative, and shall include a copy of the authorization in accordance with ITB 22.2 before 24 hours prior to the last deadline of submission of bid. The corresponding modification of the bid must accompany the respective written notice. All notices must be: (aa) prepared and submitted in accordance with ITB 22 and ITB 23,
		and in addition, the respective envelopes shall be clearly marked "WITHDRAWAL", "MODIFICATION;" and
		(bb) received by the Purchaser 24 hours prior to the deadline prescribed for submission of bids, in accordance with ITB 24.
		Bidders may withdraw or modify its Bid – Technical or Price – after it has been submitted by sending a written notice, duly signed by an authorized representative, and shall include a copy of the authorization in accordance with ITB 22.2. The corresponding modification of the Bid must accompany the respective written notice. All notices must be (aa) prepared and submitted in accordance with ITB 22 and ITB 23, and in addition, the respective envelopes shall be clearly marked "WITHDRAWAL," and "MODIFICATION;" and
		(bb) received by the Employer prior to the deadline prescribed for submission of Bids, in accordance with ITB 24.
		ii) E-submitted bids.
		a) Bidder may submit modification or withdrawal prior to the deadline prescribed for submission of bids through e-GP system by using the forms and instructions provided by the system. Once a Bid is withdrawn, bidder shall not able to submit another bid for the same bid.

Nepal ion	
vernment of Popular of Health and Popular	
of Health and pitch of Hea	26.2 Bids requested to be withdrawn in accordance with ITB 26.1 (i) shall be returned unopened to the Bidders after the end of bid opening process.
	26.3 The following provisions apply for withdrawal or modification of the Bids:
	GoN Funded
	(i) In case of bids submitted in hard copy no bid shall be withdrawn or modified in the interval between 24 hours prior time of the deadline for submission of bids and the expiration of the period of bid validity specified by the Bidder on the Letter of Bid or any extension thereof.
	(ii) In case of e-submitted bids no bids shall be withdrawn or modified in the interval between deadline for submission of bids and the expiration of the period of bid validity specified by the Bidder on the bid submission form or any extension there of.
	DP Funded
	No Bid may be withdrawn or modified in the interval between the deadline for submission of Bids and the expiration of the period of bid validity specified by the Bidder on the Letters of Technical Bid and Price Bid or any extension thereof.
	26.4 Except in case of any modification or correction in bid document made by procuring entity, Bidder may submit request for withdrawal or modification only one time.
	26.5 In case of hard copy bid, no bid may be withdrawn if the bid has already been modified; except in case of any modification or correction in bid document by procuring entity.
	26.6 Request for withdrawal or modification must be made through the same medium of submission. Request for withdrawal or modifications through different medium shall not be considered.
27.Bid Opening	27.1 The Purchaser's bid opening committee shall conduct the opening of Technical Bids in public in the presence of bidder or its representative who choose to attend at the address, date and time <i>specified in the BDS</i> . The Price Bids will remain unopened and will be held in custody of the Employer until the specified time of their opening. If the Technical Bid and Price Bid are submitted together in one envelope, the Employer shall reject the entire Bid.
	27.2 The opening committee shall download the e-submitted Technical Bid files. The e-procurement system allows the Purchaser to download the

Government of Nephricon

Ministry of Health and population

Director

e-submitted Technical Bid files (report) only after bid opening date and time after login simultaneously by two members of the Bid Opening Committee.

- 27.3 Electronically submitted Technical Bid shall be opened at first in the same time and date as *specified above*. Electronic Bids shall be opened one by one and read out. The e-submitted Technical Bids must be readable through open standards interfaces. Unreadable and or partially submitted bid files shall be considered incomplete.
- 27.4 Before opening the bids, the opening committee shall separate the envelopes of the bids received after the deadline of bid submission, the envelopes containing an application given for WITHDRAWAL, MODIFICATION of bids and the envelopes of bids duly registered. The bids received after the deadline of submission shall be returned to concerned bidder unopened. Then envelopes "WITHDRAWAL" shall be opened first, read out, and recorded, and the envelope containing the corresponding Bid shall not be opened, but returned to the Bidder. If the withdrawal notice is not accompanied by a copy of the valid authorization pursuant to ITB 22.2, the withdrawal shall not be permitted and the corresponding Bid will be opened. Next, envelopes marked "MODIFICATION" shall be opened, read out, and recorded with the corresponding Bid. No Bid shall be modified unless the corresponding Modification Notice contains a valid authorization to request the modification and is read out and recorded at bid opening. Only envelopes that are opened, read out, and recorded at bid opening shall be considered further. Price Bids, both Original and Modification, will remain unopened in accordance with ITB 27.1.
- 27.5 All other envelopes holding the Technical Bid shall be opened one at a time, and the following read out and recorded: the name of the Bidder; whether there is a modification; the presence of a Bid Security, and any other details as the Purchaser may consider appropriate. Only Technical Bids read out and recorded at bid opening shall be considered for evaluation. No Bid shall be rejected at bid opening except for late bids, in accordance with ITB 25.1.
- 27.6 The opening committee shall prepare a record of the opening of Technical Bids that shall include, as a minimum: the name of the Bidder and whether there is a withdrawal, or modification; and the presence or absence of a Bid Security. The Bidders' representatives who are present shall be requested to sign the record. The omission of a Bidder's signature on the record shall not invalidate the contents and effect of the record. A copy of the record shall be distributed to all Bidders who submitted bids in time, and posted online when electronic bidding is permitted. The Bidders' representatives who are

Government of Nepaliton

Ministry of Health and population

Director

present shall also be requested to sign an attendance sheet.

- 27.7 At the end of the evaluation of the Technical Bids, the purchaser will invite bidders who have submitted substantially responsive Technical Bids and who have been determined as being qualified for award to attend the opening of the Price Bids. The date, time, and location of the opening of Price Bids will be advised in writing by the purchaser. Bidders shall be given at least 15 days' notice for the opening of Price Bids.
- 27.8 The purchaser will notify Bidders in writing who have been rejected on the grounds of their Technical Bids being substantially nonresponsive to the requirements of the Bidding Document and return their Price Bids unopened.
- 27.9 The purchaser shall conduct the opening of Price Bids of all Bidders who submitted substantially responsive Technical Bids, in the presence of Bidders' representatives who choose to attend at the address, on the date, and time specified by the Employer. The Bidder's representatives who are present shall be requested to sign a register evidencing their attendance.
- 27.10 All envelopes containing Price Bids shall be opened one at a time and the following read out and recorded:
 - (a) the name of the Bidder;
 - (b) whether there is a modification;
 - (c) the Bid Prices, including any discounts and alternative offers; and
 - (d) any other details as the purchaser may consider appropriate.

Only Price Bids, discounts, modifications, and alternative offers read out and recorded during the opening of Price Bids shall be considered for evaluation. No Bid shall be rejected at the opening of Price Bids.

27.11 The purchaser shall prepare a record of the opening of Price Bids that shall include, as a minimum, the name of the Bidder, the Bid Price (per lot/package if applicable), any discounts, modifications and alternative offers. The Bidders' representatives who are present shall be requested to sign the record. The omission of a Bidder's signature on the record shall not invalidate the contents and effect of the record.



E. Evaluation and Comparison of Bids

28.Confidentiality	28.1 Information relating to the examination, evaluation, comparison, and post-qualification of Bids, and recommendation of contract award, shall not be disclosed to Bidders or any other persons not officially concerned with such process until publication of the Contract award; thereafter, information will be disclosed in accordance with ITB 43.1.
	28.2 Any attempt by a Bidder to influence the Purchaser in the examination, evaluation, comparison, and post-qualification of the Bids or Contract award decisions may result in the rejection of its Bid.
	28.3 Notwithstanding ITB 28.2, from the time of bid opening to the time of Contract award, if any Bidder wishes to contact the Purchaser on any matter related to the bidding process, it should do so in writing.
29.Clarification of Bids	 29.1 To assist in the examination, evaluation, comparison and post-qualification of the Technical and Price Bids, the Purchaser may, at its discretion, ask any Bidder for a clarification of its Bid. Any clarification submitted by a Bidder with regard to its Bid and that is not in response to a request by the Purchaser shall not be considered. The Purchaser's request for clarification and the response shall be in writing. No change in the prices or substance of the Technical Bid or prices in the Price Bid shall be sought, offered, or permitted, except to confirm the correction of arithmetic errors discovered by the Purchaser in the evaluation of the Bids, in accordance with ITB 36. In case of e-submission of bid, upon notification from the purchaser, the bidder shall also submit the original of documents comprising the Technical and Price Bids as per ITB 12.2 and ITB 12.3 for verification of submitted documents for acceptance of the e-submitted bid. 29.2 If a Bidder does not provide clarifications of its Bid by the date and time set in the purchaser's request for clarification, its Bid may be
20 D	rejected.
30.Deviations, Reservations, and Omissions	30.1 During the evaluation of bids, the following definitions apply:(a) "Deviation" is a departure from the requirements specified in the Bidding Document;
	(b) "Reservation" is the setting of limiting conditions or withholding from complete acceptance of the requirements specified in the Bidding Document; and
	(c) "Omission" is the failure to submit part or all of the information or documentation required in the Bidding Document.

Nepal ion 2	
yof Health and Popular Direct	or
39: Examination of Technical Bid	31.1 The purchaser shall examine the Technical Bid to confirm that al documents and technical information requested in ITB 12.2 have been submitted. If any of these documents or information (except alternative Technical Bid which is optional) is missing, the bid shall be rejected.
	31.2 In case of e-submission bids, the Employer shall confirm that all the documents and information requested in ITB 23.1 have been submitted. If any of these documents or information is missing, the bid shall be rejected.
32.Determination of Responsiveness of Technical	32.1 The Purchaser's determination of the responsiveness of a Bid is to be based on the contents of the Technical Bid itself, as defined in ITB 12.2.
Bid	32.2 A substantially responsive Technical Bid is one that meets the requirements of the Bidding Document without material deviation, reservation, or omission. A material deviation, reservation, or omission is one that,
	(a) if accepted, would:
	(i) affect in any substantial way the scope, quality, or performance of the Goods and Related Services specified in Section V, Schedule of Requirements; or
	(ii) limits in any substantial way, inconsistent with the Bidding Document, the Purchaser's rights or the Bidder's obligations under the proposed Contract; or
	(b) if rectified, would unfairly affect the competitive position of other Bidders presenting substantially responsive bids.
	32.3 The Purchaser shall examine the technical aspects of the bid in particular, to confirm that all requirements of Section V, Schedule of Requirements have been met without any material deviation or reservation.
	32.4 If a bid is not substantially responsive to the requirements of the Bidding Document, it shall be rejected by the Purchaser and may not subsequently be made responsive by correction of the material deviation, reservation, or omission.
	32.5 In case of e-submission bids, the purchaser evaluates the bid on the basis of the information in the electronically submitted bid files. If the Bidde cannot substantiate or provide evidence to establish the information provided in e-submitted bid through documents/ clarifications as per ITB Clause 29.1, the bid shall not be considered for further evaluation.

Nepalion - m	<u></u>
of Health and Popular Direct	or
Geta, Kailali	32.6 In Case, a corruption case is being filed to Court against the Natural Person or Board of Director of the firm/institution /company or any partner of JV, such Natural Person or Board of Director of the firm/institution /company or any partner of JV such bidder's bid shall be excluded from the evaluation, if public entity receives instruction from Government of Nepal.
	32.7 Except in case of e-submission, the Financial Bid of the bidder, which is evaluated as substantially non-responsive in technical bid, shall be returned to the respective bidders.
33.Non-material Non- conformities	33.1 The Purchaser may regard a Bid as responsive even if it contains minor deviations that do not materially alter or depart from the characteristics, terms, conditions and other requirement set forth in the Bidding Document or if it contains errors or oversights that are capable of being corrected without affecting the substance of the Bid.
	33.2 Provided that a Bid is substantially responsive, the Purchaser may request that the Bidder submit the necessary information or documentation, within a reasonable period of time, to rectify non-material non-conformities or omissions in the Bid related to documentation requirements. Requesting information or documentation on such non-conformities shall not be related to any aspect of the price of the Bid. Failure of the Bidder to comply with the request may result in the rejection of its Bid.
	33.3 Provided that a Bid is substantially responsive, the Purchaser shall rectify non-material non-conformities or omissions. To this effect, the Bid Price shall be adjusted, for comparison purposes only, to reflect the price of the missing or non-conforming item or component. The adjustment shall be made using the method indicated in Section III, Evaluation and Qualification Criteria.
	33.4 If small differences are found such as in technical specification, description, feature which does not make the bid to be rejected, then the cost, which is calculated to the extent possible due to such differences, shall be included while evaluating bid.
	33.5 If the value is found fifteen percent more than the quoted amount of the bidder on account of small differences pursuant to ITB 32.4, such bid shall be considered irresponsive in substance and shall not be considered for evaluation.
34.Qualification of the Bidder	34.1 The Employer shall determine to its satisfaction during the evaluation of Technical Bids whether Bidders meet the qualifying criteria specified in Section III (Evaluation and Qualification Criteria).

Meddin Inc	<u>7 . </u>
Py of Health and Popular Director	
Geta HOSP Geta, Kailali	34.2 The determination shall be based upon an examination of the documentary evidence of the Bidder's qualifications submitted by the Bidder, pursuant to ITB 19.1.
	34.3 An affirmative determination shall be a prerequisite for the opening and evaluation of a Bidder's Price Bid. A negative determination shall result into the disqualification of the Bid, in which event the Employer shall return the unopened Price Bid to the Bidder.
35.Examination of Price Bids	35.1 The purchaser shall examine the Price Bid to confirm that all documents and financial information requested in ITB 12.3 have been submitted. If any of these documents or information (except alternative Price Bid which is optional) is missing, the bid shall be rejected.
	35.2 In case of e-submission bids, the purchaser shall confirm that all the documents and financial information requested in ITB 23.1 have been submitted. If any of these documents or information is missing, the bid shall be rejected.
36.Correction of Arithmetical	36.1 During the evaluation of Price Bids, the Purchaser shall correct arithmetical errors on the following basis:
Errors	a) if there is a discrepancy between the unit price and the total price that is obtained by multiplying the unit price and quantity, the unit price shall prevail and the total price shall be corrected, unless in the opinion of the Purchaser there is an obvious misplacement of the decimal point in the unit price, in which case the total price as quoted shall govern and the unit price shall be corrected;
	b) if there is an error in a total corresponding to the addition or subtraction of subtotals, the subtotals shall prevail and the total shall be corrected; and
	c) If there is a discrepancy between the bid price in the Summary of price schedule and the bid amount in item (c) of the Letter of Price Bid, the bid price in the Summary of price schedule will prevail and the bid amount in item (c) of the Letter of Price Bid will be corrected; and
	d) if there is a discrepancy between words and figures, the amount in words shall prevail, unless the amount expressed in words is related to an arithmetic error, in which case the amount in figures shall prevail subject to (a), (b) and (c) above.
	36.2 If the Bidder that submitted the lowest evaluated Bid does not accept the correction of errors, its Bid shall be rejected and the bid security shall be forfiteed.

y of Health and ital Directo	or
37 Conversion to Single Currency	37.1 For evaluation and comparison of Price Bids, all bid prices expressed in the amounts in various currencies shall be converted into Nepalese Rupees using the selling exchange rates established by Nepal Rastra Bank and on the on the date specified in the BDS .
38. Goods manufactured in Nepal to be procured	38.1 If the price of goods manufactured in Nepal, are higher up to fifteen percent than that of manufactured in foreign countries, the goods manufactured in Nepal shall be prefered in the evaluation of the Bids as specified in BDS.
•	38.2 for granting such preference pursuant to 38.1, the bidder must submit the country of origin issued by competent authority.
39.Evaluation and Comparison of	39.1 The Purchaser shall evaluate and compare all substantially responsive Bids to determine the lowest evaluated bid.
Price Bids	39.2 To evaluate a Price Bid, the Purchaser shall only use all the criteria and methodologies defined in this Clause and in Section III, Evaluation and Qualification Criteria. No other criteria or methodology shall be permitted.
	39.3 To evaluate a Price Bid, the Purchaser shall consider the following:
	(a) the bid price as quoted in accordance with ITB 15 as specified in BDS;
	(b) adjustment for correction of arithmetic errors in accordance with ITB 36.1;
	(c) adjustment due to discounts offered in accordance with ITB 15.7;
	(d) adjustment for nonmaterial nonconformities in accordance with ITB 33.3;
	(e) adjustment due to application of the evaluation criteria specified in the BDS from amongst those set out in Section III (Evaluation and Qualification Criteria). These criteria may include factors related to the characteristics, performance, and terms and conditions of purchase of the Goods and Related Services which shall be expressed to the extent practicable in monetary terms to facilitate comparison of bids unless otherwise specified in Section III; and
	(f) adjustment due to the application of a margin of preference in accordance with ITB clause 38.
	39.4 The Purchaser's evaluation of a bid will exclude and not take into account:
	(a) in the case of Goods offered from within Nepal, all sales tax and all

Moral of Harl	
overnment of the popular overnment of Health and popular of Health and popular over Hospital Director	
Geta, Kailali	other taxes, applicable in Nepal and payable on the Goods if the Contract is awarded to the Bidder;
	(b) in the case of Goods offered from outside Nepal, all customs duties, sales tax, and other taxes, applicable in Nepal and payable on the Goods if the Contract is awarded to the Bidder; and
	(c) any allowance for price adjustment during the period of performance of the Contract, if provided in the Bid.
	39.5 If this Bidding Document allows Bidders to quote separate prices for different lots/packages, and to award multiple Contracts to a single Bidder, the methodology to determine the lowest evaluated price of the Contract combinations, including any discounts offered in the Letter of Price Bid, is specified in Section III (Evaluation and Qualification Criteria).
	39.6 In Case, a corruption case is being filed to Court against the Natural Person or Board of Director of the firm/institution /company or any partner of JV, such Natural Person or Board of Director of the firm/institution /company or any partner of JV such bidder's bid shall be excluded from the evaluation, if public entity receives instruction from Government of Nepal.
40.Purchaser's Right to Accept Any Bid, and to Reject Any or All Bids	40.1 The Purchaser reserves the right to accept or reject any Bid, and to cancel the bidding process and reject all Bids at any time prior to Contract award, without thereby incurring any liability to the Bidders.

F. Award of Contract

41.Award Criteria	41.1 The Purchaser shall select to award the Contract to the Bidder whose offer has been determined to be the lowest evaluated Bid and is substantially responsive to the Bidding Document, provided further that the Bidder is determined to be qualified to perform the Contract satisfactorily.					
42.Purchaser's Right to Vary Quantities at Time of Award	42.1 At the time the Contract is awarded, the Purchaser reserves the right to increase or decrease the quantity of Goods and Related Services originally specified in Section V, Schedule of Requirements, provided this does not exceed the percentages <i>indicated in the BDS</i> , and without any change in the unit prices or other terms and conditions of the Bid and the Bidding Document.					
43.Notification of Intention to Award	43.1 The Purchaser shall notify the concerned Bidder whose bid has been selected in accordance with ITB 41.1 within seven days of the selection of the bid, in writing that the Purchaser has intention to					

word ion In	7
overnment of Popular Director of Health and Popular Director	A MI
Geta, Kailali	accept his/her bid and shall Inform via the Letter of Intention included in the Contract Forms and the information of name, address and amount of selected bidder shall be given to all other bidders who submitted the bid.
	43.2 If no bidder submits an application pursuant to ITB 46.1 within a period of seven days of providing the notice under ITB 43.1, the Purchaser shall accept the bid selected in accordance with ITB 41.1 prior to the expiry of bid validity period, and notification of award shall be communicated to the bidder to furnish the performance security and sign the contract within fifteen days.
	43.3 In Case, a corruption case is being filed to Court against the Natural Person or Board of Director of the firm/institution /company or any partner of JV, such Natural Person or Board of Director of the firm/institution /company or any partner of JV such bidder's bid shall be excluded from the evaluation, if public entity receives instruction from Government of Nepal.
44.Performance Security	44.1 Within fifteen (15) days of the receipt of notification of award from the Purchaser, the successful Bidder shall furnish the Performance Security in accordance with the GCC, as specified below from Commercial Bank or Financial Institution eligible to issue Bank Guarantee as per prevailing Law in Nepal using Sample Form for the Performance Security included in Section VII (Contract Forms or another form acceptable to the Purchaser.
	i) If bid price of the bidder selected for acceptance is up to 15 (fifteen) percent less than the approved cost estimate, the performance security amount shall be 5 (five) percent of the bid price.
	ii) For the bid price of the bidder selected for acceptance is more than 15 (fifteen) percent below of the cost estimate, the performance security amount shall be determined as follows:
	Performance Security Amount = [(0.85 x Cost Estimate –Bid Price) x 0.5] + 5% of Bid Price.
	The Bid Price and Cost Estimate shall be exclusive of Value Added Tax.
	44.2 Failure of the successful Bidder to submit the above-mentioned Performance Security or sign the Contract Agreement shall constitute sufficient grounds for the annulment of the award and forfeiture of the Bid Security and black listing. In that event the Purchaser shall award the Contract to the next lowest evaluated Bidder whose offer is substantially responsive and is determined by the Purchaser to be qualified to perform the Contract satisfactorily.

Periment of Nephrital Pirec	7 <u>-</u>
45. Signing of Contract	45.1 The successful Bidder shall sign the contract in the form included section VIII after the submission of performance security in accordar with ITB 44.
	45.2 At the same time, the Employer shall affix a public notice on the result the award on its notice board and make arrangement for causing so notice to be affixed on the notice board also of the <i>District Coordinate Committee</i> , <i>District Administration Office</i> , <i>Provincial Treasury a Controller Office and District Treasury and Controller Office</i> . Employer may make arrangements to post the notice into its website it has; and if it does not have, into the website of the Public Procurem Monitoring Office, identifying the bid and lot/package numbers and following information: (i) the result of evaluation of bid; (ii) date publication of notice inviting bids; (iii) name of newspaper; (reference number of notice; (v) item of procurement; (vi) name a address of bidder making contract and (viii) contract Price.
	45.3 The Purchaser shall promptly respond in writing to any unsuccess Bidder who, within thirty days from the date of issuance of notificate pursuant to ITB 43.1, requests in writing the grounds on which its bid was not selected.
	45.4 If the bidder whose bid has been accepted fails to sign the contract as sta ITB 45.1, the Public Procurement Monitoring Office shall blacklist bidder on recommendation of the Public Entity.
46.Complaint and Review	46.1 If a Bidder dissatisfies with the Procurement proceedings or decision made by the Purchaser in opening of the price bid or intention to award the Contract, it may file an application to the Ch of the concerning Public Entity of the Purchaser within seven (7) day of providing the notice under ITB 27.8 and ITB 43.1 by the Public Entity, for review of the proceedings stating the factual and legrounds.
	46.2 An application filed after the deadline pursuant ITB 46.1 shall not processed.
	46.3 The chief of Public Entity of the Purchaser shall, within five (5) da after receiving the application, give its decision with reasons, writing pursuant to ITB 46.1:
	(a) whether to suspend the procurement proceeding and procedure for further proceedings to be adopted; or

(b) whether or not to reject a application.

No application can be submitted before the Review Committee for review against the decision made by the chief of the Public Entity for Government of Nephricon

Ministry of Health and population

Director

the Bid amount up to the value as stated in BDS.

- 46.4 If the Bidder is not satisfied with the decision of the Public Entity in accordance with ITB 46.3, or the decision by the Public Entity is not given within five (5) days of receipt of application pursuant to ITB 46.1, it can, within seven (7) days of receipt of such decision, file an application to the Review Committee of the GoN, stating the reason of its disagreement on the decision of the chief of Public Entity and furnishing the relevant documents, provided that its Bid amount is above the amount as stated in ITB 46.3. The application may be sent by hand, or by post, or by courier, or by electronic media at the risk of the Bidder itself.
- 46.5 Late application filed after the deadline pursuant to ITB 46.4 shall not be processed.
- 46.6 Within three (3) days of the receipt of application from the Bidder, pursuant to ITB 46.4, the Review Committee shall notify the concerning Public Entity of the Purchaser to furnish its procurement proceedings and comments on the issue, pursuant to ITB 46.3.
- 46.7 Within three (3) days of receipt of the notification pursuant to ITB 46.6, the Public Entity shall furnish the copy of the related documents along with its comment or reaction of complaint to the Review Committee.
- 46.8 The Review Committee, after inquiring from the Bidder and the Public Entity, if needed, shall give its decision within one (1) month after receiving the application filed by the Bidder, pursuant to ITB 46.4.
- 46.9 The Bidder, filing application pursuant to ITB 46.4, shall have to furnish a cash amount or Bank guarantee from Commercial Bank or Financial Institution eligible to issue Bank Guarantee as per prevailing Law *as stated in BDS* with the validity period of at least ninety (90) days from the date of the filing of application pursuant to ITB 46.4. Application filed without furnishing the security deposit shall not be processed.
- 46.10 If the claim made by the Bidder pursuant to ITB 46.4 is justified, the Review Committee shall have to return the security deposit to the applicant, pursuant to ITB 46.9, within seven (7) days of such decision made.
- 46.11 If the claim made by the Bidder pursuant to ITB 46.4 is rejected by the Review Committee, the security deposit submitted by the Bidder pursuant to ITB 46.9 shall be forfeited.



Section II. Bid Data Sheet

	A. Introduction
ITB 1.1	The number of the Invitation for Bids (IFB) is: GH/G/ICB-06/2081-82
ITB 1.1	Name of the Purchaser:
	Government of Nepal
	Ministry of Health and Population
	Geta Hospital, Geta Kailali
ITB 1.1	Name and Identification number of the Contracts (Packages/Lots): GH/G/ICB-06/2081-82/Supply, Installation and Commissioning of Infertility Treatment Equipments
ITB 2.1	Source of Fund: GoN Funded
	Name of contract/s: Supply, Installation and Commissioning of Infertility Treatment Equipments
ITB 4.1	Bidders from the following countries are not eligible:
	Not Applicable
ITB 4.2	For GoN Funded:
	Maximum number of partner in a joint venture shall be : 3 (three)
ITB 4.4	A list of debarred firms is available at http://www.ppmo.gov.np
ITB 4.9	The foreign Bidder at the time of bid submission:
	 shall submit Original Manufacture Authorization Letter
	shall declare to submit at the time of contract agreement:
	 TAX Registration Certificate.
ITB 5.1	Goods and related services to be supplied from following countries are not eligible: <i>Not Applicable</i>
	B. Bidding Document
ITB 8.1	For clarification purposes only, the Purchaser's address is:
	Name of the Purchaser: Geta Hospital
	City/Town: <i>Geta</i>
	District: Kailali

y of Health and Poppa Seta Hospital Directo							
Geta, Kan	Country: Nepal						
	Telephone:						
	Electronic Mail Address: procurement.getahospital@gmail.com						
ITB 8.1	The purchaser will respond in writing to any request for clarification provided that such request is received no later than 10 days prior to the deadline date for submission of bid.						
ITB 8.2	Pre-Bid meeting "shall" be organized at the following date, time and place:						
	Date: 2081/01/30 (13 May 2024)						
	Time: 13:00 Hour						
	Place: Geta Hospital, Geta Kailali						
	C. Preparation of Bids						
ITB 11.1	The language of the Bid is: English						
ITB 12.2 (h)	The Bidder shall submit the following additional documents with its Bid: (i) Manufacturer's Authorization Letter OR Authorized Dealer's Authorization Letter						
	(ii) Catalogue and Data Sheets						
	(iii) Required Quality Certificates (ISO/CE/USFDA) as mentioned in Technical Specification						
ITB 12.3 (c)	The Bidder shall submit with its Price Bid the following additional documents: <i>N/A</i>						
ITB 14.1	Alternative Bids shall not be permitted						
ITB 15.7	The Incoterms edition is: 2020						
ITB 15.7 (a) iii, (b) ii & (c) v	Final Destination (Project Site): Geta Hospital, Kailali, Nepal						
ITB 15.7 (b) i	For Goods offered from outside the Purchaser's country, the Bidder shall quote prices using the following Incoterm: <i>CIP</i>						
ITB 15.7 (b) iii	In addition to the CIP price specified in ITB 15.7 (b)(i), the price of the Goods manufactured outside Nepal may be quoted: <i>Geta Hospital, Kailali, Nepal</i>						
ITB 15.8	The prices quoted by the Bidder shall: not be Adjustable.						
ITB 18.3	Period of time the Goods are expected to be functioning (for the purpose of spare parts): 2 years						
ITB 19.2	(i) Manufacturer's Authorization Letter OR Authorized Dealer's Authorization Letter is required for all the items listed in Section V						

Geta, Kailall	Schedule of Requirements
ITB 19.3	The Bidder <i>is</i> required to include with its bid, evidence that it will be represented by an Agent in Nepal.
ITB 20.1	The bid validity period shall be 90 days.
ITB 21.1	The bid must be accompanied by bid security, amounting to a minimum of <i>NPR</i> . 658,000 which shall be valid for minimum 30 days beyond the bid validity period (i.e. 2082/06/20 (06 October 2025).
ITB 21.2	If the Bidder wishes to submit the Bid Security in the form of cash, the cash should be deposited in Deposit Account No.: Name of the Bank: Rastriya Banijya Bank, Dhangadhi Name of the Office: Geta Hospital Geta, Kailali Office Code No.: 370617701 Account No.: 41701001020300000 Name of Account: Kosh Tatha Lekha Niyntrak Karyalaya. Kailali (Dharauti Khata)
	and submit the receipt of the deposited amount of cash along with the bid.
ITB 22.2	The written confirmation of Authorization to sign on behalf of the Bidder shall consist of: Power of Attorney to sign the bid
	D. Submission and Opening of Bids
ITB 23.1	Bidders shall have the option of submitting their bids by electronic only
ITB 24.1	For bid submission purposes only, the Purchaser's address is: <i>Geta Hospital Geta, Kailali, Nepal</i>
ITB 24.1	The deadline for bid submission is: Date: 2082/02/23 (06 June 2025) Time: 12.00 Hours
ITB 24.1	If the last date of purchasing, submission and opening of Bid falls on a government holiday then the next working day shall be considered as the last day without any change in the time and place as fixed.
ITB 27.1	The bid opening shall take place at: Geta Hospital Geta, Kailali, Nepal Date: 2082/02/23 (06 June 2025)
	Time: 13.00 Hours
	E. Evaluation and Comparison of Bids
ITB 37.1	The date for the selling exchange rate shall be: 2082/02/23 (06 June 2025)
ITB 38.1	Domestic preference shall apply and the application methodology shall be as stipulated in Section III (Evaluation and Qualification Criteria).
ITB 39.3 (a)	Bids will be evaluated lot by lot/package (may be multiple or single). If a Price Schedule shows items listed but not priced, their prices shall be

try of Health and popular Dire	ector						
Geta, Kailali	assumed to be included in the prices of other items. An item not listed in the Price Schedule shall be assumed to be not included in the bid, and provided that the bid is substantially responsive, the average price of the item quoted by substantially responsive bidders will be added to the bid price and the equivalent total cost of the bid so determined will be used for price comparison.						
ITB 39.3 (e)	The adjustments shall be determined using the following criteria, from amongst those set out in Section III, Evaluation and Qualification Criteria:						
	(a) Deviation in Delivery schedule: <i>No</i>						
	(b) Deviation in payment schedule: No						
	(c) the cost of major replacement components, mandatory spare parts, and service: <i>No</i>						
	(d) the availability of spare parts in Nepal and after-sales services for the equipment offered in the bid : <i>No</i>						
	(e) the projected operating and maintenance costs during the life of the equipment: <i>No</i>						
	(f) the performance and productivity of the equipment offered: No						
	F. Award of Contract						
ITB 42.1	The maximum percentage by which quantities may be increased is: 15%						
	The maximum percentage by which quantities may be decreased is: 15%						
ITB 46.3	No application can be submitted before the Review Committee for review against the decision made by the chief of the Public Entity for the bid amount less than the value of Twenty Million (NRs. 20,000,000)						
ITB 46.9	The bidder, filling application pursuant to ITB 46.4, shall have to furnih a cash amount of Bank guarantee equal to ten percent (10 %) of amount of bid security in case of complaint against decision pursuant to ITB 27.8 and one percent (1%) of its quoted Bid amount in case of complaint against decision pursuant to ITB 43.1						



Section III. Evaluation and Qualification Criteria

The purpose of the Evaluation and Qualification Criteria (EQC) is to specify the criteria that the Purchaser will use to evaluate the Bids to determine the lowest evaluated substantially responsive bid and post-qualify the lowest-evaluated Bidder. The Purchaser must prepare the EQC and include it as a part of the Bidding Document. The EQC is not a Contract document and, therefore, it is not a part of the Contract.

Table of Criteria

- 1. Evaluation Criteria
 - 1.1 Technical Criteria
 - **1.2Domestic Preference**
 - 1.3 Economic Criteria
 - **1.4Multiple Contracts**
- 2. Qualification Criteria



Evaluation Criteria

Criteria for Bid evaluation are to be determined case by case basis. Select from following the appropriate criteria according to the provisions specified in ITB 39. Retain only the relevant parameters and evaluation methods to apply corresponding to the retained criteria.

1.1 Technical Criteria

These criteria should specify the minimum technical level that the Goods and Related Services shall have in order to comply with the Section V. Schedule of Requirements. Whenever possible, these criteria should be evaluated on a pass–fail system, with a minimum acceptable level for each criteria enumerated.

However, a minor deficiency in technical compliance may not be cause for rejection of the Bid. The cost of making good any deficiency should likewise be added to the Bid Price concerned. The most frequently used methods assign to the non-conforming items or components, prices based on similar methods described above under Scope, with the price of the nonconforming items or components deducted.

1.2 Domestic Preference

If the BDS so specifies, the Procuring Entity will grant a margin of preference to goods manufactured in Nepal. The nationality of the manufacturer or supplier is not a condition for such eligibility. The methods and stages set forth hereunder shall be followed in the evaluation and comparison of bids.

For comparison, responsive bids shall be classified in one of the following three groups:

- (a) **Group A:** bids exclusively offering goods manufactured in for which (i) labor, raw material, and component from within Nepal account for 30 percent or more of the EXW price of the product offered, and (ii) the production facility in which those goods will be manufactured or assembled has been engaged in manufacturing/ assembling such goods at least since the time of bid submission.
- (b) **Group B:** all other bids offering goods manufactured in Nepal.
- (c) **Group C:** bids offering goods manufactured outside Nepal that have been already imported or that will be directly imported.

Insert **one** of the following adjustment methods.

- 1. The price quoted for goods in bids of groups A and B shall include all duties and taxes paid or payable on the basic materials or component purchased in the domestic market or imported, but shall exclude the value added tax and similar taxes on the finished product. The price quoted for goods in bids of group C shall be on CIP (place of destination), which is exclusive of customs duties andother import taxes already paid or to be paid.
- 2. In the first step, all evaluated bids in each group shall be compared to determine the lowest bid in each group. Such lowest evaluated bids shall be compared with each other and if, as a result of this comparison, a bid from group A or group B is the lowest, it shall be selected for the award.
- 3. If as a result of the comparison under paragraph three above, the lowest evaluated bid is a bid from group C, the lowest evaluated bid from group C shall be further compared with the lowest evaluated bid from group A after adding to the evaluated price of goods offered in the bid from group C, for the purpose of this further comparison only, an amount equal to fifteen (15%) percent of the CIP bid price. The lowest evaluated bid determined from this last comparison shall be selected.

1.2.2 Method B

- 1. In supply, delivery and installation contracts in which there are a number of items of Goods and Related Services, a margin of preference shall be granted to eligible domestically produced Goods in accordance with the following provisions.
 - (a) The preference margin shall not be applied to the whole package but only to the eligible domestically produced Goods within the package;
 - (b) Goods offered from outside Nepal shall be quoted CIF or CIP (place of destination), which is exclusive of customs duties and other import taxes already paid or to be paid; and Goods offered locally shall be quoted EXW (ex-works free of sales and similar taxes) which shall include all duties and taxes paid or payable on the basic materials or component purchased in the domestic market or imported, but shall exclude the value added tax and similar taxes on the finished product
 - (c) all other cost components for services and works such as costs for local handling, transportation, storage, installation, and commissioning shall be quoted separately;
 - (d) in the comparison of Bids, only the CIF or CIP price in each Bid of the Goods offered from outside Nepal shall be increased by fifteen (15%) percent;
 - (e) no margin of preference shall be applied to any associated services or works



(f) Bidders shall not be permitted or required to modify the mix of local and foreign Goods after bid opening.

Method A

1.3 Economic Criteria

The economic criteria are most important when evaluating a Bid. In most cases, they are the only criteria for evaluating Bids that have passed the technical evaluation. Price, however, may not be the only criterion, as there could be other criteria that may be expressed in monetary terms. For energy consuming equipment and facilities, adjustment for efficiency over and above the minimum functional guarantees specified in the specifications (e.g. generators, pumps), losses (e.g. transformers), and future operating costs of the equipment may be taken into account in the determination of the evaluated Bid Price. The financial cost for these adjustments (added to or deducted from the Bid Price as the case may be) shall be made only when it is specified in the Bidding Document that these functional guarantees and projected operating costs are factors in bid evaluation. The methods of calculation for these evaluation factors shall be clearly specified in the Bidding Document. Deviations from the specified manner of cost calculation shall not be introduced.

Examples of such criteria appear below:

1.3.1 Adjustment for Scope

1.3.1.1 Local Handling and Inland Transportation

If the Procuring Entity wishes to consider during bid evaluation the costs for inland transportation, insurance, and other services within Nepal incidental to delivery of the goods to their final destination, then the Procuring Entity must define these service items in Section V. Schedule of Requirements, List of Goods and Related Services; and

Bidders must be required to quote for these service items as part of their bid price in the PRICE SCHEDULE FOR RELATED SERVICES TO BE OFFERED FROM OUTSIDE AND WITHIN NEPAL provided in Section IV, Bidding Forms.

In such case, the following provision should be used:

CIP price.

1.3.1.2 .Minor Omissions or Missing Items

The cost of minor omissions or missing items in the scope of supply, services, etc. should be added to the Bid Price to allow for Bid comparison on an equal basis. The price adjustment should be

Government of Nepalion

Ministry of Health and population

Director

based on a reasonable estimate of the cost by the executing agency, engineer, consultant or bid evaluation committee, taking into consideration the corresponding quoted prices from other conforming Bids. The price adjustment should be based on the fair price of the omitted item. The most frequently used methods assign to the missing item a price:

- (i) equal to the highest price quoted for the same item by the other Bidders; or
- (ii) equal to the average price quoted for the same item by the other Bidders; or
- (iii) estimated by the Procuring Entity.

Of these three methods the Procuring Entity should preferably use (i) or (ii), as Bidders frequently challenge (iii) because of its lack of transparency.

Not Applicable

1.3.2 Adjustment for Deviations from the Terms of Payment

The Procuring Entity must state here whether deviations from the terms of payment as specified in Special Conditions of Contract, Sub-Clause 15.1, are permitted or not. If permitted, the Procuring Entity shall evaluate deviations from the terms of payment in the following manner. The Procuring Entity shall first evaluate the Bids based on the terms of payment specified in the Special Conditions of Contract, Clause 15.1. The Procuring Entity shall then add an adjustment to the Bid Price to take into account the differences in cash flows. The adjustment shall be calculated as the discounted cash flow of the incremental payments of the alternative compared with those of the terms of payment specified by the Procuring Entity.

Deviations from the Terms of Payment as specified in SCC 15.1 are not permitted.

1.3.3 Adjustment for Deviations in the Delivery and Completion Schedule

Bidders are required to base their prices on the Delivery and Completion Schedule specified in Section V. Schedule of Requirements. The Procuring Entity must state here whether deviations from the specified Delivery and Completion Schedule are permitted or not. If permitted, the Procuring Entity shall evaluate deliveries by adding the corresponding price adjustment in accordance with the procedure outlined below.

Deviations from the Delivery and Completion Schedule specified in Section V (Schedule of Requirements) are not permitted.

1.3.4 Operating and Maintenance Costs

The Operating and Maintenance costs (O&M) need to be taken into account for bid evaluation

Government of Nepal Honor Commission of Health and Option Director

purposes when such costs over the life cycle of the Goods represent an important cost in relation to the capital or investment cost of the Goods. Different technologies may involve large variations in the capital costs of the Goods and the costs associated with their O&M. Normally, more elaborate technologies and materials used in the manufacturing of the Goods involve higher investment costs and lower O&M costs. O&M costs are evaluated at their present value over the life cycle of the Goods and then added to the price of the Goods.

Typical O&M cost factors for calculation are:

- (a) Number of years for initial period of operation. It is recommended that the initial period of operation does not exceed the usual period before a major overhaul of the Goods is required, usually between five and ten years. The load or working cycle (hourly, daily, monthly, seasonal) of the Goods shall be as specified by the Procuring Entity).
- (b) Operating costs such as fuel, electricity, spare parts, labour and/or other inputs required for the operation of the Goods.
- (c) Rate (normally prevailing commercial interest rate), in percent, to be used to discount to present value, all of the annual future costs calculated under (b) above for the period specified in (a).

[Insert procedures to be used]

1.3.5 Spare Parts and after Sales Service Facilities

Only those spare parts and tools which are specified on an item-wise basis in the List of Goods and Related Services in Section V. Schedule of Requirements, shall be taken into account in the bid evaluation. Supplier-recommended spare parts for a specified operating requirement shall not be considered in bid evaluation.

Not Applicable

1.3.6 Performance and Productivity of the Goods

The adjustment factor for the performance or productivity of the Goods shall be calculated based on the difference between the reference value or norm (i.e. the efficiency) as specified in Section V. Schedule of Requirements, and the corresponding value guaranteed by the Bidder in its Bid. Choose and insert one of the following:

(i) Performance and productivity of the equipment. An adjustment representing the capitalized cost of additional operating costs over the life of the plant will be added to the bid price, for

Government (Nevalion Proposition Propositi

Geta, Kallal evaluation purposes if specified in the BDS Sub-Clause 39.3 (e). The adjustment will be evaluated based on the drop in the guaranteed performance or efficiency offered in the bid below the norm of 100, using the methodology specified in BDS Sub-Clause 39.3(e).

or

(ii) An adjustment to take into account the productivity of the goods offered in the bid will be added to the bid price, for evaluation purposes only, if specified in BDS Sub-Clause 39.3 (e). The adjustment will be evaluated based on the cost per unit of the actual productivity of goods offered in the bid with respect to minimum required values, using the methodology specified in BDS Sub-Clause 39.3 (e).

1.3.7 Specific additional Criteria

Other specific additional criteria to be considered in the evaluation, and the evaluation method shall be detailed in BDS sub douse 39.3(e)

1.4 Multiple Contracts

If Goods and Related Services are grouped in multiple lots/packages, the following provision must be used: *Not Applicable*

If a Bidder submits successful Bids for multiple lots/packages (lowest evaluated substantially responsive Bids), the evaluation will also include an assessment of the Bidder's capacity to meet the aggregated qualifying requirements relating to [Purchaser to list here the qualifying requirements set for individual contracts/lots/packages, i.e., production capacity, size of operation].



Qualification Criteria

Except Qualification Requirements mentioned as optional, which may be specified as per requirements, the Procuring Entity shall specify the following Qualification Requirements without any substantial deviation.

2.1 Eligibility

Criteria	Compliance Requirements				Documents
Requirement	Single Entity	All Partners Combined	Joint Venturo Each Partner	One Partner	Submission Requirements

2.1.1 Conflict of Interest

No conflicts of interest in accordance with ITB 4.3.	Must meet requirement	Not applicable	Letter of Technical Bid

2.1.2 Government/DP Eligibility

2.1.3 Government-Owned Entity

Bidder required to meet conditions of ITB 4.5.		Must meet requirement	Must meet requirement	Not applicable	Letter of Technical Bid; Forms ELI – 1 and ELI - 2
--	--	-----------------------	-----------------------	-------------------	---

2.1.4 UN Eligibility

Not having been excluded by an act of compliance with a United	Must meet requirement	Must meet requirement	Not applicable	Technical Bid Submission Sheet
Nations Security Council				
resolution in accordance with ITB				
4.7.				

2.1.5 Nationality

Nationality in accordance with ITB 4.8.		ll .	Must meet requirement	Not applicable	Letter of Technical Bid; Forms ELI – 1 and ELI – 2 with attachments
---	--	------	-----------------------	-------------------	---



Firm Registration Certificate	must meet requirement	not applicable	must meet requirement	not applicable	Document attachment
Business Registration Certificate	must meet requirement	not applicable	must meet requirement	not applicable	Document attachment
VAT and PAN Registration certificate (only for domestic bidders)	must meet requirement	not applicable	must meet requirement	not applicable	Document attachment
Tax Clearance Certificate/Tax return submission evidence/ evidence of time extension for the <i>F/Y 2080-81 (Only for domestic bidders)</i>	must meet requirement	not applicable	must meet requirement	not applicable	Document attachment

2.2 Pending Litigation.

Criteria		Compliance Requirements				
	Entity		Submission			
Requirement		All Partners Combined	Each Partner	One Partner	Requirements	
All pending litigation, arbitration or other material events impacting the net worth and/or liquidity of the bidder, if any, shall be treated as resolved against the Bidder and so shall in total not represent more than 50 (Fifty) percent of the Bidder's net worth calculated as the difference between total assets and total liabilities.	Must meet requirement	Not applicable	Must meet requirement	Not applicable	Form LIT - 1	

2.3 Financial Situation

2.3.1 Historical Financial Performance

Criteria		Compliance Requirements				
			Submission			
Requirement	Single Entity	All Partners Combined	Each Partner	One Partner	Requirements	
Submission of audited balance sheets and income statements for the <i>last 3 years</i> to demonstrate the current soundness of the Bidder's	Must meet requirement	Not applicable	Must meet requirement	Not applicable	Form FIN – 1	

	11 220	
Medal ion	· = 4 = 1 .	
	Aller A	
"Inistry of Health and ital	Director	
	Director Criteria	ı

Criteria		Documents		
Requirement	Single Entity	Joint Ventur	Submission	
financial position. As a minimum, the Bidder's net worth for the last year calculated as the difference between total assets and total liabilities should be positive.				

2.3.2 Average Annual Turnover

Criteria		Documents			
Doguinoment	Single Entity	Joint Venture			Submission
Requirement		All Partners Combined	Each Partner	One Partner	Requirements
Minimum average annual turnover of <i>NRs. 19,800,000</i> calculated as total payments received by the Bidder for supply contracts completed or under execution over the last three years.	Must meet requirement	Must meet requirement	Must meet 25 % of the requirement	Must meet 40% of the requirement	Form FIN – 2

Only the net amount shall be calculated after deducting the amount for VAT and such amount shall be adjusted to present value by applying wholesale price index of Nepal Rastra Bank.



2.3.3 Financial Resources

Criteria		Documents			
Requirement	Doguinoment Single Entity		Joint Venture All Partners Each One		
Requirement	Single Entity	Combined	Each Partner	Partner	Requirements
The Bidder must demonstrate access to, or availability of, financial resources such as liquid assets ² unencumbered real assets, and other financial resources, (other than any contractual advance payments) to meet the cash-flow requirement of <i>NRs</i> . 19,800,000	Must meet requirement	Must meet requirement	Must meet 25 % of the requirement	Must meet 40 % of the requirement	Form FIN - 3

2.4 Experience

2.4.1 General Experience

Criteria	Compliance Requirements				Documents	
		Joint Venture				
Requirement	Single Entity	All Partners Combined	Each Partner	One Partner	Submission Requirements	
Experience under supply contracts in the role of prime supplier (single entity or JV member) or subcontractor for at least the last 1 (One) year prior to the applications submission deadline.	Must meet requirement	Not applicable	Must meet requirement	Not applicable	Form EXP – 1	

Liquid Assets mean cash and cash equivalents, short-term financial instruments, short term available-for-sale-securities, marketable securities, trade receivables, short-term financing receivables and other assets that can be converted into cash within ONE YEAR.



2.4.2 Specific Experience : Not Applicable

Criteria	Compliance Requirements				Documents
Requirement	Single	J All Partners	Joint Venture Each	One	Submission
requirement	Entity	Combined	Partner Partner	Partner	Requirements
Experience under supply contracts in the role of prime supplier (single entity or JV member) or subcontractor in at least	Must meet requirement	Must meet requirement	Not applicable	Not applicable	Form EXP – 2

Only the net amount shall be calculated after deducting the amount for VAT and such amount shall be adjusted to present value by applying wholesale price index of Nepal Rastra Bank.

2.4.3 Technical Experience ³ *Not Applicable*

	Criteria		Compliance F	Requirements		Documents
		Single	Joint Venture			Submission
	Requirement	Entity	All Partners Combined	Each Partner	One Partner	Requirements
	Bidder shall demonstrate that the ds offered have:	Must meet requirement	Must meet requirement	Not applicable	Not applicable	Form EXP – 3
(i)	been in production for at least(3) years or if the offered model is a new the manufacturer must have experience in producing the similar model for a mínimum of(3)years;					
	been sold a minimum of(4) units of similar type and specification over the last three years; and been in operation for a minimum of(3) years ⁴ .					

³ For procurement that involves high-value and technically complex items only

⁴ The bidder shall furnish a list of users who had purchased same/similar goods/equipment during the period stated,



Note:

- (3) Insert value 1 (one) to 3 (three) years
- (4) Insert number of units which is 3 x Q/T, where Q is the value which is 80 % of the number of units specified in Price Schedule (in above round figure), T is contract duration in years. Contract duration less than one year shall be considered one year. The multiplier 3 (three) can be reduced to 1 (one) in accordance with complexity of the supply.

2.3.3 Production Capacity 5 Not Applicable

Criteria	Compliance Requirements			Documents	
Requirement	Single Entity	J All Partners Combined	oint Venture Each Partner	One Partner	Submission Requirements
The Bidder or manufacturer shall demonstrate ⁶ that it can supply the type, size, and quantity of the goods as required by Purchaser in accordance with the Delivery and Completion Schedule in Section V (Schedule of Supply).	Must meet requirement	Must meet requirement	Not applicable	Not applicable	Form EXP – 4

number of equipment sold to them and the contract amount. They will be used as references to check the performance of the offered model, if necessary.

⁶ Bidder or Manufacturer shall provide evidence of production output.

⁵ For procurement that involves high-value and technically complex items only



Section IV. Bidding Forms

Table of Forms

Bid Submission Form	. 63
Bidder's Information Form	67
Joint Venture Information Form	. 68
Pending Litigation Form	. 69
Financial Situation Form	. 70
Average Annual Turnover Form	. 72
Financial Resources Form	. 73
General Experience Form	. 74
Specific Experience Form	. 75
Technical Experience Form	. 76
Production Capacity Form	. 77
Bidder's Proposed Specification Form	. 78
Price Schedule For Goods	. 79
Bid Security	. 84
Manufacturer's Authorization Letter	. 86



Letter of Technical Bid

(The Bidder shall accomplish the Letter of Technical Bid in its Letter Head Clearly showing the Bidders Complete name and address.)

	Date:
	Contract No.:
	Invitation for Bid No.:
То:	

We, the undersigned, declare that:

- (a) We have examined and have no reservations to the Bidding Document, including Addenda issued in accordance with Instructions to Bidders (ITB) Clause 9.
- (b) We offer to supply in conformity with the Bidding Document and in accordance with the delivery schedule specified in the Section V (Schedule of Requirements), the following Goods and Related Services: [insert a brief description of the goods and related services]
- (c) Our Bid consisting of the Technical Bid and the Price Bid shall be valid for a period of [insert validity period as specified in ITB 20.1 of the BDS] days from the date fixed for the bid submission deadline in accordance with the Bidding Document, and it shall remain binding upon us and may be accepted at any time before the expiration of that period;
- (d) Our firm, including any subcontractors or suppliers for any part of the Contract, have nationalities from eligible countries in accordance with ITB 4.8 and meet the requirements of ITB 3.4 & 3.5
- (e) We are not participating, as a Bidder or as a subcontractor/supplier, in more than one Bid in this bidding process in accordance with ITB 4.3(e), other than alternative Bids in accordance with ITB 14;
- (f) Our firm, its affiliates or subsidiaries, including any Subcontractors or Suppliers for any part of the contract, has not been declared ineligible by DP, under the Purchaser's country laws or official regulations or by an act of compliance with a decision of the United Nations Security Council;
- (g) We are not a government owned entity/We are a government owned entity but meet the requirements of ITB 4.5;⁷

⁷ Slect one of the options

- (h) We declare that, we including any subcontractors or suppliers for any part of the contract do not have any conflict of interest in accordance with ITB 4.3 and we have not been punished for an offense relating to the concerned profession or business.
- (i) The following commissions, gratuities, or fees, if any, have been paid or are to be paid with respect to the bidding process or execution of the Contract:

Name of Recipient	Address	Reason	Amount

(If none has been paid or is to be paid, indicate "none.")

- (j) We declare that we are solely responsible for the authenticity of the documents submitted by us. The document and information submitted by us are true and correct. If any document/information given is found to be concealed at a later date, we shall accept any legal actions by the purchaser.
- (k) We agree to permit GoN/DP or its representative to inspect our accounts and records and other documents relating to the bid submission and to have them audited by auditors appointed by the GoN/DP.



Letter of Price Bid

The Bidder must accomplish the Letter of Price Bid in its letterhead clearly showing the Bidder's complete name and address.

	Date:
	Name of the contract:
	Invitation for Bid No.:
То:	
We, the under	rsigned, declare that:
	examined and have no reservations to the Bidding Documents, including Addenda accordance with Instructions to Bidders (ITB) Clause 9;
delivery s	to supply in conformity with the Bidding Document and in accordance with the schedule specified in the Section V (Schedule of Requirements), the following Goods ed Services: [insert a brief description of the goods and related services]
(c) The total 1	price of our Bid, excluding any discounts offered in item (d) below, is:
[Incase o	f only one lot/package, insert the total Bid Price in words and figures];
[Incase o	f multiple lots/packages, insert the total price of each lot/package]
(d) The disco	unts offered and the methodology for their application are:
The disco	ounts offered are: [specify in detail each discount offered]
	t method of calculations to determine the net price after application of discounts is low: [specify in detail the method that shall be used to apply the discounts];
(e) Our bid sl	nall be valid for a period of <i>linsert validity period as specified in ITR 2011</i> days from

- the date fixed for the bid submission deadline in accordance with the Bidding Documents, and it shall remain binding upon us and may be accepted at any time before the expiration of that period;
- (f) If our bid is accepted, we commit to obtain a performance security in accordance with the



- (g) We understand that this bid, together with your written acceptance thereof included in your notification of award, shall constitute a binding contract between us, until a formal contract is prepared and executed;
- (h) We understand that you are not bound to accept the lowest evaluated bid or any other bid that you may receive;
- (i) We agree to permit the Employer/DP or its representative to inspect our accounts and records and other documents relating to the bid submission and to have them audited by auditors appointed by the Employer.
- (j) We confirm and stand by our commitments and other declarations made in connection with the submission of our Letter of Technical Bid.

Name:
In the capacity of
Signed
Duly authorized to sign the Bid for and on behalf of
Date



ELI-1: Bidder's Information Form

[The Bidder shall fill in this Form. No alterations to its format shall be permitted and no substitutions shall be accepted.]

1.	Bidder's Legal Name:	[insert full name]	
2.	In case of JV, legal name of the representative member and of each member:	[insert full name of each member in the JV and specify the representative member]	
3	Bidder's Country of Registration:	[insert country of registration]	
4.	Bidder's Year of Registration:	[insert year of incorporation]	
5.	Bidder's Legal Address in Country of Registration	[insert street/number/town or city/country]	
6.	Bidder's trading address:	[insert street/number/town or city/country]	
7.	Bidder's Telephone/Fax numbers:	[insert telephone/fax numbers, including country and city codes]	
8.	Bidder's Email Address:	[insert email address]	
9.	Bidder's Authorized Representative Information	on:	
	Name:	[insert full name]	
	Address:	[insert street/number/town or city/country]	
	Telephone/Fax numbers:	[insert telephone/fax numbers, including country and city codes]	
	Email Address:	[insert email address]	

Attached are copies of the following documents:

- 1. In case of a single entity, articles of incorporation or constitution and company incorporation/registration of the legal entity named above, in accordance with ITB 4.2 and ITB 4.8
- 2. Authorization to represent the firm or Joint Venture named above, in accordance with ITB 22.2
- 3. In case of a Joint Venture, a letter of intent to form a Joint Venture or Joint Venture agreement, in accordance with ITB 4.2
- 4.In case of a government-owned enterprise, any additional documents not covered under 1 above required to comply with ITB 4.5



ELI-2: Joint Venture Information Form

Each member of the Joint Venture must fill out this form separately to provide information relating to each JV member.

1.	Bidder's legal name:	[insert full name]		
2.	Joint Venture Partner's legal name:	[insert full name of Joint Venture Partner]		
3.	Joint Venture Partner's Country of Registration:	[insert country of registration]		
4.	Joint Venture Partner's Legal Address in Country of Registration:	[insert street/number/town or city/country]		
5.	Joint Venture Partner's Trading address	[insert street/number/town or city/country]		
6.	Joint Venture Partner's Year of Registration:	[insert year of registration]		
7.	Joint Venture Partner's Telephone/Fax numbers:	[insert telephone/fax numbers, including country and city codes]		
8.	Joint Venture Partner's Email Address:	[insert email address]		
9.	Joint Venture Partner's Authorized Represe	entative Information:		
	Name:	[insert full name]		
	Address:	[insert street/number/town or city/country]		
	Telephone/Fax numbers:	[insert telephone/fax numbers, including country and city codes]		
Email Address: [insert email address]		[insert email address]		

^{1.} Articles of incorporation or constitution and company incorporation/registration of the legal entity named above, in accordance with ITB ITB 4.2 and ITB 4.8

^{2.} Authorization to represent the firm named above, in accordance with ITB 22.2

^{3.} In the case of a government-owned enterprise, documents establishing legal and financial autonomy and compliance with commercial law, in accordance with ITB 4.5

Form LIT 1:Pending Litigation Form

Each Bidder or member of a JV must fill in this form

Choose one of the following:

- No pending litigation, arbitration or any other material events impacting the net worth and/or liquidity of the bidder.
- Below is a description of all pending litigation, arbitration involving the Bidder or any other material events impacting the net worth and/or liquidity of the bidder (or each Joint Venture partner if Bidder is a Joint Venture).

Year	Matter in Dispute	Value of Pending Claim in NRs	Value of Pending Claim as a Percentage of Net Worth
[insert year]	• Contract Identification: [indicate complete Contract name, number, and any other identification]	[insert amount]	[insert amount]
	• Name of Purchaser: [insert full name]		
	• Address of Purchaser: [insert street/city/country]		
	• Matter in dispute: [indicate main issues in dispute]		
	• Party who initiated the dispute: [indicate "Purchaser" or "Supplier"]		
	• Status of dispute: [indicate if it is being treated by under Arbitration or being dealt with by the Judiciary]		



Form FIN-1: Financial Situation Form

Each Bidder must fill out this form.

In case of a Joint Venture, each Joint Venture Partner must fill out this form separately and provide the Joint Venture Partner's name:

Joint Venture Partner: [Insert Name of Joint Venture Partner]

Financial Data for PreviousYears (in NRs)				
Year 1:	Year 2:	Year:		

Information from Balance Sheet

Total Assets (TA)		
Total Liabilities (TL)		
Net Worth = TA-TL		
Current Assets (CA)		
Current Liabilities (CL)		
Working Capital = CA-CL		

Information from Income Statement

Total Revenues		
Profits Before Taxes		
Profits After Taxes		



Attached are copies of financial statements (balance sheets including all related notes, and income statements) for number of years, as indicated above, complying with the following conditions:

- All such documents reflect the standalone financial situation of the legal entity or entities comprising the Bidder and not the Bidder's parent companies, subsidiaries, or affiliates.
- Historic financial statements must be audited by a certified accountant.
- Historic financial statements must be complete, including all notes to the financial statements.
- Historic financial statements must correspond to accounting periods already completed and audited (no statements for partial periods shall be requested or accepted).

Form FIN-2: Average Annual Turnover Form

Each Bidder must fill out this form.

The information supplied should be the Annual Turnover of the Bidder or each partner of a Joint Venture in terms of the amounts billed to clients for each year for work in progress or completed to NRs at the end of the period reported.

In case of a Joint Venture, each Joint Venture Partner must fill out this form separately and provide the Joint Venture Partner's name:

Joint Venture Partner: [Insert Name of Joint Venture Partner]

Annual Turnover Data for the Last Years					
Year	Amount (in NRs)				
Average Annual Turnover					



Form FIN-3: Financial Resources Form

Specify proposed sources of financing, such as liquid assets⁸, unencumbered real assets, and other financial means (other than any contractual advance payments) available to meet the total cash flow requirements of the subject contract

Financial Resources					
No.	Source of financing	Amount (in NRS)			
1					
2					
3					

⁸ Liquid assets mean cash and cash equivalents, short-term financial instruments, short-term available-for-sale-securities, marketable securities, trade receivables, short-term financing receivables, and other assets that can be converted into cash within one (1) year.



Form EXP-1: General Experience

[The following table shall be filled in for the Bidder and for each member of a JV. Each contract shall be supported by Signed Contract Agreement or any other relevant evidence.]

[Identify contracts that demonstrate continuous supply over the past [number] years. List contracts chronologically, according to their commencement (starting) dates.]

General Experience				
Starting Year	Ending Year	Contract Identification	Role of Bidder	
[indicate year]	 [indicate year] Brief description the Bidder: [destrictly] Amount of continuous Name of Purcha 	 Brief description of the supply performed by the Bidder: [describe supply performed briefly] Amount of contract: [insert amount] Name of Purchaser: [indicate full name] Address: [indicate street/number/town or 		



Form EXP-2: Specific Experience Form

[The following table shall be filled in for contracts performed by the Bidder and by each member of a JV. Each contract shall be supported by Signed Contract Agreement or any other relevant evidence.]

Contract	of Similar Na	ture			
Item	Information				
Contract Identification	[insert contract name and reference identification number, if applicable]				
Award date	[insert day, month, year, e.g., 10 January, 2022]				
Completion date	[insert day, month, year, e.g., 12 July, 2022]				
Role in Contract	Prime	supplier			
[check the appropriate box]	Single entity JV Member		Sub contractor		
Total Contract amount	[insert totamount]	al contract	[Insert Currency]		
Description of the contract performed by the Bidder	[Insert brief description of contract to justify similarity]				
If partner in a JV or subcontractor, specify participation of total Contract amount		cent of total]	[Insert Amount]		
Purchaser's Name:	[Insert full name]				
Purchaser's Address:	[indicate street / number / town or city / country]				
Purchaser's Telephone/fax number:	[insert telephone/fax numbers, including country and city area codes]				
Purchaser's E-mail:	[insert E-mail address, if available]				

The Bidder shall complete this form for each contract completed.



Form EXP - 3: Technical Experience

Fill out one (1) form per contract. Each contract shall be supported by Signed Contract Agreement or any other relevant evidence.

Technical Experience						
Name of Product						
Manufacturer:			Address and Nationality:			
(i) Product has been in production for at least	Description of Goods		Year of Production		Number of units produced	
years.	[insert description of Goods]		[insert years]		[insert number]	
(ii) Product (or equipment) has been sold a minimum of units of similar type and specification over the last	Description of Goods		Year of Production		Number of units that has been sold	
	[insert description of Goods]		[insert years]		[insert number]	
three (3) years.						
(iii) Product has been in operation for a minimum	Description of Goods		Number of units that has been sold and years		Details of purchasers	
of years.	[insert description of God	ods]	[insert number]	[insert year]	[insert name, adress (street / number / town or city / country), telephone/fax numbers with country and city	
					code, email adresses if available]	



Form EXP - 4: Production Capacity

Fill out one (1) form per product and manufacturer.

	Production	n Capacity
Name of Product		
Manufacturer:		Address and Nationality:
Production facility 1 (include location):	[Insert description of g	roods and production capacity]
Production facility 2 (include location):	[Insert description of g	roods and production capacity]
Production facility 3 (include location):	[Insert description of g	roods and production capacity]



Form Spec-1: Bidder's proposed Specification Form

The bidder shall fill this form to provide the information on technical specifications and standards of offered goods. Bidder's Proposed Technical Specifications and Standards in column 4 shall comply with the Purchaser Requirement (Specifications and standards) specified by the Purchaser in the Schedule of Requirements.

Item No.	Name of Goods or Related Service	Country of Origin and Brand	Bidder's Proposed TS and Standards	Reference page/ no. / Catalogue page/ no.	Status of compliance ⁹
1	2	3	4	5	6
[insert item No.]	[insert name]	[insert country of Origin and Brand name]	[insert TS and Standards]	[insert reference page or no. / catalogue page or no. if any]]	[insert status]

⁹ The bidder shall state as Fully compliance or Partially compliance or Compliance



Price Schedule Forms

[The Bidder shall fill in these Price Schedule Forms in accordance with the instructions indicated. The list of line items in column 1 of the **Price Schedules** shall coincide with the List of Goods and Related Services specified by the Purchaser in the Schedule of Requirements.]



Price Schedules Price Schedule: Goods Manufactured in Nepal

	[ITB 15.7 (a), Goods manufactured in Nepal] Date: ICB No: GH/G/ICB-06/2081-82											
1	2	3	4	5	6	7	8	9	10			
Line Item N°	Description of Goods	Delivery Date as defined by Incoterms	Quantity and physical unit	Unit price EXW	Total EXW price per line item (Col. 4×5) 15.7(a)(i)	Price per line item for inland transportation and other services required to convey the Goods to their final destination ITB 15.7(a)(iii)	Cost of local labor, raw materials and components from with origin in Nepal % of Col. 5	Sales and other taxes payable per line item if Contract is awarded (in accordance with ITB 15.7(a)(ii)	Total Price per line item (Col. 6+7) 15.7(a)(iv)			
1	Centrifuge		1 Set									
2	Refrigerated Centrifuge		1 Set									
3	Incubator		1 Set									
4	CO2 Incubator		1 Set									
5	Microscope W/Stereo Zoom		1 Set									
6	Computer Assisted Semen Analyzer		1 Set									
7	Laminar Flow		1 Set									
8	Ultrasound Machine		1 Set									
9	4K Hysteroscopy System		1 Set									
								Total Price				

Name of Bidder	Signature of Bidder	 Date	
tunie of Diager	 DIGITATE OF DIGGET	 Date	

Note: Column 5 and 6: Currencies in accordance with ITB 16 Column 7 and 9: In Nepali Currency



Goods Manufactured Outside the Nepal, to be Imported

	[ITB 15.7(b), goods to be imported Date: ICB No: GH/G/ICB-06/2081-82											
1	2	3	4	5	6	7	8	9				
Line Item N°	Description of Goods	Country of Origin	Delivery Date as defined by Incoterms	Quantity and physical unit	Unit price CIP [insert place of destination] in accordance with ITB 15.7(b)(i)	CIP Price per line item (Col. 5x6) 15.7(b)(i)	Price per line item for inland transportation and other services required in Nepal to convey the Goods to their final destination specified in BDS 15.7(b)(ii)	Total Price per Line item (Col. 7+8) 15.7(b)(iv)				
1	Centrifuge			1 Set								
2	Refrigerated Centrifuge			1 Set								
3	Incubator			1 Set								
4	CO2 Incubator			1 Set								
5	Microscope W/Stereo Zoom			1 Set								
6	Computer Assisted Semen Analyzer			1 Set								
7	Laminar Flow			1 Set								
8	Ultrasound Machine			1 Set								
9	4K Hysteroscopy System			1 Set								
							Total Price					

NT 070'11	G' (B) 11	D :
Name of Bidder	Signature of Bidder	Date
r tuille of Diagor	 Digitatare of Blader	Date

Note: Column 6 and 7: Currencies in accordance with ITB 16 Column 8: In Nepali Currency



Director Price Schedule: Goods Manufactured Outside Nepal, already imported

	[ITB 15.7 (c)), Goods already imported] Date: ICB No: GH/G/ICB-06/2081-82											
1	2	3	4	5	6	7	8	9	10	11	12	
Line Item N°	Description of Goods	Country of Origin	Delivery Date as defined by Incoterms	Quantity and physical unit	Unit price including Custom Duties and Import Taxes paid, in accordance with ITB 15.7(c)(i)	Custom Duties and Import Taxes paid per unit in accordance with ITB 15.7(c)(ii), [to be supported by documents]	Unit Price net of custom duties and import taxes, in accordance with ITB 15.7(c) (iii) (Col. 6 minus Col.7)	Price per line item net of Custom Duties and Import Taxes paid, in accordance with ITB 15.7 (c)(iii) (Col. 5×8)	Price per line item for inland transportation and other services required in Nepal to convey the goods to their final destination, as specified in BDS in accordance with ITB 15.7 (c)(v)	Sales and other taxes paid or payable per item if Contract is awarded (in accordance with ITB 15.7 (c)(iv)	Total Price per line item (Col. 9+10) ITB 15.7 (c)(vi)	
1	Centrifuge			1 Set								
2	Refrigerated Centrifuge			1 Set								
3	Incubator			1 Set								
4	CO2 Incubator			1 Set								
5	Microscope W/Stereo Zoom			1 Set								
6	Computer Assisted Semen Analyzer			1 Set								
7	Laminar Flow			1 Set								
8	Ultrasound Machine			1 Set								
9	4K Hysteroscopy System			1 Set								
										Total Bid Price		

Name of Bidder	 Signature of Bidder	Date

Note: Column 6,7, 8 and 9: Currencies in accordance with ITB 16 Column 10 and 11: In Nepali Currency



Price and Completion Schedule - Related Services

					Date:ICB No:	
1	2	3	4	5	6	7
Service No.	Description of Services (excludes inland transportation and other services required in Nepal to convey the goods to their final destination)	Country of Origin	Delivery Date at place of Final destination	Quantity and physical unit	Unit price	Total Price per Service (Col. 5*6 or estimate)
[insert number of the Service]	[insert name of Services]	[insert country of origin of the Services]	[insert delivery date at place of final destination per Service]	[insert number of units to be supplied and name of the physical unit]		[insert total priceand currency per item]
		<u></u>		Total Bid Price		

Name of Bidder [insert complete name of Bidder]

Signature of Bidder [signature of person signing the Bid]

Date [insert date]

Notes:

Column 6 and 7: Currencies in accordance with ITB Clause 16

Prices are to be quoted inclusive of all custom duties, sales and other similar taxes applicable in Nepal and payable on the Related Services, if the Contract is awarded to the Bidder



Bid Security

Bank Guarantee

Bank's Name, and Address of Issuing Branch or Office (On Letter head of the Bank)

[This is the format for the Bid Security to be issued on the letterhead by a Commercial Bank or Financial Institution eligible to issue Bank Guarantee as per prevailing Law of Nepal]

[insert Bank's Name, and Address of Issuing Branch or Office]

Date: [insert date]

Beneficiary: [insert Name and Address of Purchaser]

BID GUARANTEE No.: [insert number]

We have been informed that *[insert name of the Bidder]* (hereinafter called "the Bidder") intends to submit its bid to you (hereinafter called "the Bid") for the execution of *[insert name of contract]* under Invitation for Bids No. *[insert IFB number]* ("the IFB").

Furthermore, we understand that, according to your conditions, bids must be supported by a bid guarantee.

At the request of the Bidder, we *[insert name of Bank]* hereby irrevocably undertake to pay you any sum or sums not exceeding in total an amount of *[insert amount in figures, (insert amount in words)]* upon receipt by us of your first demand in writing accompanied by a written statement stating that the Bidder is in breach of its obligation(s) under the bid conditions, because the Bidder:

- (a) has withdrawn or modifies its Bid:
 - i) during the period of bid validity specified by the Bidder on the Letter of Bid, in case of electronic submission
 - (ii) from the period twenty-four hours prior to bid submission deadline up to the period of bid validity specified by the Bidder on the Letter of Bid, in case of hard copy submission; or
- (b) does not accept the correction of errors in accordance with the Instructions to Bidders (hereinafter "the ITB"); or
- (c) changes the prices or substance of the bid while providing information pursuant to clause 29.1 of ITB; or
- (d) having been notified of the acceptance of its Bid by the Purchaser during the period of bid validity, (i) fails or refuses to execute the Contract Agreement, or (ii) fails or refuses to furnish the performance security, in accordance with the ITB.
- (e) is involved in fraud and corruption in accordance with the ITB.



This guarantee will expire: (a) if the Bidder is the successful Bidder, upon our receipt of copies of the contract signed by the Bidder and the performance security issued to you upon the instruction of the Bidder; and (b) if the Bidder is not the successful Bidder, upon the earlier of (i) our receipt of a copy of your notification to the Bidder of the name of the successful Bidder; or (ii) thirty (30) days after the expiration of the Bidder's bid which comes to be *[insert the date]*.

Consequently, any demand for payment under this guarantee must be received by us at the office on or before that date.

This guarantee is subject to the Uniform Rules for Demand Guarantees, ICC Publication No. 758¹⁰.

... Bank's seal and authorized signature(s) ...

[Note: All italicized text is for use in preparing this form and shall be deleted from the final product.]

_

¹⁰ As the case may be, ICC Publication No. 758 (or subsequent ICC Publications) may be used. In such cases, modify the Publication number.



Manufacturer's Authorization Letter

[This letter of authorisation should be on the letterhead of the manufacturer and should be signed by the person with the proper authority to sign documents that are binding on the manufacturer]

Date: [insert date of Bid Submission]

IFB No.: [insert number]

To: [insert complete name and address of Purchaser]

WHEREAS [insert complete name of Manufacturer or Manufacturer's authorized agent] who are official manufacturers of [insert type of goods manufactured] having factories at [insert full address of Manufacturer's factories], do hereby authorize [insert complete name of Bidder] exclusively to submit a Bid in relation to the Invitation for Bids indicated above, the purpose of which is to provide the following Goods, manufactured by us [insert name and/or brief description of the goods] and to subsequently sign the Contract.

We hereby extend our full guarantee and warranty in accordance with Clause 27 of the General Conditions of Contract, with respect to the Goods offered by the above firm.

Name: [insert complete name of person signing the Authorization]

In the capacity of: [insert legal capacity of person signing the Authorization]

Signed: [insert signature of person whose name and capacity are shown above]

Duly authorized to sign the Authorization for and on behalf of: [insert complete name of Manufacture]

Date: [insert date of signing]



Section V. Schedule of Requirements

The Schedule of Requirements shall be included in the Bidding Documents by the Purchaser, and shall cover, at a minimum, a description of the Goods and Related Services to be supplied and the Delivery Schedules.

The objective of the Schedule of Requirements is to provide sufficient information to enable Bidders to prepare their Bids efficiently and accurately, in particular, the Price Schedule, for which a form is provided in Section IV.

Contents

1.	List of Goods and Related Services	88
2.	Delivery and Completion Schedule	90
3.	Technical Specifications	92
4.	Drawings	95



List of Goods and Related Services

The purpose of the List of Goods and Related Services (LGRS) is to briefly describe and specify the quantities of each of the Goods and Related Services that the Purchaser requires the Bidder to include in its Bid. As a part of the SR, the LGRS constitutes a Contract document and, therefore, it is a part of the Contract. The Purchaser must prepare the LGRS and include it as a part of the SR.

If the Goods and Related Services are grouped in lots/packages, the Purchaser must state here whether Bidders are permitted to submit Bids for individual lots or not. For example:

[Alternative 1. Insert: "The Goods and Related Services are grouped in lots. Bids on individual lots are not permitted. Only Bids for all of the lots/Packages will be accepted."]

[Alternative 2. Insert: "The Goods and Related Services are grouped in lots. Bidders shall have the option of submitting Bids for one, any combination, or all of the lots/Packages. Lots/Packages shall not be further sub-divided for the purpose of bidding."]

Lot/Pac	Lot/Package No.: [if applicable]								
Lot/Pac	Lot/Package Name : [if applicable]								
Item No.	Name of Goods or	Description ¹¹	Unit of	Quantity					
	Related Services		Measurement						

Lot/Pac	Lot/Package No.: [if applicable]									
Lot/Slic	Lot/Slice/Package Name: [if applicable]									
Item No.	Item No. Name of Goods or Related Services Description Unit of Measurement Quantity									

¹¹ Description of Goods or related sevices is to be inserted in this column



Delivery and Completion Schedule

Delivery shall take place in compliance with the dates, duration, and locations indicated below:

(I) in case of Delivery schedule is not determined as evaluation criteria

[The Purchaser shall fill in this table from column 1 to 6]

Line Item No	Description of Goods	Quantity	Physical Unit	Final Destination	Delivery Date (as per Incoterms)		
<u>1.</u>	<u>2</u>	3	4	<u>5</u>	<u>6</u>		
1	Centrifuge	1	Set				
2	Refrigerated Centrifuge	1	Set		Within <i>90 Days</i> From the date of		
3	Incubator	1	Set				
4	CO2 Incubator	1	Set	Geta Hospital,			
5	Microscope W/Stereo Zoom	1	Set	Geta, Kailali,	siging the		
6	Computer Assisted Semen Analyzer	1	Set	Nepal	Contract /LC Opening		
7	Laminar Flow	1	Set		opening		
8	Ultrasound Machine	1	Set				
9	4K Hysteroscopy System	1	Set				

(II) in case of Delivery schedule is determined as evaluation criteria: Not Applicable

[The Purchaser shall fill in this table from column 1 to 7]

Line Item			.	Final	Delivery (as per Incoterms) Date		
No No	Description of Goods	Quantity	Physical unit	Destination as specified in BDS	Earliest Delivery Date	Final Delivery Date	
1.	2	3	4	5	6	7	
[insert item No]	[insert description of Goods]	[insert quantity of item to be supplied]	[insert physical unit for the quantity]	[insert place of delivery]	[insert the number of days following the date of effectiveness the Contract]	[insert the number of days following the date of effectiveness the Contract]	

Note: The date of effectiveness of contract shall be as of signing the contract



List of Related Services and Completion Schedule: Not Applicable

[The Purchaser shall fill in this table from column 1 to 6. The Required Completion Dates should be realistic, and consistent with the required Goods Delivery Dates (as per Incoterms)]

Service	Description of Service	Quantity ¹²	Physical unit	where Services shall be performed	Final Completion Date(s) of Services
1.	2	3	4	5	6
[insert service no.]	[insert description of Related Services]	[insert quantity of item to be supplied]	[insert physical unit for the quantity]	[insert name of the Place]	[insert required Completion Date(s)]

¹² If applicable



Technical Specification

A. GENERAL TECHNICAL SPECIFICATION



	Specification of Centrifuge			
		Bidder's C	Compliance Shee	t
	Purchaser's Technical Specifications	Yes/No	Page No. in Catalogue	Remarks
CN	Centrifuge for IVF			
S.N.	Manufacturer			
	Brand			
	Type/Model			
	Country of Origin			
1	Description of Function			
1.1	A versatile high performing small benchtop centrifuge with a hybrid design			
2	System Configuration			
2.1	Dual spin centrifuge with fix angle buckets and swinging buckets			
3	Technical Specification			•
2.1	Maximum capacity Rotor: 8 x 15 ml fixed angle, 8x10ml swinging bucket			
2.2	Max. speed: 300 to 4900 with adjustable increments of 100 rpm			
2.3	LED screen large			
2.4	Shows RPM/ RCF and time.			
2.5	Speed programming in 50 RPM/ 10 xg steps.			
2.6	Time Setting: Time setting capability from 1 to 99 minutes with 1-minute increments, including a HOLD function.			
2.7	Must have interchangeable fixed angle & swinging buckets			
2.8	Acceleration and Deceleration: Should offers at least 1 acceleration and 3 deceleration profiles with standard, soft, and brake-off options.			
2.9	Acoustic and visual messages on screen warning the user the equipment status.			
2.10	Must be Microprocessor controlled with large led display user interface			
2.11	Should have an Brushless Induction motor that is maintenance free			
2.12	Memory function must have Rotors list 4 program memory			
2.13	Start, stop, open lid and short spin with adjustable speed buttons.			
2.14	there must be an Option of free/locked adjustment of RPM/RCF along the run.			
2.15	The system must have automatic lid opening and must be programmable.			
2.16	the Last values must remain in memory for easy of use			
2.17	there should be an Over-speed protection function			

	is you		
overnment or the lity of Health and Programme Geta Hospi	Director		-
Ge2a,1811a	Lid provided with security systems:		
2.19	Automatic lid lock system, Emergency lid-lock release in case of power failure		
2.20	– Locking and protection against opening along the run.		
2.21	- Lid dropping protection.		
2.22	 Port in the lid for calibration and operation checking. 		
2.23	Max g force for fixed angle rotors 3115 & for swinging rotors 3490		
2.24	it must have Protection safety steel armored ring between the centrifuge chamber and the housing.		
2.25	Chamber of centrifugation in stainless steel (easy cleaning).		
2.26	Rotors and adapters autoclavable, easy to install by the user		
2.27	There must be Forced ventilation to reduce increase of temperature		
2.28	Rotor locking system must be knob type		
2.29	Temperature range 2-40deg celsius		
2.30	Noise level: below 60 dB		
3	Accessories, Spare Parts and Consumables included	 1	
3.1	Should be supplied with standard accessories.		
4	Operational Environment	I	
4.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power supply, climate, temperature, humidity, etc.		
4.2	Power supply: 220-240 VAC, 50 Hz single phase fitted with appropriate 3 pin plug (flat). The power cable must be at least 3m long		
4.3	Power must be 130w or lower		
5	Standards, Medical Device Regulation and Safety Requirements		
5.1	Must submit a valid "Medical devices – Quality management systems – Requirements for regulatory purpose"; ISO 13485:2016.		
5.2	Must submit automatic Renewable European CE marked certificate (Directive 93/42 EEC on Medical Devices) from notified body or Medical Device Regulations - MDR (EU) 2017/745 or Declaration of conformity if it falls a risk-based class I.		
5.5	Must submit US-FDA approved product. Non submission leads to rejection the bid.		

handling, keep hygiene the equipment).

hospital technician and engineers.

Must provide user training (including how to use and safe

Must provide the first step technical training to troubleshooting and maintenance of the machine to the

Up time Service Backup, Warranty/Guarantee

User and Technical Training

6

6.1

6.2

7

of Health and Pitcheta Hospit	Director		
Geta, Kailan	Bidders must provide a comprehensive warranty for two (2) years after acceptance, along with an additional one (1) year of free service warranty. A commitment letter from the		
7.1	manufacturer confirming the comprehensive warranty must be submitted with the bid. Failure to provide this document will result in bid rejection.		
7.2	The availability of the spare parts for the next 10 years must be declared by the manufacturer and bidder. Non submission leads to reject the bid.		
7.3	During the warranty period supplier must ensure preventive maintenance and corrective/ breakdown maintenance whenever required.		
8	Installation and commissioning		
8.1	The bidder must arrange for the quoted equipment with all the listed accessories and consumables to be installed on site and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.		
9	Documentation on site		
9.1	User (Operating) manual in English must be submit during the bid.		
	Non submission leads to rejection the bid.		
9.2	Service (Technical/ Maintenance) manual in English		
9.3	Spare parts and consumables price list should be provided in the separate sheet, non-submission leads to the rejection of the bid.		
9.4	Must submit the authorisation letter or linking to the manufacturer for the sub-distributor authorisation of all the quoted items.		
9.5	Non submission leads to rejection the bid. Must submit the commitment letter for the availability of spare parts for next 10 years from the manufacturer. Non submission leads to rejection the bid.		
9.6	Note: Bidder must completely fill the Technical Specification Form (TSF). Only yes/no/all complies should not be written. Page number in the catalogue of all the required parameters must be clearly mentioned and highlighted. Failure in doing so may lead to rejection of bid from technical committee.		



Item	No. 2 :- Refrigerated Centrifuge			
	9	Bio	dder's Complian	ce Sheet
	Purchaser's Technical Specifications	Yes/No	Page No. in Catalogue	Remarks
O NI	Refrigerated Centrifuge			•
S.N.	Manufacturer			
	Brand			
	Type/Model			
	Country of Origin			
1	Description of Function			
1.1	The centrifuge must have swinging bucket, fixed angle and microplate rotors to meet current and future sample processing needs of the lab/ IVF			
2	Technical Specification			
2.1	Centrifuge should have Max RCF: 30000 x g or better			
2.2	Centrifuge should have Max Rpm: 17750 RPM or better			
2.3	Temperature Range of the centrifuge should be : -10 to + 40 degree Celsius			
2.4	System should be Microprocessor controlled with direct brushless induction low profile motor			
2.5	Programs: 3 direct program buttons, plus 50 additional programs accessible.			
	Imbalance Detection System of the Centrifuge should be			
2.6	Continuous vibration measurement, with rotor mass correction.			
2.7	Centrifuge should have Acceleration / Deceleration Rates: 2			
2.8	Centrifugation Chamber Stainless Steel			
2.9	Max Timer Range 9h, 99min + continuous			
2.10	Auto-Lock System			
2.11	The buckets and rotor sealing lids must be certified for biocontainment by a 3rd party lab of worldwide recognition			
2.12	Rotor shall be installed and removed with no tools in less than 5 seconds			
2.13	Machine should be compatible with 4x100ml Swing bucket rotor for future requirement.			
2.14	Rotor TO BE SUPPLIED:- Swing out Rotor for 15 and 50 ml Conical Tubes:			
2.15	Machine should be compatible with 4x100ml Swing bucket rotor for future requirement			
2.16	Type- Swing Bucket Rotor with 15 ml and 50 ml Adapters			
2.17	Tube Capacity- 8 x15/4x50 ml			
2.18	RPM- Max 4,500]		

Menalian Ang	
Ministry of Health and Popular Geta Hospital Director	
Geta Hoskii Dilector	

Geta	HOSP: DIECLOF		
2.19	Optional Adapter for 2ml, 5/7ml, 30ml, & 100 should available		
2.20	Rotor Capacity- 4 x145 ml		
3	Accessories, Spare Parts and Consumables included		
3.1	Should be supplied with standard accessories.		
4	Operational Environment		
-	The product offered shall be designed to be stored and to		
4.1	operate normally under the conditions of the purchaser's		
4.1	country. The conditions include Power supply, climate,		
	temperature, humidity, etc.		
	Power supply: 220-240 VAC, 50 Hz single phase fitted with		
4.2	appropriate 3 pin plug (flat). The power cable must be at least		
	3m long		
5	Standards, Medical Device Regulation and Safety		
	Requirements		
5.1	Must submit a valid "Medical devices – Quality management		
3.1	systems – Requirements for regulatory purpose"; ISO 13485:2016.		
	Must submit automatic Renewable European CE marked		
	certificate (Directive 93/42 EEC on Medical Devices) from		
5.2	notified body or Medical Device Regulations - MDR (EU)		
	2017/745 or Declaration of conformity if it falls a risk-based		
	class I.		
5.3	Must submit US-FDA approved product.		
3.3	Non submission leads to rejection the bid.		
6	User and Technical Training		
6.1	Must provide user training (including how to use and safe		
0.1	handling, keep hygiene the equipment).		
6.2	Must provide the first step technical training to		
0.2	troubleshooting and maintenance of the machine to the		
	hospital technician and engineers.		
7	Up time Service Backup, Warranty/Guarantee		
	Bidders must provide a comprehensive warranty for two (2)		
	years after acceptance, along with an additional one (1) year		
7.1	of free service warranty. A commitment letter from the manufacturer confirming the comprehensive warranty must		
	be submitted with the bid. Failure to provide this document		
	will result in bid rejection.		
	The availability of the spare parts for the next 10 years must		
7.2	be declared by the manufacturer and bidder. Non submission		
	leads to reject the bid.		
	During the warranty period supplier must ensure preventive		
7.3	maintenance and corrective/ breakdown maintenance		
	whenever required.		
8	Installation and commissioning		

Mendin Inly
Ministry of Health and popular

Gets, The bidder must arrange for the quoted equipment with all the listed accessories and consumables to be installed on site 8.1 and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail. 9 **Documentation on site** User (Operating) manual in English must be submit during the bid. 9.1 Non submission leads to rejection the bid. Service (Technical/ Maintenance) manual in English 9.2 Spare parts and consumables price list should be provided in the separate sheet, non-submission leads to the rejection of 9.3 the bid. Must submit the authorisation letter or linking to the 9.4 manufacturer for the sub-distributor authorisation of all the quoted items. Non submission leads to rejection the bid. Must submit the commitment letter for the availability of 9.5 spare parts for next 10 years from the manufacturer. Non submission leads to rejection the bid. Note: Bidder must completely fill the Technical Specification Form (TSF). Only yes/no/all complies should not be written. Page 9.6 number in the catalogue of all the required parameters must be clearly mentioned and highlighted. Failure in doing so may lead to rejection of bid from technical committee.

Item :	No. 3 :- Incubator			
		Bidder's	Compliance Sheet	
ON	Purchaser's Technical Specifications	Yes/No	Page No. in Catalogue	Remarks
S.N.	Manufacturer:			
	Brand:			
	Type /Model:			
	Country Of Origin:			
1	Description of Function			
1.1	Must be for use in IVF application			
1.2	The unit shall have internal glass door for easy viewing of samples without disturbing internal environment.			
1.3	The unit shall provide automatic over-temperature alarm, without need for manual setting.			
1.4	Temperature Range: ambient +5 °C to 75 °C. Temperature stability 0.2C			
1.5	The unit should have Gravity convection			
1.6	The temperature uniformity (spatial deviation) shall be $\leq \pm 0.6^{\circ}$ C at 37°C.			
1.7	The temperature Stability (deviation over time): $\leq \pm 0.2$ °C at 37°C.			
1.8	The unit should have Intuitive user interface for easy temperature setting			
1.9	The unit's usable space shall be approximately 110 Liters with inner dimension of 464mm (W) X 414mm (D) X 608mm (H).			
1.10	The exterior dimension is 640mm (W) X 565mm (D) X 855mm (H). (APPROXIMATE)			
1.11	The interior surface shall be made from 1.4016 grade stainless steel.			
1.12	The chamber shall have rounded corners for easy cleaning and to avoid contamination.			
1.13	The shelf system shall be easy to fit and remove easily without the use of tools.			
1.14	The user interface shall be microprocessor controlled and shall have a large vacuum fluorescent display and touch button control.			
1.15	The unit shall be provided with 2 shelves as standard and shall be capable of accommodating a maximum of 16 shelves.			
1.16	The shelf shall be capable of carrying a load of 25 kg.			
1.17	Maximum loading capacity for the unit is 50kg.			
1.18	The unit shall be provided with a standard 42mm wide access port for independent monitoring of sample temperature at the back side of the unit.	,		
1.19	Should have timer function			
1.20	System should be stackable			

- Jonly		
Course of Ne Palation		
Ministry of Health and Popul		
Geta Hospital Director		
GETT HOW	-	

eta Hosp	Director		
G1t2 1/31/3	Footprint shall not exceed 3.9 sq ft.		
1.22	The unit shall have RS 232 Interface.		
2	Accessories, consumables, & spare parts		
2.1	Should be supplied with standard accessories.		
3	Operating Environment	'	
3.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power supply, climate, temperature, humidity, etc.		
3.2	Power supply: 220-240 VAC, 50 Hz single phase fitted with appropriate 3 pin plug (flat). The power cable must be at least 3m long		
4	Standards, Medical Device Regulation and Safety Requirements		
4.1	Must submit a valid "Medical devices – Quality management systems – Requirements for regulatory purpose"; ISO 13485:2016.		
4.2	Must submit automatic Renewable European CE marked certificate (Directive 93/42 EEC on Medical Devices) from notified body or Medical Device Regulations - MDR (EU) 2017/745 or Declaration of conformity if it falls a risk-based class I. or US-FDA registered/approved		
5	User and Technical Training		
5.1	Must provide user training (including how to use and safe handling, keep hygiene the equipment).		
5.2	Must provide the first step technical training to troubleshooting and maintenance of the machine to the hospital technician and engineers.		
6	Up time Service Backup, Warranty/Guarantee		
6.1	Bidders must provide a comprehensive warranty for two (2) years after acceptance, along with an additional one (1) year of free service warranty. A commitment letter from the manufacturer confirming the comprehensive warranty must be submitted with the bid. Failure to provide this document will result in bid rejection.		
6.2	The availability of the spare parts for the next 10 years must be declared by the manufacturer and bidder. Non submission leads to reject the bid.		
6.3	During the warranty period supplier must ensure preventive maintenance and corrective/ breakdown maintenance whenever required.		
7	Installation and commissioning		
7.1	The bidder must arrange for the quoted equipment with all the listed accessories and consumables to be installed on site and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.		
8	Documentation on site		

Comme of Nevidation	
Ministra Popul	
Thistry of Health allowital Director	

eta Host	Silector		
Geta, Kail	User (Operating) manual in English must be submit during		
8.1	the bid.		
	Non submission leads to rejection the bid.		
8.2	Service (Technical/ Maintenance) manual in English		
	Spare parts and consumables price list should be provided in		
8.3	the separate sheet, non-submission leads to the rejection of		
	the bid.		
	Must submit the authorisation letter or linking to the		
8.4	manufacturer for the sub-distributor authorisation of all the		
0.4	quoted items.		
	Non submission leads to rejection the bid.		
	Must submit the commitment letter for the availability of		
8.5	spare parts for next 10 years from the manufacturer.		
	Non submission leads to rejection the bid.		
	Note:		
	Bidder must completely fill the Technical Specification Form		
8.6	(TSF). Only yes/no/all complies should not be written. Page		
0.0	number in the catalogue of all the required parameters must		
	be clearly mentioned and highlighted. Failure in doing so may		
	lead to rejection of bid from technical committee.		



		Bidder's Compliance Sheet		t
	Purchaser's Technical Specifications	Yes/No	Page No. in Catalogue	Remarks
C N	CO2 INCUBATOR		_	1
S.N.	Manufacturer			
	Brand			
	Type/Model			
	Country of Origin			
1	Description of Function			
1.1	CO2 Incubator for use in IVF application & must provide a stable culture environment by maintaining precise temperature & gas concentration			
2	System Configuration			
2.1	CO2 incubator with size of 150L (5.3 cu ft) and capacity to hold three standard SS shelves and maximum capacity of 10 shelves.			
3	Technical Specification			
3.1	Interactive navigator touchscreen for trend analysis for convenient evaluation of unit's performance, View instructions, monitor progress, decontamination routine and running log of all user interactions with the incubator			
3.2	Pre-heated gas should be mixed to eliminate any temperature fluctuations			
3.3	CO2 concentration range should be 0-20 % with CO2 sensor technology dual beam IR / TC			
3.4	O2 control for hypoxic condition with O2 measure and control range vol% choice of 121 / 590			
3.5	Unique integral humidity 3 L water reservoir that provides a high relative humidity (rH) with water level sensor which indicates when a refill is needed.			
3.6	Possibility of automatically calibrating all electronic measurement and control functions.			
3.7	Relative humidity should be >90% at 37°			
3.8	Data output RS 232			
3.9	Unique integral humidity 3 L water reservoir that provides a high relative humidity (rH) with water level sensor which indicates when a refill is needed			
3.10	Unique gas-tight segmented three door			
3.11	90°C moist heat decontamination			

Nepalion An	767.
Ministry of Health and Popular Direction Geta Hospital	tor

Geta H	osbi. Director	ı	T	1
	Glass door should have reliable door preventing the inner			
3.12	door from accidentally not being closed and compromising			
	culture conditions			
2.12	42 mm (1.6 in) standard access port which allow cables,			
3.13	plugs and tubing to be easily inserted into or out of the chamber.			
3.14	Visual alert on touchscreen display when gas supply is low and needs changing			
	System should show graph performance trends over			
3.15	established timeframes and run event history logs—protected			
3.10	with user passcodes and control lockouts.			
	CO2 incubator with size of 150L (5.3 cu ft) and capacity to			
3.16	hold three standard SS shelves and maximum capacity of 10	`		
	shelves.			
3.17	Must have intelligent control system automatically detects lid			
3.17	opening/closing			
3.18	Must have capacity of minimum 5-12 individual chambers	`		
3.19	Chamber dimension: approx. External 637× 867 × 782 mm,			
3.19	approx. Internal 470 x 607 x 530			
3.20	Must have Six sides heating with pre-heated gas for uniform			
	temperature			
3.21	Must be Highly efficient thermal insulation secures the inner			
	lid surface temperature when opened			
3.22	Must have Three - door chamber			
3.23	Must have Intelligent lid detection and precise gas supply to			
J.23	significantly reduce gas consumption			
	Temperature Control			
	• Ambient Environment: 18–33°C			
3.24	• Chamber Temperature: 3–55°C			
	• Temperature Deviation: ≤±0.1°C/ 0.5			
	• Temperature Control Precision: ±0.1°C			
	Gas Control			
	Must have Stable Gas Concentration maintained by advanced			
	gas system			
	CO ₂ and O ₂ Concentration Fluctuation: ≤±0.1%			
	Gas Concentration Recovery Time: <2.5 min after a 10-			
3.25	second lid opening			
	CO ₂ Control Range: 0% – 20%			
	Inlet Pressure 12-15 PSIG			
	O ₂ Control Range: 1% – 20%/ 5 to 90			
	Gas purity 99.5 %			
4	Accessories, consumables, & spare parts			

nepal in tarly
(3010
Winis and Hilliens and Boby
"MISTRY of Health alle : 431
Ministry of Health and population Director

	osbi, puectol	<u> </u>	
4. f ^{eta}	Should be supplied with standard accessories.		
5	Operating Environment		
5.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power supply, climate, temperature, humidity, etc.		
5.2	Power supply: 220-240 VAC, 50 Hz single phase fitted with appropriate 3 pin plug (flat). The power cable must be at least 3m long		
6	Standards, Medical Device Regulation and Safety Requirements		
6.1	Must submit a valid "Medical devices – Quality management systems – Requirements for regulatory purpose"; ISO 13485:2016		
6.2	Must submit automatic Renewable European CE marked certificate and USFDA approved certificate (Directive 93/42 EEC on Medical Devices) from notified body or Medical Device Regulations - MDR (EU) 2017/745 or Declaration of conformity if it falls a risk-based class I.		
7	User and Technical Training		
7.1	Must provide user training (including how to use and safe handling, keep hygiene the equipment).		
7.2	Must provide the first step technical training to troubleshooting and maintenance of the machine to the hospital technician and engineers.		
8.	Up time Service Backup, Warranty/Guarantee		
8.1	Bidders must provide a comprehensive warranty for two (2) years after acceptance, along with an additional one (1) year of free service warranty. A commitment letter from the manufacturer confirming the comprehensive warranty must be submitted with the bid. Failure to provide this document will result in bid rejection.		
8.2	The availability of the spare parts for the next 10 years must be declared by the manufacturer and bidder. Non submission leads to reject the bid.		
8.3	During the warranty period supplier must ensure preventive maintenance and corrective/ breakdown maintenance whenever required.		
9	Installation and commissioning		
9.10	The bidder must arrange for the quoted equipment with all the listed accessories and consumables to be installed on site and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.		
10	Documentation on site		
10.1	User (Operating) manual in English must be submit during the bid.		

nepalion Inch
(2010) 1/2 1/3/
Winis and boby
"Mistry of Health all 142
Ministry of Health and populated Director

Geta H	oshi, puectol		
Geta,	Non submission leads to rejection the bid.		
10.2	Service (Technical/ Maintenance) manual in English		
10.3	Spare parts and consumables price list should be provided in the separate sheet, non-submission leads to the rejection of the bid.		
10.4	Must submit the manufacturer authorisation letter or linking to the manufacturer for the sub-distributor authorisation of all the quoted items. Non submission leads to rejection the bid.		
10.5	Must submit the commitment letter for the availability of spare parts for next 10 years from the manufacturer. Non submission leads to rejection the bid.		
10.6	Note: Bidder must completely fill the Technical Specification Form (TSF). Only yes/no/all complies should not be written. Page number in the catalogue of all the required parameters must be clearly mentioned and highlighted. Failure in doing so may lead to rejection of bid from technical committee.		



Item No. 5:- Stereo Zoom Microscope for IVF

		Bi	dder's Complia	ince	Sheet
	Purchaser's Technical Specifications	Yes/No	Page No. Catalogue	in	Remarks
C N	Stereo Zoom Microscope for IVF				
S.N.	Manufacturer				
	Brand				
	Type/Model				
	Country of Origin				
1	Description of Function				
1.1	A versatile high performing stereo zoom microscope				
2	System Configuration				
2.1	Modular stereo zoom microscope with imaging				
3	Technical Specification				
3.1	Shall be Binocular/Trinocular and Gallelian apochromatic				
3.1	optical system, multicoated high-performance lenses				
3.2	Illumination Incident light, co axial light with intensity				
3.2	control and goose neck for extra illuminations				
	The transmitted light must have two options: Detachable BF				
3.3	(bright field) Dark Field base and Preconfigured mirror base				
	for transparent object				
3.4	Reticules for length measurement & counting				
3.5	Mangnification Zoom: 6:1 continuous zoom, convertible to 8 step magnification (8x-50x)				
3.6	Eyepieces WF 12.5X/22mm with foldable eye guards, +/-5mm D adjustments (optionally 16x/20x eyepieces must be available)				
3.7	IPD must be 49-78mm or better				
3.8	Viewing Angle must be 0-210deg tiltable ergonomic head				
3.9	Objectives must be 1.0x apochormatic and plan 1.25x				
3.10	Working Distance must be 106mm or better				
3.11	Magnification range must be equal to 6.4x-40x or better				
3.12	Object field diameter must range from 5.5-34 mm or better				
3.13	Must be equipped with Double Iris, Double beam splitter				
	Must have camera adapters CMOS Camera adapter, DSLR				
3.14	Camera, Iphone Adapter (with iphone or DSL camera)				
4	Accessories, consumables, & spare parts		1		
4.1	Should be supplied with standard accessories.				
5	Operating Environment		I		
	The product offered shall be designed to be stored and to				
5.1	operate normally under the conditions of the purchaser's country. The conditions include Power supply, climate,				
	temperature, humidity, etc.				
5.2	Power supply: 220-240 VAC, 50 Hz single phase fitted with appropriate 3 pin plug (flat). The power cable must be at least 3m long				
6	Standards, Medical Device Regulation and Safety				

Meralian Harey
Ministry of Health and Popular Geta Hospital Director

<u></u>	eta Hose	
	Requirements	
	Must submit a valid "Medical devices – Quality management	
6.1	systems – Requirements for regulatory purpose"; ISO 9001/13485:2016.	
	Must submit automatic Renewable European CE marked	
	certificate (Directive 93/42 EEC on Medical Devices) from	
	notified body or Medical Device Regulations - MDR (EU)	
6.2	2017/745 or Declaration of conformity if it falls a risk-based	
	class I and/or	
	US-FDA registered/approved	
7.	User and Technical Training	
7.1	Must provide user training (including how to use and safe	
7.1	handling, keep hygiene the equipment).	
	Must provide the first step technical training to	
7.2	troubleshooting and maintenance of the machine to the	
	hospital technician and engineers.	
8.	Up time Service Backup, Warranty/Guarantee	
	Bidders must provide a comprehensive warranty for two (2)	
	years after acceptance, along with an additional one (1) year	
8.1	of free service warranty. A commitment letter from the	
	manufacturer confirming the comprehensive warranty must	
	be submitted with the bid. Failure to provide this document	
	will result in bid rejection. The availability of the spare parts for the next 10 years must	
8.2	be declared by the manufacturer and bidder. Non submission	
0.2	leads to reject the bid.	
	During the warranty period supplier must ensure preventive	
8.3	maintenance and corrective/ breakdown maintenance	
0.2	whenever required.	
9	Installation and commissioning	
	The bidder must arrange for the quoted equipment with all	
	the listed accessories and consumables to be installed on site	
9.1	and commissioned by certified or qualified personnel; any	
	prerequisites for installation to be communicated to the	
	purchaser in advance, in detail.	
10	Documentation on site	
46.4	User (Operating) manual in English must be submit during	
10.1	the bid.	
10.2	Non submission leads to rejection the bid.	
10.2	Service (Technical/ Maintenance) manual in English	
10.2	Spare parts and consumables price list should be provided in	
10.3	the separate sheet, non-submission leads to the rejection of the bid.	
	Must submit the manufacturer authorisation letter or linking	
	to the manufacturer for the sub-distributor authorisation of	
10.4	all the quoted items.	
	Non submission leads to rejection the bid.	
	Must submit the commitment letter for the availability of	
10.5	spare parts for next 10 years from the manufacturer.	
	part parts for ment to Jeans from the management.	

weed in the
(2010 1100 000 0) (10 11/2)
Ministry of Health and Popular

Ge	Non submission leads to rejection the bid.		
10.6	Note: Bidder must completely fill the Technical Specification Form (TSF). Only yes/no/all complies should not be written. Page number in the catalogue of all the required parameters must be clearly mentioned and highlighted. Failure in doing so may lead to rejection of bid from technical committee.		

	Purchaser's Technical Specifications	Bidder's Compliance Sheet			
S.N.		Yes/No	Page No. Catalogue	in	Remarks
	Computer-Assisted Sperm Analyzer				
	Manufacturer				
	Brand				
	Type/Model				
	Country of Origin				
1	Description of Function				
1.1	Measures sperm concentration , motility , morphology , and DNA fragmentation 2. Suitable for infertility clinics, hospitals, research laboratories . Complies with WHO 5th edition standards for semen analysis				
2	System Configuration				
2.1	Complete set: Suitable, camera, software, microscope and computer				
3	Technical Specifications				
3.1	Capacity				
	Rapid scanning (~30s for concentration/motility, ~30s for DNA fragmentation, ~60s for morphology).				
3.2	Design				
	- Three-axis linked automatic scanning platform (X-Y-Z integrated, offset-type design) - Large heated glass plate (~130 mm x 75 mm), ±0.1°C accuracy - Can hold 1–4 slides at once				
3.3	Construction				
	- High-quality microscope (e.g., Nikon Ci-L) with 10x, 20x, and 100x (oil) plan achromatic objectives - Robust mechanical design ensures stable imaging and reliable autofocus				
3.4	Туре				
	Fully automatic, software-driven semen analysis system				
3.5	Loading				
	- 4 slide loading tray. Compatible with disposable double-chamber slides or metal disc sperm counting chambers				
3.6	Operation				

Menal	in the
Ministry of Health and Popular	
Geta HOSP!	Director

Gena PAT-based software automatically detects impurities, bubbles, and debris Microprocessor-controlled interface for real-time data analysis		ospita Director	1	
Analysis for positive phase contrast image: recognize tail, identify and filter impurities, agglutination areas and large bubbles in semen. The sperm c apture rate ≥95%. B Accurate motility analysis: PR ac accordance rate ≥ 95%. NR accordance rate ≥90%, IM accordance rate ≥ 95%. Parameters of sperm movement: three colored lines describe the straight line, curved line and average movement trajectory of any single motile sperm which displays 9 movement parameter values. 3.8 Morphological analysis module a Normal / abnormal sperm morphologic al accordance rate ≥ 98%. b Head shape discrimination accordance rate ≥ 95%. Automatically identify and analyze defects in sperm acrosome, head, neck and midpiece and tail, count TZI and SDI. d Accurately segment all parts of sperm morphology and support color and edge filling. c Independently display the head morphological analysis pictures and measurement data of each stained sperm. Available for samples with Papanicolaou, Diff-quik and other staining. Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. b Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. Accurately sament the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. Independent display of any individual sperm DNA fragmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	Geta,	1		
A identify and filter impurities, agglutination areas and large bubbles in semen. The sperm c apture rate ≥95%. B Accurate motility analysis: PR ac accordance rate ≥ 95%. NR accordance rate ≥90%, IM accordance rate ≥ 98%. Parameters of sperm movement: three colored lines describe the straight line, curved line and average movement trajectory of any single motile sperm which displays 9 movement parameter values. 3.8 Morphological analysis module Normal / abnormal sperm morphologic al accordance rate ≥ 98%. b Head shape discrimination accordance rate ≥ 95%. Automatically identify and analyze defects in sperm acrosome, head, neck and midpiece and tail, count TZI and SDI. Accurately segment all parts of sperm morphology and support color and edge filling. c Independently display the head morphological analysis pictures and measurement data of each stained sperm. Available for samples with Papanicolaou, Diff-quik and other staining. Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. b Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. Accurately asgment the halo and c ell nucleus of each sperm and support color filling. Accurately asgment the halo and c ell nucleus of each sperm and support color filling. Accurately asgment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. Independent display of any individual sperm DNA fragmentation analysis map and measurement data. 3.10 Display	3.7	Concentration /motility test module		
NR accordance rate ≥90%, IM accordance rate ≥ 98%. Parameters of sperm movement: three colored lines describe the straight line, curved line and average movement trajectory of any single motile sperm which displays 9 movement parameter values. 3.8 Morphological analysis module Normal / abnormal sperm morphologic al accordance rate ≥ 98%. b Head shape discrimination accordance rate ≥ 95%. Automatically identify and analyze defects in sperm a carosome, head, neck and midpiece and tail, count TZI and SDI. d Accurately segment all parts of sperm morphology and support color and edge filling. lindependently display the head morphological analysis pictures and measurement data of each stained sperm. Available for samples with Papanicolaou, Diff-quik and other staining. Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. lindependent display of any individual sperm DNA fragmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	A	identify and filter impurities, agglutination areas and		
the straight line, curved line and average movement trajectory of any single motile sperm which displays 9 movement parameter values. 3.8 Morphological analysis module a Normal / abnormal sperm morphologic al accordance rate ≥ 98%, b Head shape discrimination accordance rate ≥ 95%. Automatically identify and analyze defects in sperm acrosome, head, neck and midpiece and tail, count TZl and SDI. d Accurately segment all parts of sperm morphology and support color and edge filling. e Independently display the head morphological analysis pictures and measurement data of each stained sperm. Available for samples with Papanicolaou, Diff-quik and other staining. Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. Accurately segment the halo and e ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ca, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. c and support color filling. Independent display of any individual sperm DNA fragmentation analysis map and measurement data. 3.10 Display	В			
a Normal / abnormal sperm morphologic al accordance rate ≥ 98%, b Head shape discrimination accordance rate ≥ 95%. Automatically identify and analyze defects in sperm acrosome, head, neck and midpiece and tail, count TZl and SDl. d Accurately segment all parts of sperm morphology and support color and edge filling. e Independently display the head morphological analysis pictures and measurement data of each stained sperm. Available for samples with Papanicolaou, Diff-quik and other staining, Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. b Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFL. c Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. e Independent display of any individual sperm DNA fraugmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	С	the straight line, curved line and average movement trajectory of any single motile sperm which displays 9		
a 98%, b Head shape discrimination accordance rate ≥ 95%. Automatically identify and analyze defects in sperm acrosome, head, neck and midpiece and tail, count TZl and SDl. d Accurately segment all parts of sperm morphology and support color and edge filling. e Independently display the head morphological analysis pictures and measurement data of each stained sperm. Available for samples with Papanicolaou, Diff-quik and other staining. Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. b Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. c Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. e Independent display of any individual sperm DNA fragumentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	3.8	Morphological analysis module		
acrosome, head, neck and midpiece and tail, count TZl and SDI. d Accurately segment all parts of sperm morphology and support color and edge filling. e Independently display the head morphological analysis pictures and measurement data of each stained sperm. Available for samples with Papanicolaou, Diff-quik and other staining. Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. b Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. c Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. e Independent display of any individual sperm DNA fraugmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	a			
c acrosome, head, neck and midpiece and tail, count TZl and SDl. d Accurately segment all parts of sperm morphology and support color and edge filling. e Independently display the head morphological analysis pictures and measurement data of each stained sperm. Available for samples with Papanicolaou, Diff-quik and other staining. Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with 295% accuracy in identifying the number of sperm DNA fragmentation. b Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. c Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. e Independent display of any individual sperm DNA fraugmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	b	Head shape discrimination accordance rate $\geq 95\%$.		
support color and edge filling. Independently display the head morphological analysis pictures and measurement data of each stained sperm. Available for samples with Papanicolaou, Diff-quik and other staining. Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. c Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. c Independent display of any individual sperm DNA fraugmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	с	acrosome, head, neck and midpiece and tail, count TZl and SDl.		
Pictures and measurement data of each stained sperm. Available for samples with Papanicolaou, Diff-quik and other staining. Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. b Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. c Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. e Independent display of any individual sperm DNA fraugmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	d			
ther staining. Different parameter thresholds of each staining could be set to meet different experimental requirements. 3.9 DNA fragmentation analysis module Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. b Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. c Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. c Independent display of any individual sperm DNA fr augmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	e			
Using Sperm Chromatin Dispersion (SCD) to analyze, with ≥95% accuracy in identifying the number of sperm DNA fragmentation. b Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. c Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. e Independent display of any individual sperm DNA fr augmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real- time measurement and analysis displayed	f	other staining. Different parameter thresholds of each staining could be set to meet different experimental		
a ≥95% accuracy in identifying the number of sperm DNA fragmentation. b Automatically recognize DNA-intact and DNA-fragmented spermatozoa and automatically calculates DFI. c Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. e Independent display of any individual sperm DNA fraugmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	3.9	DNA fragmentation analysis module		
spermatozoa and automatically calculates DFI. c Accurately segment the halo and c ell nucleus of each sperm and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. e Independent display of any individual sperm DNA fr augmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, real-time measurement and analysis displayed	a	≥95% accuracy in identifying the number of sperm DNA		
and support color filling. Accurately analyzes detailed parameters such as halo ar ea, head area, unilateral halo thickness and tr ansverse head diameter of each sperm. Independent display of any individual sperm DNA fr augmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, realtime measurement and analysis displayed	b			
d head area, unilateral halo thickness and tr ansverse head diameter of each sperm. e Independent display of any individual sperm DNA fr augmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, realtime measurement and analysis displayed	с	and support color filling.		
a ugmentation analysis map and measurement data. 3.10 Display Computer-based user interface with color camera feed, realtime measurement and analysis displayed	d	head area, unilateral halo thickness and tr ansverse head diameter of each sperm.		
a Computer-based user interface with color camera feed, realtime measurement and analysis displayed	e			
a time measurement and analysis displayed	3.10	Display		
b Timer ~6.5s for initial auto-focus, <1s to switch from one	a	· ·		
	b	Timer ~6.5s for initial auto-focus, <1s to switch from one		

Nephalian - 4 meg	
Min Government of bobbs	
Geta Hospital Director	
Geta noviii	

	lospic Difector	
Geta	field of view to another	
3.11	Hardware	
a	Camera : 1/1.8-inch image size 3.15 Megapixel; 85 or 119 fps, clearly delineate the trajectory of sperm movement	
3.12	Safety - Software alerts if total sperm count <400 (sampling error >5%) - Automatically excludes large debris/bubbles	
3.13	Speed: Concentration/Motility: ~30s/sample - DNA Fragmentation: ~30s/sample - Morphology: ~60s/sample	
4	Accessories & consumables	
4.1	Disposable Double-Chamber slides (10 µm depth) - Metal Disc sperm counting chamber (optional) - Morphological Standard Smear, Microsphere Suspension (QC beads) for quality control	
5	Operating Environment	
5.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power supply, climate, temperature, humidity, etc.	
5.2	Power supply: 220-240 VAC, 50 Hz single phase fitted with appropriate 3 pin plug (flat). The power cable must be at least 3m long	
6	Standards, Medical Device Regulation and Safety Requirements	
6.1	Must submit ISO 13485:2003/AC:2007 for Medical Devices	
6.2	CE(93/42 EEC Directives) or USFDA registered product for the system.	
6.3	Electrical safety per relevant medical standards Alarm system for temperature deviations for additional sample and operator protection Low noise level for operator comfort.	
7	User and Technical Training	
7.1	Must provide user training (including how to use and safe handling, keep hygiene the equipment).	
7.2	Must provide the first step technical training to troubleshooting and maintenance of the machine to the hospital technician and engineers.	
8	Up time Service Backup, Warranty/Guarantee	
8.1	Bidders must provide a comprehensive warranty for two (2) years after acceptance, along with an additional one (1) year of free service warranty. A commitment letter from the manufacturer confirming the comprehensive warranty must be submitted with the bid. Failure to provide this document will result in bid rejection.	
8.2	The availability of the spare parts for the next 10 years must	

Han Han	Cy
Course of Nepalation	
Ministry of Health and Popular Direct	1
Gots Hospital Direct	'Or

	be declared by the bidder		
Geta	be declared by the bidder		
8.3	During the warranty period supplier must ensure preventive maintenance and corrective/ breakdown maintenance whenever required.		
9	Installation and commissioning		
9.1	The bidder must arrange for the quoted equipment with all the listed accessories and consumables to be installed on site and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.		
10	Documentation on site		
10.1	User (Operating) manual in English must be submit during the bid. Non submission leads to rejection the bid.		
10.2	Service (Technical/ Maintenance) manual in English		
10.3	Spare parts and consumables price list should be provided in the separate sheet.		
10.4	Must submit the manufacturer authorisation letter or linking to the manufacturer for the sub-distributor authorisation of all the quoted items. Non submission leads to rejection the bid.		
10.5	Must submit the commitment letter for the availability of spare parts for next 10 years. Non submission leads to rejection the bid.		
10.6	Note: Bidder must completely fill the Technical Specification Form (TSF). Only yes/no/all complies should not be written. Page number in the catalogue of all the required parameters must be clearly mentioned and highlighted.		

S.N.	Purchaser's Specification		Bidder's Compliance sheet		
	Technical Specification of IVF Workstation (Laminar flow)	Yes/No	Page No in Catalogue	Remarks	
	Manufacturer				
	Brand				
	Type/Model				
	Country of Origin				
1	Description of Function				
1.1	IVF workstation with laminar flow for IVF related work				
2	System Configuration				
2.1	State of the art IVF workstation for IVF operations				
3	Technical Specifications				
3.1	The IVF Workstation is designed to provide an optimal environment for handling oocytes, sperm, and embryos during ART procedures. It must offer a laminar flow work area, precise temperature control, an integrated CO ₂ incubator, and an anti-vibration system, ensuring stable conditions for sensitive micromanipulations.				
3.2	Capacity: Must be Suitable for a setrozzom microscope setups with an 800 mm × 500 mm heated work surface. Intended for use in IVF laboratories, embryology, and related clinical settings.				
3.3	Design: Must have Vertical laminar flow with ISO Class 5 clean air (HEPA + VOC filters). Integrated CO ₂ incubator for embryo culture Built-in filtration system to maintain a sterile environment.				
3.4	Air Flow				
a	Less than 0.5 cfu depositing bacterial concentration.				
b	Special designed duct that provides steady laminar flow to cover all working area.				
c	Average airflow: 0.2~0.5m/s; high speed mode at 0.32m/s;				
d	low speed mode at 0.25m/s				
3.5	Construction				
a	The Main housing must be constructed for ease of cleaning and maintenance U-shaped heating surface with high thermal conductivity 23.8-inch built-in monitor and 10.1-inch multi-function screen Integrated stainless-steel work area				
b	Internal Work Space Dimension Minimum 1800 mm (W) X 568mm (D) x 792 mm (H)				
c	Heating Table with Glass Heating Plate 600x500 mm or better				

of Health and in	Director		
Geta, Kailali	Type Tabletop IVF workstation with integrated CO ₂		
a	incubator and laminar airflow.		
	Operation Must have Microprocessor-based control system		
3.6	for temperature, airflow speed, CO ₂ concentration (in		
	incubator), and other critical parameters.		
	Display Must have 23.8-inch built-in monitor (for		
	microscope/camera display or PC) 10.1-inch		
3.7	multi-function touchscreen for real-time settings and system		
	monitoring. Must have port to connect Stereo Zoo		
	Microscope		
3.8	Timer Must have Built-in timer for booting, shutdown, and		
2.0	procedural timing.		
3.9	Controller		
a	Intuitive multi-function screen can achieve real-time		
	temperature control, function setting		
ь	Speed Airflow speed: Approximately 0.2–0.5 m/s (high/low		
	modes). No mechanical spin cycle applies.		
	Heater Integrated heated glass plate and heated tabletop (37		
С	\pm 0.1 °C) Temperature uniformity \pm 0.2 °C.		
_	Accessories CO ₂ gas connections Built-in waterproof		
4	sockets.		
5	Standard & Safety Requirements		
5.1	Must submit ISO 13485:2003/AC:2007 for Medical Devices		
5.2	Must Submit IEC 60601-1 for Medical Devices		
5.3	CE (93/42 EEC Directives) and/or USFDA registered		
5.5	product for the system.		
	- Electrical safety per relevant medical standards Alarm		
5.4	system for temperature deviations for additional sample		
	and operator protection Low noise level for operator		
	comfort.		
6	Operating Environment	Т	1
	The product offered shall be designed to be stored and to		
6.1	operate normally under the conditions of the purchaser's		
	country. The conditions include Power supply, climate, temperature, humidity, etc.		
	Power supply: 220-240 VAC, 50 Hz single phase fitted with		
6.2	appropriate 3 pin plug (flat). The power cable must be at		
0.2	least 3m long		
_	Standards, Medical Device Regulation and Safety		
7	Requirements		
	Must submit a valid "Medical devices – Quality management		
7.1	systems - Requirements for regulatory purpose"; ISO		
	13485:2016.		
	Must submit automatic Renewable European CE marked		
	certificate (Directive 93/42 EEC on Medical Devices) from		
7.2	notified body or Medical Device Regulations - MDR (EU)		
	2017/745 or Declaration of conformity if it falls a risk-based		
	class I.		

and a series	
Mi Government of Nepplation	
Geta Hospital Director	

eta HOSP		
8 eta, Kailal	User and Technical Training	
8.1	Must provide user training (including how to use and safe	
0.1	handling, keep hygiene the equipment).	
	Must provide the first step technical training to	
8.2	troubleshooting and maintenance of the machine to the	
	hospital technician and engineers.	
9	Up time Service Backup, Warranty/Guarantee	
	Bidders must provide a comprehensive warranty for two (2)	
	years after acceptance, along with an additional one (1) year	
9.1	of free service warranty. A commitment letter from the	
7.1	manufacturer confirming the comprehensive warranty must	
	be submitted with the bid. Failure to provide this document	
	will result in bid rejection.	
0.2	The availability of the spare parts for the next 10 years must	
9.2	be declared by the bidder. Non submission leads to reject the	
	bid.	
9.3	During the warranty period supplier must ensure preventive maintenance and corrective/ breakdown maintenance	
9.3	whenever required.	
10	Installation and commissioning	
10	<u> </u>	
	The bidder must arrange for the quoted equipment with all the listed accessories and consumables to be installed on site	
10.1	and commissioned by certified or qualified personnel; any	
10.1	prerequisites for installation to be communicated to the	
	purchaser in advance, in detail.	
11	Documentation on site	
	User (Operating) manual in English must be submit during	
11.1	the bid.	
1111	Non submission leads to rejection the bid.	
11.2	Service (Technical/ Maintenance) manual in English	
	Spare parts and consumables price list should be provided in	
11.3	the separate sheet, non-submission leads to the rejection of	
1110	the bid.	
	Must submit the manufacturer authorisation letter or linking	
11.4	to the manufacturer for the sub-distributor authorisation of	
11.4	all the quoted items.	
	Non submission leads to rejection the bid.	
	Must submit the commitment letter for the availability of	
11.5	spare parts for next 10 years from the bidder.	
	Non submission leads to rejection the bid.	
	Note:	
	Bidder must completely fill the Technical Specification	
11.6	Form (TSF). Only yes/no/all complies should not be written.	
	Page number in the catalogue of all the required parameters	
	must be clearly mentioned and highlighted. Failure in doing so may lead to rejection of bid from technical committee.	
	Lag many land to margation at hid those tackers of comments.	



Item	Item No. 8:- High End Colour Doppler USG Machine					
	Purchaser's Technical Specifications	Bidder's Compliance Sheet				
		Yes/No	Page No. Catalogue	in Remarks		
C N	High End Color Doppler Usg Machine	•				
S.N.	Manufacturer					
	Brand					
	Type/Model					
	Country of Origin					
1	Description of Function					
1.1	A versatile high performing small centrifuge					
2	System Configuration					
2.1	COLOR DOPPLER USG MACHINE WITH 3 PROBES					
3	Technical Specifications					
I	System should be following features:					
3.1	21" or more high-resolution Color LED with resolution of 1,920 x 1,080 and with articulating freely mobile arm for up/down side to side movement.					
3.2	10" or more LED Touch pad with resolution of 1,280 x 800 for better user control.					
3.3	System Should have min. 850,000 of processing channels or above.					
3.4	System must have an acquisition frame rate in 2D of min. 2000 frames/sec.					
3.5	System Should have higher Dynamic range 256db and above.					
3.6	Minimum 4 active probe ports with capability of connecting any of the probes in any port.					
3.7	System should support Frequency range between approx.1Mhz to 16MHz.					
3.8	Maximum Scanning depth should be 38 CM					
3.9	Must have fully Digital TGC.					
3.10	System should have upgradability option for structured reporting of Breast lesion using BIRADS scoring.					
3.11	System should have Biometry Assist Software in option					
3.12	System can able to adjust the height up/down features					
3.13	System should have min. ten parameter for minimize the noise reduction technique for better resolution					
3.14	System should have advance function with a single key which can able to adjust the 2D resolution and auto Doppler adjustment without manipulating baseline, scale etc.					

AND THE REAL PROPERTY.			
	Han ly		
Medal,	on 2 9 -1.		
Government of langual	3*		_
Ministry of Health and Polital	Dime		
Geta Hospit	Director		

	sb. puector	
Geta, V	Dual Live Mode Should support in the following mode of	
3.15	operation.	
	- B+B, B+B/C, B+B/PD, B+B/S-Flow	
(II)	- B+ElastoScan	
(II)	Imaging Modes	
3.16	B, 2B, 4B, M, CDI, CW, PWD, PDI, 3D/4D, Dual live	
3.17	System should have cine memory of at 500 MB with image storage of 45,000 frames and Cine Loop of 14,000 lines	
3.18	One button on control panel for speckle reduction imaging for soft and smooth image quality with variable values/steps.	
3.19	One button on control panel for optimization of both image quality.	
3.20	One button on control panel for optimization of color Doppler.	
3.21	One touch on the touch panel to select the most common transducer and preset combination. At least 4 quick touch preset can be selectable with different probe application	
3.22	User Programmable Preset	
3.23	User Programmable control panel Keys, as minimum four control panel key can be customizable for user comfort.	
3.24	System should have Anatomical M- Mode features	
3.25	Should have min. 4 no. of focus and 5 steps of frequency that can be selectable for differenct application preset.	
3.26	Chroma Map & Gray Map can be selectable upto 11 steps for different viewing resolution of 2D image	
(III)	PRF should be upto approx.19 KHz.	
3.27	Auto calculation for Doppler tracing on real time and post freeze.	
3.28	Pan Zoom and HD zoom facility on real time and post freeze upto 8 times or more	
3.29	Power Doppler and bidirectional color flow with HD Flow.	
3.30	General measurements and calculations and report format for Abdomen, Obstetrics, Gynaecology, Cardiac, Pelvis and Prostate.	
3.31	Annotation and directory of word.	
3.32	Minimum 5 or more USB Port	
3.33	System should have 8GB of RAM and Inbuilt SSD of 512 GB, capable of storing 3, 50,000 images.	
(IV)	Special Imaging Features:	
3.34	Tissue Harmonic Imaging	
3.35	Trapezoid Imaging on linear Probe	

Meradian - Haney	
Ministry of Health and popular Director	
Gota Hospita Director	

	Solo Director	
3.36	Steering upto approx. 20 degree independent of B Mode/PW on linear probe.	
3.37	3D/4D Mode and should be supplied the releastic face type 4D software, 5D Follice software should be supplied in standard package	
(V)	Must Provide:	
3.38	Full screen zoom	
3.39	Adaptive Color Doppler	
(VI)	Advanced upgradability should be available for future up-grade (Bidders are requested to offer their price in separate sheet for below). Also, please clearly highlighted for the same software in datasheet	
3.40	a) 2D NT b) Nerve Track c) 5D NT measurement d) 3MHz to 16 MHz Hockey Stick Linear probe for MSK Scanning. e) single crystal convex probe f) Mobile Export Software	
3.41	Advanced Software of volume follicle with 4D Endocavity Probe, please mentioned the frequency for the same probe.	
3.42	Strain Based Elastography for Breast & Thyroid.	
(VII)	Probe Details:	
3.43	All probes should have broad band/wide band technology, ultra-light weight with small foot print.	
(VIII)	One Set of Probe	
3.44	Convex Probe with approx.2 MHz to 8 MHz., should supply bopsy guide.	
3.45	4D Convex probe with approx. 4 MHz to 8 MHz	
3.46	TVS Probe with approx. 4 MHz to 9 MHz	
3.47	Phase array probe with approx: 3 MHz to 9 MHz.	
3.48	Liner array probe with aprox: 3 MHz to 16 MHz.	
(IX)	Accessories:	
3.47	Digital B/W thermal paper printer.	
3.48	Should provide online UPS 30 mins battery backup	
3.49	Thermal Paper – 10 Piece	
4	Operating Environment	
4.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power supply, climate, temperature, humidity, etc.	
4.2	Power supply: 220-240 VAC, 50 Hz single phase fitted with appropriate 3 pin plug (flat). The power cable must be at least 3m long	

	Harla	
of Nepalati	ion Z	
Winisi Covernment and Popul	/	
Ministry of Health and Portal	Director	
Gota HOSP!	Pilector	

Geta H	Psh. Director	
5 ^{Geta, 1}	Standards, Medical Device Regulation and Safety Requirements	
5.1	Must submit a valid "Medical devices – Quality management systems – Requirements for regulatory purpose"; ISO 13485:2016.	
5.2	Must submit automatic Renewable European CE marked certificate (Directive 93/42 EEC on Medical Devices) from notified body or Medical Device Regulations - MDR (EU) 2017/745 or Declaration of conformity if it falls a risk-based class I & US-FDA approved product	
6	User and Technical Training	
6.1	Must provide user training (including how to use and safe handling, keep hygiene the equipment).	
6.2	Must provide the first step technical training to troubleshooting and maintenance of the machine to the hospital technician and engineers.	
7	Up time Service Backup, Warranty/Guarantee	
7.1	Bidders must provide a comprehensive warranty for two (2) years after acceptance, along with an additional one (1) year of free service warranty. A commitment letter from the manufacturer confirming the comprehensive warranty must be submitted with the bid. Failure to provide this document will result in bid rejection.	
7.2	The availability of the spare parts for the next 10 years must be declared by the manufacturer and bidder. Non submission leads to reject the bid.	
7.3	During the warranty period supplier must ensure preventive maintenance and corrective/ breakdown maintenance whenever required.	
8	Installation and commissioning	
8.1	The bidder must arrange for the quoted equipment with all the listed accessories and consumables to be installed on site and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.	
9	Documentation on site	
9.1	User (Operating) manual in English must be submit during the bid. Non submission leads to rejection the bid.	
9.2	Service (Technical/ Maintenance) manual in English.	
9.3	Spare parts and consumables price list should be provided in the separate sheet, non-submission leads to the rejection of the bid.	
9.4	Must submit the authorisation letter or linking to the manufacturer for the sub-distributor authorisation of all the quoted items. Non submission leads to rejection the bid.	

Menal	on - Unity.
Government of Novikan Ministry of Health and Popular	Director

Geta no			
Geta, K	Must submit the commitment letter for the availability of		
9.5	spare parts for next 10 years from the manufacturer.		
	Non submission leads to rejection the bid.		
	Note:		
	Bidder must completely fill the Technical Specification		
9.6	Form (TSF). Only yes/no/all complies should not be written.		
	Page number in the catalogue of all the required parameters		
	must be clearly mentioned and highlighted. Failure in doing		
	so may lead to rejection of bid from technical committee.		



Item No. 9:- 4K Hysteroscopy System

	Purchaser's Specification		Bidder's Offer		
S.N.	4K Hysteroscopy System	Yes/No Page no. in catalogue Remains			
	Manufacturer:	1	g	1	
	Brand:				
	Type/Model:				
	Country of Origin:				
1	Description of Function				
1.1	Operative/Diagnosite Hysterescopy system used for the treatment or examine the inside of the uterus (womb) using a thin, lighted tube called a hysteroscope, which is inserted through the cervix.				
2	Operational Requirement				
2.1	Hysteroscopy camera system with video processor and light source, suction and irrigation pump, insufflator and scope.				
3	System Configuration				
3.1	4k Video Processor – 1 unit				
3.2	4K CMOS Camera head – 1 unit				
3.3	LED Light Source – 1 unit				
3.4	Monitor- 1 unit				
3.5	Hysteroscopy Pump system- 1 unit				
3.6	Hysteroscope set- 1 unit				
3.7	Insufflator- 1 unit				
3.8	Trolley- 1 unit				
4	Technical specifications				
4.1	Video Processor- 1 unit				
4.1.1	4K UHD resolution video processor with image enhancement, gain setting and zoom/freeze function.				
4.1.2	Processor should have Class I type of protection against electric shock and IP rating of at least IP20				
4.1.3	Signal Output: HDMI-4K, HDMI-Full HD, SDI-4K, SDI Full HD				
4.1.4	Should have facilities for USB memory with capture module and user preset for surgical application.				
4.1.5	Power Supply: rated voltage- 100-240 VAC; within ± 10 % Rated frequency- 50/60 Hz; within ± 1 Hz				
4.1.6	System must have the option for white balance and black balance.				

Merajon - Hanly	_
Government of Popular Ministry of Health and Popular School Hospital Director	

seta Hosp.	Director			
4.1.7	Processor should have touch screen that should be displaying the image as secondary monitor during the case			
4.1.8	Should have image resolution of UHD 3840x2160, 1.600 lines and Signal to noise ratio of minimum 54 dB			
4.2	LED Light Source			
4.2.1	LED light source with high intensity and low heat conduction.			
4.2.2	LED Light with minimum 50,000 hours of life			
4.2.3	LED Light source should have Class I type of protection against electric shock			
4.2.4	Automatic and manual LED brightness adjustment			
4.2.5	LED light guide cable of diameter 3.5mm and length of 300cm suitable for 2.9mm scope should be provided.			
4.3	Camera Head- 1 unit			
4.3.1	CMOS image sensor			
4.3.2 Aspect ratio of at least 16:9				
4.3.3 Weight should not be more than 250g				
4.3.4	3.4 Camera head with minimum 3.0 m of cable length			
4.3.5	4.3.5 3 Programmable camera Head Push Buttons			
4.3.6	.6 Manual/ Automated Cleaning			
4.3.7	7 Should be able to control light intensity via camera head			
4.4	4.4 Monitor- 1 unit			
4.4.1	4K Ultra HD Medical grade monitor			
4.4.2	Screen size: 31 inches or more			
4.4.3	Display devices: TFT Active Matrix LCD			
4.4.4	Resolution: 4K ultra-HD (4096x2160) pixels			
4.4.5	Aspect ratio: 17:9			
4.4.6	Viewing angle: 89° or more			
4.4.7	4.4.7 Number of display colors: approximately 1 billion colors			
4.4.8	4.4.8 Maximum power consumption of 180W			
4.4.9	Video Output/ Input > DVI-D, SDI (BNC), HDMI, Composite video(1xBNC), 3G-SDI(1xBNC), RGB			
4.5	Hystersocopy Pump- 1 unit			
4.5.1	The system should offer pressure regulated irrigation function along with suction.			
4.5.2	The system should be equipped with high-resolution color touchscreen for operation.			

y of Health and Seta Hosp!	Director			
4.5.3	The system should be compatible for various procedures such as laparoscopy, arthroscopy, hysteroscopy and urology.			
С	Suction capacity of maximum 2,200 ml/min and vacuum of maximum 480 mmHg $\pm10~\%$			
4.5.5	The system should not weight more than 10kg.			
4.5.6	Power supply: 100-240 V/AC, 50-60 Hz			
4.6	Trolley- 1 unit			
4.6.1	Bidders should provide high quality trolley with gas bottle holder, camera head fixing device and holder for mounting monitor.			
4.7	Hysteroscope- 1 set			
4.7.1	Hysteroscope with 2.9 mm diameter, 30-degree AOV and 302 mm length - 1 unit			
4.7.2	Scope should be autoclavable.			
4.7.3	Should come with following instruments.			
4.7.4	Diagnostic sheath of diameter 4mm for 2.9mm scopes with 1 rotatable stopclock of working length 270mm, rounded for 30 degree scopes should be provided 1 unit			
4.7.5	Obturator for diagnostic sheath should be provided 1 unit			
4.7.6	Biopsy foreceps oval, double action, flexible, 3 charr diameter and WL of 40cm should be provided 1 unit			
4.7.7	Grasping Forceps alligator jaw flexible, double action 3 charr diameter and WL of 40 cm should be provided 1unit			
4.7.8	Scissors blunt tip, 3 charr, 400mm length should be provided 1unit			
4.8	Insufflator- 1 unit			
4.8.1	It should be minimum 45 Liters High flow insufflator with LCD display measurement and gas heating mechanism should be provided.			
4.8.2	Should be fully automatic			
4.8.3	It should be used with Gas Cylinder or Central Gas Supply			
4.8.4	It should have visible or audible notification to notify actual vs required flow rate			
4.8.5	It should have internal venting system for safety.			
5	Accessories, consumables, & spare parts			
5.1	Should be supplied with standard accessories.			
6	Operating Environment			
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power supply, climate, temperature, humidity, etc.			

of Health and Po	Director			
eta nozra	Power supply: 220-240 VAC, 50 Hz single phase fitted with			
6.2	appropriate 3 pin plug (flat). The power cable must be at			
	least 3m long			
7	Standards, Medical Device Regulation and Safety Requirements			
7.1	Must submit a valid "Medical devices – Quality management			
7.1	systems – Requirements for regulatory purpose"; ISO 13485:2016.			
	Must submit automatic Renewable European CE marked			
7.2	certificate (Directive 93/42 EEC on Medical Devices) from notified body or Medical Device Regulations - MDR (EU)			
1.2	2017/745 or Declaration of conformity if it falls a risk-based			
	class I, and/or US FDA listing or registration.			
8	User and Technical Training			
0.4	Must provide user training (including how to use and safe			
8.1	handling, keep hygiene the equipment).			
8.2	Must provide the first step technical training to troubleshooting and maintenance of the machine to the			
8.2	hospital technician and engineers.			
9	Up time Service Backup, Warranty/Guarantee			
	Bidders must provide a comprehensive warranty for two (2)			
	years after acceptance, along with an additional one (1) year			
9.1	of free service warranty. A commitment letter from the			
7.1	manufacturer confirming the comprehensive warranty must			
	be submitted with the bid. Failure to provide this document will result in bid rejection.			
	The availability of the spare parts for the next 10 years must			
9.2	be declared by the bidder. Non submission leads to reject the			
	bid.			
0.0	During the warranty period supplier must ensure preventive			
9.3	maintenance and corrective/ breakdown maintenance whenever required.			
10	Installation and commissioning			
10	The bidder must arrange for the quoted equipment with all			
	the listed accessories and consumables to be installed on site			
10.1	and commissioned by certified or qualified personnel; any			
	prerequisites for installation to be communicated to the			
	purchaser in advance, in detail.			
11	Documentation on site			
11.1	User (Operating) manual in English must be submit during			
11.1	the bid. Non submission leads to rejection the bid.			
11.2	Non submission leads to rejection the bid. Service (Technical/ Maintenance) manual in English			
11.2	Spare parts and consumables price list should be provided in			
11.3	the separate sheet, non-submission leads to the rejection of			
	the bid.			
11.4	Must submit the manufacturer authorisation letter or linking			
11.1	to the manufacturer for the sub-distributor authorisation of all			

Varmont Of Inc			
y of Health and Po	Director		
eta Hosy			,
Geta, Kallan	1		
	Non submission leads to rejection the bid.		
	Must submit the commitment letter for the availability of		
11.5	spare parts for next 10 years from the manufacturer.		
	Non submission leads to rejection the bid.		
	Note:		
	Bidder must completely fill the Technical Specification Form		
11.6	(TSF). Only yes/no/all complies should not be written. Page		
11.0	number in the catalogue of all the required parameters must		
	be clearly mentioned and highlighted. Failure in doing so		
	may lead to rejection of bid from technical committee.		



Drawings N/A

The purpose of drawings is to specify locations, dimensions, materials to be used, stages of manufacturing, and other characteristics of the Goods and Related Services. The Purchaser should prepare such drawings, as needed, and include them in the Procurement Document. Such drawings, as part of the SR, are Contract documents and, therefore, shall be part of the Contract. Similarly, the Purchaser may request the Supplier to provide drawings either with its Bid or for approval during Contract execution.

It is essential that the Purchaser prepares a **List of Drawings** showing all drawings it supplied and issued as part of the Procurement Document.



Inspection and Test

As per SCC

Section VI. General Conditions of Contract

The GCC contain standard provisions that have been designed to remain unchanged and **to be used without modifying their text**. The GCC clearly identify the provisions that may normally need to be specified for a particular bidding process and require that such specification be introduced **through the SCC**.

The GCC are a Contract document and, therefore, are a part of the Contract.

Table of Clauses

1.	Definitions
2.	Contract Documents
3.	Fraud and Corruption
4.	Interpretation
5.	Language
6.	Joint Venture, Consortium or Association
7.	Notices 104
8.	Governing Law
9.	Settlement of Disputes
10.	Scope of Supply
11.	Delivery
12.	Supplier's Responsibilities
13.	Purchaser's Responsibilities
14.	Contract Price
15.	Terms of Payment
16.	Taxes and Duties
17.	Performance Security
18.	Copyright107
19.	Confidential Information
20.	Subcontracting
21.	Specifications and Standards
22.	Packing and Documents
23.	Insurance
24.	Transportation
25.	Inspections and Tests
26.	Liquidated Damages111
27.	Warranty111

28.	Patent Indemnity	112
29.	Limitation of Liability	113
30.	Change in Laws and Regulations	114
31.	Force Majeure	114
32.	Change Orders and Contract Amendments	115
33.	Extensions of Time	115
34.	Termination	116
35.	Assignment	117

Section VI. General Conditions of Contract

1. Definitions

- 1.1. The following words and expressions shall have the meanings hereby assigned to them:
 - (a) "Contract" means the Agreement entered into between the Purchaser and the Supplier, together with the Contract Documents referred to therein, including all attachments, appendices, and all documents incorporated by reference therein.
 - (b) "Contract Documents" means the documents listed in the Agreement, including any amendments thereto.
 - (c) "Contract Price" means the price payable to the Supplier as specified in the Agreement, subject to such additions and adjustments thereto or deductions there from, as may be made pursuant to the Contract.
 - (d) "Day" means calendar day.
 - (e) "Delivery" means the transfer of the Goods from the Supplier to the Purchaser in accordance with the terms and conditions set forth in the Contract.
 - (f) "Completion" means the fulfillment of the Related Services by the Supplier in accordance with the terms and conditions set forth in the Contract.
 - (g) "GCC" means the General Conditions of Contract.
 - (h) "Goods" means all of the commodities, raw material, machinery and equipment, and/or other materials that the Supplier is required to supply to the Purchaser under the Contract.
 - (i) "Purchaser's Country" is the country specified in the Special Conditions of Contract (SCC).
 - (j) "Purchaser" means the entity purchasing the Goods and Related Services, as specified in the SCC.
 - (k) "Related Services" means the services incidental to the supply of the goods, such as insurance, installation, training and initial maintenance and other similar obligations of the Supplier under the Contract.
 - (1) "SCC" means the Special Conditions of Contract.

- (m) "Subcontractor" means any natural person, private or government entity, or a combination of the above, including its legal successors or permitted assigns, to whom any part of the Goods to be supplied or execution of any part of the Related Services is subcontracted by the Supplier.
- (n) "Supplier" means the natural person, private or government entity, or a combination of the above, whose bid to perform the Contract has been accepted by the Purchaser and is named as such in the Agreement, and includes the legal successors or permitted assigns of the Supplier.
- (o) "GoN" means the Government of Nepal.
- (p) "The Site," where applicable, means the place named in the SCC.

2. Contract Documents

2.3 Subject to the order of precedence set forth in the Agreement, all documents forming the Contract (and all parts thereof) are intended to be correlative, complementary, and mutually explanatory.

3. Fraud and Corruption

3.1 If the Purchaser determines that the Supplier has engaged in corrupt, fraudulent, collusive, coercive or obstructive practices, in competing for or in executing the Contract, then the Purchaser may, after giving 14 days notice to the Supplier, terminate the Supplier's employment under the Contract and the provisions of GCC Clause 34.1 shall apply.

For the purposes of this Sub-Clause:

- (i) "corrupt practice" is the offering, giving, receiving or soliciting, directly or indirectly, of anything of value to influence improperly the actions of another party;
- (ii) "fraudulent practice" 13 is any act or omission, including a misrepresentation, that knowingly or recklessly misleads, or attempts to mislead, a party to obtain a financial or other benefit or to avoid an obligation;
- (iii) "collusive practice" 14 is an arrangement between two or more parties designed to achieve an improper purpose, including to influence improperly the actions of another party;

bid prices at artificial, non competitive levels.

a "party" refers to a public official; the terms "benefit" and "obligation" relate to the procurement process or contract execution; and the "act or omission" is intended to influence the procurement process or contract execution.

"parties" refers to participants in the procurement process (including public officials) attempting to establish

- (iv) "coercive practice" 15 is impairing or harming, or threatening to impair or harm, directly or indirectly, any party or the property of the party to influence improperly the actions of a party;
- (v) "obstructive practice" is
- (aa) deliberately destroying, falsifying, altering or concealing of evidence material to the investigation or making false statements to investigators in order to materially impede a GoN/DP investigation into allegations of a corrupt, fraudulent, coercive or collusive practice; and/or threatening, harassing or intimidating any party to prevent it from disclosing its knowledge of matters relevant to the investigation or from pursuing the investigation; or
- (bb) acts intended to materially impede the exercise of the GoN/DP's inspection and audit rights provided for under ITB Clause 3.5 and GCC Clause 25.
- 3.2 Without prejudice to any other rights of the Purchaser under this Contract, on the recommendation of the Purchaser, Public Procurement Monitoring Office may **blacklist** a Bidder/Supplier for its conduct for a period of one (1) to three (3) years on the following grounds and seriousness of the act committed by the bidder:
 - (a) if it is established that the Supplier committed acts specified in ITB 3.2,
 - (b) if it is established later that the Bidder has committed substantial defect in implementation of the contract or has not substantially fulfilled its obligations under the contract or the completed work is not of the specified quality as per the contract.

3.3 Incase of DP funded bid, DP:

(a) will cancel the portion of the loan allocated to a contract if it determines at any time that representatives of the Borrower or of a beneficiary of the loan engaged in corrupt, fraudulent, collusive or coercive practices during the procurement or the execution of that contract, without the Borrower having taken timely and appropriate action satisfactory to the Bank to remedy the situation;

a "party" refers to a participant in the procurement process or contract execution.

- (b) will sanction a firm or individual, including declaring them ineligible, either indefinitely or for a stated period of time, to be awarded a Bank-financed contract if it at any time determines that they have, directly or through an agent, engaged, in corrupt, fraudulent, collusive or coercive practices in competing for, or in executing, a Bank-financed contract; and
- (c) will have the right to require that Suppliers to permit the Bank to inspect their accounts and records and other documents relating to the bid submission and contract performance and to have them audited by auditors appointed by the Bank.

4. Interpretation

4.1 If the context so requires it, singular means plural and vice versa.

4.2 Incoterms

- (a) The meaning of any trade term and the rights and obligations of parties thereunder shall be as prescribed by Incoterms.
- (b) EXW, CIF, CIP, and other similar terms, shall be governed by the rules prescribed in the current edition of Incoterms, published by the International Chamber of Commerce at the date of the Invitation for Bids or as specified in the SCC.

4.3 Entire Agreement

The Contract constitutes the entire agreement between the Purchaser and the Supplier and supersedes all communications, negotiations and agreements (whether written or oral) of parties with respect thereto made prior to the date of Contract.

4.4 Amendment

No amendment or other variation of the Contract shall be valid unless it is in writing, is dated, expressly refers to the Contract, and is signed by a duly authorized representative of each party thereto.

4.5 Nonwaiver

(a) Subject to GCC Sub-Clause 4.5(b) below, no relaxation, forbearance, delay, or indulgence by either party in enforcing any of the terms and conditions of the Contract or the granting of time by either party to the other shall prejudice, affect, or restrict the rights of that party under the Contract, neither shall any waiver by either party of any breach of Contract operate as waiver of any subsequent or

continuing breach of Contract.

(b) Any waiver of a party's rights, powers, or remedies under the Contract must be in writing, dated, and signed by an authorized representative of the party granting such waiver, and must specify the right and the extent to which it is being waived.

4.6 Severability

If any provision or condition of the Contract is prohibited or rendered invalid or unenforceable, such prohibition, invalidity or unenforceability shall not affect the validity or enforceability of any other provisions and conditions of the Contract.

5. Language

- 5.1 The Contract as well as all correspondence and documents relating to the Contract exchanged by the Supplier and the Purchaser, shall be written in the language **specified in the SCC**. Supporting documents and printed literature that are part of the Contract may be in another language provided they are accompanied by an accurate translation of the relevant passages in the language specified in the SCC, in which case, for purposes of interpretation of the Contract, this translation shall govern.
- 5.2 The Supplier shall bear all costs of translation to the governing language and all risks of the accuracy of such translation.

6. Joint Venture, Consortium or Association

- 6.1 If the Supplier is a joint venture, consortium, or association, all of the parties shall be jointly and severally liable to the Purchaser for the fulfillment of the provisions of the Contract and shall designate one party to act as a leader with authority to bind the joint venture, consortium, or association. A bidder can submit only one bid either as a partner of the joint venture or individually. The composition or the constitution of the joint venture, consortium, or association shall not be altered without the prior consent of the Purchaser.
- 6.2 The contractor shall not handover the responsibility of the contract to any one member or some members of Joint Venture or any other parties, not involved in the contract.

7. Notices

7.1 Any Notice given by one party to the other pursuant to the Contract shall be in writing to the address **specified in the SCC**. The term "in writing" means communicated in written

- form with proof of receipt.
- 7.2 A Notice shall be effective when delivered or on the Notice's effective date, whichever is later.
- 8. Governing Law
- 8.1 The Contract shall be governed by and interpreted in accordance with the laws of Nepal.
- 9. Settlement of Disputes
- 9.1 The Purchaser and the Supplier shall make every effort to settle amicably by direct negotiation any disagreement or dispute arising between them under or in connection with the Contract.
- 9.2 Any dispute between the Parties as to matters arising pursuant to this Contract which cannot be settled amicably within thirty (30) days after receipt by one Party of the other Party's request for such amicable settlement may be referred to Arbitration within 30 days after the expiration of amicable settlement period as specified in SCC.
- 10. Scope of Supply
- 10.1 Subject to the SCC, the Goods and Related Services to be supplied shall be as specified in Section V, Schedule of Requirements.
- 10.2 Unless otherwise stipulated in the Contract, the Scope of Supply shall include all such items not specifically mentioned in the Contract but that can be reasonably inferred from the Contract as being required for attaining Delivery and Completion of the Goods and Related Services as if such items were expressly mentioned in the Contract.
- 11. Delivery
- 11.1 Subject to GCC Sub-Clause 32.1, the Delivery of the Goods and Completion of the Related Services shall be in accordance with the Delivery and Completion Schedule specified in the Section V, Schedule of Requirements. The details of documents to be furnished by the Supplier are **specified in the SCC**.
- 12. Supplier's Responsibilities
- 12.1 The Supplier shall supply all the Goods and Related Services included in the Scope of Supply in accordance with GCC Clause 10, and the Delivery and Completion Schedule, as per GCC Clause 11.
- 13. Purchaser's Responsibilities
- 13.1 Whenever the supply of Goods and Related Services requires that the Supplier obtain permits, approvals, and import and other licenses from public authorities in Nepal, the Purchaser shall, if so required by the Supplier, make its best effort to assist the Supplier in complying with such requirements in a timely and expeditious manner.

13.2 The Purchaser shall pay all costs involved in the performance of its responsibilities, in accordance with GCC Sub-Clause 13.1.

14. Contract Price

- 14.1 The Contract Price shall be as specified in the Agreement subject to any additions and adjustments thereto, or deductions there from, as may be made pursuant to the Contract.
- 14.2 Prices charged by the Supplier for the Goods delivered and the Related Services performed under the Contract shall not vary from the prices quoted by the Supplier in its bid, with the exception of any price adjustments authorized in the SCC.

15. Terms of Payment

- 15.1 The Contract Price shall be paid as specified in SCC.
- 15.2 The Supplier's request for payment shall be made to the Purchaser in writing, accompanied by invoices describing, as appropriate, the Goods delivered and Related Services performed, and by the documents submitted pursuant to GCC Clause 11.1 and upon fulfillment of all the obligations stipulated in the Contract.
- 15.3 Unless otherwise specified in the SCC, the Purchaser shall retain from each payment due to the Contractor the proportion stated in the SCC. Half of the total amount retained shall be repaid to the Supplier not later than thirty (30) days following the date of completion of the Supplier's performance obligations under the Contract, including any warranty obligations and half when the supplier has submitted the evidence of submission of tax return to the concerned Internal Revenue Office.
- 15.4 Payments shall be made promptly by the Purchaser, no later than thirty (30) days after submission of an invoice or request for payment by the Supplier, and the Purchaser has accepted it.
- 15.5 The currency or currencies in which payments shall be made to the Supplier under this Contract shall be as specified in the SCC.
- 15.6 In the event that the Purchaser fails to pay the Supplier any payment by its due date or within the period set forth in the GCC 15.4, the Purchaser shall pay to the Supplier interest on the amount of such delayed payment at the rate shown in the SCC, for the period of delay until due payment has been made.

16. Taxes and Duties

- 16.1 For goods supplied from outside Nepal, the Supplier shall be entirely responsible for all taxes, stamp duties, license fees, and other such levies imposed outside Nepal. However, Tax deduction at source shall be applied as per taxation laws of Nepal.
- 16.2 For goods supplied from within the Nepal, the Supplier shall be entirely responsible for all taxes, duties, license fees, etc., incurred until delivery of the contracted Goods to the Purchaser. Tax deduction at source shall be applied as per taxation law of Nepal.
- 16.3 If any tax exemptions, reductions, allowances or privileges may be available to the Supplier in Nepal, the Purchaser shall use its best efforts to enable the Supplier to benefit from any such tax savings to the maximum allowable extent.

17. Performance Security

- 17.1 The Supplier shall, within fifteen (15) days of the receipt of notification of Contract award, provide a Performance Security for the due performance of the Contract in the amounts and currencies specified in the SCC.
- 17.2 The proceeds of the Performance Security shall be payable to the Purchaser as compensation for any loss resulting from the Supplier's failure to complete its obligations under the Contract.
- 17.3 The Performance Security shall be denominated in the currency of the Contract, or in a freely convertible currency acceptable to the Purchaser, and shall be in one of the forms stipulated by the Purchaser in the SCC, or in another form acceptable to the Purchaser.
- 17.4 The Performance Security shall be discharged by the Purchaser and returned to the Supplier not later than thirty (30) days following the date of completion of the Supplier's performance obligations under the Contract, including any warranty obligations.

18. Copyright

18.1 The copyright in all drawings, documents, and other materials containing data and information furnished to the Purchaser by the Supplier herein shall remain vested in the Supplier, or, if they are furnished to the Purchaser directly or through the Supplier by any third party, including suppliers of materials, the copyright in such materials shall remain vested in such third party.

19. Confidential

19.1 The Purchaser and the Supplier shall keep confidential and

Information

shall not, without the written consent of the other party hereto, divulge to any third party any documents, data, or other information furnished directly or indirectly by the other party hereto in connection with the Contract, whether such information has been furnished prior to, during or following completion or termination of the Contract. Notwithstanding the above, the Supplier may furnish to its Subcontractor such documents, data, and other information it receives from the Purchaser to the extent required for the Subcontractor to perform its work under the Contract, in which event the Supplier shall obtain from such Subcontractor an undertaking of confidentiality similar to that imposed on the Supplier under GCC Clause 19.

- 19.2 The Purchaser shall not use such documents, data, and other information received from the Supplier for any purposes unrelated to the Contract. Similarly, the Supplier shall not use such documents, data, and other information received from the Purchaser for any purpose other than the design, procurement, or other work and services required for the performance of the Contract.
- 19.3 The obligation of a party under GCC Sub-Clauses 19.1 and 19.2 above, however, shall not apply to information that:
 - (a) the Purchaser or Supplier need to share with the Donor for Donor funded project or other institutions participating in the financing of the Contract;
 - (b) now or hereafter enters the public domain through no fault of that party;
 - (c) can be proven to have been possessed by that party at the time of disclosure and which was not previously obtained, directly or indirectly, from the other party; or
 - (d) otherwise lawfully becomes available to that party from a third party that has no obligation of confidentiality.
- 19.4 The above provisions of GCC Clause 19 shall not in any way modify any undertaking of confidentiality given by either of the parties hereto prior to the date of the Contract in respect of the Supply or any part thereof.
- 19.5 The provisions of GCC Clause 19 shall survive completion or termination, for whatever reason, of the Contract.

20. Subcontracting

20.1 The Supplier shall notify the Purchaser in writing of all subcontracts awarded under the Contract if not already

specified in the Bid. Subcontracting shall in no event relieve the Supplier from any of its obligations, duties, responsibilities, or liability under the Contract.

20.2 Subcontracts shall comply with the provisions of GCC Clauses 3.

21. Specifications and Standards

21.1 Technical Specifications and Drawings

- (a) The Supplier shall ensure that the Goods and Related Services comply with the technical specifications and other provisions of the Contract.
- (b) The Supplier shall be entitled to disclaim responsibility for any design, data, drawing, specification or other document, or any modification thereof provided or designed by or on behalf of the Purchaser, by giving a notice of such disclaimer to the Purchaser.
- (c) The Goods and Related Services supplied under this Contract shall conform to the standards mentioned in Section V, Schedule of Requirements and, when no applicable standard is mentioned, the standard shall be equivalent or superior to the official standards whose application is appropriate to the country of origin of the Goods.
- 21.2 Wherever references are made in the Contract to codes and standards in accordance with which it shall be executed, the edition or the revised version of such codes and standards shall be those specified in the Section V, Schedule of Requirements. During Contract execution, any changes in any such codes and standards shall be applied only after approval by the Purchaser and shall be treated in accordance with GCC Clause 32.

22. Packing and Documents

- 22.1 The Supplier shall provide such packing of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the Contract. During transit, the packing shall be sufficient to withstand, without limitation, rough handling and exposure to extreme temperatures, salt and precipitation, and open storage. Packing case size and weights shall take into consideration, where appropriate, the remoteness of the final destination of the Goods and the absence of heavy handling facilities at all points in transit.
- 22.2 The packing, marking, and documentation within and

outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified in the SCC, and in any other instructions ordered by the Purchaser.

23. Insurance

23.1 Unless otherwise specified in the SCC, the Goods supplied under the Contract shall be fully insured, in a freely convertible currency from an eligible country, against loss or damage incidental to manufacture or acquisition, transportation, storage, and delivery, in accordance with the applicable Incoterms or in the manner specified in the SCC.

24. Transportation

24.1 Obligations for transportation of the Goods shall be in accordance with the Incoterms specified in Sections V, Schedule of Requirements.

25. Inspections and Tests

- 25.1 The Supplier shall at its own expense and at no cost to the Purchaser carry out all such tests and/or inspections of the Goods and Related Services as are specified in Sections V, Schedule of Requirements.
- 25.2 The inspections and tests may be conducted on the premises of the Supplier or its Subcontractor, at point of delivery, and/or at the final destination of the Goods, or in another place in Nepal as specified in the SCC. Subject to GCC SubClause 25.3, if conducted on the premises of the Supplier or its Subcontractor, all reasonable facilities and assistance, including access to drawings and production data, shall be furnished to the inspectors at no charge to the Purchaser.
- 25.3 The Purchaser or its designated representative shall be entitled to attend the tests and/or inspections referred to in GCC Sub-Clause 25.2, provided that the Purchaser bear all of its own costs and expenses incurred in connection with such attendance including, but not limited to, all traveling and board and lodging expenses.
- 25.4 Whenever the Supplier is ready to carry out any such test and inspection, it shall give a reasonable advance notice, including the place and time, to the Purchaser. The Supplier shall obtain from any relevant third party or manufacturer any necessary permission or consent to enable the Purchaser or its designated representative to attend the test and/or inspection.
- 25.5 The Purchaser may require the Supplier to carry out any test and/or inspection not required by the Contract but deemed necessary to verify that the characteristics and performance of

the Goods comply with the technical specifications, codes and standards under the Contract, provided that the Supplier's reasonable costs and expenses incurred in the carrying out of such test and/or inspection shall be added to the Contract Price. Further, if such test and/or inspection impede the progress of manufacturing and/or the Supplier's performance of its other obligations under the Contract, due allowance will be made in respect of the Delivery Dates and Completion Dates and the other obligations so affected.

- 25.6 The Supplier shall provide the Purchaser with a report of the results of any such test and/or inspection.
- 25.7 The Purchaser may reject any Goods or any part thereof that fail to pass any test and/or inspection or do not conform to the specifications. The Supplier shall either rectify or replace such rejected Goods or parts thereof or make alterations necessary to meet the specifications at no cost to the Purchaser, and shall repeat the test and/or inspection, at no cost to the Purchaser, upon giving a notice pursuant to GCC Sub-Clause 25.4.
- 25.8 The Supplier agrees that neither the execution of a test and/or inspection of the Goods or any part thereof, nor the attendance by the Purchaser or its representative, nor the issue of any report pursuant to GCC Sub-Clause 25.6, shall release the Supplier from any warranties or other obligations under the Contract.

26. Liquidated Damages

26.1 Except as provided under GCC Clause 31, if the Supplier fails to deliver any or all of the Goods or perform the Related Services within the period specified in the Contract, the Purchaser may without prejudice to all its other remedies under the Contract, deduct from the Contract Price, as liquidated damages, a sum equivalent to the percentage specified in the SCC of the Contract Price for each day of delay until actual delivery or performance, up to a maximum deduction of the percentage specified in the SCC. Once the maximum is reached, the Purchaser may terminate the Contract pursuant to GCC Clause 34.

27. Warranty

- 27.1 The Supplier warrants that all the Goods are new, unused, and of the most recent or current models, and that they incorporate all recent improvements in design and materials, unless provided otherwise in the Contract.
- 27.2 Subject to GCC Sub-Clause 21.1, the Supplier further warrants that the Goods shall be free from defects arising

- from any act or omission of the Supplier or arising from design, materials, and workmanship, under normal use in the conditions prevailing in Nepal.
- 27.3 Unless otherwise specified in the SCC, the warranty shall remain valid for twelve (12) months after the Goods, or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the SCC.
- 27.4 The Purchaser shall give Notice to the Supplier stating the nature of any such defects together with all available evidence thereof, promptly following the discovery thereof. The Purchaser shall afford all reasonable opportunity for the Supplier to inspect such defects.
- 27.5 Upon receipt of such Notice, the Supplier shall, within the period specified in the SCC, expeditiously repair or replace the defective Goods or parts thereof, at no cost to the Purchaser.
- 27.6 If having been notified, the Supplier fails to remedy the defect within the period specified in the SCC, the Purchaser may proceed to take within a reasonable period such remedial action as may be necessary, at the Supplier's risk and expense and without prejudice to any other rights which the Purchaser may have against the Supplier under the Contract.

28. Patent Indemnity

- 28.1 The Supplier shall, subject to the Purchaser's compliance with GCC Sub-Clause 28.2, indemnify and hold harmless the Purchaser and its employees and officers from and against any and all suits, actions or administrative proceedings, claims, demands, losses, damages, costs, and expenses of any nature, including attorney's fees and expenses, which the Purchaser may suffer as a result of any infringement or alleged infringement of any patent, utility model, registered design, trademark, copyright, or other intellectual property right registered or otherwise existing at the date of the Contract by reason of:
 - (a) the installation of the Goods by the Supplier or the use of the Goods in the country where the Site is located; and
 - (b) the sale in any country of the products produced by the Goods.

Such indemnity shall not cover any use of the Goods or any part thereof other than for the purpose indicated by or to be reasonably inferred from the Contract, neither any infringement resulting from the use of the Goods or any part thereof, or any products produced thereby in association or combination with any other equipment, plant, or materials not supplied by the Supplier, pursuant to the Contract.

- 28.2 If any proceedings are brought or any claim is made against the Purchaser arising out of the matters referred to in GCC Sub-Clause 28.1, the Purchaser shall promptly give the Supplier a notice thereof, and the Supplier may at its own expense and in the Purchaser's name conduct such proceedings or claim and any negotiations for the settlement of any such proceedings or claim.
- 28.3 If the Supplier fails to notify the Purchaser within thirty (30) days after receipt of such notice that it intends to conduct any such proceedings or claim, then the Purchaser shall be free to conduct the same on its own behalf.
- 28.4 The Purchaser shall, at the Supplier's request, afford all available assistance to the Supplier in conducting such proceedings or claim, and shall be reimbursed by the Supplier for all reasonable expenses incurred in so doing.
- 28.5 The Purchaser shall indemnify and hold harmless the Supplier and its employees, officers, and Subcontractors from and against any and all suits, actions or administrative proceedings, claims, demands, losses, damages, costs, and expenses of any nature, including attorney's fees and expenses, which the Supplier may suffer as a result of any infringement or alleged infringement of any patent, utility model, registered design, trademark, copyright, or other intellectual property right registered or otherwise existing at the date of the Contract arising out of or in connection with any design, data, drawing, specification, or other documents or materials provided or designed by or on behalf of the Purchaser.

29. Limitation of Liability

- 29.1 Except in cases of gross negligence or willful misconduct :
 - (a) neither party shall be liable to the other party for any indirect or consequential loss or damage, loss of use, loss of production, or loss of profits or interest costs, provided that this exclusion shall not apply to any obligation of the Supplier to pay liquidated damages to

the Purchaser; and

(b) the aggregate liability of the Supplier to the Purchaser, whether under the Contract, in tort, or otherwise, shall not exceed the total Contract Price, provided that this limitation shall not apply to the cost of repairing or replacing defective equipment, or to any obligation of the Supplier to indemnify the Purchaser with respect to patent infringement.

30. Change in Laws and Regulations

30.1 Unless otherwise specified in the Contract, if after the date of the Invitation for Bids, any law, regulation, ordinance, order or bylaw having the force of law is enacted, promulgated, abrogated, or changed in Nepal where the Site is located (which shall be deemed to include any change in interpretation or application by the competent authorities) that subsequently affects the Delivery Date and/or the Contract Price, then such Delivery Date and/or Contract Price shall be correspondingly increased or decreased, to the extent that the Supplier has thereby been affected in the performance of any of its obligations under the Contract. Notwithstanding the foregoing, such additional or reduced cost shall not be separately paid or credited if the same has already been accounted for in the price adjustment provisions where applicable, in accordance with GCC Clause 14.

31. Force Majeure

- 31.1 The Supplier shall not be liable for forfeiture of its Performance Security, liquidated damages, or termination for default if and to the extent that its delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.
- 31.2 For purposes of this Clause, "Force Majeure" means an event or situation beyond the control of the Supplier that is not foreseeable, is unavoidable, and its origin is not due to negligence or lack of care on the part of the Supplier. Such events may include, but not be limited to, acts of the Purchaser in its sovereign capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions, and freight embargoes.
- 31.3 If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such condition and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for

performance not prevented by the Force Majeure event.

32. Change Orders and Contract Amendments

- 32.1 The Purchaser may at any time order the Supplier through Notice in accordance GCC Clause 7, to make changes within the general scope of the Contract in any one or more of the following:
 - (a) drawings, designs, or specifications, where Goods to be furnished under the Contract are to be specifically manufactured for the Purchaser:
 - (b) the method of shipment or packing;
 - (c) the place of delivery; and
 - (d) the Related Services to be provided by the Supplier.
- 32.2 If any such change causes an increase or decrease in the cost of, or the time required for, the Supplier's performance of any provisions under the Contract, an equitable adjustment shall be made in the Contract Price or in the Delivery and Completion Schedule, or both, and the Contract shall accordingly be amended. Any claims by the Supplier for adjustment under this Clause must be asserted within thirty (30) days from the date of the Supplier's receipt of the Purchaser's change order.
- 32.3 Prices to be charged by the Supplier for any Related Services that might be needed but which were not included in the Contract shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

33. Extensions of Time

- 33.1 If at any time during performance of the Contract, the Supplier or its Subcontractors should encounter conditions impeding timely delivery of the Goods or completion of Related Services pursuant to GCC Clause 11, the Supplier shall promptly, and at least twenty one (21) days before the expiry of procurement contract, notify the Purchaser in writing of the delay, its likely duration, and its cause. As soon as practicable after receipt of the Supplier's notice, the Purchaser shall evaluate the situation and may at its discretion extend the Supplier's time for performance, in which case the extension shall be ratified by the parties by amendment of the Contract.
- 33.2 Except in case of Force Majeure, as provided under GCC Clause 31, a delay by the Supplier in the performance of its Delivery and Completion obligations shall render the

Supplier liable to the imposition of liquidated damages pursuant to GCC Clause 26, unless an extension of time is agreed upon, pursuant to GCC Sub-Clause 34.1.

34. Termination

34.1 Termination for Default

- (a) The Purchaser, without prejudice to any other remedy for breach of Contract, by Notice of default sent to the Supplier, may terminate the Contract in whole or in part:
 - (i) if the Supplier fails to deliver any or all of the Goods within the period specified in the Contract, or within any extension thereof granted by the Purchaser pursuant to GCC Clause 33;
 - (ii) if the Supplier fails to perform any other obligation under the Contract.
 - (iii) The supplier uses the advance payment for matters other than the contarctual obligations.
 - (iv) The purchaser may terminate the contract at any time in the following condition.
 - (a) does not commence the work as per the contract.
 - (b) abandons the contract with out completing,
 - (c) fails to achieve progress as per the contract.
- (b) In the event the Purchaser terminates the Contract in whole or in part, pursuant to GCC Clause 34.1(a), the Purchaser may procure, upon such terms and in such manner as it deems appropriate, Goods or Related Services similar to those undelivered or not performed, and the Supplier shall be liable to pay to the Purchaser for any whole costs for such similar Goods or Related Services. However, the Supplier shall continue performance of the Contract to the extent not terminated.
- (c) if the Supplier, in the judgment of the Purchaser has engaged in corrupt, fraudulent, collusive, coercive or obstructive practices, as defined in GCC Clause 3, in competing for or in executing the Contract

34.2 Termination for Insolvency

The Purchaser may at any time terminate the Contract by giving Notice to the Supplier if the Supplier becomes bankrupt or otherwise insolvent. In such event, termination will be without compensation to the Supplier, provided that such termination will not prejudice or affect any right of action or remedy that has accrued or will accrue thereafter to the Purchaser.

34.3 Termination for Convenience

- (a) The Purchaser, by written Notice sent to the Supplier, may terminate the Contract, in whole or in part, at any time for its convenience. The Notice of termination shall specify that termination is for the Purchaser's convenience, the extent to which performance of the Supplier under the Contract is terminated, and the date upon which such termination becomes effective.
- (b) The Goods that are complete and ready for shipment within seven (7) days after the Supplier's receipt of the Notice of termination shall be accepted by the Purchaser at the Contract terms and prices. For the remaining Goods, the Purchaser may elect:
 - (i) To have any portion completed and delivered at the Contract terms and prices; and/or
 - (ii) To cancel the remainder and pay to the Supplier an agreed amount for partially completed Goods and Related Services and for materials and parts previously procured by the Supplier.

35. Assignment

35.1 Neither the Purchaser nor the Supplier shall assign, in whole or in part, their obligations under this Contract, except with prior written consent of the other party.

Section VII. Special Conditions of Contract

The following Special Conditions of Contract (SCC) shall supplement the General Conditions of Contract (GCC). Whenever there is a conflict, the provisions herein shall prevail over those in the GCC.

address of Suppliers including telephone number, facsimile number and electronic mail address (if applicable)] Name and Address of the Supplier: Telephone number: Facsimile number: e-mail Address: In case of arbitration, the arbitration shall be conducted in accordant with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall	GCC 1.1(i)	The Purchaser's country is: Nepal
GCC 4.2 (b) The version of Incoterms shall be: 2020 GCC 5.1 The language shall be: English GCC 7.1 For notices, the Purchaser's address shall be: Name and Address of the Purchaser: Geta Hospital Geta Kailali, Telephone number: Facsimile number: NA e-mail Address: procurement.getahospital@gmail.com For notices, the Suppliers's address shall be: [insert full name a address of Suppliers including telephone number, facsimile numb and electronic mail address (if applicable)] Name and Address of the Supplier: Telephone number: Facsimile number: e-mail Address: GCC 9.2 In case of arbitration, the arbitration shall be conducted in accordan with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall	GCC 1.1(j)	The Purchaser is: Geta Hospital Geta, Kailali
GCC 5.1 The language shall be: English GCC 7.1 For notices, the Purchaser's address shall be: Name and Address of the Purchaser: Geta Hospital Geta Kailali, Telephone number: Facsimile number: NA e-mail Address: procurement.getahospital@gmail.com For notices, the Suppliers's address shall be: [insert full name a address of Suppliers including telephone number, facsimile number and electronic mail address (if applicable)] Name and Address of the Supplier: Telephone number: Facsimile number: e-mail Address: GCC 9.2 In case of arbitration, the arbitration shall be conducted in accordan with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall	GCC 1.1 (p)	The Site is: Geta Hospital Geta, Kailali
GCC 7.1 For notices, the Purchaser's address shall be: Name and Address of the Purchaser: Geta Hospital Geta Kailali, Telephone number: Facsimile number: NA e-mail Address: procurement.getahospital@gmail.com For notices, the Suppliers's address shall be: [insert full name a address of Suppliers including telephone number, facsimile numb and electronic mail address (if applicable)] Name and Address of the Supplier: Telephone number: Facsimile number: e-mail Address: GCC 9.2 In case of arbitration, the arbitration shall be conducted in accordan with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall	GCC 4.2 (b)	The version of Incoterms shall be: 2020
Name and Address of the Purchaser: Geta Hospital Geta Kailali, Telephone number: Facsimile number: NA e-mail Address: procurement.getahospital@gmail.com For notices, the Suppliers's address shall be: [insert full name a address of Suppliers including telephone number, facsimile numb and electronic mail address (if applicable)] Name and Address of the Supplier: Telephone number: Facsimile number: e-mail Address: GCC 9.2 In case of arbitration, the arbitration shall be conducted in accordan with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall	GCC 5.1	The language shall be: English
Telephone number: Facsimile number: NA e-mail Address: procurement.getahospital@gmail.com For notices, the Suppliers's address shall be: [insert full name a address of Suppliers including telephone number, facsimile numb and electronic mail address (if applicable)] Name and Address of the Supplier: Telephone number: Facsimile number: e-mail Address: GCC 9.2 In case of arbitration, the arbitration shall be conducted in accordan with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall	GCC 7.1	For <u>notices</u> , the Purchaser's address shall be:
Facsimile number: NA e-mail Address: procurement.getahospital@gmail.com For notices, the Suppliers's address shall be: [insert full name a address of Suppliers including telephone number, facsimile numb and electronic mail address (if applicable)] Name and Address of the Supplier: Telephone number: Facsimile number: e-mail Address: GCC 9.2 In case of arbitration, the arbitration shall be conducted in accordan with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall		Name and Address of the Purchaser: Geta Hospital Geta Kailali,
e-mail Address: procurement.getahospital@gmail.com For notices, the Suppliers's address shall be: [insert full name a address of Suppliers including telephone number, facsimile numb and electronic mail address (if applicable)] Name and Address of the Supplier: Telephone number: Facsimile number: e-mail Address: GCC 9.2 In case of arbitration, the arbitration shall be conducted in accordan with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall		Telephone number:
For notices, the Suppliers's address shall be: [insert full name a address of Suppliers including telephone number, facsimile numb and electronic mail address (if applicable)] Name and Address of the Supplier: Telephone number: Facsimile number: e-mail Address: GCC 9.2 In case of arbitration, the arbitration shall be conducted in accordan with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall		Facsimile number: NA
address of Suppliers including telephone number, facsimile number and electronic mail address (if applicable)] Name and Address of the Supplier: Telephone number: Facsimile number: e-mail Address: In case of arbitration, the arbitration shall be conducted in accordant with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall		e-mail Address: procurement.getahospital@gmail.com
Telephone number: Facsimile number: e-mail Address: In case of arbitration, the arbitration shall be conducted in accordant with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall		For notices, the Suppliers's address shall be: [insert full name and address of Suppliers including telephone number, facsimile number and electronic mail address (if applicable)]
Facsimile number: e-mail Address: In case of arbitration, the arbitration shall be conducted in accordant with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall		Name and Address of the Supplier:
e-mail Address: GCC 9.2 In case of arbitration, the arbitration shall be conducted in accordant with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall		Telephone number:
GCC 9.2 In case of arbitration, the arbitration shall be conducted in accordance with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: [insert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall		Facsimile number:
with the arbitration procedures in accordance with law of Nepal Lalitpur Nepal GCC 10.1 The Scope of Supply shall be defined in: finsert "Section V, Schedule Requirements" or indicate where the Scope of Supply shall		e-mail Address:
Requirements" or indicate where the Scope of Supply shall	GCC 9.2	In case of arbitration, the arbitration shall be conducted in accordance with the arbitration procedures in accordance with law of Nepal at Lalitpur Nepal
defined.]	GCC 10.1	The Scope of Supply shall be defined in: [insert "Section V, Schedule of Requirements" or indicate where the Scope of Supply shall be defined.]
Details of shipping and documents to be furnished by the Supplier sh be:	GCC 11.1	Details of shipping and documents to be furnished by the Supplier shall be:

"For Goods supplied from abroad as per Incoterms CIP:

Upon shipment, the Supplier shall notify the Purchaser and the Insurance Company by telex or fax the full details of the shipment, including Contract number, description of Goods, quantity, the vessel, the bill of lading number and date, port of loading, date of shipment, port of discharge, etc. The Supplier shall send the following documents to the Purchaser, with a copy to the Insurance Company:

- a) *One* copy of the Supplier's invoice showing the description of the Goods, quantity, unit price, and total amount;
- **b)** Original and *One* copy of the negotiable, clean, on-board bill of lading marked "freight prepaid" and *One* copy of nonnegotiable bill of lading;
- c) One copy of the packing list identifying contents of each package;
- d) Insurance certificate;
- e) Manufacturer's or Supplier's warranty certificate;
- f) Inspection certificate, issued by the nominated inspection agency, and the Supplier's factory inspection report; and
- g) Certificate of origin.

The Purchaser shall receive the above documents at least one week before arrival of the Goods at the port or place of arrival and, if not received, the Supplier will be responsible for any consequent expenses.

For Goods from within the Purchaser's country as per Incoterm EXW:

Upon delivery of the Goods to the transporter, the Supplier shall notify the Purchaser and send the following documents to the Purchaser:

- a) Copies of the Supplier's invoice showing the description of the Goods, quantity, unit price, and total amount;
- b) Copy of packing list indentifying the contents of each package;
- c) Delivery note, railway receipt, or truck receipt;
- d) Manufacturer's or Supplier's warranty certificate;
- e) Certificate of origin; and
- f) Inspection certificate issued by the nominated inspection agency, and the Supplier's factory inspection report;

The Purchaser shall receive the above documents before the arrival of the Goods and, if not received, the Supplier will be responsible for any

	consequent expenses.
GCC 14.2	The prices charged for the Goods delivered and the Related Services to be performed shall be fixed for the duration of the contract. ¹⁶ N/A
GCC 15.1	The terms of payment to be made to the Supplier under the contract shall be as follows:
	 The payment shall be made: (a) Through an irrevocable confirmed letter of credit opened in favour of the Supplier or (b) Through accounts division/unit of the Purchaser or (c) Through the District Treasury Controller Office.
	(i) Advance Payment: Not Applicable
	(ii) On Shipment: Eighty (80) percent of the Contract Price of the Goods shipped shall be paid through irrevocable letter of credit opened in favor of the Supplier in a bank in its country, upon submission of documents specified in GCC Clause 11.
	(iii) On Acceptance: Twenty (20) percent of the Contract Price of Goods received shall be paid within thirty (30) days of receipt of the Goods upon submission of claim supported by the acceptance certificate issued by the Purchaser.
	Payment of local currency portion shall be made in Nepalese Rupees within thirty (30) days of presentation of claim supported by a certificate from the Purchaser declaring that the Goods have been delivered and that all other contracted Services have been performed.
	 B. If the Bidder is Nepali bidder: Payment of foreign currency portion shall be made in Nepalese Rupees converted at the exchange rate declared by NRB for the date of bid opening in the following manner: (i) Advance Payment: Ten (10) percent of the Contract Price shall be paid within thirty (30) days of signing of the Contract against a simple receipt and a bank guarantee for the equivalent amount and in the form provided in the bidding documents or another form acceptable to the Purchaser.
	(ii) On Delivery: Seventy (70) percent of the Contract Price

 $\underline{^{16}}$ If duration of contract is less than 12 months price adjustment shall not be applicable

145

GCC 15.3	of original invoice upon which the goods have been received at the custom point and the custom bill of the GoN. (iii) On Acceptance: Twenty (20) percent of the Contract Price of Goods received shall be paid within thirty (30) days of receipt of the Goods upon submission of claim supported by the acceptance certificate issued by the Purchaser. Payment of local currency portion shall be made in Nepalese Rupees within thirty (30) days of presentation of claim supported by a certificate from the Purchaser declaring that the Goods have been delivered and that all other contracted Services have been performed. Payment for Goods and Services supplied from within the Purchaser's country (In case of goods already imported or goods manufactured within the purchaser's country): Payment for Goods and Services supplied from within the Purchaser's country shall be made in Nepalese Rupees, as follows: (i) Advance Payment: Ten (10) percent of the Contract Price shall be paid within thirty (30) days of signing of the Contract against a simple receipt and a bank guarantee for the equivalent amount and in the form provided in the bidding documents or another form acceptable to the Purchaser. (ii) On Delivery: Seventy (70) percent of the Contract Price shall be paid on receipt of the Goods and upon submission of the documents specified in GCC Clause 11. (iii) On Acceptance: The remaining twenty (20) percent of the Contract Price shall be paid to the Supplier within thirty (30) days after the date of the acceptance certificate for the respective delivery issued by the Purchaser The proportion of payments retained is: 5 (five) percent (Note: Mention not applicable only if supplier's performance and/or warranty obligations under the Contract do not remain after payment.)
GCC 15.5	The currencies for payments shall be: Nepalese Rupees
CCC 15.5	
GCC 15.3	The proportion of payments retained is: 5 (five)) percent (Note: Mention not applicable only if supplier's performance and/or warranty
	(iii) On Acceptance: The remaining twenty (20) percent of the Contract Price shall be paid to the Supplier within thirty (30) days after the date of the acceptance certificate for the
	shall be paid on receipt of the Goods and upon submission of
	shall be paid within thirty (30) days of signing of the Contract against a simple receipt and a bank guarantee for the equivalent amount and in the form provided in the bidding documents or
	the Purchaser's country (In case of goods already imported
	within thirty (30) days of presentation of claim supported by a certificate from the Purchaser declaring that the Goods have been
	Price of Goods received shall be paid within thirty (30) days of receipt of the Goods upon submission of claim supported by the

GCC 17.1	The Supplier shall provide a Performance Security as follows:	
	I. If bid price of the bidder selected for acceptance is up to 15 (fifteen) percent less than the approved cost estimate, the performance security amount shall be 5 (five) percent of the bid price.	
	II. For the bid price of the bidder selected for acceptance is more than 15 (fifteen) percent below of the cost estimate, the performance security amount shall be determined as follows:	
	Performance Security Amount = [(0.85 x Cost Estimate –Bid Price) x 0.5] + 5% of Bid Price.	
	The Bid Price and Cost Estimate shall be exclusive of Value Added Tax.	
	The Performance Security shall be valid for the period of warranty period of Goods + 30 (Thirty) Days	
	The performance security shall be forfeited, in case the Supplier fails to complete the contractual obligation and rectify the defects within warranty period.	
GCC 17.3	The types of acceptable Performance Securities are: A bank guarantee issued by Commercial Bank or Financial Institution eligible to issue Bank Guarantee as per prevailing Law located in Nepal or commercial bank located abroad, acceptable to the Purchaser, in the format included in Section VIII, Contract Forms. Performance Security issued by foreign Bank must be counter – guaranteed by Commercial Bank or Financial Institution eligible to issue Bank Guarantee as per prevailing Law in Nepal.	
GCC 22.2	The packing, marking, and documentation within and outside the packages shall be: [insert in detail the type of packing, the markings, and all documentation required. Consideration should be given to whether, during sea transport, the Goods should be transported under deck. The provision must further expressly provide details for markings and documentation to be required within and outside the packing cases.]	
GCC 23.1	The insurance coverage shall be in accordance with: The Supplier must insure the Goods in an amount equal to 110 percent of the CIP price of the Goods from "Warehouse" to "Warehouse" on "All Risks" basis, including War Risks, Arson and Strikes."	

GCC 25.2	Tests and Inspections specified in Section V, Schedule of Requirements, shall be carried out at the following times or milestones, and places: Post-shipment Inspection at Geta Hospital, Nepal after Installation		
GCC 26.1	The applicable rate of liquidated damages shall be: 0.05 percent of the Contract Price per day		
GCC 26.1	The maximum amount of liquidated damages shall be: ten (10) percent of the Contract Price.		
	The contract shall be terminated, if liquidated damages exceeds percent of the Contract Price and blacklisting process shall be initial for the Supplier's failure to complete the contractual obligations.		
GCC 27.3	The period of validity of the Warranty shall be: As mentioned in technical specification		
	For the purposes of the Warranty, the place of final destination shall be: Geta Hospital Geta, Kailali		
GCC 27.5 & 27.6	The Supplier shall correct any defects covered by the Warranty within: 30 days of being notified by the Purchaser of the occurrence of such defects		

Section VIII. Contract Forms

Table of Forms

Letter of Intent	126
Letter of Acceptance	127
Agreement Form	128
Performance Security	129
Advance Payment Security	130

Letter of Intent

[on letterhead paper of the Purchaser]

. date.

Notes on Letter of Intent			
The issuance of Letter of Intent is the information of the selection of the bid of the successful bidder by the Purchaser and for providing information to other unsuccessful bidders who participated in the bid as regards to the outcome of the procurement process. This standard form of Letter of Intent to Award should be filled in and sent to the successful Bidder only after evaluation and selection of substantially responsible lowest evaluated bid.			
To:			
This is to notify you that, it is our intention to award the contract. for execution of the			
of the contract and identification number, as given in the Contract Data/SCC to you as your bid price			
Authorized Signature:			
Name:			
Title:			
CC: Insert name and address of all other Bidders, who submitted the bid]			

Letter of Acceptance [on letterhead paper of the Purchaser]

date	
To: name and address of the Supplier	
Subject: . <u>Notification of Award</u>	
This is to notify that your Bid dated date for execution of the name of the contract and identification number, as given in the Contract Data/SCC for the Contract price of Nepalese Rupees [insert amount in figures and words in Nepalese Rupees], as corrected in accordance with the Instructions to Bidders is hereby accepted in accordance with the Instruction to Bidders.	
You are hereby instructed to contract this office to sign the formal contract agreement within 15 days. As per the Conditions of Contract, you are also required to submit Performance Security, as specified in SCC, consisting of a Bank Guarantee in the format included in Section VIII (Contract Forms) of the Bidding Document.	
The amout of performance security shall be NRs[Insert amount] and validity period of performance security shall be[insert validity period].	
The Purchaser shall forfeit the bid security, in case you fail to furnish the Performance Security and to sign the contract within specified period.	
Authorized Signature:	
Name and Title of Signatory:	

Agreement Form

THIS AGREEMENT made on the [insert number] day of [insert month], [insert year], between [insert complete name of Purchaser] of [insert complete address of Purchaser] (hereinafter "the Purchaser"), of the one part, and [insert complete name of Supplier] of [insert complete address of Supplier] (hereinafter "the Supplier"), of the other part:

WHEREAS the Purchaser invited Bids for certain Goods and Related Services, viz., [insert brief description of the Goods and Related Services] and has accepted a Bid by the Supplier for the supply of those Goods and Related Services in the sum of NRs[insert amount of contract price in words and figures including taxes] (hereinafter "the Contract Price").

NOW THIS AGREEMENT WITNESSETH AS FOLLOWS:

- 1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Conditions of Contract referred to.
- 2. The following documents shall be deemed to form and be read and construed as part of this Agreement, viz.:
 - (a) the Purchaser's Notification to the Supplier of Award of Contract;
 - (b) the Bid Submission Form and the Price Schedules submitted by the Supplier;
 - (c) the Special Conditions of Contract;
 - (d) the General Conditions of Contract;
 - (e) the Schedule of Requirements; and
 - (f) [indicate any other documents required as appropriate]

This Contract shall prevail over all other Contract documents. In the event of any discrepancy or inconsistency within the Contract documents, then the documents shall prevail in the order listed above.

- 3. In consideration of the payments to be made by the Purchaser to the Supplier as indicated in this Agreement, the Supplier hereby covenants with the Purchaser to provide the Goods and Related Services and to remedy defects therein in conformity in all respects with the provisions of the Contract.
- 4. The Purchaser hereby covenants to pay the Supplier in consideration of the provision of the Goods and Related Services and the remedying of defects therein, the Contract Price or such other sum as may become payable under the provisions of the Contract at the times and in the manner prescribed by the Contract.

IN WITNESS whereof the parties hereto have caused this Agreement to be executed in accordance with the laws of "Nepal" on the day, month, and year indicated above.

Signed by [insert authorized signature for the Purchaser] (for the Purchaser)

Signed by [insert authorized signature for the Supplier] (for the Supplier)

Performance Security

[insert complete name and number of Contract]

To: [insert complete name of Purchaser]

WHEREAS [insert complete name of Supplier] (hereinafter "the Supplier") has received the notification of award for the execution of [insert identification number and name of contract] (hereinafter "the Contract").

AND WHEREAS it has been stipulated by you in the aforementioned Contract that the Supplier shall furnish you with a security [*insert type of security*] issued by a reputable guarantor for the sum specified therein as security for compliance with the Supplier's performance obligations in accordance with the Contract.

AND WHEREAS the undersigned [insert complete name of Guarantor], legally domiciled in [insert complete address of Guarantor], (hereinafter the "Guarantor"), have agreed to give the Supplier a security:

THEREFORE WE hereby affirm that we are Guarantors and responsible to you, on behalf of the Supplier, up to a total of [insert currency and amount of guarantee in words and figures] and we undertake to pay you, upon your first written demand declaring the Supplier to be in default under the Contract, without cavil or argument, any sum or sums within the limits of [insert currency and amount of guarantee in words and figures] as aforesaid, without your needing to prove or to show grounds or reasons for your demand or the sum specified therein.

This security is valid until the [insert day, month, year].

Name: [insert complete name of person signing the Security]

In the capacity of: [insert legal capacity of person signing the Security]

Signed: [insert signature of person whose name and capacity are shown above]

Duly authorized to sign the security for and on behalf of: [insert seal and complete name of Guarantor]

Date: [insert date of signing]

Advance Payment Security

[insert complete name and number of Contract]

To: [insert complete name of Purchaser]

In accordance with the payment provision included in the Contract, in relation to advance payments, [insert complete name of Supplier] (hereinafter called "the Supplier") shall deposit with the Purchaser a security consisting of [indicate type of security], to guarantee its proper and faithful performance of the obligations imposed by said Clause of the Contract, in the amount of [insert currency and amount of guarantee in words and figures].

We, the undersigned [insert complete name of Guarantor], legally domiciled in [insert full address of Guarantor] (hereinafter "the Guarantor"), as instructed by the Supplier, agree unconditionally and irrevocably to guarantee as primary obligator and not as surety merely, the payment to the Purchaser on its first demand without whatsoever right of objection on our part and without its first claim to the Supplier, in the amount not exceeding [insert currency and amount of guarantee in words and figures].

This security shall remain valid and in full effect from the date of the advance payment being received by the Supplier under the Contract until [(insert day, month, year) Contract completion date may be a basis for this date].

Name: [insert complete name of person signing the Security]

In the capacity of: [insert legal capacity of person signing the Security]

Signed: [insert signature of person whose name and capacity are shown above]

Duly authorized to sign the security for and on behalf of: [insert seal and complete name of Guarantor]

Date: [insert date of signing]

SECTION-VI

Bill of Quantities

Notes for Unit Rate Contracts:

Objectives

The objectives of the Bill of Quantities are

- (a) to provide sufficient information on the quantities of Works to be performed to enable Bids to be prepared efficiently and accurately; and
- (b) when a Contract has been entered into, to provide a priced Bill of Quantities for use in the periodic valuation of Works executed.

In order to attain these objectives, Works should be itemized in the Bill of Quantities in sufficient detail to distinguish between the different classes of Works, or between Works of the same nature carried out in different locations or in other circumstances which may give rise to different considerations of cost. Consistent with these requirements, the layout and content of the Bill of Quantities should be as simple and brief as possible.

Content

The Bill of Quantities should be divided generally into the following sections:

- (a) Preamble;
- (b) Work Items (grouped into parts);
- (c) Day works Schedule;
- d) Provisional Sums; and
- (d) Summary.

Preamble

The Preamble should indicate the inclusiveness of the unit prices, and should state the methods of measurement which have been adopted in the preparation of the Bill of Quantities and which are to be used for the measurement of any part of the works.

Work Items

The items in the Bill of Quantities should be grouped into sections to distinguish between those parts of the Works which by nature, location, access, timing, or any other special characteristics may give rise to different methods of construction, or phasing of the Works, or considerations of cost. General items common to all parts of the works may be grouped as a separate section in the Bill of Quantities. Day work Schedule

A Day work Schedule should be included only if the probability of unforeseen work, outside the items included in the Bill of Quantities, is high. To facilitate checking by the Employer of the realism of rates quoted by the Bidders, the Day work Schedule should normally comprise the following:

- (a) A list of the various classes of labor, materials, and Constructional Plant for which basic day work rates or prices are to be inserted by the Bidder, together with a statement of the conditions under which the Contractor will be paid for work executed on a day work basis.
- (b) Nominal quantities for each item of Day work, to be priced by each Bidder at Day work rates as bid. The rate to be entered by the Bidder against each basic Day work item should include the Contractor's profit, overheads, supervision, and other charges.

Provisional Sums

A general provision for physical contingencies (quantity overruns) may be made by including a provisional sum in the Summary Bill of Quantities. Similarly, a contingency allowance for possible price increases should be provided as a provisional sum in the Summary Bill of Quantities. The inclusion of such provisional sums often facilitates budgetary approval by avoiding the need to request periodic supplementary approvals as the future need arises. Where such provisional sums or contingency allowances are used, the Contract Data should state the manner in which they will be used, and under whose authority (usually the Project Manager's).

Summary

The Summary should contain a tabulation of the separate parts of the Bill of Quantities carried forward, with provisional sums for Day work, for physical (quantity) contingencies, and for price contingencies (upward price adjustment) where applicable.

These Notes for Preparing Specifications are intended only as information for the Employer or the person drafting the Bidding documents. They should not be included in the final documents.

Bill of Quantities

Total of Procument Items	
Total Item Price	
VAT	
Grand Total	