
(a) detector, C-ArmLSpine held out
% found on 56 held-out frames (28 cases)
confidence threshold
vertebra recall, outline truth
sacrum found
default 0.25
cross-device 0.74
(b) paired comparisons
zero-shot naming, lateral minus AP
593 vertebrae, 140 cases
+57.7 [48.9, 66.4], Holm p 0.002
real-film detector, film-trained
minus render-trained*
233 bodies, 25 spines
+27.5 [20.5, 35.2], Holm p 0.002
sacrum found at 0.25, same-device
minus held-out detector
56 frames, 28 cases
+14.3 [5.4, 23.2], Holm p 0.002
device card minus uncalibrated
physics, R-squared x 100
24 held-out folds, four devices
+3.8 [2.0, 5.7], Holm p 0.002
paired difference, percentage points
(95 % cluster bootstrap of the difference)
* the film-trained detector's number is not held out
(196 of 233 bodies sit in its training bank)
(c) device cards: gain over uncalibrated physics
R-squared gain, card minus physics (held-out fold)
one held-out fold
mean over folds, 95 % bootstrap over folds
0/6 negative
1/6 negative
1/6 negative
2/6 negative
Philips BV Pulsera
Siemens Cios Fusion
Elekta XVI
PIXIUM 4600