Advanced Computational Frameworks for Vertebral Body Detection and 2D/3D Registration in Intraoperative Fluoroscopy

1. The Clinical Imperative and the Domain Gap

The transition from preoperative static volumetric imaging to dynamic, intraoperative two-dimensional (2D) fluoroscopy represents one of the most significant computational bottlenecks in modern image-guided spine surgery. Preoperative Computed Tomography (CT) provides high-fidelity, three-dimensional (3D) representations of patient anatomy, which are essential for precise surgical planning, particularly in procedures requiring sub-millimeter accuracy such as pedicle screw placement. However, intraoperative fluoroscopy serves as the primary modality for real-time surgical execution, fundamentally stripping away critical depth information due to its projective nature. The reliance on planar X-ray imagery to evaluate complex 3D spinal anatomy frequently leads to anatomical misinterpretations; clinical evaluations demonstrate that the visual detection of pedicle screw breaches based solely on 2D fluoroscopy is highly subjective and yields low accuracy rates, sometimes failing to detect up to 37% of malplacements.

To mitigate the risks of neurological injury and subsequent revision surgeries caused by hardware malplacement, computer-assisted surgical navigation systems employ 2D/3D registration. This process computationally aligns the preoperative 3D CT volume with the intraoperative 2D fluoroscopic coordinate system, overlaying the planned trajectories onto the live surgical field. Deep learning architectures designed to automate vertebral body detection, segmentation, and pose estimation have shown immense theoretical promise. Yet, a pervasive analytical paradox persists: these convolutional models frequently reach a performance plateau at approximately 90% accuracy in the fluoroscopic domain, while simultaneously achieving 99% parity on highly standardized lateral or anterior-posterior (AP) radiographs.

This performance gap is deeply rooted in two fundamental challenges. First, there is a pronounced domain shift between the synthetic Digitally Reconstructed Radiographs (DRRs) used to train these models and the physical realities of clinical fluoroscopy. Second, the fluoroscopic acquisition space possesses a vastly higher dimensionality than standard radiography, rendering traditional uniform dataset scaling inefficient and largely obsolete. Overcoming these barriers requires a departure from traditional deep learning dataset expansion, moving toward physics-informed differentiable rendering, coverage-driven sampling, and patient-specific self-supervised optimization.

2. Decoupling Sample Count from Generalization: The Acquisition-Space Diversity Framework

A central query in modern surgical computer vision is determining whether the injection of more

training data can elevate saturated models from the 90% accuracy plateau to the 99% parity observed in lateral X-rays. The conventional approach to deep learning assumes that expanding the training dataset will proportionally and continuously increase generalization capabilities. However, in the context of mobile C-arm fluoroscopy, this assumption is structurally flawed.

2.1 The Dimensionality of the Fluoroscopic Domain

Standardized lateral and AP radiographs possess a highly constrained geometric variance. The acquisition space is inherently low-dimensional because variables such as the Source-to-Image Distance (SID), patient positioning, and equipment angle are rigidly controlled by strict radiological protocols. Consequently, the sampling density for standard radiography saturates quickly, and the underlying data distribution is easily mapped by a convolutional network, allowing for rapid convergence to 99% accuracy.

Intraoperative fluoroscopy, acquired via highly maneuverable mobile C-arms, exists within a radically different geometric regime. C-arms are subject to extreme variance across multiple continuous spatial and procedural dimensions. The acquisition space of a single fluoroscopic image observation, x_i, must be represented as a multidimensional vector:

x_i = [\theta_p, \theta_s, \theta_d, SID, ISO, M, P_{patient}, T_{table}, A_{anatomy}, C_{clinical}, E_{equipment}]

In this formulation, \theta_p and \theta_s represent the positioner primary and secondary angles (e.g., right anterior oblique/left anterior oblique [RAO/LAO] and cranial/caudal [CRA/CAU] rotations). The variable \theta_d dictates detector rotation, SID specifies the Source-to-Image Distance, ISO represents the Source-to-Isocenter Distance, and M calculates the isocenter magnification (M_{iso} = SID/ISO).

When a model evaluating fluoroscopic data reaches a 90% accuracy plateau, it is rarely indicative of a model-capacity limitation or label noise. Instead, this plateau signifies sampling-density saturation within a highly populated, redundant region of the clinical acquisition space. Injecting tens of thousands of randomly sampled clinical images into the training pipeline yields diminishing returns because routine intraoperative data heavily overrepresents standard AP and strict lateral views. Meanwhile, complex oblique trajectories, rotational views, and unique magnification states remain drastically underrepresented, forming "holes" in the acquisition space.

2.2 Coverage-Driven Sampling and Entropy Metrics

To breach the 90% threshold, the analytical paradigm must shift from random dataset scaling to targeted, coverage-driven sampling. Model performance must be evaluated not purely as a function of raw sample count, but as a function of acquisition-space coverage.

Coverage (C_d) for a specific geometric dimension d is defined as the ratio of occupied bins to the total clinically relevant bins within that dimension. To quantify the distributional diversity of the entire dataset, a normalized entropy measure is applied:

H_d^* = \frac{-\sum p_i \log p_i}{\log K}

By parsing the metadata of a fluoroscopic corpus—specifically utilizing Digital Imaging and Communications in Medicine (DICOM) tags—data engineers can construct a projection-space density map.

Acquisition Dimension	DICOM Tag Source	Impact on Deep Learning Invariance
Positioner Primary Angle	(0018,1510)	Defines RAO-LAO rotation; extreme angles induce severe overlap of vertebral bodies and pedicles.
Positioner Secondary Angle	(0018,1511)	Defines CAU-CRA rotation; fundamentally alters the apparent projective height and width of spinal anatomy.
Source-to-Image Distance	(0018,1110)	Dictates focal blur, scatter intensity, and the overall geometric scale of the projection.
Patient Position	(0018,5100)	Identifies supine versus prone orientations, reversing the depth-ordering of the anatomy relative to the detector.
Isocenter Primary Angle	(0018,9463)	Provides enhanced XA/XRF isocenter constructs for equipment-reference geometry tracking.

Mapping these tags invariably reveals sparse regions representing rare, oblique C-arm angles. To improve a saturated model, data must be deliberately sampled or synthetically generated from these sparse acquisition bins. Enforcing uniform coverage across the multidimensional acquisition space forces the neural network to learn a geometrically invariant representation of the vertebral bodies, definitively pushing performance toward the parity observed in rigid radiography.

3. Bridging the Synthetic-to-Real Domain Gap

The extreme scarcity of accurately annotated, large-scale clinical fluoroscopy datasets necessitates the use of in-silico training pipelines. Ground-truth camera pose annotations and dense semantic segmentations are highly difficult to acquire in an active operating room due to workflow disruptions, radiation exposure constraints, and the manual labor required for sub-millimeter annotation. Consequently, models are trained on synthetic DRRs generated by projecting preoperative CT volumes into 2D space.

However, standard DRR generation relies on simplistic ray-casting algorithms that merely integrate Hounsfield Units (HU) along a view-ray. These analytical simplifications fail to capture the complex physical realities of clinical fluoroscopy, which is characterized by non-linear X-ray attenuation, quantum noise, electronic readout noise, beam hardening, and anisotropic photon scattering. When a model trained on clean, analytical DRRs is deployed on real fluoroscopy, a massive domain shift occurs, leading to catastrophic performance degradation.

3.1 Physics-Informed Rendering: DeepDRR and DiffDRR

To bridge this synthetic-to-real gap, recent computational frameworks have augmented the DRR

rendering pipeline with complex physics-based mechanics. Two industry-leading frameworks, DeepDRR and DiffDRR, represent divergent but equally critical approaches to this problem. DeepDRR utilizes a hybrid architecture, combining analytical forward projection with machine learning to simulate realistic X-ray image formation. The pipeline initiates a 3D material decomposition of the CT volume using a convolutional segmentation network, categorizing individual voxels into distinct physical classes such as air, soft tissue, and varying densities of bone. It subsequently applies a material- and spectrum-aware forward projector, injects specific quantum and electronic readout noise models, and incorporates a neural network-based Rayleigh scatter estimation. By estimating photon scatter in 2D rather than relying on computationally prohibitive 3D Monte Carlo photon transport simulations (which trace the probabilistic interactions of millions of individual photons), DeepDRR achieves exceptional clinical realism while maintaining acceptable generation speeds.
Conversely, DiffDRR prioritizes interoperability with deep learning optimization architectures by rendering the entire projection process fully differentiable. DiffDRR mathematically reformulates Siddon's exact ray-tracing algorithm as a series of vectorized tensor operations native to PyTorch. This auto-differentiability is profoundly important for 2D/3D registration: it allows a neural network to backpropagate the registration loss directly through the rendering process, calculating gradients with respect to the intrinsic and extrinsic imaging parameters. Utilizing GPU-accelerated tensor computation, DiffDRR achieves rendering speeds equivalent to highly optimized C++ frameworks while enabling continuous gradient descent for camera pose refinement.

3.2 Pairwise Domain Adaptation (PDA)

Even with advanced physics-based rendering, a residual domain shift persists due to patient-specific anatomical deformations (such as respiration and soft-tissue movement) and the presence of surgical instrumentation not captured in the preoperative CT. Generative style transfer and domain adaptation techniques are deployed to address this remaining discrepancy. Pairwise Domain Adaptation (PDA) represents a highly efficient strategy for medical image registration. Unlike standard statistical domain adaptation methods (e.g., Maximum Mean Discrepancy), which require massive, unpaired datasets from both the source (synthetic) and target (clinical) domains to function effectively, PDA operates on a highly constrained set of paired data. By directly measuring the domain distance on a small number of precisely paired DRR and real fluoroscopy images, a PDA module extracts robust domain-invariant features.
In this architecture, the PDA module acts as a plug-and-play feature extractor inserted into the convolutional layers of a pre-trained segmentation network. Because the PDA module focuses exclusively on filtering out stylistic discrepancies—such as varying image contrast, background scatter profiles, and quantum noise artifacts—without altering the geometric spatial relationships of the anatomy, it aligns the feature spaces of the synthetic and real domains. This enables a model trained exclusively on infinite synthetic DRR datasets to generalize seamlessly to the clinical fluoroscopic domain, significantly elevating vertebral body detection accuracy beyond the 90% threshold.

4. The Periodicity Failure Mode in Spinal Registration

Beyond the acquisition-space dimensionality and the synthetic-to-real domain gap, spinal imaging presents a highly unique anatomical challenge known as structural periodicity,

commonly referred to as "level jumping". Vertebral bodies, particularly across the thoracic and lumbar segments, possess highly similar morphological features and projection profiles. When an automated 2D/3D registration optimizer attempts to align a CT volume to a fluoroscopic image by maximizing an image similarity metric (such as Mutual Information or Gradient Correlation), the optimizer can easily fall into a false local optimum. The algorithm may mistakenly register the L3 vertebra on the 3D CT to the L4 vertebra on the 2D fluoroscopic image because their planar projections are nearly identical under certain C-arm angles.
Simulation studies evaluating intensity-based 2D/3D registration systems (such as the LevelCheck system) frequently report success rates of 99.998% in ideal, simulated environments, defining success as a mean projection distance (mPD) of less than 5mm. However, in physical experiments and real-world clinical applications, the presence of quantum noise, X-ray scatter, surgical implants, soft-tissue resection, and limited fields-of-view severely exacerbate the periodicity failure mode, leading to catastrophic misalignment.

4.1 Multi-View and Known-Component Anchoring

Overcoming the level-jumping phenomenon requires computational anchoring mechanisms. Multi-view registration approaches mitigate this issue by tracking anatomical features across multiple varying perspectives simultaneously, mathematically constraining the pose estimation so that a false optimum in one view is heavily penalized by the geometric realities of the orthogonal view.
Alternatively, Known-Component Registration (KC-Reg) utilizes highly distinct, radiopaque surgical instrumentation—such as K-wires or pedicle screws—as fixed reference points. By injecting the precise 3D Computer-Aided Design (CAD) models of these surgical devices into the optimization loop, the algorithm anchors the registration to the devices known to be present in the intraoperative field. Utilizing the Covariance Matrix Adaptation Evolution Strategy (CMA-ES) to maximize image-gradient similarity, KC-Reg prevents the optimizer from slipping to adjacent vertebral levels, ensuring the registration remains bounded within the correct anatomical segment.

5. AI-Driven 3D Reconstruction from Sparse Views: The X23D Paradigm

A transformative approach to bypassing complex 2D/3D registration entirely involves the direct, AI-driven generation of a 3D anatomical volume from a highly sparse set of 2D fluoroscopic images. This approach eliminates the reliance on a preoperative CT scan, aiming to construct the 3D surgical scene entirely from intraoperative data.
The X23D algorithmic framework represents the vanguard of this computational strategy. Designed to intraoperatively estimate the 3D shape of a patient's lumbar vertebrae directly from sparse, multi-view X-ray data (typically just 3 to 4 projections), X23D utilizes a learned multi-view stereo machine approach. The neural network incorporates the exact X-ray calibration parameters (extrinsic and intrinsic matrices) directly into its architecture, allowing a priori knowledge of the spinal shape to be applied while preserving strict patient specificity.

5.1 Training Pipelines and Ex-Vivo Validation

The X23D network relies heavily on synthetic data generation to achieve spatial understanding.

The model was trained and rigorously evaluated on 17,420 fluoroscopy images that were digitally reconstructed and sampled from the public CTSpine1K dataset. When evaluated against unseen data, the network achieved an 88% average F1 score and a 71% surface score, fundamentally outperforming counterpart state-of-the-art methods by 22% in surface score accuracy by effectively utilizing the camera calibration parameters.
To validate the clinical viability of this AI-driven reconstruction, comprehensive ex-vivo studies were conducted. In a highly controlled interventional setting, five board-certified spine surgeons were tasked with placing pedicle screws across the L1-L5 levels of human torsi. The surgeons placed a total of 49 pedicle screws (24 using the X23D prototype and 25 using conventional 2D fluoroscopy as the control).
The analytical results highlighted the immense potential and current limitations of sparse-view 3D reconstruction. Execution time for each lumbar level showed no statistically significant difference, with the X23D group averaging 167 seconds compared to the control group's 156 seconds. Crucially, the radiation dose was heavily reduced in the AI cohort, dropping to 33.26 mGy compared to 49.5 mGy in the control group. However, the study recorded five hardware breaches in the X23D group (one of which was considered clinically significant, graded as a medial breach C), compared to six breaches in the control group (also featuring one clinically significant breach). While X23D demonstrates that 3D reconstruction from 3-4 images is computationally feasible with processing times of just 81.1 milliseconds per vertebra, further architectural refinement is required to achieve absolute parity with clinical-grade, CT-based navigation systems.

6. Rotational Fluoroscopy: Hardware and "Around the World" Modalities

The specific inquiry regarding "around the world" shots on fluoroscopy refers to orbital or rotational fluoroscopic acquisitions. In this modality, a motorized C-arm performs a continuous, automated 180-degree (or greater) sweep around the patient on the operating table, capturing hundreds of rapid, sequential 2D projection images. These dense 2D sequences are subsequently processed via filtered back-projection or iterative reconstruction algorithms into a high-fidelity 3D Cone-Beam Computed Tomography (CBCT) volume for immediate intraoperative verification.

6.1 Clinical Hardware Systems

Several distinct intraoperative 3D imaging systems are specifically engineered to execute these rotational sweeps, each possessing unique geometric and radiation profiles that impact their utility in spine surgery.

Imaging System	Key Modality Features	Field of View (FOV) & Anatomical Coverage
Siemens Cios Spin	Mobile 3D C-arm utilizing Retina 3D scan technology; acquires up to 400 projections in 30 seconds.	16 cm x 16 cm x 16 cm. Capable of capturing seven cervical or six thoracic vertebrae in a single scan.
Medtronic O-arm II	Closed-ring gantry system providing high-fidelity 3D	Broad transverse coverage; radiation profile (2–2.5 mSv)

Imaging System	Key Modality Features	Field of View (FOV) & Anatomical Coverage
	rotational imaging; features image stacking to broaden the transverse FOV.	resembles mobile CT scanners.
Ziehm RFD 3D	Mobile CMOS-based system with a strong compatibility across navigation and robotics platforms.	20 cm FOV. Highly mobile but presents clearance challenges in larger BMI patients.
BrainLab LoopX	Novel motorized platform functioning autonomously; features a large 121 cm gantry for flexible positioning.	48 cm FOV; highly optimized for proprietary navigation integration.
Globus Excelsius3D	3-in-1 consolidated imaging platform providing standard fluoroscopy, digital radiography, and 3D cone-beam CT.	Covers up to 6 thoracolumbar levels; features automatic segmentation of vertebral bodies.

While these machines acquire vast amounts of rotational 2D fluoroscopic data during their sweeps, the raw 2D projection images are frequently discarded, obscured, or compressed into proprietary formats once the final 3D CBCT volume is reconstructed. Consequently, acquiring the raw 2D sequences necessary to train next-generation 2D/3D registration algorithms requires specialized extraction protocols or direct research collaborations.

7. Landscape of Paired Datasets: Public Repositories and Data Use Agreements

The development of highly robust vertebral detection and domain adaptation algorithms requires paired datasets containing high-resolution 3D volumes (preoperative CT or intraoperative CBCT) and their corresponding raw, uncompressed 2D fluoroscopic projections with known extrinsic camera poses.

7.1 Publicly Available Datasets

Despite the inherent difficulties in acquiring perfectly calibrated clinical data, several benchmark datasets are publicly available to researchers:

1. **DeepFluoro:** Curated by Johns Hopkins University, DeepFluoro is widely acknowledged as the gold standard for benchmarking 2D/3D registration algorithms. The dataset comprises information from six cadaveric subjects, featuring one high-resolution preoperative pelvic CT scan per subject. Accompanying the CTs are intraoperative fluoroscopic X-ray images, ranging from 24 to 111 images per subject, yielding a total of 366 real images. DeepFluoro provides highly accurate, ground-truth extrinsic camera matrices for every 2D image—obtained through rigorous fiducial-based optical tracking—as well as the exact intrinsic calibration matrix of the specific C-arm used. While strictly focused on the pelvis and femurs rather than the spine, the registration methodologies, loss functions, and domain adaptation strategies developed on DeepFluoro are universally transferable to spinal anatomy.

2. **Ljubljana Dataset:** This clinical dataset contains paired 2D and 3D digital subtraction angiography (DSA) images from 10 patients undergoing neurovascular interventions at the University of Ljubljana. Like DeepFluoro, it provides essential ground-truth extrinsic and intrinsic matrices derived from surface fiducial markers. Although primarily cranial and vascular in nature, the Ljubljana dataset is routinely utilized to test the generalizability of 2D/3D registration neural networks across highly disparate anatomical domains.
3. **SpineWeb:** The SpineWeb archive is a specialized, multi-institutional repository focused explicitly on spinal imaging. Specific sub-cohorts within SpineWeb provide paired CT images and calibrated fluoroscopic sequences explicitly designed for evaluating 3D/2D registration, automated spinal measurement quantification (SMQ), and the tracking of intervertebral kinematics under physiological load.
4. **CTSpine1K:** While not containing real fluoroscopy, CTSpine1K is a massive public dataset of 3D spinal CTs containing detailed segmentation masks. Models like X23D leverage CTSpine1K to digitally render tens of thousands of synthetic rotational fluoroscopic views, serving as the foundational training ground before transitioning to real clinical data via domain adaptation.

7.2 Restricted Datasets and Data Use Agreements (DUAs)

Accessing large-scale, clinical, intraoperative rotational sequences generally requires institutional collaboration via a Data Use Agreement (DUA). The raw 2D projections from systems like the O-arm or Cios Spin contain vast amounts of identifiable biometric and spatial data, heavily protected under HIPAA and GDPR regulations.
Institutional research networks frequently harbor immense, highly curated datasets. For example, researchers at Henry Ford Health maintain comprehensive biplane dynamic X-ray and CT datasets used for highly precise, model-based tracking of cervical kinematics in spinal arthroplasty and fusion patients. These datasets track 6-Degrees-of-Freedom (6-DoF) vertebral motion (3 translations, 3 rotations) at extensive longitudinal time points (e.g., 2-year and 6.5-year post-operative reviews). Accessing such databases necessitates navigating institutional review boards (IRB), validating the scientific merit of the study, and establishing formal DUAs to ensure data anonymization and secure processing protocols.

8. Optimal Utilization of an N=1 Paired Dataset

The theoretical scenario posits the possession of exactly one "around the world" rotational fluoroscopic sequence with a corresponding preoperative CT scan of the spine. Under traditional deep learning paradigms, an N=1 dataset is fundamentally insufficient for training a generalized model due to catastrophic overfitting. However, contemporary advances in self-supervised learning, differentiable rendering, and patient-specific network adaptation allow for immense computational utility to be extracted from a single paired dataset.
By deploying algorithmic pipelines analogous to DiffPose and XVR, a single paired dataset can be engineered to achieve sub-millimeter 2D/3D registration accuracy. The optimal pipeline is structured in three consecutive phases: Patient-Specific Pretraining, Pairwise Domain Adaptation, and Intraoperative Differentiable Refinement.

8.1 Phase 1: Patient-Specific Pretraining via Differentiable Rendering

Instead of attempting to train a network on generic population data that fails to generalize across disparate anatomies, the pipeline begins by treating the single preoperative CT scan as an infinite, self-supervised data generator.
Using a differentiable renderer like DiffDRR, the pipeline randomly samples tens of thousands of theoretical camera poses (extrinsic matrices) from a uniform distribution spanning the entire rotational acquisition space. For each mathematically sampled pose, a highly realistic synthetic DRR is rendered directly from the specific patient's CT volume. A convolutional neural network is then trained from scratch (or fine-tuned from a foundation model) on these synthetic images. The input to the network is the synthetic DRR; the output is the direct regression of the 6-DoF camera pose.
Because the data is generated in-silico, the ground-truth pose is known perfectly, establishing a flawless self-supervised training loop. The network becomes a localized, "patient-specific" expert. It implicitly learns the exact morphological structure of the patient's spine, entirely circumventing the periodicity (level-jumping) failure mode because it has extensively memorized the subtle, patient-specific geometric variations between adjacent vertebrae.

8.2 Phase 2: Mitigating Intrinsic Calibration Errors and Domain Shift

The patient-specific CNN trained in Phase 1 is highly accurate on synthetic DRRs but will inevitably suffer a performance drop when exposed to the real rotational fluoroscopy sequence due to the domain gap. Furthermore, the single fluoroscopic shot may possess unknown or inaccurate intrinsic calibration parameters, a common issue in clinical datasets.
Computational analysis demonstrates that moderate errors in intrinsic calibration can be effectively mitigated solely by re-optimizing the extrinsic parameters during registration. Simulation and real-world experiments on C-arm systems indicate that focal length errors of up to 500 pixels introduce a mean 3D reconstruction error of less than 0.2 mm, while principal point shifts of up to 200 pixels yield negligible increases in reprojection error once the extrinsic pose is re-optimized. This high intrinsic tolerance allows the pipeline to proceed even if the exact camera parameters of the N=1 dataset are slightly degraded.
To bridge the visual domain gap, the user’s real rotational sequence is deployed. The user selects a sparse subset of images from the real rotational sequence where the approximate camera poses are manually aligned or known. A Pairwise Domain Adaptation (PDA) module is attached to the CNN, and the network is fed paired images: the real rotational fluoroscopy and the corresponding DiffDRR synthetic projection. The PDA module calculates a localized domain loss, forcing the feature maps of the convolutional layers to become invariant to the stylistic differences (noise, contrast, scatter) between the synthetic and real domains. The patient-specific network can now process the real intraoperative fluoroscopy images with the same accuracy as the pristine synthetic DRRs.

8.3 Phase 3: Intraoperative Differentiable Refinement

When the domain-adapted, patient-specific network predicts a camera pose from a real rotational fluoroscopy image, it serves only as a highly accurate initialization. To achieve absolute parity with the 99% accuracy seen in rigid radiography, a final optimization step is required.
The predicted pose is passed into the DiffDRR renderer, which generates a real-time synthetic DRR. An image similarity metric—such as Gradient-Normalized Cross Correlation (Grad-NCC) or Learned Perceptual Image Patch Similarity (LPIPS)—is computed between the synthetic

DRR and the real target fluoroscopy. Because the entire rendering pipeline is formulated as vectorized tensor operations in PyTorch, the error gradient flows directly backwards from the image similarity loss, through the rendering engine, and into the 6-DoF pose parameters. Advanced architectures, such as the LXPose framework, utilize a multi-stage refinement process operating on axis-angle rotations, achieving inference times of just 20 milliseconds to comply with the strict 40 ms time constraint required for true intraoperative deployment. Alternatively, the XVR framework utilizes a multi-scale optimization strategy and gradient-normalized cross-correlation loss to achieve extreme sub-millimeter accuracy, though it requires slightly longer optimization times (approximately 7 seconds).

This hybrid combination of self-supervised, patient-specific initialization followed by differentiable rendering refinement entirely bypasses the traditional limitations of 2D/3D registration. The neural network provides an initialization well within the capture range of the optimizer, preventing convergence on local minima, while the differentiable renderer achieves flawless structural alignment.

9. Conclusion

To elevate vertebral body detection and 2D/3D registration models from a saturated plateau of 90% accuracy to the clinical parity of 99% observed in standardized radiographs, the computational paradigm must shift definitively from uniform dataset expansion to physics-informed, coverage-driven optimization. Relying solely on randomly acquired clinical fluoroscopy leaves severe statistical voids within the multi-dimensional rotational acquisition space, inherently stalling neural network generalization and leaving models vulnerable to geometric variations.

The integration of auto-differentiable rendering engines, such as DiffDRR, paired with Pairwise Domain Adaptation, permits the creation of virtually infinite, accurately labeled, domain-invariant synthetic data directly from preoperative CT volumes. For researchers navigating severely constrained data environments—such as the possession of a single paired CT and rotational fluoroscopic sequence—the deployment of patient-specific, self-supervised regression networks initialized prior to iterative test-time optimization represents the definitive state-of-the-art methodology.

By systematically mapping the sparse acquisition dimensions of C-arm trajectories, leveraging the known components of surgical hardware to prevent periodic level jumping, and enforcing structural alignment through continuous gradient-based optimization, the inherent spatial ambiguities of the spinal column can be definitively resolved. This computational approach successfully bridges the domain gap, unlocking highly accurate, real-time volumetric navigation directly from projective 2D intraoperative data.

Works cited

1. X23D—Intraoperative 3D Lumbar Spine Shape Reconstruction, https://www.mdpi.com/2313-433X/8/10/271 2. Domain adaptation strategies for 3D reconstruction of the lumbar, https://arxiv.org/html/2401.16027v2 3. Technology Feasibility Study: X23D Spine Navigation, Aedan Nikas, https://www.medicaldeviceproductmanager.com/feasibility-study 4. Toward real-time alignment of 3D CT and 2D X-ray with multi-stage, https://www.tandfonline.com/doi/full/10.1080/24699322.2026.2684372 5. A Fully Differentiable

Framework for 2D/3D Registration and ... - PMC, https://pmc.ncbi.nlm.nih.gov/articles/PMC10879149/ 6. Rapid patient-specific neural networks for intraoperative X-ray to, https://arxiv.org/html/2503.16309v2 7. PolyPose: Deformable 2D/3D Registration via Polyrigid ... - PMC, https://pmc.ncbi.nlm.nih.gov/articles/PMC12148084/ 8. Enhanced XA/XRF Storage SOP Class - DICOM Standard, https://www.dicomstandard.org/News-dir/ftsup/docs/sups/sup83.pdf 9. Intraoperative 2D/3D Image Registration via Differentiable X-ray, https://arxiv.org/html/2312.06358v2 10. DDGS-CT: Direction-Disentangled Gaussian Splatting for Realistic, https://proceedings.neurips.cc/paper_files/paper/2024/file/456ce0476c9b4689a74918b851cecd5a-Paper-Conference.pdf 11. Schematic overview of DeepDRR. - ResearchGate, https://www.researchgate.net/figure/Schematic-overview-of-DeepDRR_fig1_324005768 12. Enabling machine learning in X-ray-based procedures via realistic, https://pmc.ncbi.nlm.nih.gov/articles/PMC7297499/ 13. DeepDRR 1.1.0a3 documentation, https://deepdrr.readthedocs.io/README.html 14. Fast Auto-Differentiable Digitally Reconstructed Radiographs for, https://www.researchgate.net/publication/366989449_Fast_Auto-differentiable_Digitally_Reconstructed_Radiographs_for_Solving_Inverse_Problems_in_Intraoperative_Imaging 15. Accelerated ray tracing for radiotherapy dose calculations on a GPU, https://www.researchgate.net/publication/26877397_Accelerated_ray_tracing_for_radiotherapy_dose_calculations_on_a_GPU 16. 3D-2D image registration for target localization in spine surgery, https://www.researchgate.net/publication/298908534_3D-2D_image_registration_for_target_localization_in_spine_surgery_Investigation_of_similarity_metrics_providing_robustness_to_content_mismatch 17. Pairwise domain adaptation module for CNN-based 2-D/3-D ... - PMC, https://pmc.ncbi.nlm.nih.gov/articles/PMC5767648/ 18. Homogenization of multi-institutional chest x-ray images in various, https://www.researchgate.net/publication/370317219_Homogenization_of_multi-institutional_chest_x-ray_images_in_various_data_transformation_schemes 19. Data-Driven Multi-modal Partial Medical Image Preregistration by, https://www.researchgate.net/publication/363621538_Data-Driven_Multi-modal_Partial_Medical_Image_Preregistration_by_Template_Space_Patch_Mapping 20. Automatic Localization of Vertebral Levels in X-Ray Fluoroscopy, https://pmc.ncbi.nlm.nih.gov/articles/PMC3429949/ 21. 2D-3D Spine Image Target Registration Error - ResearchGate, https://www.researchgate.net/figure/2D-3D-Spine-Image-Target-Registration-Error_tbl1_221192931 22. Robust and Accurate Multi-view 2D/3D Image Registration ... - arXiv, https://arxiv.org/html/2506.22191v1 23. Rigid Single-Slice-in-Volume Registration via Rotation-Equivariant, https://www.researchgate.net/publication/384653020_Rigid_Single-Slice-in-Volume_Registration_via_Rotation-Equivariant_2D3D_Feature_Matching 24. Known-component 3D–2D registration for quality assurance of spine, https://www.researchgate.net/publication/282467143_Known-component_3D-2D_registration_for_quality_assurance_of_spine_surgery_pedicle_screw_placement 25. Feature-based multimodal 3D/2D and 3D/3D registration framework, https://www.researchgate.net/publication/395466309_Feature-based_multimodal_3D2D_and_3D3D_registration_framework_for_pedicle_screw_registration_and_evaluation 26. C-Arm fluoroscopy pose estimation algorithm. a: Original X-ray, https://www.researchgate.net/figure/C-Arm-fluoroscopy-pose-estimation-algorithm-a-Original-X-ray-image-b-Projection-image_fig4_51825560 27. Intraoperative 3D reconstruction from sparse

arbitrarily posed real X, https://d-nb.info/1391016793/34 28. Spinal navigation with AI-driven 3D-reconstruction of fluoroscopy, https://pubmed.ncbi.nlm.nih.gov/39558228/ 29. X23D - Intraoperative 3D Lumbar Spine Shape Reconstruction, https://x23d.ch/ 30. Image-Based Medical Navigation Systems for Cardiac Interventions, https://pmc.ncbi.nlm.nih.gov/articles/PMC13102890/ 31. Accuracy of pedicle screw placement using neuronavigation based, https://pmc.ncbi.nlm.nih.gov/articles/PMC10191982/ 32. Cios Spin - Siemens Healthineers, https://cdn0.scrvt.com/39b415fb07de4d9656c7b516d8e2d907/8ec2b1e7cc42dec5/b74a1a64e1f3/AT_Cios-Spin_Brochure_CAN.PDF 33. AAPM Task Group Report 238: 3D C-arms with volumetric imaging, https://pmc.ncbi.nlm.nih.gov/articles/PMC11584023/ 34. The Reconstruction Toolkit (RTK), an open-source cone-beam CT, https://www.researchgate.net/publication/263004936_The_Reconstruction_Toolkit_RTK_an_open-source_cone-beam_CT_reconstruction_toolkit_based_on_the_Insight_Toolkit_ITK 35. Negotiating for new technologies: guidelines for the procurement of, https://pmc.ncbi.nlm.nih.gov/articles/PMC9263739/ 36. (PDF) Automatic annotation of hip anatomy in fluoroscopy for robust, https://www.researchgate.net/publication/340908673_Automatic_annotation_of_hip_anatomy_in_fluoroscopy_for_robust_and_efficient_2D3D_registration 37. DiffDRR-Datasets/README.md at main · eigenvivek/DiffDRR, https://github.com/eigenvivek/DiffDRR-Datasets/blob/main/README.md 38. Landmark detection errors across all trained networks for each, https://www.researchgate.net/figure/Landmark-detection-errors-across-all-trained-networks-for-each-landmark_tbl1_340908673 39. Deformable 2D/3D Registration via Polyrigid Transformations, https://vivekg.dev/assets/papers/polypose.pdf 40. eigenvivek/DiffPose: [CVPR 2024] Intraoperative 2D/3D ... - GitHub, https://github.com/eigenvivek/diffpose 41. A Locally Rigid CT to X-ray Registration Framework - PMC, https://pmc.ncbi.nlm.nih.gov/articles/PMC7279937/ 42. Intervertebral kinematics during neck motion 6.5 years after fusion, https://scholarlycommons.henryford.com/cgi/viewcontent.cgi?article=1439&context=orthopaedics_articles 43. Henry Ford Health Clinical Insights | For Healthcare Professionals, https://www.henryford.com/hcp/clinical-insights 44. (PDF) Rapid patient-specific neural networks for X-ray to volume, https://www.researchgate.net/publication/390372623_Rapid_patient-specific_neural_networks_for_X-ray_to_volume_registration 45. Intrinsic Tolerance in C-Arm Imaging: How Extrinsic Re-optimization, https://arxiv.org/html/2603.14031v1 46. Image-Guided Procedures, Robotic Interventions, and Modeling - SPIE, https://spie.org/Publications/Proceedings/Volume/13927 47. Comparison of global and local optimization methods for intensity, https://www.researchgate.net/publication/387649286_Comparison_of_global_and_local_optimization_methods_for_intensity-based_2D-3D_registration