
two consecutive exposures




one move cranial — at the anchor

shaded: both frames show L2 and L3


The matcher has nothing else to work with:
no labels, no anatomy outside the field.

the two windows, in millimetres



verse108 · 1.00 mm per row · p SD 2.4 mm
field 153 mm holds four bodies, keeps two
box extent 48 mm = 1.5 × p, not the body

one body is what can be confused



Δ is commanded, not measured.
A move off by more than 13 mm lets the alias in: the frame locks one body off, the names above it are all one off, and that is a wrong-level operation.