MILLIGAL VALUES FOR LACOSTE & ROMBERG, INC. MODEL G GRAVITY METER #G- 352

TABLE 1

COUNTER	VALUE IN	WACTOU TOD	COMMED	**** YTH* Thr	TACTION DON
READING*	MILLIGAL	FACTOR FOR INTERVAL	COUNTER	VALUE IN	FACTOR FOR
KENDTHG	PULLITORL	INIBRVAL	READING*	MILLIGAL	INTERVAL
000.	000.00	1.06293			
100	106.29	1.06287	3600	3828.71	1.06451
200	212.58	1.06283	3700	3935.16	1.06449
300	318.86	1.06280	3800	4041.61	1.06447
400	425.14	1.06277	3900	4148.05	1.06446
500	531.42	1.06275	4000	4254.50	1.06444
600	637.70	1.06275	4100	4360.94	1.06443
700	743.97	1.06276	4200	4467.39	1.06442
800	850.25	1.06280	4300	4573.83	1.06442
900	956.53	1.06284	4400	4680.27	1.06442
1000	1062.81	1.06289	4500	4786.71	1.06443
1100	1169.10	1.06294	4600	4893.16	1,06445
1200	1275.39	1.06300	4700	4999.60	1.06448
1300	1381.69	1.06305	4800	5106.05	1,06452
1400	1488.00	1.06312	4900	5212.50	1.06455
1500	1594.31	1.06317	5000	5318.96	1.06457
1600	1700.63	1.06324	5100	5425.41	1.06458
1700	1806.95	1.06332	5200	5531.87	1.06458
1800	1913.28	1.06339	5300	5638.33	1,06454
1900	2019,62	1.06346	5400	5744.78	1.06449
2000	2125.97	1.06355	5500	5851.23	1.06444
2100	2232.32	1.06363	5600	5957.68	1.06437
2200	2338.69	1.06373	5700	6064.11	1.06429
2300	2445.06	1.06383	5800	6170.54	1.06419
2400	2551.44	1.06395	5900	6276.96	1.06409
2500	2657.84	1.06408	6000	6383.37	1,06397
2600	2764.25	1.06418	6100	6489.77	1.06382
2700	2870.66	1.06428	6200	6596.15	1.06365
2800	2977.09	1.06437	6300	6702.51	1.06345
2900	3083.53	1.06445	6400	6808.86	1,06323
3000	3189.97	1.06452	6500	6915.18	1.06300
3100	3296.43	1.06456	6600	7021.48	1.06276
3200	3402.88	1.06458	6700	7127.76	1.06251
3300	3509.34	1.06458	6800	7234.01	1.06224
3400	3615.80	1.06456	6900	7340.23	1.06196
3500	3722.25	1.06454	7000	7446,43	

<sup>\*</sup> Note: Right-hand wheel on counter indicates approximately 0.1 milligal.

JS

<sup>2-14-74</sup>