





RAZOR RECOMMENDED INSTALLATIONS


INSTALLATIONS IN PIPELINES




Mounted on a horizontal pipe with a diameter of 8 inches or larger



Mounted horizontally on an elbow on pipes with a diameter of 2 inches or larger




Mounted a vertical pipe with a diameter of 8 inches or larger. Mounted at 45° it can be mounted on smaller pipes. The flow can be downwards only if the pipe is completely full (No waterfall effect).




Mounted vertically on an elbow on pipes with diameter of 2 inches or larger. Notice that the cage has to be located deep into the vertical line to avoid the "Gas Trap" section at the top.

INSTALLATIONS IN PIPELINES (Cont.)




Mounted horizontally on a 2 inch or larger "T". On 2" pipes the top portion of the cage has to be right at the opening of the "T" to avoid pressure drop



Mounted vertically on a 2 inch or larger "T". On 2" pipes the top portion of the cage has to be right at the opening of the "T" to avoid pressure drop

INSTALLATION ON TANKS



Mounted on the side of a tank. It can be mounted horizontally or at an angle.