# PCR RUN REPORT

## Test1

Generated: 02/11/2026 17:33:32

## Run Summary

| Parameter | Value |
|---|---|
| Program | Test1 |
| Run Started | 02/11/2026 15:59:22 |
| Report Generated | 02/11/2026 17:33:32 |
| Run Duration | 1 h 34 min 9 sec |
| Number of Cycles | 30 |
| Report Cycle Profile | Average of 30 cycles |
| Final ESP32 State | Running |

## PCR Program Settings

| Stage | Temperature | Hold Time |
|---|---|---|
| Initial Denaturation (T1) | 98.00 ° C | 3 min 0 sec |
| Denaturation (T2) | 95.00 ° C | 30 sec |
| Annealing (T3) | 60.00 ° C | 30 sec |
| Extension (T4) | 72.00 ° C | 1 min 0 sec |
| Final Extension (T5) | 72.00 ° C | 5 min 0 sec |

## Run Status

| Parameter | Value |
|---|---|
| Current Temperature | 5.18 ° C |
| Target Temperature | 4.00 ° C |

| Parameter | Value |
|---|---|
| Current Stage | Final Hold |
| Phase | Holding |
| Cycle | 30/30 |
| Hold Remaining | 0 sec |
| Heater | OFF |
| Cooling | ON |

## Temperature Profile

Initial denaturation, averaged cycling profile across N cycles, final extension, and actual cooling to the final hold.




The cycling section is aligned stage-by-stage and averaged across 30 recorded cycle(s). Raw ESP32 readings remain separate from the display-only HOLD smoothing used by the live UI.

Report note: PCR execution, stage transitions, cycle timing, heating and cooling are controlled by the ESP32. Flutter only records ESP32 status/temperature data and performs report post-processing. The averaged graph does not alter PCR control or timing on the ESP32.