NORTHSTAR ENVIRONMENTAL
Water and environmental testing
Bloomington, Minnesota · (952) 555-0122 · northstarenviro.com

LABORATORY REPORT
NS-4471
Issued 10/22/2026

Report of Analysis: Chloride in Water

Four client-collected water samples, chloride by ion-selective electrode

REPORT TO	REPORT DETAILS
Owen Brandt 4127 34th Ave S Minneapolis, MN 55406 (612) 555-0183 owen.brandt.writes@gmail.com	**Report no.** NS-4471 **Project** Salt in the Lakes **Collected** 10/10/2026, by client **Received** 10/12/2026 11:02 **Analyzed** 10/19/2026 **Reported** 10/22/2026

TEMPERATURE ON RECEIPT	CONTAINERS	PRESERVATION	CUSTODY FORM	HOLDING TIME
4.1 °C, on ice	4 of 4, 250 mL plastic	None required	Signed at drop-off	Within 28 days

Results

Lab ID	Client sample ID	Matrix	Collected	Analyzed	Dilution	Chloride, mg/L	RL, mg/L
NS-4471-1	**COR-1**	Water	10/10/2026 09:12	10/19/2026	None	**241**	10
NS-4471-2	**PEN-OUT**	Water	10/10/2026 10:33	10/19/2026	None	**229**	10
NS-4471-3	**ODG-IN**	Water	10/10/2026 12:20	10/19/2026	None	**297**	10
NS-4471-4	**SUM-1**	Water	10/10/2026 13:41	10/19/2026	None	**139**	10

mg/L: milligrams per liter. RL: reporting limit. Results are reported as whole numbers.

Method and preparation

Method NE-CL-ISE-02, chloride by ion-selective electrode (method sheet revised 2024-03-12)

Sample prep Read as received, unfiltered and undiluted: 50 mL of sample plus 1.0 mL of ionic strength adjuster, stirred 2 minutes

Electrode Solid-state chloride ISE, double-junction reference

Calibration and quality control

Calibration Two-point, 10 mg/L and 100 mg/L chloride standards, start of analysis day 10/19/2026

Electrode slope -57.4 mV per decade (accepted -54 to -60)

Matrix spike Not performed; not part of this method

Duplicate Not performed; not part of this method

Notes

1. Client comment on the custody form: split samples.
2. Screening method. Not accredited for chloride. Not intended for drinking-water compliance or permit reporting.
3. These results relate only to the samples as received. Reproduce this report only in full.

APPROVED

Gary Lindqvist 10/22/2026
Laboratory Director
(952) 555-0122
glindqvist@northstarenviro.com