View_measurement_IV.vi
C:\Users\root\Desktop\CMS_Current\Istrip-measurement\View_measurement_IV.vi
Last modified on 11/5/2015 at 2:33 PM
Printed on 4/4/2016 at 1:50 PM

Strip View

Bad Sensor

IV   33305010000619

Strip Total

StripTot
StripSur

Output Array

0	-5.00E+0	1.46E-11
0	-1.00E+1	2.46E-11
0	-1.50E+1	5.61E-11

Iavg(good)@70V
0.0

Iavg(good)@300V
0.0

Qualification String

Itot(70V) HIGH = 10.0E+3
Itot(300V) HIGH = 10.0E+3
2 strips above 1uA at70V.
The rest have avg currentt of0.0E+0
2 strips above 5uA at 300V.
The rest have average current of 0.0E+0