



# Low-Latency EEG Marker Integration in Serious Games for Neurocognitive Assessment

Julian Andres Salazar Parias

Universidad Nacional de Colombia  
Facultad de ingeniería y arquitectura  
Departamento de ingeniería eléctrica, electrónica y computación  
Manizales, Colombia  
2025

# **Low-Latency EEG Marker Integration in Serious Games for Neurocognitive Assessment**

**Julian Andres Salazar Parias**

Tesis presentada como requisito parcial para optar por el título de:  
**Maestría en ingeniería - Automatización Industrial**

Director(a):

MSc Bernardo Andrés Cardona Noreña

Línea de investigación:

Inteligencia Artificial

Grupo de investigación:

Grupo de Control y Procesamiento Digital de Señales (GCPDS)

Universidad Nacional de Colombia

Facultad de ingeniería y arquitectura

Departamento de ingeniería eléctrica, electrónica y computación

2025

"The concern for man and his destiny must always be the primary interest of any technical effort. Never forget this between your diagrams and equations"

*Albert Einstein*

"Believe you can, and you are halfway there."

*Theodore Roosevelt*

# **Declaración**

Me permito afirmar que he realizado esta tesis de manera autónoma y con la única ayuda de los medios permitidos. Todos los pasajes que se han tomado de manera textual o figurativa de textos publicados y no publicados, los he reconocido en el presente trabajo. Ninguna parte del presente trabajo se ha empleado en ningún otro tipo de tesis.

Manizales, 2025

---

Julian Andres Salazar Parias

# List of Figures

<b>1-1.</b> Comparison of neuroimaging modalities by spatial resolution, temporal resolution, and cost. EEG stands out for its affordability, portability, and millisecond-level responsiveness. . . . .	2
<b>1-2.</b> Core cognitive models targeted by serious games in ADHD interventions: attention, working memory, inhibition, and planning. Each model maps to a specific set of game dynamics and EEG markers. . . . .	3
<b>1-3.</b> Challenges in EEG systems and serious games for ADHD evaluation. . . . .	7
<b>2-1.</b> Characteristic morphology of an Event-Related Potential (ERP), highlighting exogenous and endogenous components such as the N200 and P300. . . . .	15
<b>2-2.</b> Simplified functional scheme of the modulation and filtering stage in a Delta-Sigma architecture ADC. . . . .	16
<b>3-1.</b> Block diagram of the MONEEE architecture, evidencing the segregation between the deterministic acquisition (MCU) and high-level processing (MPU) domains. . . . .	19
<b>3-2.</b> Schematic design for the module responsible for acquiring EEG signals. . . .	21
<b>3-3.</b> Schematic design of coupling filters. . . . .	22
<b>3-4.</b> Schematic design of the module responsible for communicating the collected data to another device or to the cloud. . . . .	22
<b>3-5.</b> Module for impedance visualization of the electrodes. . . . .	23
<b>3-6.</b> Connection between the different ADS1299 acquisition modules. . . . .	23
<b>3-7.</b> Motherboard for microcontroller and microprocessor. . . . .	24
<b>4-1.</b> Flowchart of the Interrupt Service Routine (ISR) associated with the Data Ready signal (DRDY). . . . .	26

## **List of Tables**

<b>1-1.</b> Acquisition devices used for BCI. The table provides an overview of the different hardware devices, their specifications, and communication protocols. . . . .	11
<b>2-1.</b> Comparative analysis of synchronization methods for BCI systems. . . . .	17
<b>4-1.</b> Definition of the Serial Event Transmission Protocol. . . . . . . . . . .	27

# Content

<b>List of figures</b>	ii
<b>List of tables</b>	iii
<b>Content</b>	v
<b>1. Preliminaries</b>	1
1.1. Motivation . . . . .	1
1.2. Problem statement . . . . .	5
1.2.1. Unpredictable Latency and Jitter Between Game Events and EEG Recordings . . . . .	5
1.2.2. Resource and power constraints in embedded EEG platforms . . . . .	6
1.3. Investigation question . . . . .	8
1.4. Objectives . . . . .	9
1.4.1. General Objective . . . . .	9
1.4.2. Specific Objectives . . . . .	9
1.5. State of art . . . . .	10
<b>2. Theoretical Framework</b>	14
2.1. Neurophysiology and Event-Related Potentials (ERPs) . . . . .	14
2.2. Hardware Architecture for Signal Acquisition . . . . .	15
2.3. Digital Synchronization Protocols . . . . .	17
<b>3. Hardware Architecture (The MONEEE System)</b>	18
3.1. System Topology and Data Flow . . . . .	18
3.2. Analog Front-End (AFE) and Biomedical Interface . . . . .	19
3.3. The Digital Core: Heterogeneous Processing . . . . .	20
3.4. Event Synchronization Interface (USB-C) . . . . .	21
<b>4. Firmware Architecture and Temporal Synchronization Strategy</b>	25

## **Low-Latency EEG Marker Integration in Serious Games for Neurocognitive Assessment**

---

4.1.	Deterministic Firmware Design on the TM4C1294 . . . . .	25
4.1.1.	Hardware Event Injection Strategy . . . . .	26
4.2.	Integration Protocol with the Simulation Environment . . . . .	27
4.3.	Processing in the Compute Module (Raspberry Pi CM4) . . . . .	27
<b>5.</b>	<b>Validation and Results</b>	<b>29</b>
<b>6.</b>	<b>Conclusion and Future Work</b>	<b>30</b>
	<b>References</b>	<b>31</b>

# 1 Preliminaries


## 1.1 Motivation

Brain–Computer Interfaces (BCIs) have emerged as a powerful class of technologies that enable direct communication between the brain and external devices. These systems are increasingly being applied in neurorehabilitation, education, and clinical diagnosis due to their ability to monitor and interpret neural activity in real time. BCIs have the potential to revolutionize the way cognitive states are assessed and modulated by offering closed-loop interaction mechanisms that adapt to the user’s brain dynamics [Lim et al., 2023, Lin and Chang, 2025]. Central to this capability is the choice of neuroimaging modality, which must meet strict criteria in temporal resolution, portability, and cost-effectiveness—especially in applications involving children or naturalistic settings.

Several neuroimaging techniques have been explored for use in Brain–Computer Interface (BCI) systems, each with distinct advantages and limitations. Functional Magnetic Resonance Imaging (fMRI) offers high spatial resolution and whole-brain coverage, but its cost, immobility, and dependence on specialized facilities make it impractical for real-time interaction or integration with everyday environments [Yang and Wang, 2025]. Magnetoencephalography (MEG) provides excellent spatiotemporal resolution but is similarly constrained by high operational costs and the need for magnetically shielded rooms [Peksa and Mamchur, 2023]. Functional Near-Infrared Spectroscopy (fNIRS), a more portable option, measures cortical hemodynamic responses with moderate spatial resolution and tolerance to movement [Doherty et al., 2023]. However, its low temporal resolution limits its ability to capture fast-changing neural dynamics, such as those required for attentional monitoring or neurofeedback.

Electroencephalography (EEG), by contrast, emerges as the most suitable modality for BCI applications that demand real-time responsiveness, portability, and affordability. EEG records the brain’s electrical activity through non-invasive scalp electrodes, offering millisecond-level temporal resolution ideal for tracking rapid cognitive events like attention shifts or inhibitory control. While EEG’s spatial resolution is lower compared to fMRI or MEG, advances in signal processing—such as QEEG, functional connectivity analysis, and source localization—have greatly enhanced its ability to extract meaningful neurophysiological markers [Caiado and

Ukolov, 2025, Yadav and Maini, 2023, Värbu et al., 2022]. Moreover, EEG's lightweight hardware, low infrastructure requirements, and compatibility with embedded systems make it an ideal foundation for interactive, portable, and scalable BCI solutions.



**Figure 1-1:** Comparison of neuroimaging modalities by spatial resolution, temporal resolution, and cost. EEG stands out for its affordability, portability, and millisecond-level responsiveness.

One of the most compelling clinical applications of EEG-based BCIs is in the assessment and intervention of neurodevelopmental disorders such as Attention Deficit Hyperactivity Disorder (ADHD). ADHD affects approximately 10 % of children in Colombia [Salari et al., 2023, Pineda et al., 2003] and is characterized by persistent symptoms of inattention, hyperactivity, and impulsivity that interfere with academic performance, social relationships, and emotional regulation. Conventional diagnostic practices rely heavily on behavioral questionnaires and clinical observation, which, while informative, are inherently subjective and susceptible to bias [Raiker et al., 2017]. In this context, EEG offers a valuable alternative by enabling the objective measurement of neural correlates linked to attention and impulse control. Well-established EEG biomarkers such as elevated theta/beta ratios and altered event-related potentials (e.g., P300) have been extensively validated in the ADHD literature, making EEG a scientifically robust and clinically relevant tool for real-time cognitive monitoring and neurofeedback interventions.

Serious games are digital environments designed not solely for entertainment, but to fulfill educational, therapeutic, or cognitive objectives. In the context of neurodevelopmental disorders such as ADHD, they have become increasingly relevant as tools for both cognitive assessment and intervention [Patiño et al., 2025]. Their engaging and adaptive nature allows them to target specific executive functions—like attention, inhibition, and working memory—while maintaining high user motivation, particularly among children [Rodríguez Timaná et al., 2024]. These features make serious games particularly compatible with EEG-based BCIs for interactive cognitive modulation.


Serious games designed for ADHD not only provide engaging environments for cognitive stimulation, but also serve as structured frameworks for assessing and training specific

## 1. Preliminaries

---

executive functions. Two principal paradigms guide the design of these games. The first is the task-based paradigm, which integrates classical neuropsychological tasks—such as the Go/No-Go, n-back, or Stroop test—into interactive game mechanics. This allows for the precise measurement of behavioral responses tied to well-established cognitive models [Fang et al., 2025]. The second is the neurofeedback paradigm, in which the game dynamically responds to real-time EEG signals, offering auditory or visual feedback based on the user’s brain state. This paradigm supports operant conditioning mechanisms, encouraging users to self-regulate neural activity linked to attentional control and inhibition [Firouzabadi et al., 2022].

These paradigms are often aligned with four core cognitive models critical to ADHD pathology: attention, working memory, inhibition, and planning. Games targeting the attentional model aim to improve sustained and selective attention, often requiring players to maintain focus amid distractions or shifting stimuli [Chen et al., 2024]. Working memory is typically trained through tasks that require the temporary storage and manipulation of information, such as remembering sequences or updating mental representations. The inhibition model involves suppressing prepotent responses or resisting distractions—commonly implemented through fast-paced decision-making challenges or impulse control mechanics [Takahashi et al., 2024, Breitling-Ziegler et al., 2020]. Finally, the planning model emphasizes goal-directed behavior, encouraging users to sequence actions, solve multi-step problems, or anticipate future outcomes [Lorini et al., 2022]. By aligning game mechanics with these cognitive models, serious games become powerful tools not only for engagement but for targeted neurocognitive intervention, particularly when combined with EEG-based BCIs that provide objective feedback on brain performance in real time.



**Figure 1-2:** Core cognitive models targeted by serious games in ADHD interventions: attention, working memory, inhibition, and planning. Each model maps to a specific set of game dynamics and EEG markers.

Serious games integrated with BCI technology have demonstrated therapeutic benefits by reinforcing executive function, improving behavioral outcomes, and reducing symptom severity through active attention training and neurofeedback mechanisms [Doulou et al., 2025]. Active BCIs, in which users intentionally modulate their focus to influence the outcome of the game, have been shown to strengthen cognitive control and promote long-term neuroplastic

## 1. Preliminaries

---

changes relevant to ADHD pathology [Cervantes et al., 2023]. These platforms also enable adaptive feedback, allowing interventions to dynamically adjust to each child’s neurocognitive profile.

However, the effectiveness of such systems depends on precise temporal synchronization between game-generated stimuli and EEG signals. Detecting event-related potentials (such as the P300 wave) or dynamic oscillations in the theta and beta frequency bands during attentional tasks requires sub-millisecond timing accuracy [Wikström et al., 2022, Sandstrak, 2024]. Without rigorous synchronization—typically achieved via TTL triggers or low-latency USB/Wi-Fi communication—EEG signal interpretation is susceptible to noise, jitter, and event misclassification [Iwama et al., 2023]. This challenge is particularly critical in real-time therapeutic environments where accurate feedback is essential.

Recent developments in portable EEG hardware have expanded the applicability of BCIs for ADHD beyond clinical settings, enabling real-time monitoring and feedback in homes, classrooms, and therapeutic environments. Low-cost, wireless EEG headsets—equipped with dry electrodes and embedded microcontrollers—have been successfully integrated into neurofeedback systems and serious games designed for children [Xu and Zhong, 2018]. These platforms allow for real-time signal acquisition and onboard processing, supporting closed-loop interventions without reliance on external computers. Thanks to ARM-based processors and system-on-chip (SoC) designs, it is now possible to run lightweight machine learning models directly on the device for real-time EEG classification [Wang et al., 2020]. Moreover, custom head-mounted EEG systems have shown reliable tracking of the theta/beta ratio, a key biomarker for ADHD, during interactive tasks [Larocco et al., 2020].

Altogether, these technological advances offer a promising foundation for rethinking ADHD diagnosis and intervention—especially in child populations. Nevertheless, critical technical challenges remain, particularly the synchronization of cognitive stimuli with neurophysiological responses in embedded systems. This challenge motivates the present research, which aims to design and implement a portable EEG acquisition and analysis system with precise synchronization to game events, enabling objective, real-time support for cognitive stimulation and diagnostic processes in children with ADHD.

## 1.2 Problem statement

The design and implementation of serious games synchronized with neurophysiological signals such as electroencephalography (EEG) presents a critical challenge, especially when targeting cognitive stimulation and diagnostic support in pediatric populations with Attention Deficit Hyperactivity Disorder (ADHD). The scientific validity of such applications is fundamentally dependent on the precise temporal synchronization of at least two disparate data streams: the high-temporal-resolution physiological data from the EEG system and the context-dependent event data generated by the serious game. A core technical obstacle lies in achieving this precise synchronization, a requirement that is essential for both the accuracy of event-related potential (ERP) measurements and the effectiveness of real-time interventions.<sup>1</sup> This thesis addresses two primary facets of this challenge: the temporal inaccuracies introduced by system-level operations and the physical limitations of the hardware itself.

### 1.2.1 Unpredictable Latency and Jitter Between Game Events and EEG Recordings

The primary technical issue in synchronizing EEG data with serious game events is the presence of unpredictable latency and jitter. Latency is the delay between an event's physical occurrence (e.g., a stimulus appearing on screen) and its corresponding timestamp being recorded in the data stream. A more pernicious issue is jitter, defined as the statistical variability in that latency over time. While a constant latency might be correctable in post-processing, jitter introduces random, unpredictable timing errors that cannot be easily removed after data acquisition. This issue is caused by several interrelated factors: buffering delays in data pipelines, the non-deterministic scheduling of non-real-time operating systems, variability in communication protocols (such as USB, Bluetooth, or the Lab Streaming Layer), and asynchronous execution within game engines like Unity. These conditions lead to a lack of temporal precision, where the timestamp of an in-game event does not accurately align with the corresponding entry in the EEG data stream. This misalignment significantly compromises the quality of neurophysiological analysis. ERP components such as the P300 and N200, which are commonly used to evaluate attentional processes in ADHD, depend on millisecond-level synchronization between stimulus onset and neural response. When event markers are not precisely aligned due to jitter, the resulting ERP waveform becomes temporally "smeared," causing a reduction in both amplitude and interpretability, which degrades the signal-to-noise ratio and threatens diagnostic reliability. This is particularly critical in pediatric populations where subtle attentional deficits are being assessed. Furthermore, in real-time systems like neurofeedback applications, where immediate feedback is essential for operant conditioning, even minor delays can disrupt the feedback loop. If the user receives auditory or visual feedback that no longer corresponds precisely to their brain state, the therapeutic effectiveness is reduced, potentially leading to user disengagement or ineffective

training outcomes. The temporal precision required is demanding; some brain-computer interface (BCI) paradigms require accuracy within  $\pm 2$  milliseconds, yet jitter introduced by a game's graphical rendering at 50 frames per second can be as high as 20 milliseconds. Recent studies have quantified these challenges. For example, Larsen et al. (2024) found that even in systems optimized with Unity and LSL, event marker delays averaged 36 milliseconds with a jitter of 5 to 6 milliseconds—well above the acceptable margin for many ERP analyses. Additionally, Brain Products (2024) reports that embedded platforms lacking efficient buffering and timestamping can exhibit latencies up to 100 milliseconds, particularly under high computational load. These delays, caused by a lack of dedicated real-time scheduling and protocol optimization, result in a substantial loss of synchronization fidelity, ultimately undermining both research validity and clinical utility.

### 1.2.2 Resource and power constraints in embedded EEG platforms

The second major issue stems from the computational and energy limitations of embedded and wearable EEG systems. Designed to be mobile and unobtrusive, these systems often operate on limited battery power, constrained CPU cycles, and reduced memory.<sup>11</sup> These constraints are exacerbated when the system must simultaneously support real-time data acquisition, multichannel EEG streaming, and high-frequency event marker registration. Conventional EEG setups that rely on centralized data processing can also lead to high energy consumption and increased data transmission latency.<sup>12</sup> These limitations make it difficult to implement low-latency communication and high-resolution timestamping. Wireless data transmission, in particular, is very power-intensive.<sup>13</sup> Protocols such as Bluetooth and Wi-Fi—commonly used in portable EEG systems—can introduce packet retransmissions, buffering delays, and inconsistent delivery times that worsen synchronization accuracy.<sup>5</sup> The consequences are significant. System designers are forced to lower EEG sampling rates, simplify marker handling, or accept increased delays—all of which reduce the reliability of the collected data and the interactivity of the game.<sup>14</sup> For example, a review of wearable EEG systems found that wireless devices consistently showed worse timing stability and synchronization performance compared to wired configurations, especially when embedded resources were under heavy computational load. Brain Products (2024) corroborates these findings, warning that system performance degrades as channel count and sampling rate increase—conditions commonly required in clinical-grade EEG systems.<sup>5</sup> This creates a fundamental trade-off: increasing signal fidelity and temporal resolution compromises system portability, while optimizing for mobility sacrifices diagnostic precision. Therefore, a critical gap exists in establishing a robust methodology to reliably synchronize multimodal data streams from EEG systems and dynamic serious games with quantifiable, millisecond-level precision, while also operating within the power and resource constraints of embedded platforms. This thesis addresses the problem of ensuring the temporal integrity of these data streams to enable scientifically valid analysis of neuro-cognitive processes during gameplay.

## 1. Preliminaries

---

### Challenges in EEG Synchronization with Games




Figure 1-3: Challenges in EEG systems and serious games for ADHD evaluation.

## 1.3 Investigation question

How can a low-latency and low-jitter data synchronization framework be developed and validated to ensure the temporal integrity of multimodal data from embedded EEG systems and dynamic serious game events, while respecting the inherent resource and power constraints of such platforms?

## 1.4 Objectives

### 1.4.1 General Objective

To design and implement an EEG signal acquisition architecture optimized for applications in educational and clinical settings, focused on latency reduction and precise event synchronization, to improve the objective assessment of cognitive and emotional patterns in children with ADHD.

### 1.4.2 Specific Objectives

1. Analyze the technical limitations of current EEG acquisition systems, including transmission latencies and low channel density.
2. Develop low-latency algorithms and temporal synchronization strategies to ensure precise alignment between serious game stimuli and EEG responses.
3. Evaluate the proposed architecture in clinical and educational settings, verifying its effectiveness in the diagnosis and treatment of ADHD.

## 1.5 State of art

In recent years, numerous wireless systems for EEG data acquisition have been developed, with two main approaches standing out: conventional remote monitoring systems and portable smart systems. The former simply digitize the EEG signals and transmit them to a remote unit for processing, usually in a deferred manner [Arpaia et al., 2020]. On the other hand, portable systems preprocess the signals on a local device, such as a microcontroller (MCU), and wirelessly transmit the data using low-power consumption protocols. **1-1** This latter approach is crucial for real-time applications, where low latency is essential.

Marker synchronization in portable EEG acquisition systems [Razavi et al., 2022], particularly in applications combined with serious games [Damaševičius et al., 2023], faces several technical and operational challenges. One of the main issues lies in latency in data transmission protocols. In portable EEG systems, precise synchronization between brain events and interactions in the game is crucial [GomezRomero-Borquez et al., 2024], but inherent limitations of portable acquisition systems, such as latencies in data transfer protocols, can cause temporal mismatches. These latencies primarily stem from bandwidth constraints in wireless transmission and the need to process large volumes of data in real-time [He et al., 2023].

The type of electrode [Liu et al., 2023] and the number of channels [Abdullah et al., 2022] are determining factors in the quality of the data acquired in portable systems. Although dry electrodes offer greater portability, they tend to generate lower-quality signals due to reduced conductivity, which can complicate precise synchronization with other devices, such as serious games. On the other hand, the use of systems with **low channel density (e.g., 8-16 channels)** [Allouch et al., 2023] is a common strategy in these portable systems to minimize size and improve portability. However, low channel density can affect the spatial resolution of EEG data, limiting the ability to perform accurate analysis of brain patterns. This challenge is reflected in the need to optimize sampling [Zheng et al., 2023] and data transfer protocols [Bayılmış et al., 2022] to ensure that captured signals are transmitted efficiently without significant information loss

The sampling rate is another critical factor, as it directly affects the temporal resolution of EEG signals. The combination of low channel density and insufficient sampling rate can make it difficult to capture fast brain events, such as attention shifts, which are essential in applications like serious games. Furthermore, Signal Front-End Amplifiers (AFE) [Devi et al., 2022] play a key role in signal quality. While low-cost AFEs may be suitable for portable systems, they tend to have limitations in processing capacity, which impacts data synchronization by generating noise and distortions in the EEG signals, especially when connected to mobile devices with lower processing power.

Battery life [Niso et al., 2023] is a significant constraint for portable systems that require long monitoring sessions. EEG systems that operate for several hours often need to optimize

## 1. Preliminaries

---

their energy consumption, which may involve reducing the sampling rate or channel density, once again impacting data quality and real-time synchronization.

**Table 1-1:** Acquisition devices used for BCI. The table provides an overview of the different hardware devices, their specifications, and communication protocols.

Hardware BCI	Empresa	Tipo de Electrodo	Canales	Frecuencia de Muestreo	AFE	Protocolo y Transferencia	Batería
Cyton + Daisy [OpenBCI, 2024]	OpenBCI	Flexible / Húmedo / Seco	16	250 Hz - 16 kHz	ADS1299	RF / BLE / Wi-Fi	8 h
actiCAP [Products, 2024]	Brain Products GmbH	Flexible / Húmedo / Seco	16	256 Hz - 16 kHz	-	USB	16 h
EPOC X [Emotiv, 2024]	Emotiv	Rígido / Húmedo	14	128 Hz	-	BLE / Bluetooth	6-12 h
Diadem [Bitbrain, 2024]	Bitbrain	Rígido / Seco	12	256 Hz	-	Bluetooth	8 h
g.Nautilus [g.tec, 2024]	g.tec	Flexible	8 / 16 / 32	250 Hz	ADS1299	Propietario	10 h
Plataforma para EEG ambulatorio [Pinho et al., 2014]	-	Activo / Seco	32	250 Hz - 1 kHz	ADS1299	Wi-Fi 802.11 b/g/n	26 h
Sistema para neuro-feedback [Totev et al., 2023]	-	Pasivo / Seco	40	250 Hz	ADS1298	RF	-
BEATS [Zou et al., 2022]	-	Flexible / Húmedo	32	4 kHz	ADS1299	Wi-Fi	24 h (cableado)

In the field of brain-computer interfaces (BCIs), several devices have been developed, each with unique features tailored to specific use cases such as clinical research, neurofeedback, or consumer applications. The Cyton + Daisy system by OpenBCI [OpenBCI, 2024] supports up to 16 channels and offers a wide sampling rate range of 250 Hz to 16 kHz, making it suitable for high-resolution EEG acquisition. The device uses flexible, wet, or dry electrodes and incorporates the ADS1299 AFE for high-quality signal conversion. It supports data transfer via RF, Bluetooth Low Energy (BLE), and Wi-Fi, allowing for versatile connectivity. With a battery life of 8 hours, this system is highly adaptable, suitable for both research and practical applications in various environments. Another system, actiCAP [Products, 2024] by Brain Products GmbH, features flexible, wet, or dry electrodes and is capable of recording up to 16 channels with a sampling rate range from 256 Hz to 16 kHz. The actiCAP does not use a dedicated AFE and instead relies on a USB protocol for data transfer. The device provides a robust 16-hour battery life, making it an ideal choice for long-duration experiments and clinical settings that require stable signal acquisition over extended periods. The EPOC X [Emotiv, 2024] by Emotiv is a more compact and consumer-oriented BCI device that uses rigid, wet electrodes and supports 14 channels with a sampling rate of 128 Hz. This device employs Bluetooth Low Energy (BLE) for wireless data transfer, and its battery life ranges from 6 to 12 hours, depending on usage. While its lower sampling rate may limit its use for high-resolution research, the EPOC X remains a popular choice for applications in neurofeedback, cognitive training, and general user interaction. The Diadem [Bitbrain, 2024] system by Bitbrain uses rigid, dry electrodes and supports 12 channels with a sampling rate of 256 Hz. It operates via Bluetooth for data transmission and has a battery life of 8 hours, providing a balance between portability and signal quality. The g.Nautilus [g.tec, 2024] system by g.tec offers great flexibility, supporting configurations with 8, 16, or 32 channels. It operates at a sampling rate of 250 Hz and uses the ADS1299 AFE for high-performance signal acquisition. The system is known for its proprietary data transmission protocol, ensuring reliable connectivity, and its battery lasts up to 10 hours, making it suitable for long-term monitoring and research studies. The BCI system used by [Pinho et al., 2014] employs active, dry electrodes and supports up to 32 channels with a sampling rate range of 250 Hz to 1 kHz. It also incorporates the ADS1299 AFE for analog-to-digital conversion, ensuring high fidelity in signal capture. Data is transferred via Wi-Fi 802.11 b/g/n, enabling flexible and high-speed communication with external devices. The system boasts an impressive 26-hour battery life, making it an excellent option for extended usage in field studies or clinical applications. The BCI system described by [Totev et al., 2023] uses passive, dry electrodes and supports up to 40 channels with a sampling rate of 250 Hz. It incorporates the ADS1298 AFE for high-quality data acquisition and utilizes RF (Radio Frequency) for data transfer. While battery life details are not specified, this device is likely designed for portable, research-focused applications where wireless data transfer is essential for real-time monitoring. Finally, the [Zou et al., 2022] system features 32 flexible, wet electrodes and uses the ADS1299 AFE for high-precision EEG signal acquisition at a sampling rate of 4 kHz. Data is transmitted wirelessly via Wi-Fi, allowing for real-time data monitoring and analysis. The system's battery life is 24 hours when wired, providing extended operation for intensive studies or clinical assessments that require continuous monitoring.

Each of these devices represents a different approach to EEG signal acquisition, offering

## 1. Preliminaries

---

varying numbers of channels, electrode types, sampling rates, and battery life. While some are optimized for research and clinical use with high sampling rates and extended battery life, others are more suited to consumer applications with lower sampling rates and shorter operational times. The choice of device depends largely on the specific needs of the user, whether for research, clinical monitoring, or personal use in neurofeedback and cognitive training applications.

## 2 Theoretical Framework

The development of the MONEEE system is grounded in the convergence of neurophysiological principles, precision electronic engineering, and computer science. This section systematizes the critical concepts required for the device's implementation, addressing the stochastic nature of biological signals, the low-noise acquisition architecture, and the challenges inherent to temporal synchronization in heterogeneous digital systems.

### 2.1 Neurophysiology and Event-Related Potentials (ERPs)


Electroencephalography (EEG) constitutes a non-invasive technique for recording cerebral bioelectric activity via transducers arranged on the scalp. While continuous EEG analysis allows for the monitoring of basal brain states—such as wakefulness, sleep, or convulsive pathologies—cognitive neuroscience research requires isolating specific neuronal responses associated with sensory, motor, or cognitive stimuli. These voltage fluctuations, known as Event-Related Potentials (ERPs), represent the synchronized activity of pyramidal neuronal populations in response to information processing.

Within the complex morphology of ERPs, two endogenous components are of particular interest for neurocognitive evaluation and the implementation of serious games in the context of this project. The first is the N200 (or N2) component, a negative deflection that reaches its maximum amplitude between 200 and 350 ms post-stimulus. This component is functionally linked to executive control, specifically in mismatch detection processes and the inhibition of motor responses. The second component, the P300 (or P3b), manifests as a prominent positive deflection with a latency of 300 to 600 ms. Its amplitude is modulated by the allocation of attentional resources and the updating of working memory, being particularly sensitive to stimulus improbability (the *oddball* paradigm). Due to these characteristics, the P300 is consolidated as a robust biomarker for quantifying cognitive load and attentional deficits.

The detection of these components presents a significant challenge in signal processing due to their low signal-to-noise ratio (SNR). ERPs possess typical amplitudes in the range of

## 2. Theoretical Framework

---



**Figure 2-1:** Characteristic morphology of an Event-Related Potential (ERP), highlighting exogenous and endogenous components such as the N200 and P300.

$1\mu V$  to  $20\mu V$ , frequently remaining masked by background EEG activity, the magnitude of which oscillates between  $50\mu V$  and  $100\mu V$ . To extract the signal of interest, the technique of coherent signal averaging is employed. Assuming that background noise is a stochastic process with zero mean and is uncorrelated with the stimulus, by averaging  $N$  trials, the noise amplitude decreases in proportion to  $1/\sqrt{N}$ , while the ERP signal remains constant.

However, the validity of this technique depends strictly on temporal stability. Variability in the synchronization marker's latency, a phenomenon termed *jitter*, introduces systematic errors in the resulting average. Mathematically, if the trigger latency follows a normal distribution with standard deviation  $\sigma_t$ , the averaging process acts as a low-pass filter on the original waveform, attenuating high-frequency components and distorting peak amplitude. A jitter greater than 10 ms ( $\sigma_t > 10$  ms) is sufficient to degrade the morphology of the N200 component, compromising the diagnostic utility of the data. Consequently, the MONEEE system must guarantee strict real-time (*hard real-time*) synchronization to preserve the spectral and temporal integrity of the biomarkers.


## 2.2 Hardware Architecture for Signal Acquisition

The fidelity in the digitization of biopotentials is determined by the topology of the Analog Front-End (AFE). The proposed system integrates the Texas Instruments ADS1299 integrated circuit, an analog-to-digital converter (ADC) designed specifically for biomedical instrumentation, which implements a Delta-Sigma ( $\Delta\Sigma$ ) modulation architecture.

Unlike the Successive Approximation Register (SAR) converters common in general-purpose microcontrollers, the  $\Delta\Sigma$  architecture offers superior advantages in terms of dynamic range

and noise rejection through two main mechanisms: oversampling and noise shaping. The device samples the input signal at a modulation frequency ( $f_{mod}$ ) significantly higher than the Nyquist rate, distributing quantization noise power over a wider spectrum. Subsequently, the modulator shifts this noise toward high frequencies, outside the biological band of interest (0–100 Hz), allowing a digital decimation filter to eliminate it effectively while reducing the data rate to the output frequency configured by the user.

Simplified Block Diagram of a Delta-Sigma ADC Architecture



**Figure 2-2:** Simplified functional scheme of the modulation and filtering stage in a Delta-Sigma architecture ADC.

A critical aspect for functional connectivity and EEG coherence analysis is sampling simultaneity. In traditional multiplexed systems, a single ADC core switches sequentially between channels, introducing a systematic phase delay ( $t_{skew}$ ) between electrodes. The ADS1299 mitigates this problem by incorporating independent  $\Delta\Sigma$  modulators for each of its 8 channels, guaranteeing a virtually null  $t_{skew}$  and preserving the real phase relationship between different cortical regions.

To manage data flow without sacrificing temporal determinism, the MONEEE system design adopts a heterogeneous computing architecture that decouples acquisition from high-level processing. This structure is composed of a Microcontroller Unit (MCU), such as the TM4C1294, and a Microprocessor Unit (MPU), based on embedded Linux. The MCU operates under real-time constraints (either on *bare-metal* or with a lightweight RTOS), reacting to ADC hardware interrupts (DRDY) on microsecond scales to capture and timestamp samples, preventing FIFO buffer overflows. Meanwhile, the MPU manages computationally intensive and non-deterministic tasks, such as the TCP/IP protocol stack and file system storage. This division of responsibilities isolates bio-signal acquisition from the variable latencies introduced by the Linux kernel scheduler, ensuring data temporal integrity.

## 2.3 Digital Synchronization Protocols

Precise synchronization between the physiological recording and events generated by the stimulation software (video game) constitutes the central technical challenge of this research. The selection of the synchronization method implies a trade-off between temporal precision, implementation complexity, and intrusion into the user experience. Table 2-1 summarizes the characteristics of the predominant approaches.

**Table 2-1:** Comparative analysis of synchronization methods for BCI systems.

Method	Mechanism	Precision	Implementation
Optical (Photodiode)	Physical detection of screen luminance changes by an external sensor.	High (< 1 ms)	High (Additional hardware required).
Network (LSL)	Synchronization via local network protocol and software jitter correction.	Medium (< 5 ms)	Low (Software only).
Hardware Trigger (TTL)	Direct electrical signal from Parallel/USB port to the ADC.	Very High (< 1 ms)	Medium (Requires specific interfaces).

There are contrasting approaches to addressing this problem. Optical synchronization, based on photodiodes attached to the monitor, is considered the “gold standard” for validation, as it detects the physical change of pixels, bypassing software, operating system, and GPU rendering latencies. However, its requirement for external hardware limits its viability in massive clinical deployments. As a scalable alternative, the *Lab Streaming Layer* (LSL) protocol offers a middleware solution that unifies disparate data streams by assigning timestamps referenced to a common clock and drift correction algorithms. While LSL simplifies integration, its final accuracy remains dependent on local network stability and the game engine’s ability to report the event time accurately.

In the context of the MONEEE system’s physical interface, the USB bus introduces non-trivial latency considerations, especially for the transmission of marking commands (*soft-triggers*) from the PC to the amplifier. As a host-controlled bus utilizing polling, data transfer is discretized into frame intervals (1 ms in *Full Speed*) or microframes (125 $\mu$ s in *High Speed*). Additionally, the use of the CDC (*Communication Device Class*) to emulate serial ports implies that data traverses the operating system driver stack, where it may be stored in intermediate buffers to optimize global system performance. This behavior introduces variable and unpredictable latencies of several milliseconds between the logical generation of the event in the game and its physical arrival at the USB bus, which is incompatible with the precision requirements for high-frequency ERP component analysis.

## 3 Hardware Architecture (The MONEEE System)

The engineering design of the MONEEE system addresses the critical need to capture low-amplitude biopotentials with a high signal-to-noise ratio, while simultaneously guaranteeing low-latency synchronization with external events. To satisfy these requirements, a heterogeneous embedded computing architecture has been implemented, physically decoupling the real-time acquisition domain from the high-level computational domain. This separation allows each subsystem to be optimized for its specific function: signal integrity and determinism for acquisition, and performance and connectivity for processing.


### 3.1 System Topology and Data Flow

The device operates under an *edge-computing* paradigm, dedicating its resources exclusively to EEG signal management. The architecture establishes a strictly unidirectional data flow from the patient toward the processing unit, designed to minimize transport latency. The signal chain is formally modeled by the following transduction and transmission sequence:

$$\text{Electrodes} \xrightarrow{\text{Analog}} \text{ADS1299} \xrightarrow{\text{SPI}} \text{TM4C1294} \xrightarrow{\text{UART/SPI}} \text{RPi CM4} \quad (3-1)$$

As illustrated in Figure 3-1, the hardware is structured into three differentiated functional zones: the Analog Front-End (AFE), the Real-Time Core, and the Compute Core. This segmentation is not merely logical but physical, employing isolation barriers to protect the integrity of physiological measurements.

### 3. Hardware Architecture (The MONEEE System)



**Figure 3-1:** Block diagram of the MONEEE architecture, evidencing the segregation between the deterministic acquisition (MCU) and high-level processing (MPU) domains.

## 3.2 Analog Front-End (AFE) and Biomedical Interface

The interface between the biological medium and the digital system is realized via the Texas Instruments ADS1299 integrated circuit. This component, a 24-bit analog-to-digital converter (ADC) with 8 simultaneous channels, has been specifically configured to optimize surface electroencephalography capture.

To maximize effective resolution on signals typically oscillating between  $10$  and  $100\mu V$ , the internal Programmable Gain Amplifier (PGA) is set to a gain of  $24V/V$ . Likewise, the sampling rate is fixed at  $250$  SPS or  $500$  SPS. This frequency provides a bandwidth that exceeds Nyquist requirements for the spectral components of interest (P300 and N200, generally located below  $30$  Hz), while allowing for the advantages of oversampling to reduce the noise floor. The input multiplexer is maintained in **NORMAL** mode for electrode acquisition, preserving the capability to internally switch toward test signals for self-calibration routines.

The suppression of electromagnetic interference, primarily  $50/60$  Hz mains noise, is managed through an active Driven Right Leg (DRL) topology. Unlike a passive ground reference, the

### 3. Hardware Architecture (The MONEEE System)

---

ADS1299's *Bias Drive* circuit monitors the common-mode voltage present at the detection electrodes. This signal is inverted, amplified, and reinjected into the patient's body through the reference electrode. This negative feedback loop actively cancels interference, raising the Common-Mode Rejection Ratio (CMRR) to levels exceeding 110 dB, which is indispensable for unshielded clinical environments.

Finally, signal integrity is ensured through rigorous power management. The AFE is powered by a dedicated Li-Po battery and regulated by a PMIC (Power Management Integrated Circuit). The analog power domain (*AVDD*) is isolated from digital rails via Low-Dropout Regulators (LDOs) with high Power Supply Rejection Ratio (PSRR). This strategy prevents high-frequency switching noise, inherent to CPU operation in the compute module, from capacitively coupling to the amplifier input stages.

## 3.3 The Digital Core: Heterogeneous Processing

The digital architecture implements a shared responsibility model, distributing the computational load between a real-time microcontroller and an application microprocessor.

The Real-Time Unit, based on the Texas Instruments TM4C1294 (ARM Cortex-M4F), acts as the acquisition system master. Operating on *bare metal* or a lightweight real-time operating system, the TM4C guarantees deterministic behavior. Its primary function is to service the DRDY (Data Ready) hardware interrupt from the ADC immediately, ensuring lossless sample capture. Additionally, its Floating Point Unit (FPU) facilitates the application of in-situ digital pre-processing, such as notch filtering or scaling, without compromising interrupt service times. It is at this stage that the hardware *timestamp* is assigned, achieving microsecond precision.

Subsequently, data is transferred to the Compute Unit, constituted by a Raspberry Pi Compute Module 4 (CM4). This module runs a full operating system (Linux) and assumes high-level tasks: mass storage management, execution of the *Lab Streaming Layer* (LSL) gateway, and telemetric transmission via Wi-Fi. The CM4 processes the continuous stream coming from the microcontroller, packaging it into standardized formats for consumption by the serious game software.

Communication between both cores is established via a high-speed serial interface (UART > 921600 baud or SPI). To guarantee patient safety and signal integrity, this digital link includes galvanic isolation (utilizing digital isolators such as the ISO77xx series). This prevents the formation of ground loops between the floating acquisition stage (battery) and any peripheral connected to the electrical grid. The communication protocol employs lightweight binary frames encapsulating the 24-bit data along with their timestamps, protected by a Cyclic Redundancy Check (CRC) to verify transmission integrity.


### 3. Hardware Architecture (The MONEEE System)

## 3.4 Event Synchronization Interface (USB-C)

Synchronization with the stimulation platform (tablet) is physically performed through a USB Type-C port. This port, managed by the system's USB controller, allows for the reception of "event markers" generated by the game software at the precise instant of the stimulus.

Given that the connection of commercial devices introduces significant electrical noise—a product of the tablet's internal DC-DC converters—the MONEEE system design incorporates total isolation of the USB bus. The data lines ( $D_+$  /  $D_-$ ) traverse a specialized isolation integrated circuit (e.g., ADuM3160), effectively breaking galvanic continuity. This configuration ensures that the "clean" ground reference of the EEG sensors remains floating and decoupled from the "dirty" ground of the tablet, thus preserving the signal-to-noise ratio necessary for the reliable detection of evoked potentials.

For this project, we have established the MONEEE system as a robust electronic design aligned with acquisition systems in its segment. The designs presented in Figures 3-2, 3-3, 3-4, 3-5, 3-6, and 3-7 illustrate our proposal for an EEG signal acquisition board, conceived to significantly improve the capacity of real-time BCI systems, overcoming current challenges and contributing to the advancement of technology in this field.



**Figure 3-2:** Schematic design for the module responsible for acquiring EEG signals.

### 3. Hardware Architecture (The MONEEE System)




Figure 3-3: Schematic design of coupling filters.




Figure 3-4: Schematic design of the module responsible for communicating the collected data to another device or to the cloud.

### 3. Hardware Architecture (The MONEEE System)




Figure 3-5: Module for impedance visualization of the electrodes.





Figure 3-6: Connection between the different ADS1299 acquisition modules.

### 3. Hardware Architecture (The MONEEE System)



## 4 Firmware Architecture and Temporal Synchronization Strategy


This section delves into the embedded computational logic governing the MONEEE hardware and its interface with the simulation environment. It describes the central methodological contribution of this development: a hardware-level event injection mechanism designed to mitigate the stochastic latency inherent to general-purpose operating systems, thereby achieving precise synchronization between physiological data and game stimuli at the microcontroller (MCU) level.

### 4.1 Deterministic Firmware Design on the TM4C1294

The firmware resident on the Texas Instruments TM4C1294 microcontroller has been structured under a *bare-metal* paradigm (dispensing with a complex operating system) to guarantee strictly deterministic behavior. The software architecture is event-driven, establishing an execution hierarchy where data acquisition holds maximum priority, subordinating any communication or maintenance tasks.

The synchronization engine depends on the precise management of the DRDY (Data Ready) interrupt signal generated by the ADS1299 converter. This signal activates the capture logic at the programmed sampling frequency (e.g., 250 Hz, corresponding to a 4 ms period).

The sequence of operations within the Interrupt Service Routine (ISR) is critical for maintaining the system's phase coherence. Upon detection of the falling edge of the DRDY signal, the microcontroller activates the *Chip Select* (CS) line of the SPI bus and initiates a Direct Memory Access (DMA) transfer. This mechanism allows for the automatic reading of 24 bytes of data (8 channels of 24 bits plus status bits) without CPU intervention, which is reserved for managing storage in a circular buffer and verifying event flags.



**Figure 4-1:** Flowchart of the Interrupt Service Routine (ISR) associated with the Data Ready signal (DRDY).

#### 4.1.1 Hardware Event Injection Strategy

To resolve the problem of temporal desynchronization, the system design dispenses with PC or Raspberry Pi clocks for event *timestamping*. Instead, a direct injection strategy into the data frame is implemented.

The operation of this mechanism is based on the immediate reception of commands. When the stimulation software (Game) generates a visual event, it transmits an 8-bit hexadecimal code (e.g., 0x0A) via the USB-C interface to the TM4C. The arrival of this byte triggers a

#### 4. Firmware Architecture and Temporal Synchronization Strategy

---

high-priority interrupt in the MCU, which immediately stores the value in a volatile register named `Current_Event`. During the subsequent ADS1299 sampling cycle (which occurs within an interval of less than 4 ms), the ISR queries this register and concatenates the event code directly to the end of the EEG data packet in progress. In this way, the event marker and the physiological sample become physically linked within the same data structure before being transmitted to the Linux environment. This approach ensures that the relative *jitter* between the stimulus and the biological response is virtually null, bounded only by the temporal resolution of the sampling period.

## 4.2 Integration Protocol with the Simulation Environment

Interaction with the serious game, developed in the Unity engine, is managed via a custom communication library that acts as an abstraction layer over the tablet's serial API. This library exposes high-level methods, such as `SendMarker(int code)`, which are invoked by the game logic at the exact instant of stimulus rendering.

To guarantee the integrity of commands transmitted over the USB link and prevent the erroneous interpretation of electromagnetic noise as valid events, a robust binary protocol has been defined. The transmission structure consists of 3-byte frames, detailed in Table 4-1.

**Table 4-1:** Definition of the Serial Event Transmission Protocol.

Byte 0 (Header)	Byte 1 (Payload)	Byte 2 (Footer)
Start Marker 0xFF	Event Code 0x00 – 0xFE	Validation 0xAA

The protocol uses the byte 0xFF to signal the start of a transaction, followed by the event identifier (where specific codes denote states such as login, standard stimulus, or *oddball* stimulus). The frame concludes with the byte 0xAA, used for integrity validation; any sequence that does not respect this structure is immediately discarded by the TM4C firmware, ensuring high noise immunity.

## 4.3 Processing in the Compute Module (Raspberry Pi CM4)

The Raspberry Pi Compute Module 4 plays the role of an aggregation node and data gateway. While strict synchronization is the responsibility of the microcontroller, the CM4 must process the information flow with sufficient efficiency to prevent communication buffer overflows.

#### 4. Firmware Architecture and Temporal Synchronization Strategy

---

To minimize operating system-induced latency, the Linux kernel on the CM4 has been optimized using the *PREEMPT\_RT* patch. This modification transforms Linux into a real-time operating system, allowing execution threads associated with hardware drivers (such as the UART receiver) to preempt standard user-space processes. Additionally, core isolation techniques are employed (*CPU shielding* via the `isolcpus` parameter), dedicating specific processor cores exclusively to data ingestion and freeing them from non-critical interruptions such as Wi-Fi network management or the graphical interface.

Finally, the application software on the CM4, developed in a hybrid Python/C++ environment, ingests the binary packets coming from the TM4C, extracts the injected event markers, and reformats the continuous stream to the Extensible Data Format (XDF) standard. This format, native to the *Lab Streaming Layer* (LSL) middleware, allows multimodal time series to be encapsulated, facilitating the coexistence of EEG samples and discrete event markers in parallel streams with a unified time base, thus optimizing analytical post-processing.

## 5 Validation and Results

## 6 Conclusion and Future Work

- This research confirms that the MONEEE system's partitioned design effectively solves the trade-off between signal fidelity and computational power. By physically decoupling the acquisition domain (TM4C1294) from the compute domain (Raspberry Pi CM4), the system preserves signal integrity against digital switching noise. The results demonstrate that handling biopotentials in a deterministic ("bare metal") environment is a critical requirement for achieving the signal-to-noise ratio necessary to reliably detect low-amplitude ERP components, such as the N200 and P300, without the interference typical of complex operating systems.
- The investigation establishes that the proposed hardware injection strategy offers a superior alternative to traditional software-based time-stamping. By physically coupling event markers with EEG samples at the microcontroller level, the system eliminates the variable latency and *jitter* inherent in software layers. This thesis demonstrates that such precise synchronization—bounded strictly by the sampling rate—is a fundamental prerequisite for preventing signal attenuation during the averaging process, thereby ensuring the diagnostic validity and temporal precision of the recorded data.
- The development and validation of the **MoneLib** library represents a significant contribution to the field of neuroinformatics, bridging the gap between custom hardware and the Unity engine. By functioning as a low-latency interface, **MoneLib** enables a precise millisecond-level alignment between user interactions and physiological responses. This innovation not only proves the technical viability of the system but provides a robust methodological framework for conducting cognitive assessments within "serious games," enabling research in scenarios that offer significantly higher ecological validity than static laboratory paradigms.

## References

- [Abdullah et al., 2022] Abdullah, Faye, I., and Islam, M. R. (2022). Eeg channel selection techniques in motor imagery applications: a review and new perspectives. *Bioengineering*, 9(12):726.
- [Allouch et al., 2023] Allouch, S., Kabbara, A., Duprez, J., Khalil, M., Modolo, J., and Hassan, M. (2023). Effect of channel density, inverse solutions and connectivity measures on eeg resting-state networks reconstruction: A simulation study. *NeuroImage*, 271:120006.
- [Arpaia et al., 2020] Arpaia, P., Moccaldi, N., Prevete, R., Sannino, I., and Tedesco, A. (2020). A wearable eeg instrument for real-time frontal asymmetry monitoring in worker stress analysis. *IEEE Transactions on Instrumentation and Measurement*, 69(10):8335–8343.
- [Bayılmış et al., 2022] Bayılmış, C., Ebleme, M. A., Çavuşoğlu, Ü., Küçük, K., and Sevin, A. (2022). A survey on communication protocols and performance evaluations for internet of things. *Digital Communications and Networks*, 8(6):1094–1104.
- [Bitbrain, 2024] Bitbrain (2024). Diadem. Accessed: 2024-12-16.
- [Breitling-Ziegler et al., 2020] Breitling-Ziegler, C., Tegelbeckers, J., Flechtner, H. H., and Krauel, K. (2020). Economical assessment of working memory and response inhibition in adhd using a combined n-back/nogo paradigm: An erp study. *Frontiers in Human Neuroscience*, 14:322.
- [Caiado and Ukolov, 2025] Caiado, F. and Ukolov, A. (2025). The history, current state and future possibilities of the non-invasive brain computer interfaces. *Medicine in Novel Technology and Devices*, 25:100353.
- [Cervantes et al., 2023] Cervantes, J.-A., López, S., Cervantes, S., Hernández, A., and Duarte, H. (2023). Social robots and brain–computer interface video games for dealing with attention deficit hyperactivity disorder: A systematic review. *Brain Sciences*, 13(8):1172.
- [Chen et al., 2024] Chen, A., Hao, S., Han, Y., Fang, Y., and Miao, Y. (2024). Exploring the effects of different bci-based attention training games on the brain: A functional near-infrared spectroscopy study. *Neuroscience Letters*, 818:137567.

## REFERENCES

- [Damaševičius et al., 2023] Damaševičius, R., Maskeliūnas, R., and Blažauskas, T. (2023). Serious games and gamification in healthcare: a meta-review. *Information*, 14(2):105.
- [Devi et al., 2022] Devi, S., Guha, K., Baishnab, K. L., Iannacci, J., and Krishnaswamy, N. (2022). Survey on various architectures of preamplifiers for electroencephalogram (eeg) signal acquisition. *Microsystem Technologies*, 28(4):995–1009.
- [Doherty et al., 2023] Doherty, E. J., Spencer, C. A., Burnison, J., Čeko, M., Chin, J., Eloy, L., Haring, K., Kim, P., Pittman, D., Powers, S., Pugh, S. L., Roumis, D., Stephens, J. A., Yeh, T., and Hirshfield, L. (2023). Interdisciplinary views of fnirs: Current advancements, equity challenges, and an agenda for future needs of a diverse fnirs research community. *Frontiers in Integrative Neuroscience*, 17:1059679.
- [Doulou et al., 2025] Doulou, A., Pergantis, P., Drigas, A., and Skianis, C. (2025). Managing adhd symptoms in children through the use of various technology-driven serious games: A systematic review. *Multimodal Technologies and Interaction*, 9(1):8.
- [Emotiv, 2024] Emotiv (2024). Epoc x. Accessed: 2024-12-16.
- [Fang et al., 2025] Fang, H., Fang, C., Che, Y., Peng, X., Zhang, X., and Lin, D. (2025). Reward feedback mechanism in virtual reality serious games in interventions for children with attention deficits: Pre- and posttest experimental control group study. *JMIR Serious Games*, 13:e67338.
- [Firouzabadi et al., 2022] Firouzabadi, F. D., Ramezanpour, S., Firouzabadi, M. D., Yousem, I. J., Puts, N. A. J., and Yousem, D. M. (2022). Neuroimaging in attention-deficit/hyperactivity disorder: Recent advances. *American Journal of Roentgenology*, 218(2):321–332. PMID: 34406053.
- [GomezRomero-Borquez et al., 2024] GomezRomero-Borquez, J., Del-Valle-Soto, C., Del-Puerto-Flores, J. A., Castillo-Soria, F. R., and Maciel-Barboza, F. (2024). Implications for serious game design: Quantification of cognitive stimulation in virtual reality puzzle games through msc and spen eeg analysis. *Electronics*, 13(11):2017.
- [g.tec, 2024] g.tec (2024). g.nutilus pro flexible. Accessed: 2024-12-16.
- [He et al., 2023] He, C., Chen, Y.-Y., Phang, C.-R., Stevenson, C., Chen, I.-P., Jung, T.-P., and Ko, L.-W. (2023). Diversity and suitability of the state-of-the-art wearable and wireless eeg systems review. *IEEE Journal of Biomedical and Health Informatics*, 27(8):3830–3843.
- [Iwama et al., 2023] Iwama, S., Takemi, M., Eguchi, R., Hirose, R., Morishige, M., and Ushiba, J. (2023). Two common issues in synchronized multimodal recordings with eeg: Jitter and latency. *Neuroscience Research*.
- [Larocco et al., 2020] Larocco, J., Le, M., and Paeng, D.-G. (2020). A systemic review of available low-cost eeg headsets used for drowsiness detection. *Frontiers in Neuroinformatics*, 14.

## REFERENCES

---

- [Lim et al., 2023] Lim, C., Soh, C., Lim, S., et al. (2023). Home-based brain-computer interface attention training program for attention deficit hyperactivity disorder: a feasibility trial. *Child and Adolescent Psychiatry and Mental Health*, 17:15.
- [Lin and Chang, 2025] Lin, J. and Chang, W. (2025). Effectiveness of serious games as digital therapeutics for enhancing the abilities of children with attention-deficit/hyperactivity disorder (adhd): Systematic literature review. *JMIR Serious Games*, 13:e60937.
- [Liu et al., 2023] Liu, Q., Yang, L., Zhang, Z., Yang, H., Zhang, Y., and Wu, J. (2023). The feature, performance, and prospect of advanced electrodes for electroencephalogram. *Biosensors*, 13(1):101.
- [Lorini et al., 2022] Lorini, E., Sabouret, N., Ravenet, B., Davila, J. L. F., and Clavel, C. (2022). Cognitive planning in motivational interviewing. In *Proceedings of the 14th International Conference on Agents and Artificial Intelligence (ICAART 2022)*, pages 1–11, Online, Portugal. INSTICC: Institute for Systems and Technologies of Information, Control and Communication.
- [Niso et al., 2023] Niso, G., Romero, E., Moreau, J. T., Araujo, A., and Krol, L. R. (2023). Wireless eeg: A survey of systems and studies. *NeuroImage*, 269:119774.
- [OpenBCI, 2024] OpenBCI (2024). Cyton daisy biosensing boards (16-channel). Accessed: 2024-12-16.
- [Patiño et al., 2025] Patiño, J., Vega, I., Becerra, M. A., Duque-Grisales, E., and Jimenez, L. (2025). Integration between serious games and eeg signals: A systematic review. *Applied Sciences*, 15(4):1946.
- [Peksa and Mamchur, 2023] Peksa, J. and Mamchur, D. (2023). State-of-the-art on brain-computer interface technology. *Sensors*, 23(13):6001.
- [Pineda et al., 2003] Pineda, D. A., Lopera, F., Palacio, J. D., Ramirez, D., and Henao, G. C. (2003). Prevalence estimations of attention-deficit/hyperactivity disorder: differential diagnoses and comorbidities in a colombian sample. *International Journal of Neuroscience*, 113(1):49–71.
- [Pinho et al., 2014] Pinho, F., Correia, J. H., Sousa, N. J., Cerqueira, J. J., and Dias, N. (2014). Wireless and wearable eeg acquisition platform for ambulatory monitoring. In *2014 IEEE 3rd International Conference on Serious Games and Applications for Health (SeGAH)*, pages 1–7. IEEE.
- [Products, 2024] Products, B. (2024). Acticap. Accessed: 2024-12-16.
- [Raiker et al., 2017] Raiker, J. S., Freeman, A. J., Perez-Algorta, G., Frazier, T. W., Findling, R. L., and Youngstrom, E. A. (2017). Accuracy of achenbach scales in the screening of attention-deficit/hyperactivity disorder in a community mental health clinic. *Journal of the American Academy of Child & Adolescent Psychiatry*, 56(5):401–409.

## REFERENCES

- [Razavi et al., 2022] Razavi, M., Janfaza, V., Yamauchi, T., Leontyev, A., Longmire-Monford, S., and Orr, J. (2022). Opensync: an open-source platform for synchronizing multiple measures in neuroscience experiments. *Journal of neuroscience methods*, 369:109458.
- [Rodríguez Timaná et al., 2024] Rodríguez Timaná, L., Castillo García, J., Bastos Filho, T., Ocampo González, A., Hincapié Monsalve, N., and Valencia Jimenez, N. (2024). Use of serious games in interventions of executive functions in neurodiverse children: Systematic review. *JMIR Serious Games*, 12:e59053.
- [Salari et al., 2023] Salari, N., Ghasemi, H., Abdoli, N., Rahmani, A., Shiri, M. H., Hasheidian, A. H., Akbari, H., and Mohammadi, M. (2023). The global prevalence of adhd in children and adolescents: a systematic review and meta-analysis. *Italian journal of pediatrics*, 49(1):48.
- [Sandstrak, 2024] Sandstrak, G. (2024). A method for synchronized use of eeg and eye tracking in fully immersive vr. *Frontiers in Human Neuroscience*.
- [Takahashi et al., 2024] Takahashi, N., Ono, T., Omori, Y., Iizumi, M., Kato, H., Kasuno, S., and Tsuchiya, K. J. (2024). Assessment of executive functions using a 3d-video game in children and adolescents with adhd. *Frontiers in Psychiatry*, 15:1407703.
- [Totev et al., 2023] Totev, T., Taskov, T., and Dushanova, J. (2023). A wireless eeg system for neurofeedback training. *Applied Sciences*, 13(1):96.
- [Värbu et al., 2022] Värbu, K., Muhammad, N., and Muhammad, Y. (2022). Past, present, and future of eeg-based bci applications. *Sensors (Basel)*, 22(9):3331.
- [Wang et al., 2020] Wang, X., Hersche, M., Tomekce, B., Kaya, B., Magno, M., and Benini, L. (2020). An accurate eegnet-based motor-imagery brain–computer interface for low-power edge computing. In *2020 IEEE International Symposium on Medical Measurements and Applications (MeMeA)*. IEEE.
- [Wikström et al., 2022] Wikström, V., Saarikivi, K., Falcon, M., Makkonen, T., Martikainen, S., Putkinen, V., Cowley, B. U., and Tervaniemi, M. (2022). Inter-brain synchronization occurs without physical co-presence during cooperative online gaming. *Neuropsychologia*, 174:108316.
- [Xu and Zhong, 2018] Xu, J. and Zhong, B. (2018). Review on portable eeg technology in educational research. *Computers in Human Behavior*, 81:340–349.
- [Yadav and Maini, 2023] Yadav, H. and Maini, S. (2023). Electroencephalogram based brain–computer interface: Applications, challenges, and opportunities. *Multimedia Tools and Applications*, 82:47003–47047.
- [Yang and Wang, 2025] Yang, L. and Wang, Z. (2025). Applications and advances of combined fmri-fnirs techniques in brain functional research. *Frontiers in Neurology*, 16:1542075.

## REFERENCES

---

- [Zheng et al., 2023] Zheng, J., Li, Y., Zhai, Y., Zhang, N., Yu, H., Tang, C., Yan, Z., Luo, E., and Xie, K. (2023). Effects of sampling rate on multiscale entropy of electroencephalogram time series. *Biocybernetics and Biomedical Engineering*, 43(1):233–245.
- [Zou et al., 2022] Zou, B., Zheng, Y., Shen, M., Luo, Y., Li, L., and Zhang, L. (2022). Beats: An open-source, high-precision, multi-channel eeg acquisition tool system. *IEEE Transactions on Biomedical Circuits and Systems*, 16(6):1287–1298.

---

## REFERENCES