


COMPREHENSIVE EYE TEST REPORT

Patient Information

Name:	Niharika
Date of Birth:	2025-06-18
Patient ID:	4


Preliminary Test Results



Preliminary Analysis Report

Parameter	Value
HER	1.581
VER	1.172
Horizontal Squint	N/A
Vertical Squint	No

Hirschberg Test Results



Detailed Analysis Report

Measurement	Value
Left Vertical (LV)	6.08px
Left Horizontal (LH)	8.06px
Right Vertical (RV)	6.00px
Right Horizontal (RH)	6.08px
Left Iris White Spot	0.00 px
Right Iris White Spot	0.00 px
Right Vertical Distance (RVD)	N/A
Right Horizontal Distance (RHD)	N/A
Left Vertical Distance (LVD)	N/A

Left Horizontal Distance (LHD)	N/A
--------------------------------	-----

Additional Notes

Hirschberg Analysis Report
Input Image Path: hirschberg_Niharika_4_2025-06-18_2026-03-20.jpg
Reflex Light Not Detected

Gaze Test Results



9 Gaze Analysis Report

Gaze Position	Areal Ratio
gaze_1.jpg	0.475
gaze_2.jpg	0.915
gaze_3.jpg	1.714
gaze_4.jpg	0.807
gaze_5.jpg	1.029
gaze_6.jpg	1.208
gaze_7.jpg	0.765
gaze_8.jpg	0.963
gaze_9.jpg	1.006

Grade Summary Report

Image: gaze_1.jpg

Grades

L_Lateral	-4
R_Medial	-4

Distances

R_Lateral	28.16
R_Medial	90.80
R_Upper	26.42
R_Lower	46.01
L_Lateral	51.01
L_Medial	61.27
L_Upper	20.88
L_Lower	30.27

Image: gaze_2.jpg

No grade data available

Distances

R_Lateral	42.95
R_Medial	78.59
R_Upper	20.59
R_Lower	38.28
L_Lateral	50.01
L_Medial	58.69
L_Upper	19.92
L_Lower	27.02

Image: gaze_3.jpg

Grades

R_Lateral	-3
-----------	----

L_Medial	-4
----------	----

Distances

R_Lateral	59.30
R_Medial	51.43
R_Upper	20.52
R_Lower	28.28
L_Lateral	23.32
L_Medial	74.33
L_Upper	23.60
L_Lower	35.47

Image: gaze_4.jpg

Grades

L_Lateral	-4
R_Medial	-4

Distances

R_Lateral	23.19
R_Medial	82.57
R_Upper	29.07
R_Lower	37.54
L_Lateral	60.13
L_Medial	43.38
L_Upper	23.43
L_Lower	26.42

Image: gaze_5.jpg

No grade data available

Distances

R_Lateral	37.05
R_Medial	74.46
R_Upper	25.08
R_Lower	32.45

L_Lateral	52.04
L_Medial	61.98
L_Upper	18.38
L_Lower	27.02

Image: gaze_6.jpg

Grades

R_Lateral	-3
L_Medial	-4

Distances

R_Lateral	55.08
R_Medial	46.96
R_Upper	21.26
R_Lower	26.57
L_Lateral	26.31
L_Medial	78.85
L_Upper	26.25
L_Lower	34.66

Image: gaze_7.jpg

Grades

L_Lateral	-4
R_Medial	-4

Distances

R_Lateral	37.34
R_Medial	55.32
R_Upper	19.42
R_Lower	20.12
L_Lateral	50.01
L_Medial	37.11
L_Upper	21.93
L_Lower	21.95

Image: gaze_8.jpg

No grade data available

Distances

R_Lateral	40.02
R_Medial	47.27
R_Upper	17.12
R_Lower	13.60
L_Lateral	38.33
L_Medial	35.36
L_Upper	16.16
L_Lower	13.00

Image: gaze_9.jpg**Grades**

R_Lateral	-3
L_Medial	-4

Distances

R_Lateral	47.00
R_Medial	38.83
R_Upper	14.14
R_Lower	18.03
L_Lateral	29.15
L_Medial	49.82
L_Upper	15.23
L_Lower	18.87

Strabismus Pattern Analysis

To determine pattern strabismus, we compare horizontal deviations across gaze directions:

Deviation is calculated for each gaze as:

$(\text{Right Lateral} - \text{Right Medial}) + (\text{Left Lateral} - \text{Left Medial})$

We average the deviations from:

- Upgaze: gaze_1.jpg, gaze_2.jpg, gaze_3.jpg
- Downgaze: gaze_7.jpg, gaze_8.jpg, gaze_9.jpg
- Primary Gaze: gaze_5.jpg

The difference between Upgaze and Downgaze helps detect patterns:

- If Up - Down ≤ -10 : A-pattern
- If Up - Down ≥ 15 : V-pattern
- If both Upgaze and Downgaze deviations are significantly greater than Primary gaze is X-pattern
- If Upgaze deviation $>$ Primary $>$ Downgaze is Y-pattern
- If Downgaze deviation $>$ Primary $>$ Upgaze is Reverse Y-pattern
- Else: No pattern detected

Calculated horizontal deviation values:

Average Upgaze Deviation: -53.45

Average Downgaze Deviation: -7.29

Primary Gaze Deviation: -47.35

Deviation Difference (Up - Down): -46.17

Pattern Result: V-pattern detected

Strabismus Pattern Analysis

To determine Strabismus pattern strabismus, we compare horizontal deviations across gaze directions:

Deviation is calculated for each gaze as:

$(\text{Right Lateral} - \text{Right Medial}) + (\text{Left Lateral} - \text{Left Medial})$

We average the deviations from:

- Upgaze (gaze_1.jpg, gaze_2.jpg, gaze_3.jpg):

Values: -72.90 -44.32 -43.14

Average: -53.45

- Downgaze (gaze_7.jpg, gaze_8.jpg, gaze_9.jpg):

Values: -5.08 -4.28 -12.50

Average: -7.29

- Primary Gaze (gaze_5.jpg):

Value: -47.35


The difference between Upgaze and Downgaze helps detect patterns:

- If Up - Down ≤ -10 : A-pattern
- If Up - Down ≥ 15 : V-pattern
- If both Upgaze and Downgaze deviations are significantly greater than Primary gaze is X-pattern
- If Upgaze deviation $>$ Primary $>$ Downgaze is Y-pattern
- If Downgaze deviation $>$ Primary $>$ Upgaze is Reverse Y-pattern
- Else: No pattern detected

Calculated Difference (Up - Down): -46.17

Final Pattern Result: V-pattern detected

Deviation Visualization



End of Report - Generated by Eye Test Analysis System