

PATIENT NAME test bbook
PATIENT ID 1002
DATE OF BIRTH 19_01_2000
EXAM DATE 09-07-2026

REFERRING PHYSICIAN sathya
EXAMINING TECH. vijay
DEVICE EyeYantra v1.0
SESSION #33

1 Preliminary Screening


CAPTURED - CLEAR

Basic eye-alignment screening - corneal landmark detection (RT/RR, LT/LR)

LANDMARK	REGION / DESCRIPTION	STATUS
RT / RR	Right eye - temporal / pupil reflex	Within normal limits
LT / LR	Left eye - temporal / pupil reflex	Within normal limits

Preliminary Clinical Impression:

No gross misalignment detected at preliminary screening. Proceed to Hirschberg and 9-Gaze testing for confirmation.



2 Hirschberg Corneal Reflex Test

CAPTURED - CLEAR

Corneal light-reflex position relative to pupil centre, both eyes

EYE	REFLEX POSITION	ESTIMATED DEVIATION
Left	Slightly nasal	-5.5°
Right	Slightly nasal	-5.4°

Reflex Localization Clinical Impression:

No gross misalignment detected. Normal corneal light reflex.



3 Nine-Position Gaze Test

DEVIATION NOTED

Ocular rotation assessed across all nine cardinal gaze positions


POS.	DIRECTION	R_LATERAL	R_MEDIAL	L_LATERAL	L_MEDIAL
1	Up-Left	-	-	-	-
2	Up	-	-	-	-
3	Up-Right	-	-	-	-
4	Left	-	-	-	-
5	Primary	-	-	-	-
6	Right	-	-	-	-
7	Down-Left	-	-	-	-
8	Down	-	-	-	-
9	Down-Right	-	-	-	-

4 Strabismus Pattern Analysis

PENDING

Horizontal deviation compared across upgaze and downgaze positions

0.00	AVG. UPGAZE DEVIATION
0.00	AVG. DOWNGAZE DEVIATION
0.00	PRIMARY GAZE DEVIATION
0.00	UP-DOWN DIFFERENCE



No significant horizontal pattern deviation detected across vertical gaze shifts.

5 Clinical Impression

Automated eye test analysis reports a 'No significant pattern' based on vertical gaze differences. Corneal light reflex landmark positions and reflex deviations measured on Hirschberg test confirm alignment details. Findings are within the range typically followed up with comprehensive orthoptic examination. This report is generated by the EyeYantra automated ocular-alignment analysis system and is intended to support -- not replace -- clinical evaluation by a qualified ophthalmologist.

vijay

Technician

sathya

Ophthalmologist


WARNING: This is an automated screening report. Clinicians must verify the annotated landmarks prior to making clinical decisions.

6 Preliminary Test -- Detailed Values

MEASURED

Corneal landmark numerical measurements

Parameter	Value
Horizontal Eye Ratio (HER)	1.141
Vertical Eye Ratio (VER)	1.044
Horizontal Squint	N/A
Vertical Squint	N/A



7 Hirschberg Test -- Detailed Measurements

MEASURED

Corneal reflex coordinates and iris distances

Measurement	Value
Left Vertical (LV)	8.00 px
Left Horizontal (LH)	11.00 px
Right Vertical (RV)	8.06 px
Right Horizontal (RH)	11.05 px
Left Iris White Spot	7.21 px
Right Iris White Spot	6.40 px
Right Vertical Dist. (RVD)	9.303 deg
Right Horizontal Dist. (RHD)	5.432 deg
Left Vertical Dist. (LVD)	11.250 deg
Left Horizontal Dist. (LHD)	5.455 deg

Additional Notes

Hirschberg Analysis Report
Input Image Path: hirschberg_test bbook_1002_19_01_2000_2026-07-09.jpg
No Strabismus Detected

8 9-Gaze Areal Ratios

MEASURED

Iris-scleral areal ratio per gaze position

No areal ratio data available

10 Pattern Analysis Methodology & Raw Values

ANALYSIS

Detailed breakdown of the pattern detection algorithm

Deviation Formula

Deviation = (R_Lateral - R_Medial) + (L_Lateral - L_Medial)

Averaged across gaze groups:

Upgaze: gaze_1, gaze_2, gaze_3

Downgaze: gaze_7, gaze_8, gaze_9

Primary: gaze_5

Pattern Detection Criteria

Diff >= 5 and Upgaze > Downgaze: V-pattern

Diff >= 5 and Upgaze < Downgaze: A-pattern

Upgaze > Primary > Downgaze: Y-pattern

Downgaze > Primary > Upgaze: Reverse Y-pattern

Otherwise: No significant pattern

Raw Per-Gaze Deviation Values

Upgaze: Avg: 0.00

Downgaze: Avg: 0.00

Primary: gaze_5: 0.00 Value: 0.00

Difference: Upgaze - Downgaze = 0.00

Final Pattern Result: No significant pattern