

PATIENT NAME Mounika
PATIENT ID 63773
DATE OF BIRTH 01/01/2006
EXAM DATE 07-07-2026

REFERRING PHYSICIAN/DOCTOR
EXAMINING TECH. vijay
DEVICE EyeYantra v1.0
SESSION #33

1 Preliminary Screening

CAPTURED - CLEAR

Basic eye-alignment screening - corneal landmark detection (RT/RR, LT/LR)

LANDMARK	REGION / DESCRIPTION	STATUS
RT / RR	Right eye - temporal / pupil reflex	Within normal limits
LT / LR	Left eye - temporal / pupil reflex	Review Required

Preliminary Clinical Impression:

Horizontal squint (Exotropia) detected. Proceed to Hirschberg and 9-Gaze testing for confirmation.



2 Hirschberg Corneal Reflex Test


REVIEW REQUIRED

Corneal light-reflex position relative to pupil centre, both eyes

EYE	REFLEX POSITION	ESTIMATED DEVIATION
Left	Within normal limits	0°
Right	Slightly nasal	-33.5°

Reflex Localization Clinical Impression:

Corneal reflex displacement noted. Further clinical correlation recommended.



3 Nine-Position Gaze Test

Ocular rotation assessed across all nine cardinal gaze positions

POS.	DIRECTION	R_LATERAL	R_MEDIAL	L_LATERAL	L_MEDIAL
1	Up-Left	-	-4	-4	-
2	Up	-	-	-	-
3	Up-Right	-2	-	-	-4
4	Left	-	-4	-4	-
5	Primary	-	-	-	-
6	Right	-3	-	-	-4
7	Down-Left	-	-4	-4	-
8	Down	-	-	-	-
9	Down-Right	-2	-	-	-4

DEVIATION NOTED




4 Strabismus Pattern Analysis

Horizontal deviation compared across upgaze and downgaze positions

43.55	AVG. UPGAZE DEVIATION
36.37	AVG. DOWNGAZE DEVIATION
53.53	PRIMARY GAZE DEVIATION
7.18	UP-DOWN DIFFERENCE

V-PATTERN DETECTED



Deviation increases in upgaze relative to downgaze beyond the primary-gaze baseline, consistent with a V-pattern horizontal deviation. Clinical correlation advised.

5 Clinical Impression

Automated eye test analysis reports a 'V-pattern detected' based on vertical gaze differences. Corneal light reflex landmark positions and reflex deviations measured on Hirschberg test confirm alignment details. Findings are within the range typically followed up with comprehensive orthoptic examination. This report is generated by the EyeYantra automated ocular-alignment analysis system and is intended to support -- not replace -- clinical evaluation by a qualified ophthalmologist.

vijay
Technician

DOCTOR
Ophthalmologist

WARNING: This is an automated screening report. Clinicians must verify the annotated landmarks prior to making clinical decisions.