

PATIENT NAME Priyanka_document
PATIENT ID 488e
DATE OF BIRTH 31_01_2000
EXAM DATE 07-07-2026

REFERRING PHYSICIAN DOCTOR
EXAMINING TECH. vijay
DEVICE EyeYantra v1.0
SESSION #33

1 Preliminary Screening

CAPTURED - CLEAR

Basic eye-alignment screening - corneal landmark detection (RT/RR, LT/LR)

LANDMARK	REGION / DESCRIPTION	STATUS
RT / RR	Right eye - temporal / pupil reflex	Within normal limits
LT / LR	Left eye - temporal / pupil reflex	Within normal limits

Preliminary Clinical Impression:

No gross misalignment detected at preliminary screening. Proceed to Hirschberg and 9-Gaze testing for confirmation.



2 Hirschberg Corneal Reflex Test

CAPTURED - CLEAR

Corneal light-reflex position relative to pupil centre, both eyes

EYE	REFLEX POSITION	ESTIMATED DEVIATION
Left	Slightly nasal	-20.5°
Right	Within normal limits	0°

Reflex Localization Clinical Impression:

No gross misalignment detected. Normal corneal light reflex.



3 Nine-Position Gaze Test

Ocular rotation assessed across all nine cardinal gaze positions

POS.	DIRECTION	R_LATERAL	R_MEDIAL	L_LATERAL	L_MEDIAL
1	Up-Left	-	-4	-4	-
2	Up	-	-	-	-
3	Up-Right	-4	-	-	-4
4	Left	-	-4	-4	-
5	Primary	-	-	-	-
6	Right	-4	-	-	-4
7	Down-Left	-	-4	-4	-
8	Down	-	-	-	-
9	Down-Right	-3	-	-	-4

DEVIATION NOTED




4 Strabismus Pattern Analysis

Horizontal deviation compared across upgaze and downgaze positions

22.25	AVG. UPGAZE DEVIATION
45.05	AVG. DOWNGAZE DEVIATION
20.81	PRIMARY GAZE DEVIATION
-22.80	UP-DOWN DIFFERENCE

A-PATTERN DETECTED



Deviation increases in downgaze relative to upgaze, consistent with an A-pattern horizontal deviation. Clinical correlation advised.

5 Clinical Impression

Automated eye test analysis reports a 'A-pattern detected' based on vertical gaze differences. Corneal light reflex landmark positions and reflex deviations measured on Hirschberg test confirm alignment details. Findings are within the range typically followed up with comprehensive orthoptic examination. This report is generated by the EyeYantra automated ocular-alignment analysis system and is intended to support -- not replace -- clinical evaluation by a qualified ophthalmologist.

vijay
Technician

DOCTOR
Ophthalmologist


WARNING: This is an automated screening report. Clinicians must verify the annotated landmarks prior to making clinical decisions.

6 Preliminary Test -- Detailed Values

MEASURED

Corneal landmark numerical measurements

Parameter	Value
Horizontal Eye Ratio (HER)	1.147
Vertical Eye Ratio (VER)	1.29
Horizontal Squint	N/A
Vertical Squint	N/A



7 Hirschberg Test -- Detailed Measurements

MEASURED

Corneal reflex coordinates and iris distances

Measurement	Value
Left Vertical (LV)	7.07 px
Left Horizontal (LH)	8.06 px
Right Vertical (RV)	7.07 px
Right Horizontal (RH)	9.06 px
Left Iris White Spot	12.53 px
Right Iris White Spot	0.00 px
Right Vertical Dist. (RVD)	0.000 deg
Right Horizontal Dist. (RHD)	0.000 deg
Left Vertical Dist. (LVD)	12.728 deg
Left Horizontal Dist. (LHD)	20.466 deg

Additional Notes
Hirschberg Analysis Report
Input Image Path: hirschberg_Priyanka_document_488e_31_01_2000_2026-07-07.jpg
No Strabismus Detected

8 9-Gaze Areal Ratios

MEASURED

Iris-scleral areal ratio per gaze position

Gaze Position	Areal Ratio
gaze_1.jpg	0.840
gaze_2.jpg	0.917
gaze_3.jpg	0.878
gaze_4.jpg	1.041
gaze_5.jpg	1.019
gaze_6.jpg	1.084
gaze_7.jpg	0.667
gaze_8.jpg	1.152
gaze_9.jpg	1.444



9 Per-Gaze Grade & Distance Breakdown

DETAILED

Grades indicate EOM involvement; distances in pixels

gaze_1.jpg -- Up-Left

Grades:	L_Lateral: -4 R_Medial: -4		
R_Lateral	50.22	L_Lateral	103.81
R_Medial	104.68	L_Medial	71.18
R_Upper	34.99	L_Upper	41.40
R_Lower	48.55	L_Lower	45.88

gaze_2.jpg -- Up

No significant grade deviation

R_Lateral	54.34	L_Lateral	63.39
R_Medial	75.43	L_Medial	67.91
R_Upper	16.03	L_Upper	21.21
R_Lower	29.61	L_Lower	29.61

gaze_3.jpg -- Up-Right

Grades:	R_Lateral: -4 L_Medial: -4		
R_Lateral	99.25	L_Lateral	49.24
R_Medial	56.80	L_Medial	111.00
R_Upper	42.72	L_Upper	35.00
R_Lower	45.49	L_Lower	51.40

gaze_4.jpg -- Left

Grades:	L_Lateral: -4 R_Medial: -4		
R_Lateral	52.61	L_Lateral	110.45
R_Medial	108.26	L_Medial	52.35
R_Upper	32.76	L_Upper	54.20
R_Lower	45.34	L_Lower	47.01

gaze_5.jpg -- Primary

No significant grade deviation

R_Lateral	54.59	L_Lateral	57.00
R_Medial	65.92	L_Medial	66.48
R_Upper	17.89	L_Upper	19.42
R_Lower	27.29	L_Lower	28.02

gaze_6.jpg -- Right

Grades:	R_Lateral: -4 L_Medial: -4		
R_Lateral	107.52	L_Lateral	52.04
R_Medial	41.48	L_Medial	94.64
R_Upper	42.72	L_Upper	19.70
R_Lower	43.14	L_Lower	33.11

gaze_7.jpg -- Down-Left

Grades:	L_Lateral: -4 R_Medial: -4		
R_Lateral	35.44	L_Lateral	76.11
R_Medial	109.07	L_Medial	50.22
R_Upper	32.56	L_Upper	35.47
R_Lower	48.85	L_Lower	30.00

gaze_8.jpg -- Down

No significant grade deviation

R_Lateral	61.85	L_Lateral	67.08
R_Medial	101.24	L_Medial	72.69
R_Upper	32.57	L_Upper	32.02
R_Lower	34.06	L_Lower	22.56

gaze_9.jpg -- Down-Right

Grades: R_Lateral: -3 | L_Medial: -4

R_Lateral	77.41	L_Lateral	43.68
R_Medial	80.01	L_Medial	83.49
R_Upper	27.73	L_Upper	25.30
R_Lower	23.09	L_Lower	25.24

10 Pattern Analysis Methodology & Raw Values

ANALYSIS

Detailed breakdown of the pattern detection algorithm

Deviation Formula

Deviation = (R_Lateral - R_Medial) + (L_Lateral - L_Medial)

Averaged across gaze groups:

Upgaze: gaze_1, gaze_2, gaze_3

Downgaze: gaze_7, gaze_8, gaze_9

Primary: gaze_5

Pattern Detection Criteria

Diff >= 5 and Upgaze > Downgaze: V-pattern

Diff >= 5 and Upgaze < Downgaze: A-pattern

Upgaze > Primary > Downgaze: Y-pattern

Downgaze > Primary > Upgaze: Reverse Y-pattern

Otherwise: No significant pattern

Raw Per-Gaze Deviation Values

Upgaze:	gaze_1: -21.83 gaze_2: -25.61 gaze_3: -19.31	Avg: 22.25
Downgaze:	gaze_7: -47.74 gaze_8: -45.00 gaze_9: -42.41	Avg: 45.05
Primary:	gaze_5: -20.81	Value: 20.81
Difference:	Upgaze - Downgaze = -22.80	

Final Pattern Result: A-pattern detected