Same patient, different number

Work in pairs. Where you see ✎ **Predict**, write your answer **before** you click. The task numbers match the tasks on the web pages.

Names _____ Specialty _____

patrikbrynolfsson.github.io/estro2026-dwi-handson

A Same brain, different ADC

ADC & IVIM sandbox · a real brain scan, 21 b-values

TAB 1 · SAME BRAIN, DIFFERENT ADC

1 Move the b-value slider from 0 to 1000. Which parts of the brain lose the most signal between b = 0 and b = 100?

_____

2 Brain tissue: read its ADC with each choice of b-values.

b 0 & 800 _____ b 200 & 800 _____ × 10^{-3} mm^2/s

3 ✎ **Predict** Which region has the biggest difference between the two ADCs?

☐ Brain tissue ☐ CSF ☐ Vessels + CSF

Then check (vs b 200 & 800): brain _____ % CSF _____ % vessels + CSF _____ %

4 A change in brain ADC is real only above **11 %** (QIBA). The first scan's ADC used b = 0, the follow-up's did not. How much of the 11 % does that alone account for? And if both scans used b = 0?

Only one used b = 0: _____ % of 11 % Both used b = 0: _____

TAB 2 · TREATMENT (a simulated tumour: blood fraction f falls from 0.25 to 0.08)

1 ✎ **Predict** Fill in the four ADC values before you press **Show the answers**. Then write the real ones.

	Before (predict)	After (predict)	Before (real)	After (real)	Change
b 0 & 800					%
b 200 & 800					%

2–3 Which b-values see a change? Is it bigger than the 11 % limit? Would you call it a response?

☐ Response ☐ No response ☐ Depends on the b-values

_____

4 Look at the b 50 & 800 row. Is a lowest b-value of 50 enough?

_____

TAB 3 · NOISE

1–2 ✎ **Predict** Which IVIM map shows clear anatomy, and which looks like noise?

Clear: ☐ D ☐ f ☐ D* Noise: ☐ D ☐ f ☐ D* Right? ☐ Yes ☐ No

3 CSF behind the midbrain: f = _____ Does that make sense? _____

4–5 Scan the same voxel 10 times. Which stays close to its true value?

SNR 50: ☐ D ☐ f ☐ D* SNR 20: ☐ D ☐ f ☐ D*

Brain data: E. Peterson, IVIM dataset, Figshare 2016, doi:10.6084/m9.figshare.3395704 (CC0). True-change limits: QIBA DWI Profile, Stage 3, 2022 (brain 11 %, prostate 47 %). Lowest b-value: Kooreman et al., Radiother Oncol 2020.

B Has the tumour responded?

Prostate, two scans 13 days apart

2 Expert ROI drawn on ADC: mean ADC on each scan.

Scan 1 _____ Scan 2 _____ $\times 10^{-3}$ mm²/s Change _____ %

3 Change the ROI. How much does the value move **within scan 1**?

ROI on scan 1	Mean	Median	10th percentile
Expert, on ADC			
Grown 2 mm			
Shrunk 2 mm			
Expert, on T2w			

Moves least: ☐ Mean ☐ Median ☐ 10th percentile

4 ✎ **Predict** Scan 1 was before treatment, scan 2 two weeks in. Has the tumour responded? Why?

☐ Yes ☐ No ☐ Cannot tell

_____

5 Does your conclusion depend on which ROI you used?

_____

C Contouring on DWI

Another man from the same study · numbers = the four steps

1 Your outline on T2w:

Volume _____ cm³

2 ✎ **Predict** On DWI it will be

☐ bigger ☐ same ☐ smaller

Your DWI volume _____ cm³

3 Compare your two outlines, then flick T2w ↔ DWI and measure the shift at the back of the prostate with the ruler.

Prediction right? ☐ Yes ☐ No Overlap (Dice) _____ Centre shift _____ mm

Ruler _____ mm

4 The radiologist: overlap (Dice) _____ DWI outline further back by _____ mm

DISCUSS

1 Would you use the DWI outline as the GTV, as a boost volume (as in FLAME), or only as a guide when you contour on T2w?

☐ GTV ☐ Boost volume ☐ Guide only

2 Margin that would cover the shift you measured: _____ mm

3 Would distortion correction or registration remove the shift? What would you check?

_____

Your checklist

before you trust an ADC · fill in at the debrief

1 · Which b-values? _____

2 · Whose ROI? _____

3 · Bigger than the repeatability? _____

4 · Geometrically trustworthy? _____

Prostate data: QIN-PROSTATE-Repeatability, TCIA, doi:10.7937/K9/TCIA.2018.MR1CKGND (CC BY 4.0); Fedorov et al., Sci Data 2018;5:180281 and Invest Radiol 2017;52:538-46 (tumour ADC limit 42 %; volume 71 % on T2w, 112 % on ADC). FLAME: Kerkmeijer et al., J Clin Oncol 2021;39:787-96.