TMS

Subject: S310luemp **DC:** Week 1 **Date:** 2022-10-11

Examiner/Recorder: AS/PS **Time:** 2:20 – 4:55

Bainsight project/session: S310luemp/week 1

Coil: Lt Blue Remote coil

TMS Target	MagSim Setting	Sample #s	Reps: +/- for MEP value above threshold
put on bicep + tricep electrodes			
TMS table at 10			
		red-light = lateral	
right Bicep m-max		m-max = 13.4	100mA, 600μS
			normal polarity
left Bicep m-max		m-max = ~7	use peak at
			100mA, 1000μS
left Bicep H-max		PW = 1000μS, stim amplitude	~ 60 - 80 mA
			55
3:10pm – 3:20 setup for TMS			
mVCs			
uninv_bicep			X ✓ ✓ 400
tricep			8
uninv_tricep			✓ ✓ ✓ 12,000
			BV sample #
inv-bicep			✓ ✓ ✓ ~40
inv_tricep			✓ ✓ ✓ ~80
Bistim - simultaneous discharge			
uninv- bicep hotspot			
sample 60	50		
9:00	50		22 – –
6:00	50		– – –

TMS Target	MagSim Setting	Sample #s	Reps: +/- for MEP value above threshold
3:00	50		— — — Sample 33
12:00			+ — — 33
Sample 60			— — + 46
Sample 60	51		37 8 22
	52		174 9 42 27
	53		100 53 20 34 105
12:00	53		29 30 25 4
9:00			11 10 Sample 54
6:00			214 34 91 213
Sample 62	53		137
	52		163 15 11 53 55
Stand + put his cushion on the chair			
raised table to 11			
foot rests under feet			
Inv_bicep hotspot			
Sample 83	80		
Sample 83	90		— —
active inv bicep goal = 40			
Sample 83	100		—
3:00	99		— — —
12:00			— — —
9:00			— —
			coil over heart → light blue coil

TMS Target	MagSim Setting	Sample #s	Reps: +/- for MEP value above threshold
Right grid 3,5	100		— —
look around			
sample 134			
look around			
uninv_bicep_rc		mvc	goal = 200
sample 60	50		x ✓ ✓ ✓ ✓ ✓
Bistim Simult discharge	60		✓ ✓ ✓ ✓ ✓
	40		✓ ✓ ✓ ✓ ✓
	30		✓ ✓ ✓ ✓ ✓
	70		✓ ✓ ✓ ✓ ✓
	80		✓ ✓ ✓ ✓ ✓
	90		✓ ✓ ✓ ✓ -
	45		✓ ✓ ✓ ✓ ✓
	65		✓ ✓ ✓ ✓ ✓
	55		✓ ✓ ✓ ✓ ✓ S50 = 58. Bistim
			goal = 40
inv_bicep_rc			
sample 134	100		✓ ✓ ✓ ✓ ✓
Bistim SD.	90		✓ ✓ ✓ ✓ ✓
	95		✓ ✓ ✓ ✓ ✓
	85		✓ ✓ ✓ ✓ ✓
	80		✓ ✓ ✓ ✓ ✓

TMS Target	MagSim Setting	Sample #s	Reps: +/- for MEP value above threshold
		Coil	overheat switch → dark blue coil
uninv_tricep_rc		goal = 300	
Sample 60			
	60		✓ ✓ ✓ ✓ ✓
	80		✓ ✓ ✓ ✓ ✓
	90		✓ ✓ ✓ ✓ ✓
	100		✓ ✓ ✓ ✓ ✓
	70		✓ ✓ ✓ ✓ ✓
	50		✓ ✓ ✓ ✓ ✓
	30		✓ ✓ ✓ ✓ ✓
	40		✓ ✓ ✓ ✓ ✓
	95		✓ ✓ ✓ ✓ ✓
	85		✓ ✓ ✓ ✓ ✓
inv_tricep_rc		goal = 80	
sample 134			
	100		✓ ✓ ✓ ✓ ✓
	95		✓ ✓ ✓ ✓ ✓
	80		✓ ✓ ✓ ✓ ✓
	85		✓ ✓ ✓ ✓ ✓
	90		✓ ✓ ✓ ✓ ✓
uninv_bicep_sici		goal = 200	
Sample 60			ignore 1st sample in Matlab + Brainvision
TS 0/73/2			~~✓~~ X ✓ → 20 more samples
sici 55/73/2			✓ → 20 samples
ICF 55/73/10			✓ → 20 samples

TMS Target	MagSim Setting	Sample #s	Reps: +/- for MEP value above threshold
inv_bicep_SICI		goal = 40	
sample134			
TS	0/100/2		✓ → 20 samples
SICI	75/100/2		✓ → 6 samples, coil overheat
~~ICF~~	~~75/100/10~~		change to light blue coil (heard 1 sample when connecting coil but didn't see any additional samples in Brain Vision or Matlab SICI window
	75/100/2		✓ → 14 more samples
ICF	75/100/10		✓ → 20 samples
4:55			