What the mockup must show: one-page digest of the specification

Sources: the shared document "V4: key specification and open questions" (text read on 29 October 2026): sections 3.2 to 3.8, table 3.3, questions Q1 to Q15; repository `PershinIlya/v4`: `intended-use.md`, ADR 0003 and 0004, questions G6, I6, I8, `briefs/deterministic-text-layer.md`. Statuses and table 3.3 values are the shared document's; the rest is a mockup assumption (`README.md`).

#	The mockup shows	Source, status
1	**Core analyses only.** Signal quality; absolute and relative spectra and maps per 1 Hz band against norms; raw EEG excerpts (display only); alpha peak frequency; theta/beta ratio and frontal alpha asymmetry index as numbers only; asymmetry map; coherence per pair and band; channel ERPs as waveforms and peak values (P1/N1, N2, P3, CNV; new in V4); VCPT omissions, commissions, mean reaction time and its standard deviation; errors per 40-trial block without norms; repeated measurements side by side.	3.2 Scope, Decided; table 3.3 Core
2	**"To confirm" in the core** (Q6), status only in annotations: global coherence; frequency markers per electrode with automatic peak suggestions; checklist over channel ERPs (neutral columns, M76). **Not shown:** later modules (questionnaires, in V3 until certified; workload, arousal and central sensitivity indices; vigilance; microstates; spike detection; the classifier; AI interpretation); research only (subtype and response classifiers, treatment recommendations); dropped (grouped summary table, bio signals, heart rate variability, external ERP program); not planned (change significance test). **Open, not shown by mockup choice:** lagged, instantaneous and total coherence (Q6); ICA component ERPs, ERPC2, ICA component maps and spectra as a result view (the cleaning shows maps only to choose artefacts, 3.4), sLORETA (Q8).	Table 3.3; Q6, Q8; 3.4
3	**No condition wording, no interpretation.** Ratios as numbers only; no labels naming a condition or a mental function ("attention", "impulsivity"); no "normal" or "abnormal".	3.2, Decided; `intended-use.md` §4
4	**Fixed sentences only**: (a) a value against the norm, (b1) condition-neutral explanation; one per value beyond a set threshold (mockup: per marked value; single values always restated with z, M6); a summary of the largest deviations with a completeness rule (Japan only if the certification body agrees); data quality and limits. No language model.	3.2, Decided; scheme to confirm by the clinical lead
5	**Multiple comparison correction**: a false discovery rate method or permutation tests, possibly per block, not Bonferroni; chosen at implementation. Mockup stand-in: M33.	3.2 Working
6	**Norms and intended use.** Existing normative recordings reprocessed with the V4 pipeline (mockup: invented "Mock norms", M2). Users: physicians, licensed psychologists or neuropsychologists; general clinical use, no condition named; age range from the usable norms.	3.2 Norms, Decided; Intended use, To confirm
7	**Recording systems**: the two amplifier families V3 reads today, each paired with its VCPT task software; others later through EDF+ (an open EEG file format), each a validated change. Task software named generically until the CTO confirms naming (M37, M38).	3.2 Working
8	**Browser only; clinician in control.** No install on the recording computer. The system proposes segments, frequency markers, peaks and the ICA components to remove (ICA in the core or the first module); the clinician confirms or corrects. Every choice is logged and shown in the report, markers and peaks included.	3.4 Working; I8
9	**Earlier measurement** uploaded again and reprocessed with the current software and norms; side by side, each against its own norm, descriptive only; the report says so. No V3 numbers.	3.4 Working
10	**Report from blocks**: system results (locked), the clinic's layout and texts, the clinician's anamnesis, comments after result sections, interpretation and conclusion (V4 report editor). Final when issued; a correction is a new version; software and norm versions on every PDF; only the PDF leaves.	3.5 Working; I6
11	**Clinic templates**: a preset in V4's template editor (result blocks and order, layout, the clinic's texts), not a Word file. Result blocks belong to the company; figure colours fixed. In the first release the company builds the templates on request; clinics edit their own later.	3.5 Working; I6
12	**Languages**: English and German first; Japanese when V4 enters Japan; others on request.	3.5 Working
13	**Privacy**: a separate, encrypted identity store; the analysis sees a code number, age and sex; names only for the treating clinic; company staff only with time-limited access the clinic grants; every access logged; consent per measurement for norms, research and model training; kept during the contract, deleted when the clinic asks.	3.6 Working
14	**Operations and quality targets**, only where the clinician meets them: the company's cloud first, data in Germany (own-server installation built and tested, offer open, Q10); processing within 10 minutes in 95% of cases; results reproducible, issued reports never change. Not shown: availability, growth, V3 comparison.	3.7, 3.8 Working

Open points the mockup shows but does not decide: severity scale and sentence threshold (G6); the correction method (3.2); ranking spectrum bins and global coherence in the summary, now only counted (M7); summary and VCPT values in Japan (3.2, table 3.3); Word export (Q14, PDF only in the first release); report service by company staff (Q9, not shown); claim words in clinic texts (I6).

Hard rules: synthetic data, "Mockup, synthetic data" on every page and screen; no real templates, clinic names, logos or texts; no condition names, interpretation or advice.