MS1 background: previous rule vs lowest 5%

NIST_POS_Samp_01-01.mzXML • scan 1823 • RT 299.839 s • m/z 100–250

Previous cutoff scale

Median = 226.7k; 29% of axis height

Before filtering · 148 centroids (new threshold overlaid)



Previous clipped rule · 21/148 kept



Lowest 5%, minimum 8 · 83/148 kept



New background zoom

Median = 28.2k; 29% of axis height

Before filtering · 148 centroids (new threshold overlaid)



Previous clipped rule · 21/148 kept



Lowest 5%, minimum 8 · 83/148 kept



Original centroids
Kept by previous rule
Additional under new rule
Subset/background mean μ
Active cutoff μ + 3σ

Identical raw intensities in both columns; shared axes within each column. ▲ marks peaks taller than the axis.

Full-scan estimates, 25 Da windows. Here supported windows use 8 weak neighbors. Removed ≠ confirmed noise.