Biological Indicator is designed for monitoring of ethylene oxide (E0) sterilization processes. The presence of Bacillus atrophaeus spores is detected by a visual color change . The final readout of a negative result is made after 24 hours of incubation. Must conform to ANSI/AAMI/EN/ISO 11140-1:2005. Conformance certificates to be provided by supplier Supplier must supply a loan incubator for use of this item
Paper, sterilization, crêped, white / Blue / Green Suitable for use as a wrapping material for sterilization of dressings, instruments, etc. Paper, mass not less than 57g/m2, moisture repellent and resistant to air and fluid borne barrier. Must comply with the latest issue of SANS 50868-2 and DIN 59953-6 Size: 1800mm x 1800mm
Wrap, Double, sterilization, medical, non-woven Spunbond-Meltblown-Spunbond (SMS) To comply with the latest issue of SANS 50868-2, DIN 58953-6, SANS 11607-1 and/or EN868-2 Regular set Size: ±600 x ±600mm
Wrap, sterilization, medical, non-woven Must be compatible to be use with steam, ethylene oxide, formaldehyde and Hydrogen peroxide gas sterilization processes Spunbond-Meltblown-Meltblown-Meltblown-Spunbond (SMMMS) To comply with the latest issue of SANS 50868-2 DIN 58953-6, SANS 11607-1 and/or EN868-2 Size: ±600 x 600mm
Wrap, sterilization, medical, non-woven Must be compatible to be use with steam, ethylene oxide, formaldehyde and Hydrogen peroxide gas sterilization processes Spunbond-Meltblown-Meltblown-Meltblown-Spunbond (SMMMS) To comply with the latest issue of SANS 50868-2 DIN 58953-6, SANS 11607-1 and/or EN868-2 Size: 900 x 900mm
Wrap, sterilization, medical, non-woven Must be compatible to be use with steam, ethylene oxide, formaldehyde and Hydrogen peroxide gas sterilization processes Spunbond-Meltblown-Meltblown-Meltblown-Spunbond (SMMMS) To comply with the latest issue of SANS 50868-2 DIN 58953-6, SANS 11607-1 and/or EN868-2 Size: 900 x 1200mm
Wrap, sterilization, medical, non-woven Must be compatible to be use with steam, ethylene oxide, formaldehyde and Hydrogen peroxide gas sterilization processes Spunbond-Meltblown-Meltblown-Meltblown-Spunbond (SMMMS) To comply with the latest issue of SANS 50868-2 DIN 58953-6, SANS 11607-1 and/or EN868-2 Size: 900 x 1500mm
* Wrap, sterilization, medical, non-woven Must be compatible to be use with steam, ethylene oxide, formaldehyde and Hydrogen peroxide gas sterilization processes Spunbond-Meltblown-Meltblown-Meltblown-Spunbond (SMMMS) To comply with the latest issue of SANS 50868-2 DIN 58953-6, SANS 11607-1 and/or EN868-2 Size: 1200 x 1200mm

Specification