



*Sensors* (Basel). 2015 Oct; 15(10): 25055–25071.

PMCID: PMC4634500

Published online 2015 Sep 29. doi: [10.3390/s151025055](https://doi.org/10.3390/s151025055)

PMID: [26426020](https://pubmed.ncbi.nlm.nih.gov/26426020/)

# Quantitative Assessment of Parkinsonian Tremor Based on an Inertial Measurement Unit

[Houde Dai](#),<sup>1,2,\*</sup> [Pengyue Zhang](#),<sup>3</sup> and [Tim C. Lueth](#)<sup>2</sup>

Vittorio M.N. Passaro, Academic Editor

<sup>1</sup>Quanzhou Institute of Equipment Manufacturing, Haixi Institutes, Chinese Academy of Sciences, Bolan Av., Jinjiang 362200, China

<sup>2</sup>Institute of Micro Technology and Medical Device Technology, Technische Universitaet Muenchen, 85748 Garching, Germany; E-Mail: [tim.lueth@tum.de](mailto:tim.lueth@tum.de)

<sup>3</sup>Medical Faculty, Kunming University of Science and Technology, Kunming 650500, China; E-Mail: [zpy19802000@163.com](mailto:zpy19802000@163.com)

\* Author to whom correspondence should be addressed; E-Mail: [dhd@fjirsm.ac.cn](mailto:dhd@fjirsm.ac.cn); Tel.: +86-595-3635-0373; Fax: +86-595-8266-8861.

Received 2015 Jul 1; Accepted 2015 Sep 14.

[Copyright](#) © 2015 by the authors; licensee MDPI, Basel, Switzerland.

This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution license (<http://creativecommons.org/licenses/by/4.0/>).

## Abstract

Quantitative assessment of parkinsonian tremor based on inertial sensors can provide reliable feedback on the effect of medication. In this regard, the features of parkinsonian tremor and its unique properties such as motor fluctuations and dyskinesia are taken into account. Least-square-estimation models are used to assess the severities of rest, postural, and action tremors. In addition, a time-frequency signal analysis algorithm for tremor state detection was also included in the tremor assessment method. This inertial sensor-based method was verified through comparison with an electromagnetic motion tracking system. Seven Parkinson's disease (PD) patients were tested using this tremor assessment system. The measured tremor amplitudes correlated well with the judgments of a neurologist ( $r = 0.98$ ). The systematic analysis of sensor-based tremor quantification and the corresponding experiments could be of great help in monitoring the severity of parkinsonian tremor.

**Keywords:** Inertial measurement unit, wearable system, tremor quantification, Parkinson's disease, technical validation

## 1. Introduction

It is estimated that 1% of 70-year-olds suffer from Parkinson's disease (PD) [1]. At present, about four to six million people are PD patients, in which about 10% of PD patients are younger than 50 years old [2]. Due to demographic increase of the elderly population, PD will occur more frequently in the future.

Parkinsonian tremor (oscillatory movement) is the central symptom of PD, presenting in about 70% of PD cases [3]. It can be judged in the patient's hands, feet, and head with a particular frequency (3.5–7.5 Hz) and amplitude (speed and range) [4]. Tremors often begin from one finger and expand to the whole arm, with the rhythmic back-and-forth action called the “pill-rolling” tremor [5]. The “pill-rolling” tremor takes its name from the action of rolling a pill with the thumb and index finger.

Parkinsonian tremor manifests in different types: rest, postural, and action tremors. Rest tremor, which is the characteristic of parkinsonian tremor, happens when a body part is relaxed. The postural tremor occurs while a body part is held straight out from the body in a stable position against gravity. The action tremor (kinetic tremor) happens when a voluntary contraction of a muscle follows a certain action—for example, holding a cup [6]. Pure rest tremor is rare. For a PD patient, postural or action tremors appear together with rest tremor, but with different frequencies. A rest tremor may be combined in a postural tremor, but disappears during an action tremor task. An action tremor occurs in most PD; however, the tremor severity is not associated with age or disease duration [7].

The precise mechanism of PD remains uncertain at present [8]. The Unified Parkinson's Disease Rating Scale (UPDRS), which is a subjective assessment performed by the qualitative judgment of neurologists, is the clinical standard for parkinsonian tremor assessment. UPDRS tremor ratings are from 0 to 4, which denote normal, slight, mild, moderate, and severe, respectively.

Spiral drawing tests by a parkinsonian patient are used to judge the severity of parkinsonian tremor [7]. However, there is no practical instrumental method for the accurate monitoring of the PD symptoms as of yet. Nevertheless, doctors and patients are eager for objective severity assessment of parkinsonian tremor [9].

In addition, about 50% of PD patients have side effects such as dyskinesia and motor fluctuations after five years of levodopa medication. The occurrence and amplitude of parkinsonian tremor are influenced by such effects [10,11].

Timmer et al. reported that parkinsonian tremor exhibits a two-order nonlinear and no-strictly periodic oscillation [12]. Because the dominant frequency of parkinsonian tremor has no direct relationship with tremor severity, the tremor severity can be rated only by tremor range (i.e., amplitude) [13].

To quantify neurological symptoms during deep-brain surgery, we have presented a wearable system for tremor quantitative assessment [14,15,16,17].

The goal of this study is to investigate the features of parkinsonian tremor and to develop a sensor-based quantitative assessment method. In this concept, the current possibilities of inertial sensor technology and motion-tracking algorithms can be used to implement quantitative assessments of parkinsonian tremor.

## 2. State of the Art and Task Description

---

### 2.1. State of the Art in Parkinsonian Tremor Quantification

Some research groups have used infrared cameras, video tracking, digital drawing tablets, laser-based displacement transducers, and electromyography (EMG) to assess tremors objectively [1,2,3,4,5,6,7,8,9,10]. Sensors were placed on the body, feet or arms of the patient. Most of these sensor-based systems could detect the change between therapy “OFF” and “ON” states during medication. However, because these systems have big dimensions or patients were uncomfortable wearing such devices, such systems have limited usability in clinical applications [7].

At present, most tremor assessment methods were based on inertial sensors and a computer-based system. There has recently been growing interest in the application of MEMS (Micro-Electro-Mechanical Systems) inertial sensors for continuous monitoring of PD symptoms. With the development of new MEMS technology, the dimension of the sensor circuit board is smaller and signal processing is easier to carry out than before [5].

Some studies have indicated that sensor data from accelerometers and gyroscopes correlate strongly with UPDRS tremor ratings [5,18,19,20,21,22].

The research carried out by Elble et al., which enrolled 928 patients, indicated that tremor amplitude is logarithmically correlated to the five-point UPDRS scale [20].

The Kinesia™ system (Great Lakes NeuroTechnologies Inc., Cleveland, OH, USA), which has been approved by the FDA (the United States Food and Drug Administration) for sale, is used to assess parkinsonian symptoms with an inertial measurement unit (IMU), which embeds a three-axis gyroscope and a three-axis accelerometer in a single chip, on the top side of a finger [20]. Giuffrida's research indicated that the logarithm of the peak powers' summation of power spectrums of both accelerometer and gyroscope signals, during rest and postural tremor tasks, correlated best with UPDRS tremor scale (the coefficient of determination  $r^2 = 0.9$ ) [23]. At the same time, the RMS summation of both accelerometer and gyroscope signals, during action tremor tasks, correlated best with UPDRS tremor ratings ( $r^2 = 0.69$ ) [21].

In addition, Niazmand et al. presented a glove system to quantitatively assess parkinsonian tremor based on accelerometers [18]. Salarian et al. presented an ambulatory tremor monitor with a three-axis gyroscope on the wrist [4]. The Motus Movement Monitor (Motus Bioengineering Inc., Benicia, CA, USA), based on a three-axis gyroscope on the top side of the patient's palm, is used to assess tremor and other symptoms [24]. Pierleoni et al. used a wrist module, consists of an accelerometer and a gyroscope, to classify and qualitatively assess tremors [25]. Khan et al. presented a wearable accelerometer system to classify tremors based on several algorithms, especially the non-linear radial basis function kernel [26]. Zhang et al. developed an accelerometer sensing system that adopted the short-time Fourier transform to reduce the instability of the tremor signals [27].

Beginning in August 2014, the Michael J. Fox Foundation for Parkinson's Research (MJFF) and Intel Corporation have joined forces to improve Parkinson's disease monitoring via big data analytics and data from wearable computing. Open-access sensor data (accelerometer, compass, etc.) are available now [28].

## 2.2. Special Concerns for Side Effects and Task Description

Motor fluctuations, without being considered by other tremor assessment systems, indicate that the effective period of certain doses is shorter all the time, with the name of end-of-dose deterioration. Motor fluctuations also represent the alternations between "ON" (a state of good response to anti-parkinsonian medications) and "OFF" (a state for patients experiencing parkinsonian symptoms). The symptoms of a PD patient may reappear unexpectedly and suddenly, a switch sensation described as "ON/OFF" syndrome. Levodopa-induced dyskinesia (LID) involves a series of hyperkinetic movements (involuntary, episodic, and irregular) such as athetosis, chorea, and dystonia [10]. The occurrence and severity of parkinsonian tremor are influenced by such side effects. Therefore, special concerns for MD and LID should be adopted in the tremor assessment.

The research conducted by Burkhard et al. shows that the RMS of the power spectrum of hand-attached gyroscope signals, from 0.25 to 3.25 Hz, during dyskinesia task correlated well with the five-point clinical ratings of dyskinesia severity [24]. The influence of dyskinesia during a tremor task can be


removed by applying a high-pass filter, with the cutoff frequency ( $f_c$ ) set at 3.25 Hz, to the measured IMU signals.

For a PD patient, only the tremor amplitude is required, because the specific tremor frequency does not change for a certain period. In this study, the logarithm of the peak powers' summation of the IMU sensor signals is regarded as the tremor amplitude. The tremor amplitudes, which are regressed from the peak powers of the IMU signals, can be gauged as the clinician ratings [29].

### 3. Quantitative Assessment of Parkinsonian Tremor

#### 3.1. Test Tasks and Relevant Parameters

According to the suggestions of neurologists, several tasks are chosen to assess the symptom severities of parkinsonian tremor. As shown in [Figure 1](#), tremor assessment includes three tasks: rest tremor, postural tremor, and action tremor. Each assessment task lasts ten seconds.




The parameters that should be extracted from sensor signal processing and displayed in the graphical user interface (GUI) are:

1. amplitude of parkinsonian tremor (R); and
2. dominant frequency of parkinsonian tremor (F).

The dominant frequency of parkinsonian tremor is only used as a reference parameter to judge whether it is the range of parkinsonian tremor (3.5–7.5 Hz).

#### 3.2. System Concept

A portable parkinsonian tremor assessment system was used in this study [14,15,16,17]. Its system diagram is shown in [Figure 2](#). The reason why use wired communication is for radio safety in the operating room. For home application, a wireless communication interface can be adopted.



[Figure 2](#)

System diagram of the tremor assessment system. This system consists of two parts: glove part (**left**) and computer part (**right**).

A six-axis IMU (MPU6050, InvenSense Inc., San Jose, CA, USA) is placed on the top of the index finger for tremor assessment. MPU6050 has very small dimensions (4 mm × 4 mm × 0.9 mm). A textile glove, incorporating a command module and an IMU circuit board, was placed on the top of middle index finger. The command module sent the sensor data to the computer via a wired USB cable.

The sampling rate of the tremor assessment system was working at 100 Hz. The sensor data from the command module are sent to the computer via a serial-to-USB transmission [30].

### 3.3. Parkinsonian Tremor Assessment Methods

The algorithm used to quantify the severity of parkinsonian tremor is very important. Tremor Quantification has been performed by signal processing methods such as time-domain analysis, time-frequency analysis, spectral analysis, and nonlinear analysis [31,32]. Currently, spectral analysis is used for the majority of these studies [33,34].

In this study, signal processing involves FIR (finite impulse response) filters and IIR (infinite impulse response) filters as well as other special algorithms such as power spectral density (PSD) analysis. To detect tremors, the signals are then processed with PSD calculation (auto power spectrum) with a certain time interval (3–10 s).

For range (displacement) analysis, the angular velocity obtained from the gyroscope needs to be integrated over time only once, but the integration of the acceleration signals requires double integration. However, the sensor bias and drifts are integrated as well. Thus, the IMU raw data were used for the tremor quantification.

The three-axis accelerometer outputs include gravitational acceleration, which equals  $9.81 \text{ m/s}^2$  in vector product, and linear acceleration. Gravitational acceleration can be removed from the accelerometer outputs with high-pass filters. Thus, there are only linear acceleration data for the following signal processing.

Quadratic mean (RMS) interprets actual vibration levels, while PSD results show dominant frequency that contribute the most to the tremor. Because the tremors are based on a dominant frequency, the advantages of PSD compared to a statistical measure (quadratic mean) are that it highlights the tremor signals from noise and other movement with analysis in the frequency dimension and the squared value of the signals.


### 3.4. PSD Analysis for Tremor Signal

Given its oscillatory nature, tremors quantification is suitable to the spectral analysis (PSD method) [34].

The dominant frequency can be calculated using PSD estimation. If the dominant frequencies in different axes are not the same, the valid dominant frequency in the axis with the highest peak power is defined as the dominant frequency of all axes.

Heldman et al. presented the discovery that the logarithm of the peak power in all triple-axis accelerations and three-axis angular velocities correlates well with the clinical scores of tremor [22]. Then the total peak power in all axes of the accelerometer and gyroscope signals is the spectral power distribution of all axes' data around the valid dominant frequency with the PSD method. The peak power in all axes after normalization is regarded as the amplitude of tremor.

For PSD estimation, peak power means the power estimation around the dominant frequency in the power spectrum of a sensor signal. The peak power in this study is the power estimation around the dominant frequency with  $\pm 0.3 \text{ Hz}$  length in the single sided power spectrum of ten-second sensor signals [21]. Figure 3 shows the system diagram of peak power, which means the power estimation around the dominant frequency in the sensor signals' power spectrum.



[Figure 3](#)

Power spectral density (PSD) estimation of ten-second inertial sensor signals.  $f_{\text{dominant}}$  represents the dominant frequency of the signals.

Equation (1) shows the calculation of the peak power of IMU signals:

$$\text{Peak power} = \int_{f_1}^{f_2} \frac{\text{FFT}^*(\text{signals}) \times \text{FFT}(\text{signals})}{N^2} df \quad (1)$$

where \* denotes the complex conjugate and  $N$  is the received sample points in the sensor signals.

The frequency resolution is  $f_{\text{sample rate}}/N$ , while  $f_{\text{sample rate}}$  is the sampling frequency.

Therefore, the power spectrum is converted into a single-side power spectrum. The power spectrum magnitude (peak power) has units of the input signal unit-RMS squared. The power estimation units of IMU signals are  $(^\circ/\text{s})^2/\text{Hz}$  and  $\text{g}^2/\text{Hz}$ , respectively. Here,  $g$  is equal to  $\text{m}/\text{s}^2$ .

### 3.5. Tremor State Detection

The occurrence of tremor in a patient also depends on many factors, especially motor fluctuations and the patient's mental and physical status. Tremor may sometimes disappear even in a patient with severe tremor. Therefore, it is important to quantify tremor severity during the stable tremor state. The tremor state is classified into two types in this study: valid state and invalid state. The sensor signals in an invalid state will be discarded because the patient's tremor signal cannot correctly be measured.

After a ten-second tremor-assessment task, the tremor signals need to be checked. Valid state means the sensor signal is stable in both the time domain and frequency domain. The judgments of valid state are listed as follows:



1. Frequency domain: In the frequency domain of the ten-second signals, the proportion of peak power to the whole power estimation should be bigger than 85%.
2. Time domain: In the time domain, the SD (standard deviation) value of ten-second angular velocity ranges (peak-to-peak values of all axes of the gyroscope) should be bigger than 70% of the mean gyroscope signal ranges.

For the valid state detection, only the gyroscope signal was utilized in present study. Parameters for the valid state detection both in the time domain and frequency domain are:

$$V_t = 1 - \frac{SD(G_{P-P})}{Mean(G_{P-P})} \quad (2)$$

$$V_f = \frac{P_{peak}}{\sum P_i} \quad (3)$$

where  $G_{P-P}$  is the matrix of peak-to-peak values of the combined triple-axis gyroscope signals during a ten-second tremor task, Mean represents the mean value,  $P_{peak}$  is the power estimation around the dominant tremor frequency ( $\pm 0.3$  Hz),  $P_i$  is the power estimation of a frequency point in the PSD estimation, while  $\sum P_i$  is the power distribution in the whole frequency domain.

### 3.6. Least-Square-Estimation Model and Signal Flow Diagram

Because accelerometers and gyroscopes are used to measure linear and rotational movement, respectively, for some tremor tasks the accelerometers present a higher correlation, while the gyroscopes perform a higher correlation during other tremor assessment tasks. Thus, a multiple linear regression model is used to fit the clinician ratings (UPDRS tremor score) and the peak powers during each tremor task [29,31,35,36].

Then the linear regression model can be expressed as:

$$R = R_0 + \ln \left( \sum_{i=1}^3 b_i \cdot PA_i + \sum_{i=1}^3 c_i \cdot PG_i \right) \quad (4)$$

where  $R$  is the predicated tremor score;  $R_0$ ,  $b_i$ , and  $c_i$  are the regression coefficients.  $i = 1, 2$ , and  $3$  denotes  $x$ ,  $y$ , and  $z$  axis, respectively;  $PA_i$  and  $PG_i$  are the single-axis peak power for the triple-axis accelerometer and triple-axis gyroscope, respectively [20].  $R_0$  and the three scaling factors ( $b_i$  and  $c_i$ ) are different for the three tremors (rest, posture, and action) [37]. These three regression coefficients (scaling factors) were obtained according to Stevens' power law in psychophysics [38].

The flowchart of the signal processing for the tremor quantification in this study is shown in [Figure 4](#). The sensor data are from the glove part, which is based on the IMU board and command module. In the computer part, the sensor data are band-pass filtered from 3.25 to 12 Hz (tremor band) before tremor amplitude quantification, in order to remove gravitational acceleration and dyskinesia components.





[Open in a separate window](#)

Figure 4

Signal processing for the tremor detection and quantification (six channels).

The signals in ten seconds are used for the PSD estimation at the end of the ten-second tremor-assessment task. Initially, the signals from the gyroscope and the accelerometer are analyzed separately. The single-sided, scaled, and auto power spectra of the signal of each gyroscope axis and each accelerometer axis, of which length is set to ten seconds, are separately computed to get the four-channel power spectra in real-time. The dominant frequencies and the estimated peak powers of these frequencies are thus obtained. Then, the estimated peak powers are multiplied by scaling factors. The channel with the highest peak power is the dominant channel, and its dominant frequency is the dominant frequency ( $F$ ) of all channels. Other channels are needed to perform PSD estimations again with the dominant frequency, and multiply the peak powers by the scaling factors again. The power estimations in all axes around the same dominant frequency ( $F$ ) are calculated and sum up to  $P'$ .

Then, the sum of peak powers, which is regarded as the tremor amplitude, and the dominant frequency can be displayed in the recoding list.

At the same time, the power spectrum of every axis in a certain period (three seconds) is calculated continuously. The dominant frequency and peak power can be obtained as soon as the tremor occurs. When the dominant frequency is in the range of 3.5–7.5 Hz and the sum of peak powers is more than a threshold, the tremor is reported.

## 4. Experimental Section

---

### 4.1. Validation of Analytical Methods

Neurologists judge tremor severity according to its position information, but IMU raw data (acceleration and angular velocity) are used for tremor quantification.

As described in [14], we have compared the tremor assessment systems based on an IMU with an electromagnetic (EM) motion tracking system (NDI Inc., Brewton, AL, Canada), whose RMS accuracy is  $0.30^\circ$  [39]. The EM sensor has small dimensions and was placed on the top of the index finger, together with the IMU board. An experiment for analytical validation of tremor quantification was then carried out. Detailed description is presented in this part.

#### 4.1.1. Hypothesis

In comparison with those from the EM system, the tremor amplitude (peak powers) and dominant frequency of the tremor assessment system should meet the following requirements [40,41]:

- Mean value and standard deviation of the differences between dominant frequencies:  $f_{md} < 1.00 \pm 0.88$  Hz;
- Correlation coefficient of peak powers:  $r > 0.95$ .

#### 4.1.2. Materials

- NDI (Northern Digital Inc., Waterloo, ON, Canada) Aurora<sup>®</sup> EM tracking system with a six degree-of-freedom (DOF) sensor;
- Tremor assessment system (including a USB cable and a command module with a sensor board);
- Laptop with the EM system (Aurora Toolbox) application software;
- Laptop with the application software of the tremor assessment system (LabVIEW-based user interface for tremor assessment; MATLAB (Natick, MA, USA) R2012b for data analysis).

#### 4.1.3. Experimental Setup and Method

[Figure 5](#) shows the assessment system for parkinsonian motor symptoms.



[Open in a separate window](#)

**Figure 5**

Prototype implementation of the assessment system (glove part) for parkinsonian motor symptoms. In addition to tremor assessment, this modified version of the assessment system can be used to assess bradykinesia, dyskinesia, and rigidity.

Nine healthy subjects performed postural simulative tremor tasks with different amplitudes, which were recorded by both the IMU-based and EM-based tracking systems. Dominant frequencies and amplitudes between the two systems were compared.

The dominant tremor frequencies calculated by the IMU-based tremor assessment system and EM system were  $f_{tb}$  and  $f_{em}$ , respectively, which were calculated by FFT method. The difference of their dominant frequencies in a single tremor task was then represented by:

$$f_d = |f_{tb} - f_{em}| \quad (5)$$

Therefore,  $f_{md}$  (mean and standard deviation of dominant frequency differences) can be calculated using the values of  $f_d$  (frequency differences) during all imitated tremor assessment tasks.

The tremor amplitude (peak powers based on the IMU signals) calculated by the tremor assessment system was  $R$ , while the tremor amplitude (peak power based on the position data) estimated by the EM system was  $E$ . Then the correlation coefficient between these two parameters was  $r$ .

#### 4.1.4. Results

The peak power (amplitude) of the tremor assessment system was calculated with both gyroscope and accelerometer signals using the PSD method, while the peak power of EM signals was calculated only from the position signals of the subject's finger.

For the postural tremor task, the differences of dominant frequencies between the two systems were small in all situations ( $f_{md} = 0.12 \pm 0.14$  Hz). The maximum difference was 0.57 Hz. Therefore,  $f_{md}$  was smaller than  $1.00 \pm 0.88$  Hz.

The correlation coefficient ( $r$ ) of dominant frequency between the tremor assessment system and EM system was 0.996.

For stable movements (i.e., means consistent frequency and range) performed by the subjects, the correlation between the two peak powers of the tremor assessment system and the EM tracking system was approximately linear. The correlation coefficient between these two systems ( $r = 0.97$ ,  $p < 0.001$ ) was larger than 0.95.

Results also show that inconsistent tremor actions resulted in smaller tremor amplitude when using the PSD method. On the other hand, tremor amplitude increased due to that fact that some sampling points were missing.

### 4.2. Clinical Experiments of Tremor Quantification

Clinical experiments of patients were carried out employing the tremor assessment system.

#### 4.2.1. Hypothesis

The tremor amplitude calculated by the tremor assessment system was  $R$ , while the tremor amplitude judged by the neurologist was  $D$ .

The tremor amplitude correlation between the tremor assessment system and the clinical ratings should meet the requirement [21]:

- Correlation coefficient:  $r > 0.84$ .

#### 4.2.2. Experiment Setup

A total of seven patients with Parkinsonian tremor were tested by the tremor assessment system. Their tremor severities (UPDRS tremor scores: D) ranged from 1 to 3. The measurements of these patients' tremors were performed after their medication had been stopped for more than 24 h.


Rest, postural, and action tremor assessment tasks were sequentially performed. However, for each patient, only the side with the more severe tremor was assessed. Each tremor assessment task lasted for ten seconds.

Time-frequency analyses and statistical analyses were carried out based on the IMU signals [42]. The valid state detection algorithm was executed for each assessment task.


#### 4.2.3. Results

[Figure 6](#) shows the IMU signals of the tremor assessment system and their power spectra for a patient with slight parkinsonian tremor ( $D = 1$ ). In [Figure 6a](#), the tremor was disappeared suddenly. However, the tremor amplitude fluctuated only on a modest scale in [Figure 6b](#).







**(a) Invalid state ( $V_t = 0.84$ )**



**(b) Valid state ( $V_t = 0.23$ )**



**(c) Invalid state ( $V_f = 0.69$ )**



**(d) Valid state ( $V_f = 0.91$ )**

[Open in a separate window](#)

**Figure 6**

Tremor state for a patient with parkinsonian tremor (UPDRS score D = 1): (a,b): The three-axis angular velocity and three-axis acceleration waveforms of rest tremor; (c,d): The three-axis combined power spectra of angular velocity signals (**bottom chart**) and acceleration signals (**upper chart**).

The validation results of tremor quantification and valid state detection algorithms demonstrate that the tremor amplitude calculated using PSD method is based on the premises that the tremor range is stable and its peak power is sharp.

When performing the signal processing algorithms, the rest and postural tremor tasks were assessed as either a valid state or an invalid state for PSD estimation. The amplitude of the action tremor is not included in the present results.

Therefore, the signals in [Figure 6b](#) could be used to calculate the tremor amplitude. The assessment tasks whose signals had invalid parameters ( $V_t > 0.7$  or  $V_f > 0.8$ ) were discarded.

[Figure 6b](#) and d show a valid tremor task (D = 1), in the time domain and frequency domain, respectively [\[35\]](#).

For the action tremor assessment, it was difficult to extract the tremor from voluntary movement because there were multiple frequency components across its power spectrum [\[39\]](#). A better signal filter should be employed in the future [\[43\]](#).

The peak power from all IMU outputs (R) had the greatest correlation with the clinical scores ( $r = 0.98$ ).

In addition to the linear regression model for tremor quantification, which involves both gyroscope and the accelerometer outputs, the tremor amplitude calculations of the gyroscope and the accelerometer were analyzed independently. [Table 1](#) indicates the calculated parameters together with the clinical scores, which resulted from a neurologist's judgments using UPDRS ratings. These parameters were the mean values from valid postural and rest tremor assessment tasks.



**Table 1**

Results for rest and postural tremor assessments

Subject	ln(acc. power)	ln(gyro.power)	ln(acc. RMS)	ln(gyro. RMS)	R	Clinical Score(D)
Patient 1	-8.52	0 (-6.27)	-2.95	0.30	0 (-5.41)	0.0
Patient 2	-7.42	0.78	-2.42	1.93	0 (-4.01)	0.5
Patient 3	-3.06	7.05	-0.92	3.92	1.27	1.0
Patient 4	-4.10	7.37	-1.40	4.14	1.36	1.5
Patient 5	-2.74	7.87	-0.58	4.31	1.99	2.0
Patient 6	-2.12	7.55	-0.49	4.22	2.06	2.0
Patient 7	-1.95	11.21	-0.22	5.61	2.81	3.0
r	0.88	0.93	0.91	0.93	0.98	/
$\alpha$ (2-tailed)	0.008	< 0.01	0.004	0.002	<0.01	/

Here acc. means accelerometer; gyro. represents gyroscope; RMS means root-mean-square; ln denotes the natural logarithm; R represents the predicated tremor score from the tremor assessment system; and r denotes the correlation coefficient between the relative parameters calculated by the IMU signals and clinical scores judged by the neurologist. For ln(gyro.power) and R, the negative values are can be regarded as zero because these values are too small.

[Table 1](#) shows that the gyroscope signal had a stronger correlation ( $r = 0.93$ ) with the clinical scores than the accelerometer signal ( $r = 0.88$  and  $0.91$  for peak powers and RMS, respectively).

[Table 1](#) also shows that the proportion of rotational parts and translation parts of a tremor (i.e., ln(gyro. power) versus ln(acc. power)) varies from patient to patient.

The difference between dominant frequencies from gyroscope and accelerometer were small ( $< 0.2$  Hz). For all the measured patients, the dominant tremor frequency was constant or had only small fluctuations, even if the tremor amplitude changed during ten-second task period.

The results indicate that the tremor frequency was stable for all patients; however, the tremor amplitude fluctuated continuously. The quantitative assessment of tremor, with an IMU and employing adaptive algorithms, provided an objective rating to classify rest or postural tremor severity even on a small scale.

## 5. Conclusions/Outlook

In this study, the features of parkinsonian tremor and its quantitative assessment methods have been presented. A time-frequency signal analysis algorithm for tremor state detection was adopted. After the ten-second tremor-assessment task, the task state is judged, form both frequency and time domains. The task with an invalid state will be discarded, while the signal with a valid state will be further processed via a least-square-estimation model. The peak powers of all axes of accelerations and angular velocities are used to the linear regression analysis. The regression coefficients are different for different tremor types.

A better cognition of the analytical methods in the system based on an IMU was achieved with the validation results. In comparative experiments with an EM tracking system, we showed that the raw data of an IMU correlated well with the severity of imitated tremor tasks ( $r = 0.996$ ). Results also show that inconsistent tremor actions resulted in smaller tremor amplitude when using the PSD method. On the other hand, tremor amplitude increased due to that fact that some sampling points were missing.

Next, the clinical experiments were performed. The clinical experiment showed the assessed severity correlated well with the judgments of a neurologist ( $r = 0.98$ ). In addition, the dominant tremor frequency was constant or had only small fluctuations even if the tremor amplitude changed during the ten-second task period.

This study provides the necessary science and engineering to guide future signal processing methods. Then future works should involve following points:

1. The relation between the peak power of IMU signals and the consistency of IMU signals in PD patients should be studied [43,44].
2. The repeatability of tremor amplitude with the same patient at different times should be investigated further [45].
3. The regression coefficients in Equation (5) are different according to tasks and tremor types. More clinical experiments are needed to modify the scale factors and coefficients in the tremor amplitude calculation. In addition, essential tremor and other types of pathological tremors need to be quantitatively assessed.
4. Together with the assessments of bradykinesia, rigidity, and the side effects of medication or surgical treatments such as dyskinesia, paresthesia, and other non-motor symptoms, more parameters should be evaluated in a single system for the overall assessment of parkinsonian symptoms [16,46].

In addition, there are two main trends in the development of MEMS motion sensors: more sensors embedded in a single chip and more accurate sensor fusion algorithms. These two points make it easier for the future system to detect the hand tremor with even smaller dimensions and greater accuracy. For example, a three-axis compass can remove the yaw drift of the IMU fusion output. Thus, more accurate orientation values, even location values, can be obtained. The new MEMS technology and motion-tracking algorithms will improve the performance of the tremor assessment system.

## Acknowledgments

---

The authors would like to thank J. Mehrkens and Joerg-Christian Tonn from the Department of Neurosurgery, University Hospital Grosshadern (LMU Muenchen) for their work in this project. The authors would like to thank all the test subjects for their kind support. The authors would also like to thank Gildas Henriet from STMicroelectronics SA (Geneva, Switzerland) for his technical support and Nucleo boards. This research was supported by “the National Natural Science Foundation of China” under the Grant No. 61501428.

## Author Contributions

---

Houde Dai conceived, designed and implemented the system, acquired the data and drafted the manuscript. Tim C. Lueth contributed to valuable discussions and suggestions. Pengyue Zhang provided suggestions for the medical issues.

## Conflicts of Interest

---

The authors declare no conflict of interest.

## References

---

1. Chaudhuri K.R., Ondo W.F. Handbook of Movement Disorders. Springer Healthcare Ltd.; London, UK: 2009. pp. 1–2.
2. Louis E.D., Ferreira J.J. How common is the most common adult movement disorder update on the worldwide prevalence of essential tremor. *Mov. Disord.* 2010;25:534–541. doi: 10.1002/mds.22838. [[PubMed](#)] [[CrossRef](#)]
3. Crawford P., Zimmerman E.E. Differentiation and diagnosis of tremor. *Am. Fam. Physician.* 2011;83:697–702. [[PubMed](#)]
4. Salarian A., Russmann H., Wider C., Burkhard P.R., Vingerhoets F.J.G., Aminian K. Quantification of tremor and bradykinesia in Parkinson's disease using a novel ambulatory monitoring system. *IEEE Trans. Biomed. Eng.* 2007;54:313–322. doi: 10.1109/TBME.2006.886670. [[PubMed](#)] [[CrossRef](#)]
5. Hurtado J.M., Charles C.M., Tamas L.B., Sigvardt K.A. Dynamics of tremor-related oscillations in the human globus pallidus: A single case study. *Proc. Nat. Acad. Sci. USA.* 1999;96:1674–1679. doi: 10.1073/pnas.96.4.1674. [[PMC free article](#)] [[PubMed](#)] [[CrossRef](#)]
6. Deuschl G., Bain P., Brin M. Consensus statement of the movement disorder society on tremor. *Mov. Disord.* 1998;13:2–23. doi: 10.1002/mds.870131303. [[PubMed](#)] [[CrossRef](#)]
7. Phwa R., Lyons K.E., Koller W. Handbook of Parkinson's Disease. 4th ed. CRC Press; New York, NY, USA: 2007. pp. 53–59.
8. Mittel C.S. Parkinson's Disease, Overview and Current Abstracts. 1st ed. Nova Science Pub. Inc.; New York, NY, USA: 2003.
9. Hess C.W., Pullman S.L. Tremor: Clinical phenomenology and assessment techniques. *Tremor Other Hyperkinet. Mov. (N.Y.)* 2012;2:1–15. [[PMC free article](#)] [[PubMed](#)]
10. Hauser R.A., Friedlander J., Zesiewicz T.A., Adler C.H., Seeberger C.H., O'Brien C.F., Molho E.S., Factor S.A. A home diary to assess functional status in patients with Parkinson's disease with motor fluctuations and dyskinesia. *Clin. Neuropharmacol.* 2000;23:75–81. doi: 10.1097/00002826-200003000-00003. [[PubMed](#)] [[CrossRef](#)]
11. Patel S., Lorincz K., Hughes R., Huggins N., Growdon J., Standaert D., Akay M., Dy J., Welsh M., Bonato P. Monitoring motor fluctuations in patients with Parkinson's disease using wearable sensors. *IEEE Trans. Inform. Technol. Biomed.* 2009;13:864–873. doi: 10.1109/TITB.2009.2033471. [[PMC free article](#)] [[PubMed](#)] [[CrossRef](#)]
12. Moore G.P., Ding L., Bronte-Stewart H.M. Concurrent Parkinson tremors. *Physiol. J.* 2000;529:273–281. doi: 10.1111/j.1469-7793.2000.00273.x. [[PMC free article](#)] [[PubMed](#)] [[CrossRef](#)]
13. Elble R.J., Deuschl G. Milestones in tremor research. *Mov. Disord.* 2011;26:1096–1105. doi: 10.1002/mds.23579. [[PubMed](#)] [[CrossRef](#)]
14. Dai H.D., Otten B., Mehrkens J.H., D'Angelo L.T., Lueth T.C. A novel glove monitoring system used to quantify neurological symptoms during deep-brain stimulation surgery. *IEEE Sens. J.* 2013;13:3193–3202. doi: 10.1109/JSEN.2013.2271775. [[CrossRef](#)]
15. Dai H.D., D'Angelo L.T. Quantitative assessment of tremor during deep-brain stimulation using a wearable glove system; Proceedings of IEEE International Workshop of Internet-of-Things Networking and Control; New Orleans, LA, USA. 24–27 June 2013; pp. 81–85.

16. Dai H.D., Otten B., Mehrkens J.H., D'Angelo L.T. A portable system for quantitative assessment of parkinsonian rigidity; Proceedings of 35th Annual International Conference of the IEEE Engineering in Medicine and Biology Society; Osaka, Japan. 3–7 July 2013; pp. 6591–6594. [[PubMed](#)]
17. Dai H.D., D'Angelo L.T. A portable system for quantitative assessment of parkinsonian bradykinesia during deep-brain stimulation surgery; Proceedings of IEEE Second International Conference on Advances in Biomedical Engineering; Tripoli, Lebanon. 11–13 September 2013; pp. 77–80.
18. Niazmand K., Tonn K., Kalaras A., Fietzek U.M., Mehrkens J.H., Lueth T.C. Quantitative evaluation of Parkinson's disease using sensor based smart glove; Proceedings of IEEE Symposium on Computer-Based Medical Systems; Bristol, UK. 27–30 June 2011; pp. 1–8.
19. Synnott J.L., Chen C.D., Moore G. WiiPD-objective home assessment of Parkinson's disease using the Nintendo Wii remote. *IEEE Trans. Inform. Technol. Biomed.* 2012;16:1304–1312. doi: 10.1109/TITB.2012.2215878. [[PubMed](#)] [[CrossRef](#)]
20. Elble R.J., Pullman S.L., Matsumoto J.Y., Raethjen J., Deuschl G., Tintner R., Tremor Research Group Tremor amplitude is logarithmically related to 4- and 5-point tremor rating scales. *Brain.* 2006;129:2660–2666. doi: 10.1093/brain/awl190. [[PubMed](#)] [[CrossRef](#)]
21. Mostile G.J., Giuffrida P., Adam O.R., Davidson A., Jankovic J. Correlation between Kinesia system assessments and clinical tremor scores in patients with essential tremor movement. *Mov. Disord.* 2010;25:1938–1943. doi: 10.1002/mds.23201. [[PubMed](#)] [[CrossRef](#)]
22. Heldman D.A., Jankovic J., Vaillancourt D.E., Prodoehl J., Elble R.J., Giuffrida J.P. Essential tremor quantification during activities of daily living. *Parkinsonism Relat. D.* 2011;17:537–542. doi: 10.1016/j.parkreldis.2011.04.017. [[PMC free article](#)] [[PubMed](#)] [[CrossRef](#)]
23. Giuffrida J.P., Riley D.E., Maddux B.N., Heldman D.A. Clinically deployable Kinesia™ technology for automated tremor assessment. *Mov. Disord.* 2009;24:723–730. doi: 10.1002/mds.22445. [[PubMed](#)] [[CrossRef](#)]
24. Burkhard P.R., Shale H., Langstom J.W., Tetrud J.W. Quantification of dyskinesia in Parkinson's disease validation of a novel instrumental method. *Mov. Disord.* 1999;14:754–27. doi: 10.1002/1531-8257(199909)14:5<754::AID-MDS1007>3.0.CO;2-1. [[PubMed](#)] [[CrossRef](#)]
25. Pierleoni P., Palma L., Belli A., Pernini L. A real-time system to aid clinical classification and quantification of tremor in Parkinson's disease; Proceedings of IEEE-EMBS International Conference on Biomedical and Health Informatics; Valencia, Italy. 1–4 June 2014; pp. 113–116.
26. Khan F.M., Barnathan M., Myers S., Loftus S.A. Wearable Accelerometer System for Unobtrusive Monitoring of Parkinson's Disease Motor Symptoms; Proceedings of IEEE International Conference on Bioinformatics and Bioengineering; Boca Raton, FL, USA. 10–12 November 2014; pp. 120–125.
27. Zhang H.H., Chen X.Y., Lin W.Y., Chou W.C., Lee M.Y. A novel accelerometer-based method for the real-time assessment of Parkinson's tremor; Proceedings of IEEE International Conference on Communication Problem-Solving; Beijing, China. 5–7 December 2014; pp. 87–90.
28. The Michael J. Fox Foundation for Parkinson's Research. [(accessed on 14 September 2015)]. Available online: <https://www.michaeljfox.org/page.html?access-parkinsons-clinical-data-and-biospecimens>.
29. Timmer J., Lauk M., Lücking C.H. Confidence regions for spectral peak frequencies. *Biometrical J.* 1997;39:849–861. doi: 10.1002/bimj.4710390713. [[CrossRef](#)]

30. Lueth T.C., D'Angelo L.T., Czabke A. Pervasive and Smart Technologies for Healthcare: Ubiquitous Methodologies and Tools. IGI Global; Hershey, PA, USA: 2010. Chapter. 14. TUM-AgeTech: A New Framework for Pervasive Medical Devices; pp. 295–321.
31. Timmer J., Häußler S., Lauk M., Lücking C.H. Pathological tremors deterministic chaos or nonlinear stochastic oscillators. *Chaos*. 2000;10:278–288. doi: 10.1063/1.166494. [[PubMed](#)] [[CrossRef](#)]
32. Veluvolu K.C., Ang W.T. Estimation of physiological tremor from accelerometers for real-time applications. *Sensors*. 2011;11:3020–3036. doi: 10.3390/s110303020. [[PMC free article](#)] [[PubMed](#)] [[CrossRef](#)]
33. Burkhard P.R., Langston J.W., Tetrud J.W. Voluntarily simulated tremor in normal subjects. *Neurophysiol. Clin.* 2002;32:119–126. doi: 10.1016/S0987-7053(02)00296-4. [[PubMed](#)] [[CrossRef](#)]
34. Narcisa V., Aguilar D., Nguyen D.V., Campos L., Brodovsky J., White S., Adams R., Tassone F., Hagerman P.J., Hagerman R.J. A quantitative assessment of tremor and ataxia in female FMR1 premutation carriers using CATSYS. *Curr. Gerontol. Geriatr. Res.* 2011;2001:1–7. doi: 10.1155/2011/484713. [[PMC free article](#)] [[PubMed](#)] [[CrossRef](#)]
35. Timmer J., Lauk M., Häußler S., Radt V. Cross-spectral analysis of tremor time series. *Int. J. Bifurcat. Chaos*. 2000;10:2595–2610. doi: 10.1142/S0218127400001663. [[CrossRef](#)]
36. Timmer J., Gantert C., Deuschl G., Honerkamp J. Characteristics of hand tremor time series. *Biol. Cybern.* 1993;70:75–80. doi: 10.1007/BF00202568. [[PubMed](#)] [[CrossRef](#)]
37. Teravainen H., Calne D.B. Action tremor in Parkinson's disease. *J. Neurol. Neurosurg. Psychiatry*. 1980;43:257–267. doi: 10.1136/jnnp.43.3.257. [[PMC free article](#)] [[PubMed](#)] [[CrossRef](#)]
38. Luce R.D., Krumhansl C. Measurement, Scaling, and Psychophysics. In: Atkinson R.C., Herrnstein R.J., Lindzey G., Luce R.D., editors. *Stevens' Handbook of Experimental Psychology*. Wiley; New York, NY, USA: 1988. pp. 1–74.
39. NDI Aurora Electromagnetic Tracking System. [(accessed on 14 September 2015)]. Available online: <http://www.ndigital.com/aurora.php>.
40. Niazmand K., Kalaras A., Dai H., Lueth T.C. Comparison of methods for tremor frequency analysis for patients with Parkinson's disease; Proceedings of International Conference on BioMedical Engineering and Informatics; Shanghai, China. 15–17 October 2011; pp. 1312–1316.
41. Bland J.M., Altman D.G. Applying the right statistics analyses of measurement studies. *Ultrasound Obstet. Gynecol.* 2003;22:85–93. doi: 10.1002/uog.122. [[PubMed](#)] [[CrossRef](#)]
42. O'Suilleabhain P.E., Matsumoto J.Y. Time-frequency analysis of tremors. *Brain*. 1998;121:2127–2134. doi: 10.1093/brain/121.11.2127. [[PubMed](#)] [[CrossRef](#)]
43. Riviere C.N., Reich S.G., Thakor N.V. Adaptive Fourier modeling for quantification of tremor. *J. Neurosci. Methods*. 1997;74:77–87. doi: 10.1016/S0165-0270(97)02263-2. [[PubMed](#)] [[CrossRef](#)]
44. Ellermeier W., Faulhammer G. Empirical evaluation of axioms fundamental to Stevens's ratio-scaling approach I. Loudness production. *Percept. Psychophys.* 2000;62:1505–1511. doi: 10.3758/BF03212151. [[PubMed](#)] [[CrossRef](#)]
45. Popovic L.Z., Sekara T.B., Popovic M.B. Adaptive band-pass filter (ABPF) for tremor extraction from inertial sensor data. *Comput. Meth. Programs Biomed.* 2010;99:298–305. doi: 10.1016/j.cmpb.2010.03.018. [[PubMed](#)] [[CrossRef](#)]

46. Cole B.T., Roy S.H., de Luca C.J., Nawab S.H. Dynamical learning and tracking of tremor and dyskinesia from wearable sensors. IEEE Trans. Neural Syst. Rehabil. Eng. 2014;22:982–991. doi: 10.1109/TNSRE.2014.2310904. [[PubMed](#)] [[CrossRef](#)]

---

Articles from Sensors (Basel, Switzerland) are provided here courtesy of **Multidisciplinary Digital Publishing Institute (MDPI)**