

Calibration Certificate

Certificate No. : TZ0325122310 Certificate Date: 12/15/2025

Reference Instrumentation

Device	Calibration device Manufacturer	Equipment No.	Accuracy
Programmable Temperature & Humidity Chamber	Donguan Kesionots Precision Instrument	KS-HW150	$\pm 0.05^{\circ}\text{C}; \pm 3\%\text{RH}$
High Accuracy RTD Thermometer	OMEGA	HH376	$\pm 0.15^{\circ}\text{C}$

We certify that the devices mentioned on this certificate have been thoroughly tested, validated and reached performance accuracy specifications over the specific ranges. This device was assembled, tested and calibrated in accordance with its specifications and prior to release for shipment. All reference instruments are calibrated accredited laboratory, and traceable to NIST according Mutual Recognition Agreement.

Product Information

Product Name	Model	Serial No.	Production Date	Expiry Date	Quantity
USB TEMP&RH DATA LOGGER	TempU03	TZ0325122310	12/15/2025	12/15/2027	1PC

Calibration Results (Environment Conditions: Air Temperature 20.5°C, Relative Humidity 53.0%)

Validation Points	Tolerance	Result
-20°C	$\pm 0.5^{\circ}\text{C}$	Pass
0°C	$\pm 0.5^{\circ}\text{C}$	Pass
30°C	$\pm 0.5^{\circ}\text{C}$	Pass
30%RH	$\pm 5\%\text{RH}$	Pass
50%RH	$\pm 5\%\text{RH}$	Pass
70%RH	$\pm 5\%\text{RH}$	Pass

Test Result Pass/Fail	Pass	Signature of Calibration Technician	Andy Z.
-----------------------	------	-------------------------------------	---------