

RIVM
Ids Ngs
Antonie van Leeuwenhoeklaan 9
3721MA Bilthoven
Netherlands

Project code/Order number

117921

Date of Report

10-Sep-2019

Authorization

Jan Peter Heutink
Manager Operations

**BaseClear B.V.**

Galileiweg 4
2333 BD Leiden
The Netherlands
T +31 71 523 3917
E info@baseclear.com

Next Generation Sequencing project

Project 117921
 Client Ids Ngs
 Date 10-Sep-2019 13:46

1. Introduction

This report contains a detailed overview of your next-generation sequencing experiment. The data and results apply to the following samples:

Sample Name	Illumina demultiplexing
11042457_Sau_BSR	X
11047029_Sau_BSR	X
11047189_Sau_BSR	X
11047201_Sau_BSR	X
11047208_Sau_BSR	X
11047214_Sau_BSR	X
11047218_Sau_BSR	X
11047233_Sau_BSR	X
11047237_Sau_BSR	X
11047239_Sau_BSR	X
11047240_Sau_BSR	X
11047245_Sau_BSR	X
11047254_Sau_BSR	X
11047258_Sau_BSR	X
11047263_Sau_BSR	X
22007926_Ecl_BSR	X
22007929_Ecl_BSR	X
22007938_Mmo_BSR	X
22007941_Kpn_BSR	X
22007944_Ecl_BSR	X
22007948_Eco_BSR	X
22007950_Pae_BSR	X
22007951_Ecl_BSR	X
22007952_Pst_BSR	X
1071900071_Shig_BBD	X
1121900489_Eco_BSR_EPI	X
1121900496_Eco_BSR_EPI	X
1121900497_Eco_BSR_EPI	X
1121900498_Eco_BSR_EPI	X
1121900505_Eco_BSR_EPI	X
1121900507_Eco_BSR_EPI	X
1121900508_Eco_BSR_EPI	X
1121900509_Eco_BSR_EPI	X
1121900510_Eco_BSR_EPI	X
1121900511_Eco_BSR_EPI	X
1121900512_Eco_BSR_EPI	X
1121900514_Eco_BSR_EPI	X
1121900515_Eco_BSR_EPI	X
1121900517_Eco_BSR_EPI	X
1131900228_Bord_BSR	X
1141900087_Lm_BSR_EPI	X
1141900088_Lm_BSR_EPI	X
1141900089_Lm_BSR_EPI	X

1141900090_Lm_BSR_EPI	X
1141900091_Lm_BSR_EPI	X
1141900092_Lm_BSR_EPI	X
1141900093_Lm_BSR_EPI	X
1141900094_Lm_BSR_EPI	X
1141900095_Lm_BSR_EPI	X
1141900096_Lm_BSR_EPI	X
1141900097_Lm_BSR_EPI	X
1141900098_Lm_BSR_EPI	X
NLA001900549	X
NLA001900575	X
NLA001900577	X
NLA001900579	X
NLA001900580	X
NLA001900581	X
NLA001900582	X
NLA001900583	X
NLA001900585	X
NLA001900586	X
NLA001900591	X
NLA001900592	X

2. Download of data and analysis files

Order placed via the order-portal

The data and analysis results can be downloaded via the [order-portal](#) after login with your credentials.

If you would like to download your data via a FTPs client (e.g. FileZilla), you will need to login to the [order-portal](#), navigate to 'Settings', check the 'Make your results also downloadable from our FTP server' box, and save your changes.

FTP connection details:

Login	order-portal username
Password	order-portal password
Host	ftps.baseclear.com
Port	990
Protocol	FTP – File Transfer Protocol
Encryption	Require implicit FTP over TLS

3. Illumina demultiplexing

Single-end or paired-end sequence reads were generated using the Illumina NovaSeq 6000 or MiSeq system. The sequences generated with the MiSeq system were performed under accreditation according to the scope of BaseClear B.V. (L457; NEN-EN-ISO/IEC 17025).

When paired-end sequencing is being performed, the "Number of reads" noted in this report is referring to read pairs. FASTQ read sequence files were generated using bcl2fastq2 version 2.18. Initial quality assessment was based on data passing the Illumina Chastity filtering. Subsequently, reads containing PhiX control signal were removed using an in-house filtering protocol. In addition, reads containing (partial) adapters were clipped (up to a minimum read length of 50 bp). The second quality assessment was based on the remaining reads using the FASTQC quality control tool version 0.11.5.

The final quality scores per sample are provided as enclosure.

4. Results

A summary of the results is provided below. Sequence and project data are recorded digitally in our secure database and stored for backup purposes only. Data is stored for a period of one month. The result files which have been generated within this project for each analysis are as follows:

Illumina demultiplexing

The files are stored in the "raw_sequences" folder and include:

- The compressed sequence reads in FASTQ format (*.filt.fastq.gz)
- A text file containing the MD5 checksum (fingerprint) of each gzipped FASTQ file ("md5sum_*.txt")

Illumina demultiplexing report

Project 117921
Client Ids Ngs
Date 10-Sep-2019 13:46



Summary of the results - Illumina demultiplexing report

* More information about the average Q-score can be found at the Illumina website (<https://www.illumina.com/science/education/sequencing-quality-scores.html>)

Statistics after demultiplexing and filtering

Sample name	Number of reads	Yield in mbp	Average quality
11042457_Sau_BSR	2602088	749	35.12
11047029_Sau_BSR	3707864	1065	34.8
11047189_Sau_BSR	2763516	794	35.16
11047201_Sau_BSR	3140126	901	35.08
11047208_Sau_BSR	2716940	781	35.14
11047214_Sau_BSR	2453069	711	35.02
11047218_Sau_BSR	1753886	510	34.64
11047233_Sau_BSR	2464617	710	34.98
11047237_Sau_BSR	2364729	684	35.1
11047239_Sau_BSR	3091317	884	35.08
11047240_Sau_BSR	2890001	824	35.19
11047245_Sau_BSR	3419828	976	35.09
11047254_Sau_BSR	2290470	659	35.09
11047258_Sau_BSR	2570537	737	35.07
11047263_Sau_BSR	3150830	907	35.14
22007926_Ecl_BSR	1521912	446	35.03
22007929_Ecl_BSR	1772732	518	35.09
22007938_Mmo_BSR	1678327	486	35.18
22007941_Kpn_BSR	1604058	466	34.94
22007944_Ecl_BSR	2001729	584	35.1
22007948_Eco_BSR	2018686	587	35.16
22007950_Pae_BSR	1868452	549	35.17
22007951_Ecl_BSR	1757344	510	35.13
22007952_Pst_BSR	1875129	550	34.97
1071900071_Shig_BBD	1868246	524	35.17
1121900489_Eco_BSR_EPI	1326785	380	34.95
1121900496_Eco_BSR_EPI	1652481	470	34.95
1121900497_Eco_BSR_EPI	1746144	502	35.12

Sample name	Number of reads	Yield in mbp	Average quality
1121900498_Eco_BSR_EPI	1620879	458	35.06
1121900505_Eco_BSR_EPI	1689512	476	35.13
1121900507_Eco_BSR_EPI	1371049	390	35.08
1121900508_Eco_BSR_EPI	1867761	536	35.06
1121900509_Eco_BSR_EPI	1494029	431	34.99
1121900510_Eco_BSR_EPI	1907923	546	34.83
1121900511_Eco_BSR_EPI	2004998	576	35.08
1121900512_Eco_BSR_EPI	1507448	433	35.17
1121900514_Eco_BSR_EPI	1704711	495	35.06
1121900515_Eco_BSR_EPI	1850793	528	35.15
1121900517_Eco_BSR_EPI	1702893	491	35.02
1131900228_Bord_BSR	1889029	542	35.13
1141900087_Lm_BSR_EPI	3405090	967	35.02
1141900088_Lm_BSR_EPI	1711286	508	34.88
1141900089_Lm_BSR_EPI	1958512	563	34.88
1141900090_Lm_BSR_EPI	1525968	452	34.6
1141900091_Lm_BSR_EPI	1618182	479	34.76
1141900092_Lm_BSR_EPI	2090962	605	34.97
1141900093_Lm_BSR_EPI	1554989	462	34.7
1141900094_Lm_BSR_EPI	1218966	363	34.53
1141900095_Lm_BSR_EPI	2669009	769	34.9
1141900096_Lm_BSR_EPI	2204181	630	34.94
1141900097_Lm_BSR_EPI	2291187	670	35.03
1141900098_Lm_BSR_EPI	2734156	779	34.92
NLA001900549	2753550	769	35.3
NLA001900575	2173294	611	35.28
NLA001900577	1964536	567	35.19
NLA001900579	2417398	697	35.22
NLA001900580	2690767	749	35.22
NLA001900581	1853361	524	35.31
NLA001900582	2355016	661	35.3
NLA001900583	2783422	769	35.21
NLA001900585	2627338	744	35.28
NLA001900586	1333919	378	35.25
NLA001900591	3002311	774	34.48
NLA001900592	2435288	685	35.27