

Product Information



Glove Name CURAD® Clear Vinyl Gloves
Description Powder-Free Vinyl Synthetic Examination Gloves
Non-Sterile, Ambidextrous

US Office: Medline Industries, Inc.
Northfield, IL, USA
1-800-MEDLINE
www.medline.com

Product Numbers		Packaging
CURCA4135R	One-Size	100 Gloves / Box- 10 Box/ Cs
		10 Boxes / Case
		1000 Gloves / Case

Product Not Made with Natural Rubber Latex

Physical Properties and Quality Characteristics

Characteristics:	Glove Color/Finish	Clear/Smooth
	Glove Length	235 mm
	Finger Thickness	0.08 mm
	Palm Thickness	0.08 mm
	Primary Material	Polyvinyl Chloride (PVC)
	Accelerators	None
	Cuff	Beaded

Pinholes:	FDA Requirement	2.5 AQL
------------------	-----------------	---------

Tensile Strength:	Before Aging/After Aging	16 MPa/14 MPa
	ASTM Standard D5250	11 MPa Min.

Ultimate Elongation:	Before Aging/After Aging	375%/355%
	ASTM Standard D5250	300% Min.

Viral Penetration Test:	ASTM Standard F1671	Complies with F1671 Test Method
--------------------------------	---------------------	---------------------------------

Natural Rubber Latex Protein Content:	ASTM Standard D5712	Not applicable; Latex-Free
--	---------------------	----------------------------

Residual Powder Content:	ASTM Standard D6124	Less than 2 mg/glove
---------------------------------	---------------------	----------------------

CURAD and Medline are registered trademarks of Medline Industries, Inc.

© Medline Industries, Inc. not to be reproduced or distributed without prior written permission of Medline Industries, Inc.

Updated 1.2021