

Calibration Report

Application No: SAMPLE-001

Date: 8/31/2026

Inspector: Anjali

Status: Draft

Instrument Details

Manufacturer: Mettler Toledo

Model: XPR205

Serial Number: MT-XPR-001

Max Capacity: 220

Test Results

Test: Repeatability

```
{  
  "readings": [  
    0,  
    0.01,  
    -0.01  
  ],  
  "sd": "0.01"  
}
```

Test: Eccentricity

```
{  
  "offsets": [  
    0.1,  
    0.2,  
    -0.1  
  ],  
  "mpe": 0.5  
}
```