


Sample: T2IS3
 Operator: LvdM, ChemEng, Stb Uni
 Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
 File: C:\...\RHASCH~1\005-955.SMP

Test Number: 3
 Analyzed: 2026/06/18 12:45:22PM
 Reported: 2026/06/29 10:51:13PM
 Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
 Material: Sand / Water
 Background: Water RI 1.331
 Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.0%

Combined Report



Summary Report

Analysis Conditions

FlowRate: 12.0 l/m
 Circulation time: 60 sec

Ultrasonic intensity: 60 %
 Ultrasonic time: 75 sec

Sample

Sample Concentration: 0.00259 %
 Obscuration: 10.4 %

Volume Distribution Arithmetic Statistics

		Std Dev of 3			Std Dev of 3
Mean	23.01	0.424	Mode	4.005	0.137
Median	6.801	0.354			

Selected Sizes

Diameter (µm)	Percent Finer
100.0	95.0
50.00	82.5
20.00	69.9
10.00	57.4
1.000	8.4

Peaks

Peak Number	% of Dist.*	% of Dist. Std Dev of 3	Mean	Mean Std Dev of 3	Median	Median Std Dev of 3	Mode
1	9.5	N/A	0.764	N/A	0.760	N/A	0.799
2	46.4	N/A	3.902	N/A	3.415	N/A	4.005
3	15.2	N/A	14.16	N/A	13.50	N/A	11.96
4	8.8	N/A	30.53	N/A	30.02	N/A	28.36
5	20.2	N/A	80.87	N/A	76.95	N/A	79.92

* Peaks must comprise at least 5.00 % of the distribution.

Sample: T2IS3
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-955.SMP

Test Number: 3
Analyzed: 2026/06/18 12:45:22PM
Reported: 2026/06/29 10:51:14PM
Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.0%

Report by Size Table

Low Diameter (μ m)	Cumulative Volume Finer (Percent)	Low Diameter (μ m)	Cumulative Volume Finer (Percent)	Low Diameter (μ m)	Cumulative Volume Finer (Percent)	Low Diameter (μ m)	Cumulative Volume Finer (Percent)
150.000	99.9	40.000	79.4	7.000	50.6	0.900	7.0
100.000	95.0	30.000	75.4	6.000	47.3	0.800	5.5
90.000	93.0	25.000	72.9	5.000	42.8	0.700	3.7
80.000	90.7	20.000	69.9	4.000	36.9	0.600	2.0
70.000	88.1	10.000	57.4	3.000	29.3	0.500	0.6
60.000	85.3	9.000	55.4	2.000	19.8	0.400	0.1
50.000	82.5	8.000	53.2	1.000	8.4		



Micromeritics Instrument Corporation

Saturn DigiSizer 5200 V1.12

Saturn DigiSizer 5200 V1.12

5200 LSHU V2.01 S/N 216

Page 3

Sample: T2IS3
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-955.SMP

Test Number: 3
Analyzed: 2026/06/18 12:45:22PM
Reported: 2026/06/29 10:51:14PM
Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.0%

Report by Volume Percent


Low Diameter (μm)	Cumulative Volume Finer (Percent)	Low Diameter (μm)	Cumulative Volume Finer (Percent)	Low Diameter (μm)	Cumulative Volume Finer (Percent)
77.281	90.0	6.801	50.0	1.138	10.0

Sample: T2IS3
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-955.SMP

Test Number: 3
Analyzed: 2026/06/18 12:45:22PM
Reported: 2026/06/29 10:51:14PM
Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.0%




Sample: T2IS3
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-955.SMP

Test Number: 3
Analyzed: 2026/06/18 12:45:22PM
Reported: 2026/06/29 10:51:14PM
Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.0%



Sample: T2IS3
 Operator: LvdM, ChemEng, Stb Uni
 Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
 File: C:\...\RHASCH~1\005-955.SMP

Test Number: 3
 Analyzed: 2026/06/18 12:45:22PM
 Reported: 2026/06/29 10:51:14PM
 Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
 Material: Sand / Water
 Background: Water RI 1.331
 Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.0%

