


Sample: T2SG2
 Operator: LvdM, ChemEng, Stb Uni
 Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
 File: C:\...\RHASCH~1\005-951.SMP

Test Number: 3
 Analyzed: 2026/06/17 16:18:18PM
 Reported: 2026/06/29 9:28:38PM
 Background: 2026/06/17 10:31:40PM

Model: (1.520, 0.1000000), 1.331
 Material: Sand / Water
 Background: Water RI 1.331
 Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 11.1%

Combined Report



Summary Report

Analysis Conditions

FlowRate: 12.0 l/m
 Circulation time: 60 sec

Ultrasonic intensity: 60 %
 Ultrasonic time: 75 sec

Sample

Sample Concentration: 0.00489 %
 Obscuration: 12.8 %

Volume Distribution Arithmetic Statistics

		Std Dev of 3			Std Dev of 3
Mean	82.74	2.080	Mode	267.7	0.000
Median	22.81	2.026			

Selected Sizes

Diameter (µm)	Percent Finer
100.0	71.1
50.00	61.9
20.00	48.2
10.00	38.1
1.000	4.9

Peaks

Peak Number	% of Dist.*	% of Dist. Std Dev of 3	Mean	Mean Std Dev of 3	Median	Median Std Dev of 3	Mode
1	5.6	N/A	0.790	N/A	0.784	N/A	0.799
2	31.5	N/A	3.955	N/A	3.513	N/A	4.243
3	10.1	N/A	13.16	N/A	12.84	N/A	11.96
4	7.0	N/A	24.18	N/A	23.93	N/A	23.86
5	14.6	N/A	51.06	N/A	48.79	N/A	47.60
6	6.8	N/A	111.1	N/A	108.6	N/A	106.6
7	24.4	N/A	259.8	N/A	259.2	N/A	267.7

* Peaks must comprise at least 5.00 % of the distribution.

Sample: T2SG2
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-951.SMP

Test Number: 3
Analyzed: 2026/06/17 16:18:18PM
Reported: 2026/06/29 9:28:38PM
Background: 2026/06/17 10:31:40PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 11.1%

Report by Size Table

Low Diameter (μm)	Cumulative Volume Finer (Percent)	Low Diameter (μm)	Cumulative Volume Finer (Percent)	Low Diameter (μm)	Cumulative Volume Finer (Percent)	Low Diameter (μm)	Cumulative Volume Finer (Percent)
300.000	94.6	60.000	64.9	8.000	35.3	0.900	4.0
250.000	86.1	50.000	61.9	7.000	33.4	0.800	3.0
200.000	78.5	40.000	57.9	6.000	31.1	0.700	1.8
150.000	75.4	30.000	53.7	5.000	28.1	0.600	0.7
100.000	71.1	25.000	51.3	4.000	23.8	0.500	0.1
90.000	69.8	20.000	48.2	3.000	18.5	0.400	0.0
80.000	68.5	10.000	38.1	2.000	12.1		
70.000	66.9	9.000	36.8	1.000	4.9		



Micromeritics Instrument Corporation

Saturn DigiSizer 5200 V1.12

Saturn DigiSizer 5200 V1.12

5200 LSHU V2.01 S/N 216

Page 3

Sample: T2SG2
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-951.SMP

Test Number: 3
Analyzed: 2026/06/17 16:18:18PM
Reported: 2026/06/29 9:28:38PM
Background: 2026/06/17 10:31:40PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 11.1%

Report by Volume Percent


Low Diameter (μm)	Cumulative Volume Finer (Percent)	Low Diameter (μm)	Cumulative Volume Finer (Percent)	Low Diameter (μm)	Cumulative Volume Finer (Percent)
271.444	90.0	22.809	50.0	1.702	10.0

Sample: T2SG2
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-951.SMP

Test Number: 3
Analyzed: 2026/06/17 16:18:18PM
Reported: 2026/06/29 9:28:38PM
Background: 2026/06/17 10:31:40PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 11.1%




Sample: T2SG2
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-951.SMP

Test Number: 3
Analyzed: 2026/06/17 16:18:18PM
Reported: 2026/06/29 9:28:38PM
Background: 2026/06/17 10:31:40PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 11.1%



Sample: T2SG2
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-951.SMP

Test Number: 3
Analyzed: 2026/06/17 16:18:18PM
Reported: 2026/06/29 9:28:38PM
Background: 2026/06/17 10:31:40PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 11.1%

