


Sample: T2SS2
 Operator: LvdM, ChemEng, Stb Uni
 Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
 File: C:\...\RHASCH~1\005-957.SMP

Test Number: 3
 Analyzed: 2026/06/18 14:24:39PM
 Reported: 2026/06/29 11:00:20PM
 Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
 Material: Sand / Water
 Background: Water RI 1.331
 Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.3%

Combined Report



Summary Report

Analysis Conditions

FlowRate: 12.0 l/m
 Circulation time: 60 sec

Ultrasonic intensity: 60 %
 Ultrasonic time: 75 sec

Sample

Sample Concentration: 0.00359 %
 Obscuration: 12.1 %

Volume Distribution Arithmetic Statistics

		Std Dev of 3			Std Dev of 3
Mean	39.01	1.741	Mode	4.243	0.145
Median	10.11	0.500			

Selected Sizes

Diameter (µm)	Percent Finer
100.0	86.3
50.00	75.4
20.00	62.1
10.00	49.8
1.000	6.6

Peaks

Peak Number	% of Dist.*	% of Dist. Std Dev of 3	Mean	Mean Std Dev of 3	Median	Median Std Dev of 3	Mode
1	7.5	N/A	0.778	N/A	0.773	N/A	0.799
2	39.9	N/A	3.854	N/A	3.450	N/A	4.243
3	15.9	N/A	13.91	N/A	13.26	N/A	11.96
4	11.8	N/A	33.77	N/A	32.97	N/A	35.70
5	13.9	N/A	79.15	N/A	76.24	N/A	75.45
6	11.0	N/A	183.8	N/A	180.7	N/A	189.5

* Peaks must comprise at least 5.00 % of the distribution.

Sample: T2SS2
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-957.SMP

Test Number: 3
Analyzed: 2026/06/18 14:24:39PM
Reported: 2026/06/29 11:00:20PM
Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.3%

Report by Size Table

Low Diameter (µm)	Cumulative Volume Finer (Percent)	Low Diameter (µm)	Cumulative Volume Finer (Percent)	Low Diameter (µm)	Cumulative Volume Finer (Percent)	Low Diameter (µm)	Cumulative Volume Finer (Percent)
250.000	99.4	50.000	75.4	7.000	43.5	0.800	4.1
200.000	96.3	40.000	72.1	6.000	40.5	0.700	2.6
150.000	91.6	30.000	67.8	5.000	36.4	0.600	1.2
100.000	86.3	25.000	65.2	4.000	31.0	0.500	0.3
90.000	84.8	20.000	62.1	3.000	24.2	0.400	0.0
80.000	82.9	10.000	49.8	2.000	16.0		
70.000	80.6	9.000	48.0	1.000	6.6		
60.000	78.1	8.000	45.9	0.900	5.5		



Micromeritics Instrument Corporation

Saturn DigiSizer 5200 V1.12

Saturn DigiSizer 5200 V1.12

5200 LSHU V2.01 S/N 216

Page 3

Sample: T2SS2
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-957.SMP

Test Number: 3
Analyzed: 2026/06/18 14:24:39PM
Reported: 2026/06/29 11:00:20PM
Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.3%

Report by Volume Percent


Low Diameter (μm)	Cumulative Volume Finer (Percent)	Low Diameter (μm)	Cumulative Volume Finer (Percent)	Low Diameter (μm)	Cumulative Volume Finer (Percent)
133.329	90.0	10.107	50.0	1.371	10.0

Sample: T2SS2
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-957.SMP

Test Number: 3
Analyzed: 2026/06/18 14:24:39PM
Reported: 2026/06/29 11:00:20PM
Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.3%




Sample: T2SS2
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-957.SMP

Test Number: 3
Analyzed: 2026/06/18 14:24:39PM
Reported: 2026/06/29 11:00:20PM
Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.3%



Sample: T2SS2
Operator: LvdM, ChemEng, Stb Uni
Submitter: Rebekah Haschick, BotZoo, Dr Ndlovu
File: C:\...\RHASCH~1\005-957.SMP

Test Number: 3
Analyzed: 2026/06/18 14:24:39PM
Reported: 2026/06/29 11:00:20PM
Background: 2026/06/18 10:38:57PM

Model: (1.520, 0.1000000), 1.331
Material: Sand / Water
Background: Water RI 1.331
Smoothing: Medium

Comments: 0.0005% TSPP. Initial obsc = 10.3%

