

# Medical Image Processing for Diagnostic Applications

## MRI – Acquisition Devices

Online Course – Unit 19

Andreas Maier, Joachim Hornegger, Markus Kowarschik, Frank Schebesch

Pattern Recognition Lab (CS 5)



# Topics

## MR Acquisition Devices

### Summary

Take Home Messages

Further Readings

# MR Acquisition Devices

MR scanners are huge and heavy systems with strong superconducting magnets:



Figure 1: Examples of 3 Tesla systems: MAGNETOM Verio (left) and Trio Scanner (right) (images courtesy of Siemens Medical Solutions)

# Physical and Mathematical Ingredients

For MRI imaging, the following physical and mathematical concepts are required to be understood:

- **nuclei** serve as objects to be imaged,
- **homogeneous magnetic fields** are generated by the scanner to **align** the nuclear **moment vectors**,
- the **resonance phenomenon** that results from the interaction of nuclei with the magnetic field enables measurements,
- **Fourier methods** are used for image reconstruction,
- **image enhancement algorithms** are applied to compensate for violations of the required homogeneity of the magnetic field (**bias field correction**).

**Note:** Details in physics are not in the focus of this course, but the algorithmic aspects.



# Components of an MR Scanner

The major four components of an MR Scanner are:

- the main magnet,
- a magnetic field gradient system,
- a radio frequency system (RF system),
- and the imaging system.




Figure 2: Main components of an MR scanner (image courtesy of the National High Magnetic Field Laboratory, Florida)

## Components of an MR Scanner: Main Magnet

The **main magnet** is required to generate a strong uniform static magnetic field for the polarization of nuclear spins.

In practice, there are several options:

- **permanent magnet**: low field applications ( $< 0.3$  T),
- **resistive magnet**: low field applications ( $< 1.5$  T),
- **superconducting magnet**: used for higher magnetic field strengths (high end research scanners, whole-body up to 11.75 T, general research  $> 20$  T).

superconducting -> hoher stromfluss ohne erhitzung

The main magnet of the MAGNETOM Verio scanner is only 6.5 tons!

## Necessity of Gradients

To distinguish between the nuclei, the idea is to associate a unique magnetic field with each type of nucleus. This can be achieved by continuous variation of the magnetic field dependent on the 3-D position of the nuclei.




Figure 3: Schematic of continuously changing the magnetic field in 3-D

**Note:** The gradient strength in 3-D can be used for motion compensation. If the patient motion is known, the gradient field can be adjusted properly!

# Components of an MR Scanner: Gradient System

The **magnetic field gradient system** is required to generate magnetic fields of well-defined and controlled spatial inhomogeneity as a function of the particular x-, y- and z-coordinates in space.

The gradient field is needed for signal localization in space.

- Gradient strength: e. g., 45 mT/m (millitesla per meter) in the Tim Trio 3T
- Rise time to ramp up gradient decides on quality of gradient: the smaller the rise time, the better (rise time today  $< 1.0$  ms).




Figure 4: Principle Structure of the gradient system (image courtesy of the National High Magnetic Field Laboratory, Florida)



# Components of an MR Scanner: RF System

The **radio frequency system** has two components:

- the **transmitter coil** generates a rotating magnetic field for the excitation of a spin system,
- the **receiver coil** converts magnetic changes into electrical signals.

In some systems transmission and receiver coils are identical which is then called a **transceiver coil**. all in one



Figure 5: Examples of RF head and body matrix coils (image Siemens Medical Solutions)

# Topics

MR Acquisition Devices

## Summary

Take Home Messages

Further Readings

## Take Home Messages

- Magnetic resonance imaging (MRI) requires large acquisition devices which consist of a **magnet**, a **gradient system**, an **RF system**, and the **imaging system**.
- Driven by concepts from quantum physics, a real magnetic field is not perfectly homogeneous which has to be compensated for by image processing.

mag field is not perfect -> artifacts -> image processing

## Further Readings

The webpage of the [National High Magnetic Field Laboratory](#) can be one starting point for more detailed information regarding MRI. For an initial overview of the technology, the following article is worth reading:

[MRI: A Guided Tour](#) by Kristen Coyne.

Another article worth reading is this survey paper on algorithms for intensity correction methods:

[Zujun Hou](#). “A Review on MR Image Intensity Inhomogeneity Correction”. In: *International Journal of Biomedical Imaging* 2006. Article ID 49515 (Feb. 2006), pp. 1–11. DOI: 10.1155/IJBI/2006/49515