GUJARAT PRINT PACK PUBLICATIONS PVT LTD

MONTHLY INTERNAL CALIBRATION RECORDS – GSM CUTTING PLATE

FORMAT NO.	REV NO.	DATE	PAGE NO.
F/QC/11	01	12-Oct-2026	1 of 1 (digital)

MONTHLY INTERNAL CALIBRATION RECORDS – GSM CUTTING PLATE. The four plates — No. 54 (20 x 20cm), No. 55 (10 x 10cm), No. 56 (5 x 5cm), No. 57 (2.5 x 2.5cm) — measured four times each, with the deviation beside every measurement, then Pass / Fail and the tester's sign for each plate.

PROCEDURE: Calibration to be done by qualified QA person as per calibration work procedure. If any deviation is found during internal calibration, it must be reported to the QA manager. Calibration records of all weights used must be available and valid.

Each Deviation % is worked out from the size written beside it, by area, against the plate's own size printed at the head of its column.

Device ID No	Location	Department	Manufacturer
WS-2026-10	QC Lab	Quality Control	ScaleCo
Serial No	**Calibration Expiry**	**Calibration Date**	**Due Date**
SC123456	12-10-2027	12-10-2026	12-10-2027

INDEX	NO. 54 (20 X 20CM)	DEVIATION %	NO. 55 (10 X 10CM)	DEVIATION %	NO. 56 (5 X 5CM)	DEVIATION %	NO. 57 (2.5 X 2.5CM)	DEVIATION %
Tasted-1	20.01	0.1%	10.02	0.4%	5.01	0.4%	2.50	0%
2								
3								
4								
Pass/Fail	Pass		Pass		Pass		Pass	
Sign.	Kapila		Kapila		Kapila		Kapila	