GUJARAT PRINT PACK PUBLICATIONS PVT LTD

WEEKLY INTERNAL CALIBRATION RECORDS - WEIGHT SCALE

FORMAT NO.	REV NO.	DATE	PAGE NO.
F/QC/12	01	12-Oct-2026	1 of 1 (digital)

WEEKLY INTERNAL CALIBRATION RECORDS - WEIGHT SCALE. One line per weekly calibration: the five test weights, the value the scale showed for each, the deviation, Pass / Fail, the tester's sign and when the next one is due.

PROCEDURE: Calibration to be done by qualified QA person as per calibration work procedure. If any deviation is found during internal calibration, it must be reported to the QA manager. Calibration records of all weights used must be available and valid.

On the paper the Deviation % of each entry is written on a second line beneath it; here every weight carries its own Deviation % column on the same line. Each Deviation % is worked out from the weight and its tested value — (tested − weight) ÷ weight × 100, in whatever unit they are written — and Pass / Fail follows the Acceptable Tolerance above.

Device ID No	Location	Department	Manufacturer	Serial No	Calibration Expiry	Minimum Reading Capacity	Maximum Reading Capacity	Acceptable Tolerance
QC-76	Laboratory	LAB	Force Strain Sensors	06	27-08-2024	0.1 gm	600 gm	0.05 %

SR. NO.	DATE	TESTED BY	WEIGHT 1	TESTED VALUE 1	DEVIATION %	WEIGHT 2	TESTED VALUE 2	DEVIATION %	WEIGHT 3	TESTED VALUE 3	DEVIATION %	WEIGHT 4	TESTED VALUE 4	DEVIATION %	WEIGHT 5	TESTED VALUE 5	DEVIATION %	PASS / FAIL	SIGN	NEXT DUE DATE
1	12-10-2026	Jeni	0.050mg	0.050mg	0%	50.000gm	50.000gm	0%	100.000gm	100.000gm	0%	200.000gm	200.000gm	0%	400.000gm	400.000gm	0%	Pass	Jeni	24-10-2026
2	12-10-2026	Jeni	0.050mg	0.050mg	0%	50.000gm	50.000gm	0%	100.000gm	100.000gm	0%	200.000gm	200.000gm	0%	400.000gm	400.000gm	0%	Pass	Jeni	24-10-2026
3	12-10-2026	Jeni	0.050mg	0.050mg	0%	50.000gm	50.000gm	0%	100.000gm	100.000gm	0%	200.000gm	200.000gm	0%	400.000gm	400.000gm	0%	Pass	Jeni	19-10-2026
4	12-10-2026	Jeni	0.050mg	0.050mg	0%	50.000gm	50.000gm	0%	100.000gm	100.000gm	0%	200.000gm	200.000gm	0%	400.000gm	400.000gm	0%	Pass	Jeni	dd-mm-yyyy