GUJARAT PRINT PACK PUBLICATION PRIVATE LIMITED	F/QC/17 (00/01.12.2021)

Measuring Scale (Ruler) Internal Verification Record

Master Measuring Scale (Ruler) Details	
Make	Kristeen
Serial number	QC-44
Range	0 - 600 mm
Least count	1 mm
Calibration date	23.09.2024
Calibration due date	22.09.2025
Calibration report number	CAL/LPI/GP-23/23

Measuring Scale (Ruler) under calibration details	
Make	Ajanta
Serial number	GP-3/MS/36
Range	0-600mm
Least count	1 mm
Calibration date	24.09.2024
Calibration due date	23.09.2025
Accetable tolerance	**+ / - 1 mm**

#	Master calibrator reading	Observed reading (mm)	Difference	Status (Pass/Fail)	Calibration performed by
1	100 mm	100	0.0	Pass	Kapila Barad
2	200 mm	200	0.0		
3	300 mm	300	0.0		
4	400 mm	400	0.0		
5	500 mm	500	0.0		
6	490 mm	590	0.0		