

AMENDMENT NO. 2 DECEMBER 2013
TO
IS 15803 : 2008 RESPIRATORY PROTECTIVE DEVICES —
SELF CONTAINED
CLOSED CIRCUIT BREATHING APPARATUS
CHEMICAL OXYGEN (KO₂) TYPE, SELF GENERATING,
SELF RESCUERS — SPECIFICATION

(Page 4, clause 5.16.5, Para 1) — Substitute the following para for the existing:

‘For this apparatus the temperature of the inhaled gas irrespective of its rated duration and humidity shall not exceed 55°C, when tested at 35 liter/min during the rated duration when tested for general laboratory performance test in accordance with A-1’.