

AMENDMENT NO. 1 JANUARY 2000
TO
IS 13293 : 1992 GAS DETECTOR TUBES — GENERAL
REQUIREMENTS AND METHODS OF TEST

(Page 1, clause 2.1.3, line 3) — Substitute '100 cc/min' for '1 cc/sec'.

(Page 1, clause 2.2.3, line 1) — Substitute '10 mm' for '15 mm'.

(CHD 08)

Reprography Unit, BIS, New Delhi, India