

AMENDMENT NO. 1 JANUARY 2006
TO
IS 2620 : 1963 SPECIFICATION FOR DISTILLING
FLASKS

(Page 5, clause 3.4) — Substitute the following for the existing clause:

'3.4 Nominal Capacity — The distilling flasks shall be made in the following nominal capacities:

- a) 30 ml, b) 50 ml, c) 100 ml, d) 125 ml, e) 150 ml, f) 250 ml,
g) 500 ml, h) 750 ml, j) 1 000 ml, and k) 2 000 ml.'

(Page 6, Table 1) — Substitute the following for the existing:

Table I Recommended Dimensions for Distilling Flasks

Sl No.	Nomina l Capacit y	Internal Diameter of Neck	Length of Side Tube	Internal Diameter of Side Tube	Bulb Capacity	External Diameter of Bulb	Wall Thickness, Min
(1)	(2)	(3)	(4)	(5)	(6)	(7)	(8)
	ml	mm	mm	mm	ml	mm	mm
i)	30	16 ± 1	90 ± 3	4 ± 0.5	39 ± 2	43.5	0.8
ii)	50	16 ± 1	100 ± 3	4 ± 0.5	65 ± 3	52	0.8
iii)	100	16 ± 1	100 ± 3	4 ± 0.5	130 ± 5	65	1.0
iv)	125	16 ± 1	100 ± 3	7 ± 0.5	163 ± 5	68	1.0
v)	150	18 ± 1	120 ± 4	5.5 ± 0.5	195 ± 7	74	1.0
vi)	250	20 ± 1.2	140 ± 5	6 ± 0.5	325 ± 10	87.5	1.0
vii)	500	24 ± 1.5	160 ± 4	7.5 ± 0.7	650 ± 15	110	1.2
viii)	750	28 ± 1.5	160 ± 5	8.5 ± 0.8	975 ± 20	125.5	1.2
ix)	1 000	30 ± 1.5	160 ± 5	9.5 ± 0.8	1 300 ± 25	138	1.3
x)	2 000	33 ± 2.5	170 ± 5	10.5 ± 0.8	2 600 ± 50	170	1.5

(CHD 10)