

AMENDMENT NO. 1 FEBRUARY 1983
TO
IS:2711-1979 SPECIFICATION FOR DIRECT
READING pH METERS

(Second Revision)

Alteration

**(Page 7, clause 5.4) - Substitute the following
for the existing clause:**

**'5.4 For moving index instruments, the scales shall
be divided and marked with the major scale marks in
series of one. Each dimension on the full scale
shall represent a minimum of 0.1 pH.'**

(ETDO 67)

Reprography Unit, ISI, New Delhi, India