

AMENDMENT NO. 1 DECEMBER 2018
TO
IS 4533 : 1995 SUCTION APPARATUS — SPECIFICATION

(Second Revision)

(Page 1, clause 5.1.1) — Substitute the following for the existing clause:

‘5.1.1 The vacuum unit for Type 1 apparatus shall be so designed and constructed that it shall be able to create a vacuum up to 65 percent of the atmospheric pressure at testing site in 10 s and 92 percent of the atmospheric pressure at testing site in not more than 70 s with a minimum free air displacement rate of 25 to 30 l/min, for evacuating two jars of 2 litre capacity each. The electrical components shall be housed in a flame-proof enclosure conforming to IS/IEC 60079-1. It shall be mounted in-such a way as to produce minimum vibrations.’

(Page 1, clause 5.1.2) — Substitute the following for the existing clause:

‘5.1.2 The vacuum unit for Type 2 apparatus shall be capable of creating a vacuum up to 59 percent of atmospheric pressure at testing site in 10 s and 79 percent of atmospheric pressure at testing site in not more than 40 s with a minimum free air displacement rate of 20 l/min. It shall be mounted in such a way as to produce minimum vibrations. The electrical components shall not produce sparks during operation.’

(Page 4, Annex A) — Substitute ‘IS/IEC 60079-1 : 2007 Explosive atmospheres : Part 1 Equipment protection by flameproof enclosures ‘d’ ’ for ‘IS 2148 : 1981 ‘Flame-proof enclosures for electrical apparatus (*second revision*).’