

IS 4610 : 1968 SPECIFICATION FOR GLASS TUBES FOR GENERAL PURPOSE AND REFERENCE THERMOMETERS

(Page 5, clause 5.4) — Substitute the following for the existing clause:

'5.4 Limit of Alkalinity — When graded according to the method prescribed in IS 2303 (Part 1/Sec 1) : 1994*, glass tubes shall conform to Class HGB 4 of the glass.'

(Page 5, footnote marked *) — Substitute the following for the existing:

*Grading glass for alkalinity : Part 1 Hydrolytic resistance, Section 1 Hydrolytic resistance of glass grains at 98°C - Method of test and classification (*first revision*).'

(Page 6, clause 6.2.1) — Substitute the following for the existing clause:

'6.2.1 BIS Certification Mark

The bags and boxes may also be marked with the Standard Mark.

6.2.1.1 The use of the Standard Mark is governed by the provisions of *Bureau of Indian Standards Act*, 1986 and the Rules and Regulations made thereunder. The details of conditions under which the licence for the use of Standard Mark may be granted to manufacturers or producers may be obtained from the Bureau of Indian Standards.'