

**AMENDMENT NO. 1 AUGUST 1999  
TO  
IS 12914 : 1990 BROMADIOLONE, TECHNICAL —  
SPECIFICATION**

( Page 2, clause A-2.1 ) — Substitute the following for the existing:

**'A-2.1 A HPLC shall be equipped with a printer-plotter cum integrator and UV detector. The suggested operating conditions are given below which can be varied provided standardization is done:**

|                  |                                                                                                                   |
|------------------|-------------------------------------------------------------------------------------------------------------------|
| Column           | Stainless steel, 250 mm length and 4.6 mm internal diameter packed with RP 18-5 micron or 10 micron particle size |
| Mobile Phase     | Methanol 75 ml + Water 25 ml + Acetic Acid 0.5 ml                                                                 |
| Detector         | UV (250 nm), solvent flow rate 0.8 ml/min                                                                         |
| Chart Speed      | 5 mm/min                                                                                                          |
| Injection Volume | 20 micro litre                                                                                                    |
| Retention Time   | Bromadiolone 3.4 min approx, di- <i>n</i> -butyl phthalate (internal standard) 2.3 min approx.'                   |

( Page 2, clause A-3 ) — Substitute the following for the existing:

**'A-3 REAGENTS**

**A-3.1 Di-*n*-Butyl Phthalate (Internal Standard) AR Grade**

**A-3.2 Methanol, HPLC grade.**

**A-3.3 Water, HPLC grade.**

**A-3.4 Bromadiolone standard, of known purity.**

**A-3.5 Acetic Acid'**

( Page 2, clause A-4.2.2, line 3 ) — Substitute '2 ml' for '4 ml'.

( Page 2, clause A-4.3.1, line 1 ) — Substitute '20 micro litre' for '5 ul'.

( FAD 1 )