

AMENDMENT NO. 1 JULY 2021

TO

**IS 12981 : 1991 INDIAN STANDARD COMMON SALT —
IRON FORTIFIED — SPECIFICATION**

(Page 2, clause A-2.1.1) — Substitute ‘Spectrophotometer’ for ‘Photoelectric Absorptiometer’

(Page 2, clause A-2.3.3) — Substitute the following for the existing clause :
‘Measure absorbance of the sample solution as prepared at **A-2.3.2** at 510 nm.
Read off milligram iron from the standard calibration graph (*see A-2.3.4*).’