

**AMENDMENT NO. 1 JANUARY 1993
TO
IS 10539:1983 SPECIFICATION FOR SICKLE
TERMINAL DEVICE FOR ARTIFICIAL LIMBS**

(Page 1, Fig. 1):

- a) (*Heigh of the sickle*) — Substitute '75 to 80' for '72 $\pm$ 2'.
- b) (*Length of the thread*) — Substitute '15 to 20' for '20 max'.

(MHD 10)

Reprography Unit, BIS, New Delhi, India