

TO

IS:10775-1984 SPECIFICATION FOR BODY LEVEL
HEARING AIDS

(Page 9, Table 1, Sl No. 2, col 2) - Substitute
'High-frequency average OSPL 90' for 'Average OSPL
90'.

(Page 9, Table 1, Sl No. 2, col 3) - Add the
following after the existing matter:

'The average value of OSPL 90 shall be taken at
the frequencies 1 000, 1 600 and 2 500 Hz.'

(Page 9, Table 1, Sl No. 6, col 4, 5 and 6) -
Substitute 'at least between' for 'between'.

(LTDC 5)

Reprography Unit, BIS, New Delhi, India