

**AMENDMENT NO. 1 FEBRUARY 2014
TO
IS 15322 : 2003 PARTICLE FILTERS USED IN
RESPIRATORY PROTECTIVE EQUIPMENT —
SPECIFICATION**

(Page 3, Annex A, clause A-4.1, line 5) — Insert the following after **A-4.1** (b):

The filter shall be allowed to return to room temperature for at least 4 h between exposures and prior to subsequent testing.'

(CHD 8)

Publication Unit, BIS, New Delhi, India